



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

LDB

Agency: Lynn Haven P.D. Instrument Serial Number: 80-000206

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____ <input type="checkbox"/> Remain in Service <input type="checkbox"/> Remove from Service	<input type="checkbox"/> Notification (Agency Inspection) <input checked="" type="checkbox"/> Contact <input type="checkbox"/> Other _____ Date: <u>1/7/08 + 1/9/08</u>	The following records were audited: <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate of Analysis
<b>Comments:</b> Spoke with both Chief Messer and Jim Smith regarding comments made on agency inspection report # dated 9/2/2006. A memo explaining the comment was completed and will be placed into the instrument file.		

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Other	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces
<b>Comments:</b>		

Recommended Action/Action Taken:
<b>RECEIVED</b> JAN 11 2008 <i>enr</i> FDLE Alcohol Testing Program

Margaret Geddings  
Signature of Alcohol Testing Program Staff Member

1/9/08  
Date

Original - FDLE

Copy - Agency

DAVID S. MESSER  
CHIEF



DENNIS L. KIAI  
ASST. CHIEF

RECEIVED

JAN 11 2008 *ang*

FDLE  
Alcohol Testing Program

# Memo

**To:** Maggie Geddings, Department Inspector, F.D.L.E. Alcohol Testing Program  
**From:** Sergeant James Smith, Agency Inspector, Lynn Haven Police Department *[Signature]*  
**Date:** January 9, 2008  
**Subject:** Agency Inspection Report, dated 9/2/2006, Instrument serial #80-000206

The remarks section of this report, when submitted electronically, stated this machine sucks (air through a bad simulator thermometer seal) and does not refer to the above referenced Intoxilyzer 8000. This unit was on loan to our department while our Intoxilyzer 8000 was in for repairs. The instrument was not in compliance and the entirety of the remarks in this field was not transmitted.

The simulator in question was removed from service pending repairs to the thermometer seal and a new simulator was purchased.

The Lynn Haven Police Department's instrument, 80-000840, has since been placed back in service and I assure you that I have the utmost confidence in the Intoxilyzer 8000.

Please feel free to call if you have any further questions or problems with this agency inspection report.

108 E. 9th St. • Lynn Haven, Florida 32444  
(850) 265-4111 • Fax (850) 265-7328

## Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source
FDLE/ATP PROC SMITH,JAMES,C	09/02/2006	15:15:17	8100.26	Agency Inspection

<b>0.05 Lot#</b> 200509A	<b>0.08 Lot#</b> 200509B	<b>0.20 Lot#</b> 20509C	<b>0.08 Gas Lot#</b>
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	

**# Sims** 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.029
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.068
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.146
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.173
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.042
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	No	0.08 g/210L Test #2	0.074
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2 (Retry)	0.080
		0.20 g/210L Test #2	0.166
		0.20 g/210L Test #2 (Retry)	0.177
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.082
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.172
		0.20 g/210L Test #3 (Retry)	0.176
		Interferent Detect Test #3	INT

**Remarks:**

THIS MACHINE SUCKS Non-compliance: .

