

Florida Department of Law Enforcement

Control Tests

Instrument Serial #: 80-000206

Location	Date	Time	Version	Target	Source
FDLE/ATP PROC	09/02/2006	12:13:57	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	12:14	
		Control Test	0.040	12:15	
		Air Blank	0.000	12:15	
FDLE/ATP PROC	09/02/2006	12:17:47	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	12:18	
		Control Test	0.044	12:18	
		Air Blank	0.000	12:19	
FDLE/ATP PROC	09/02/2006	12:19:28	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	12:19	
		Control Test	0.045	12:20	
		Air Blank	0.000	12:21	
FDLE/ATP PROC	09/02/2006	12:21:09	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	12:21	
		Control Test	0.044	12:22	
		Air Blank	0.000	12:22	
FDLE/ATP PROC	09/02/2006	12:24:26	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	12:24	
		Control Test	0.138	12:25	
		Air Blank	0.000	12:26	
FDLE/ATP PROC	09/02/2006	12:26:06	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	12:26	
		Control Test	0.169	12:27	
		Air Blank	0.000	12:27	
FDLE/ATP PROC	09/02/2006	16:04:26	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	16:04	
		Control Test	0.168	16:05	
		Air Blank	0.000	16:06	
FDLE/ATP PROC	09/02/2006	16:06:11	8100.26.00	0.080	Local Wet bath
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	16:06	

Control Tests

Instrument Serial #: 80-000206

Location	Date	Time	Version	Target	Source
FDLE/ATP PROC	09/02/2006	16:06:11	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Control Test	0.176	16:07	
		Air Blank	0.000	16:07	
FDLE/ATP PROC	09/02/2006	16:08:03	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:08	
		Control Test	0.085	16:08	
		Air Blank	0.000	16:09	
FDLE/ATP PROC	09/03/2006	15:38:09	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:38	
		Control Test	0.171	15:39	
		Air Blank	0.000	15:39	
FDLE/ATP PROC	09/03/2006	15:39:52	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:40	
		Control Test	0.173	15:41	
		Air Blank	0.000	15:41	
FDLE/ATP PROC	09/03/2006	15:41:33	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:42	
		Control Test	0.000	15:42	
		Air Blank	0.000	15:43	
FDLE/ATP PROC	09/03/2006	15:43:13	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:43	
		Control Test	0.000	15:44	
		Air Blank	0.000	15:44	
FDLE/ATP PROC	09/03/2006	15:48:43	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:49	
		Control Test	0.000	15:49	
		Air Blank	0.000	15:50	
FDLE/ATP PROC	09/03/2006	15:52:13	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:52	
		Control Test	0.000	15:53	
		Air Blank	0.000	15:53	

Control Tests

Instrument Serial #: 80-000219

Location	Date	Time	Version	Target	Source
FDLE FMROC	08/30/2006	12:41:02	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		12:41	Sequence Aborted
		Air Blank	0.000	12:41	

Control Tests

Instrument Serial #: 80-000223

<u>Location</u>	<u>Date</u>	<u>Time</u>	<u>Version</u>	<u>Target</u>	<u>Source</u>
FDLE/ATP MELBOURNE	04/14/2006	10:33:29	8100.26.00	0.080	Remote Dry gas
		<u>Test</u>	<u>Value</u>	<u>Time</u>	

Control Tests

Instrument Serial #: 80-000224

Location	Date	Time	Version	Target	Source
FDLE/ATP JROC	06/02/2006	12:36:19	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	12:36	
		Control Test	0.047	12:37	
		Air Blank	0.000	12:37	
FDLE/ATP JROC	06/02/2006	12:38:01	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	12:38	
		Control Test	0.000	12:39	Interferent Detect
		Air Blank	0.000	12:39	
FDLE/ATP JROC	07/14/2006	16:15:54	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		16:16	Sequence Aborted
		Air Blank	0.000	16:16	
FDLE/ATP JROC	07/14/2006	16:16:43	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		16:16	Sequence Aborted
		Air Blank	0.000	16:17	
FDLE/ATP JROC	07/14/2006	16:17:32	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:17	
		Control Test	0.078	16:18	Interferent Detect
		Air Blank	0.000	16:19	
FDLE/ATP JROC	07/14/2006	16:19:18	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:19	
		Control Test	0.000	16:19	RFI Detect
		Air Blank	0.000	16:20	

Control Tests

Instrument Serial #: 80-000228

Location	Date	Time	Version	Target	Source
FDLE MROC	07/11/2006	14:33:53	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:34	
		Control Test	0.198	14:35	
		Air Blank	0.000	14:35	
FDLE MROC	07/11/2006	14:36:04	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:36	
		Control Test	0.000	14:37	
		Air Blank	0.000	14:37	
CMI INC	09/11/2006	09:36:16	8100.26.00	0.000	Local Dry gas
		Test	Value	Time	
		Air Blank		09:36	Sequence Aborted
		Air Blank	0.000	09:36	

Control Tests

Instrument Serial #: 80-000230

Location	Date	Time	Version	Target	Source
CMI INC	02/27/2006	09:07:05	8100.26.00	0.000	Local ITP
		Test	Value	Time	
		Air Blank	0.000	09:07	
		ITP Check	0.079	09:07	
		Air Blank	0.000	09:08	

Control Tests

Instrument Serial #: 80-000737

Location	Date	Time	Version	Target	Source
ALACHUA COUNTY SO	04/29/2006	20:04:11	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:04	
		Control Test	0.000	20:05	
		Air Blank	0.000	20:05	
ALACHUA COUNTY SO	04/29/2006	20:07:45	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:08	
		Control Test	0.047	20:08	
		Air Blank	0.000	20:09	
ALACHUA COUNTY SO	04/29/2006	20:11:29	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:11	
		Control Test	0.000	20:12	RFI Detect
		Air Blank	0.000	20:12	
ALACHUA COUNTY SO	04/29/2006	20:12:41	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:13	
		Control Test	0.000	20:13	
		Air Blank	0.000	20:14	
ALACHUA COUNTY SO	04/29/2006	20:14:27	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:14	
		Control Test	0.000	20:15	
		Air Blank		20:15	RFI Detect
		Air Blank	0.000	20:16	
ALACHUA COUNTY SO	04/29/2006	20:19:14	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:19	
		Control Test	0.000	20:20	
		Air Blank	0.000	20:20	
ALACHUA COUNTY SO	04/30/2006	14:25:59	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:26	
		Control Test	0.000	14:27	
		Air Blank	0.000	14:27	

Control Tests

Instrument Serial #: 80-000738

Location	Date	Time	Version	Target	Source
FT. WALTON BEACH PD	05/01/2006	15:33:20	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	15:33	
		ITP Check	0.083	15:34	
		Air Blank	0.000	15:34	
FT. WALTON BEACH PD	05/01/2006	15:37:59	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:38	
		Control Test	0.078	15:38	
		Air Blank	0.000	15:39	

Control Tests

Instrument Serial #: 80-000739

Location	Date	Time	Version	Target	Source
OKALOOSA COUNTY SO	07/09/2006	21:21:39	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		21:21	RFI Detect
		Air Blank	0.000	21:22	
OKALOOSA COUNTY SO	07/09/2006	21:23:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:23	
		Control Test	0.080	21:24	
		Air Blank	0.000	21:24	
OKALOOSA COUNTY SO	07/09/2006	21:24:45	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:25	
		Control Test	0.080	21:25	
		Air Blank	0.000	21:26	
OKALOOSA COUNTY SO	07/09/2006	21:26:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:26	
		Control Test	0.081	21:27	
		Air Blank	0.000	21:27	

Control Tests

Instrument Serial #: 80-000752

Location	Date	Time	Version	Target	Source
MARION COUNTY SO	03/29/2006	08:50:58	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:51	
		Control Test	0.079	08:51	
		Air Blank	0.000	08:52	
MARION COUNTY SO	03/29/2006	08:54:09	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	08:54	
		ITP Check	0.079	08:55	
		Air Blank	0.000	08:55	

Control Tests

Instrument Serial #: 80-000760

Location	Date	Time	Version	Target	Source
ALACHUA COUNTY SO	04/30/2006	15:58:01	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:58	
		Control Test	0.000	15:59	
		Air Blank	0.000	15:59	
ALACHUA COUNTY SO	04/30/2006	16:00:03	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:00	
		Control Test	0.000	16:01	
		Air Blank	0.000	16:01	
ALACHUA COUNTY SO	04/30/2006	16:02:08	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:02	
		Control Test	0.080	16:03	
		Air Blank	0.000	16:03	
ALACHUA COUNTY SO	04/30/2006	16:05:53	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:06	
		Control Test	0.000	16:07	
		Air Blank	0.000	16:07	
ALACHUA COUNTY SO	05/31/2006	09:33:01	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:33	
		Control Test	0.048	09:34	
		Air Blank	0.000	09:34	
ALACHUA COUNTY SO	05/31/2006	09:35:14	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:35	
		Control Test	0.197	09:36	
		Air Blank	0.000	09:36	

Control Tests

Instrument Serial #: 80-000763

Location	Date	Time	Version	Target	Source
OKALOOSA COUNTY SO	04/08/2006	14:43:07	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:43	
		Control Test	0.078	14:43	
		Air Blank	0.000	14:44	
OKALOOSA COUNTY SO	04/08/2006	14:47:13	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	14:47	
		ITP Check	0.080	14:48	
		Air Blank	0.000	14:48	
OKALOOSA COUNTY SO	04/08/2006	14:50:24	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		14:50	Sequence Aborted
		Air Blank	0.000	14:51	
OKALOOSA COUNTY SO	04/08/2006	14:51:30	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	14:51	
		ITP Check	0.079	14:52	
		Air Blank	0.000	14:52	

Control Tests

Instrument Serial #: 80-000775

Location	Date	Time	Version	Target	Source
COLUMBIA COUNTY SO	05/22/2006	06:29:23	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	06:29	
		ITP Check	0.080	06:30	
		Air Blank	0.000	06:30	

Control Tests

Instrument Serial #: 80-000780

Location	Date	Time	Version	Target	Source
SUWANNEE COUNTY SO	10/17/2006	08:32:11	8100.26.00	0.080	Remote Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:32	
		Control Test	0.080	08:32	
		Air Blank	0.000	08:33	

Control Tests

Instrument Serial #: 80-000784

Location	Date	Time	Version	Target	Source
HOLMES COUNTY SO	11/12/2006	22:14:05	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:13	
		Control Test	0.079	22:13	Interferent Detect
		Air Blank	0.000	22:14	
HOLMES COUNTY SO	11/12/2006	22:16:55	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:16	
		Control Test	0.079	22:16	Interferent Detect
		Air Blank	0.000	22:16	
HOLMES COUNTY SO	11/12/2006	22:27:31	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:26	
		Control Test	0.078	22:27	Interferent Detect
		Air Blank	0.000	22:27	
HOLMES COUNTY SO	11/12/2006	22:28:57	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	22:28	
		ITP Check	0.078	22:28	
		Air Blank	0.000	22:28	
HOLMES COUNTY SO	11/13/2006	03:27:16	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	03:26	
		Control Test	0.080	03:26	
		Air Blank	0.000	03:27	
HOLMES COUNTY SO	11/15/2006	21:20:39	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:19	
		Control Test	0.079	21:20	
		Air Blank	0.000	21:20	

Control Tests

Instrument Serial #: 80-000785

Location	Date	Time	Version	Target	Source
HOLMES COUNTY SO	11/12/2006	22:08:03	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:07	
		Control Test	0.080	22:07	Interferent Detect
		Air Blank	0.000	22:08	
HOLMES COUNTY SO	11/12/2006	22:16:40	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:15	
		Control Test	0.080	22:16	Interferent Detect
		Air Blank	0.000	22:16	
HOLMES COUNTY SO	11/15/2006	21:23:42	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:22	
		Control Test	0.080	21:23	
		Air Blank	0.000	21:23	
HOLMES COUNTY SO	12/10/2006	04:29:24	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:28	
		Control Test	0.080	04:28	
		Air Blank	0.000	04:29	
HOLMES COUNTY SO	12/28/2006	19:33:55	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:33	
		Control Test	0.076	19:33	
		Air Blank	0.000	19:33	
HOLMES COUNTY SO	12/28/2006	21:14:58	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:14	
		Control Test	0.080	21:14	
		Air Blank	0.000	21:14	

Control Tests

Instrument Serial #: 80-000786

Location	Date	Time	Version	Target	Source
BROWARD COUNTY SO	04/22/2006	01:35:17	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	01:35	
		ITP Check	0.079	01:36	
		Air Blank	0.000	01:36	
BROWARD COUNTY SO	12/15/2006	12:19:29	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:18	
		Control Test	0.081	12:18	
		Air Blank	0.000	12:19	

Control Tests

Instrument Serial #: 80-000787

Location	Date	Time	Version	Target	Source
GULF BREEZE PD	05/09/2006	11:22:15	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:22	
		Control Test	0.045	11:23	
		Air Blank	0.000	11:23	
GULF BREEZE PD	11/30/2006	06:50:57	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	06:49	
		Control Test	0.050	06:50	
		Air Blank	0.000	06:50	

Control Tests

Instrument Serial #: 80-000788

Location	Date	Time	Version	Target	Source
ST LUCIE COUNTY SO	10/21/2006	20:37:53	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:38	
		Control Test	0.078	20:38	
		Air Blank	0.000	20:39	

Control Tests

Instrument Serial #: 80-000792

Location	Date	Time	Version	Target	Source
ST LUCIE COUNTY SO	04/05/2006	13:05:26	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:05	
		Control Test	0.078	13:06	
		Air Blank	0.000	13:06	
ST LUCIE COUNTY SO	04/05/2006	13:07:23	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:07	
		Control Test	0.077	13:08	
		Air Blank	0.000	13:08	
ST LUCIE COUNTY SO	04/05/2006	13:08:59	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:09	
		Control Test	0.077	13:09	
		Air Blank	0.000	13:10	

Control Tests

Instrument Serial #: 80-000796

Location	Date	Time	Version	Target	Source
LONGBOAT KEY PD	02/15/2006	16:36:26	8100.26.00	0.000	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:36	
		Control Test	0.078	16:37	
		Air Blank	0.000	16:37	
LONGBOAT KEY PD	03/22/2006	15:49:51	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	15:50	
		ITP Check	0.078	15:50	
		Air Blank	0.000	15:51	

Control Tests

Instrument Serial #: 80-000808

Location	Date	Time	Version	Target	Source
HILLSBOROUGH SO	11/03/2006	11:38:49	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:37	
		Control Test	0.081	11:38	
		Air Blank	0.000	11:38	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	05/24/2006	11:01:20	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:01	
		Control Test	0.188	11:02	
CITRUS COUNTY SO	05/24/2006	12:53:39	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	12:54	
		Control Test	0.195	12:54	
CITRUS COUNTY SO	05/24/2006	12:55:29	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	12:55	
		Control Test	0.197	12:56	
CITRUS COUNTY SO	05/24/2006	12:58:38	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	12:59	
		Control Test	0.198	12:59	
CITRUS COUNTY SO	05/24/2006	13:00:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:01	
		Control Test	0.079	13:01	
CITRUS COUNTY SO	05/24/2006	13:02:21	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:02	
		Control Test	0.080	13:03	
CITRUS COUNTY SO	05/24/2006	13:04:18	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:04	
		Control Test	0.080	13:05	
CITRUS COUNTY SO	07/18/2006	18:20:56	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:21	
		Control Test	0.078	18:21	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	07/18/2006	18:20:56	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:22	
CITRUS COUNTY SO	08/31/2006	08:36:16	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:36	
		Control Test	0.023	08:37	Interferent Detect
		Air Blank	0.000	08:37	
CITRUS COUNTY SO	08/31/2006	08:38:33	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	08:39	
		Control Test	0.000	08:39	
		Air Blank	0.000	08:40	
CITRUS COUNTY SO	08/31/2006	08:40:34	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:41	
		Control Test	0.060	08:41	
		Air Blank	0.000	08:41	
CITRUS COUNTY SO	08/31/2006	09:34:54	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:35	
		Control Test	0.078	09:35	
		Air Blank	0.000	09:36	
CITRUS COUNTY SO	08/31/2006	09:37:07	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:37	
		Control Test	0.077	09:38	
		Air Blank	0.000	09:38	
CITRUS COUNTY SO	08/31/2006	09:40:12	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:40	
		Control Test	0.184	09:41	
		Air Blank	0.000	09:41	
CITRUS COUNTY SO	08/31/2006	11:47:16	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:47	
		Control Test	0.193	11:48	
		Air Blank	0.000	11:48	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	08/31/2006	11:49:12	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:49	
		Control Test	0.194	11:50	
		Air Blank	0.000	11:50	
CITRUS COUNTY SO	08/31/2006	11:52:34	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:53	
		Control Test	0.077	11:53	
		Air Blank	0.000	11:54	
CITRUS COUNTY SO	08/31/2006	11:54:23	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:54	
		Control Test	0.048	11:55	
		Air Blank	0.000	11:56	
CITRUS COUNTY SO	08/31/2006	12:18:09	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	12:18	
		Control Test	0.000	12:19	
		Air Blank	0.000	12:19	
CITRUS COUNTY SO	09/12/2006	19:43:14	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	19:43	
		Control Test	0.196	19:44	
		Air Blank	0.000	19:44	
CITRUS COUNTY SO	09/12/2006	19:46:37	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	19:47	
		Control Test	0.078	19:47	
		Air Blank	0.000	19:48	
CITRUS COUNTY SO	09/12/2006	19:48:57	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	19:49	
		Control Test	0.048	19:50	
		Air Blank	0.000	19:50	
CITRUS COUNTY SO	09/12/2006	20:10:12	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:10	
		Control Test	0.075	20:11	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	09/12/2006	20:10:12	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:11	
CITRUS COUNTY SO	09/12/2006	20:11:57	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:12	
		Control Test	0.077	20:13	
		Air Blank	0.000	20:13	
CITRUS COUNTY SO	09/12/2006	20:18:07	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	20:18	
		Control Test	0.078	20:19	
		Air Blank	0.000	20:19	
CITRUS COUNTY SO	09/12/2006	21:29:45	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	21:30	
		ITP Check	0.079	21:30	
		Air Blank	0.000	21:31	
CITRUS COUNTY SO	10/26/2006	09:46:45	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:45	
		Control Test	0.068	09:46	
		Air Blank	0.000	09:46	
CITRUS COUNTY SO	10/26/2006	09:48:52	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:47	
		Control Test	0.071	09:48	
		Air Blank	0.000	09:48	
CITRUS COUNTY SO	10/26/2006	09:51:04	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:50	
		Control Test	0.072	09:50	
		Air Blank	0.000	09:51	
CITRUS COUNTY SO	10/26/2006	10:02:28	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:01	
		Control Test	0.070	10:01	
		Air Blank	0.000	10:02	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	10/26/2006	10:16:11	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:15	
		Control Test	0.074	10:15	
		Air Blank	0.000	10:16	
CITRUS COUNTY SO	10/26/2006	10:40:50	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:39	
		Control Test	0.074	10:40	
		Air Blank	0.000	10:40	
CITRUS COUNTY SO	10/26/2006	11:19:27	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:18	
		Control Test	0.074	11:18	
		Air Blank	0.000	11:19	
CITRUS COUNTY SO	10/26/2006	11:49:42	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:48	
		Control Test	0.076	11:49	
		Air Blank	0.000	11:49	
CITRUS COUNTY SO	10/30/2006	14:38:37	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:37	
		Control Test	0.078	14:38	
		Air Blank	0.000	14:38	
CITRUS COUNTY SO	10/30/2006	14:56:02	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:55	
		Control Test	0.078	14:55	
		Air Blank	0.000	14:56	
CITRUS COUNTY SO	10/30/2006	19:45:48	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:44	
		Control Test	0.078	19:45	
		Air Blank	0.000	19:45	
CITRUS COUNTY SO	10/30/2006	20:37:38	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:36	
		Control Test	0.077	20:37	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	10/30/2006	20:37:38	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:37	
CITRUS COUNTY SO	10/30/2006	20:46:19	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:45	
		Control Test	0.078	20:45	
		Air Blank	0.000	20:46	
CITRUS COUNTY SO	11/16/2006	12:55:51	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:54	
		Control Test	0.079	12:55	
		Air Blank	0.000	12:55	
CITRUS COUNTY SO	11/19/2006	00:04:46	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:03	
		Control Test	0.079	00:04	
		Air Blank	0.000	00:04	
CITRUS COUNTY SO	12/29/2006	13:40:54	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:39	
		Control Test	0.080	13:40	
		Air Blank	0.000	13:40	
CITRUS COUNTY SO	12/29/2006	13:43:04	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:41	
		Control Test	0.048	13:42	
		Air Blank	0.000	13:43	
CITRUS COUNTY SO	12/29/2006	13:47:38	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:46	
		Control Test	0.048	13:47	
		Air Blank	0.000	13:47	
CITRUS COUNTY SO	12/29/2006	13:49:13	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:48	
		Control Test	0.078	13:48	
		Air Blank	0.000	13:49	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	12/29/2006	14:14:38	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		14:14	Ambient Fail
		Air Blank	0.000	14:14	
CITRUS COUNTY SO	12/29/2006	14:52:43	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		14:52	Ambient Fail
		Air Blank	0.000	14:52	
CITRUS COUNTY SO	12/29/2006	14:54:21	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:53	
		Control Test	0.000	14:53	
		Air Blank	0.000	14:54	
CITRUS COUNTY SO	12/29/2006	14:55:45	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:54	
		Control Test	0.025	14:55	Interferent Detect
		Air Blank	0.000	14:55	
CITRUS COUNTY SO	12/29/2006	14:57:19	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:56	
		Control Test	0.036	14:56	
		Air Blank	0.000	14:57	
CITRUS COUNTY SO	12/29/2006	14:58:41	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:57	
		Control Test	0.048	14:58	
		Air Blank	0.000	14:58	
CITRUS COUNTY SO	12/29/2006	15:00:03	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:59	
		Control Test	0.053	14:59	
		Air Blank	0.000	15:00	
CITRUS COUNTY SO	12/29/2006	15:17:36	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:16	
		Control Test	0.000	15:17	
		Air Blank	0.000	15:17	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	12/29/2006	15:19:32	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:18	
		Control Test	0.066	15:19	
		Air Blank	0.000	15:19	
CITRUS COUNTY SO	12/29/2006	15:35:27	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:34	
		Control Test	0.064	15:34	
		Air Blank	0.000	15:35	
CITRUS COUNTY SO	12/29/2006	15:37:46	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:36	
		Control Test	0.071	15:37	
		Air Blank	0.000	15:37	
CITRUS COUNTY SO	12/29/2006	16:05:28	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:04	
		Control Test	0.072	16:04	
		Air Blank	0.000	16:05	
CITRUS COUNTY SO	12/29/2006	16:07:47	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:06	
		Control Test	0.076	16:07	
		Air Blank	0.000	16:07	
CITRUS COUNTY SO	12/29/2006	16:09:05	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:08	
		Control Test	0.000	16:08	
		Air Blank	0.000	16:09	
CITRUS COUNTY SO	12/29/2006	16:10:50	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:09	
		Control Test	0.077	16:10	
		Air Blank	0.000	16:10	
CITRUS COUNTY SO	12/29/2006	16:12:35	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:11	
		Control Test	0.073	16:12	

Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	12/29/2006	16:12:35	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:12	
CITRUS COUNTY SO	12/29/2006	16:29:25	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:28	
		Control Test	0.074	16:28	
		Air Blank	0.000	16:29	
CITRUS COUNTY SO	12/29/2006	16:31:14	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	16:30	
		ITP Check	0.080	16:30	
		Air Blank	0.000	16:31	
CITRUS COUNTY SO	12/29/2006	16:55:08	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:54	
		Control Test	0.076	16:54	
		Air Blank	0.000	16:55	
CITRUS COUNTY SO	12/29/2006	16:56:39	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:55	
		Control Test	0.077	16:56	
		Air Blank	0.000	16:56	

Control Tests

Instrument Serial #: 80-000822

Location	Date	Time	Version	Target	Source
LAKE SO	05/08/2006	13:57:06	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	13:57	
		ITP Check	0.080	13:58	
		Air Blank	0.000	13:58	

Control Tests

Instrument Serial #: 80-000824

Location	Date	Time	Version	Target	Source
LEESBURG PD	05/08/2006	16:28:43	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	16:29	
		ITP Check	0.079	16:29	
		Air Blank	0.000	16:30	

Control Tests

Instrument Serial #: 80-000830

Location	Date	Time	Version	Target	Source
HILLSBOROUGH SO	06/09/2006	22:31:07	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:31	
		Control Test	0.078	22:31	
		Air Blank	0.000	22:32	

Control Tests

Instrument Serial #: 80-000832

Location	Date	Time	Version	Target	Source
MARTIN COUNTY SO	09/09/2006	16:03:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:03	
		Control Test	0.082	16:03	
		Air Blank	0.000	16:04	
MARTIN COUNTY SO	09/09/2006	16:08:07	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:08	
		Control Test	0.081	16:08	
		Air Blank	0.000	16:09	

Control Tests

Instrument Serial #: 80-000840

Location	Date	Time	Version	Target	Source
LYNN HAVEN PD	07/13/2006	00:37:06	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:37	
		Control Test	0.073	00:37	
		Air Blank	0.000	00:38	
LYNN HAVEN PD	08/04/2006	18:18:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:19	
		Control Test	0.073	18:19	
		Air Blank	0.000	18:19	
LYNN HAVEN PD	08/04/2006	18:20:19	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:20	
		Control Test	0.072	18:21	
		Air Blank	0.000	18:21	
LYNN HAVEN PD	08/19/2006	11:53:11	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:53	
		Control Test	0.073	11:53	
		Air Blank	0.000	11:54	
LYNN HAVEN PD	08/19/2006	11:55:19	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:55	
		Control Test	0.073	11:56	
		Air Blank	0.000	11:56	
LYNN HAVEN PD	08/20/2006	10:40:37	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:41	
		Control Test	0.073	10:41	
		Air Blank	0.000	10:41	
LYNN HAVEN PD	09/18/2006	13:04:35	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:05	
		Control Test	0.080	13:05	
		Air Blank	0.000	13:05	
LYNN HAVEN PD	10/09/2006	11:35:02	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:35	
		Control Test	0.081	11:35	

Control Tests

Instrument Serial #: 80-000840

Location	Date	Time	Version	Target	Source
LYNN HAVEN PD	10/09/2006	11:35:02	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:36	
LYNN HAVEN PD	10/09/2006	11:39:32	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:39	
		Control Test	0.082	11:40	
		Air Blank	0.000	11:40	
LYNN HAVEN PD	10/09/2006	11:41:02	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:41	
		Control Test	0.081	11:41	
		Air Blank	0.000	11:42	
LYNN HAVEN PD	10/09/2006	11:43:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:44	
		Control Test	0.082	11:44	
		Air Blank	0.000	11:44	
LYNN HAVEN PD	10/09/2006	11:48:25	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:48	
		Control Test	0.082	11:49	
		Air Blank	0.000	11:49	

Control Tests

Instrument Serial #: 80-000859

Location	Date	Time	Version	Target	Source
CAPE CORAL PD	06/06/2006	16:05:40	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:06	
		Control Test	0.078	16:06	
		Air Blank	0.000	16:06	
CAPE CORAL PD	10/25/2006	11:18:37	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	11:19	
		ITP Check	0.082	11:19	
		Air Blank	0.000	11:19	

Control Tests

Instrument Serial #: 80-000861

Location	Date	Time	Version	Target	Source
SARASOTA COUNTY SO	12/12/2006	13:07:36	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		13:07	Sequence Aborted
		Air Blank	0.000	13:07	

Control Tests

Instrument Serial #: 80-000871

Location	Date	Time	Version	Target	Source
CORAL GABLES PD	03/09/2006	21:56:37	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:57	
		Control Test	0.079	21:57	
		Air Blank	0.000	21:57	
CORAL GABLES PD	03/09/2006	21:58:18	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	21:58	
		ITP Check	0.079	21:59	
		Air Blank	0.000	21:59	

Control Tests

Instrument Serial #: 80-000872

Location	Date	Time	Version	Target	Source
MIAMI PD	02/21/2006	19:32:17	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:32	
		Control Test	0.079	19:33	
		Air Blank	0.000	19:33	

Control Tests

Instrument Serial #: 80-000874

Location	Date	Time	Version	Target	Source
MIAMI PD	02/21/2006	19:32:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:33	
		Control Test	0.079	19:33	
		Air Blank	0.000	19:33	

Control Tests

Instrument Serial #: 80-000875

Location	Date	Time	Version	Target	Source
CALHOUN COUNTY SO	08/12/2006	09:00:28	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:00	
		Control Test	0.000	09:01	
		Air Blank	0.000	09:02	
CALHOUN COUNTY SO	08/12/2006	09:02:28	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	09:02	
		ITP Check	0.078	09:03	
		Air Blank	0.000	09:03	

Control Tests

Instrument Serial #: 80-000876

Location	Date	Time	Version	Target	Source
HIALEAH PD	03/08/2006	22:00:48	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:01	
		Control Test	0.080	22:01	
		Air Blank	0.000	22:02	
HIALEAH PD	03/08/2006	22:02:19	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	22:02	
		ITP Check	0.079	22:03	
		Air Blank	0.000	22:03	

Control Tests

Instrument Serial #: 80-000878

Location	Date	Time	Version	Target	Source
HOMESTEAD PD	03/09/2006	17:18:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:18	
		Control Test	0.079	17:19	
		Air Blank	0.000	17:19	

Control Tests

Instrument Serial #: 80-000879

Location	Date	Time	Version	Target	Source
MIAMI BEACH PD	03/09/2006	11:55:19	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	11:55	
		ITP Check	0.081	11:56	
		Air Blank	0.000	11:56	
MIAMI BEACH PD	04/09/2006	23:18:30	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:18	
		Control Test	0.078	23:19	
		Air Blank	0.000	23:19	
MIAMI BEACH PD	04/10/2006	02:43:34	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	02:44	
		Control Test	0.076	02:44	
		Air Blank	0.000	02:44	
MIAMI BEACH PD	07/29/2006	02:21:04	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	02:21	
		Control Test	0.080	02:21	
		Air Blank	0.000	02:22	

Control Tests

Instrument Serial #: 80-000881

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	02/21/2006	16:31:46	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:32	
		Control Test	0.078	16:32	
		Air Blank	0.000	16:33	
MIAMI-DADE PD	02/21/2006	16:33:40	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	16:34	
		ITP Check	0.080	16:34	
		Air Blank	0.000	16:35	
MIAMI-DADE PD	06/14/2006	23:05:49	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:06	
		Control Test	0.081	23:06	
		Air Blank	0.000	23:07	

Control Tests

Instrument Serial #: 80-000882

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	09/27/2006	14:01:40	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:02	
		Control Test	0.075	14:02	
		Air Blank	0.000	14:03	
MIAMI-DADE PD	09/27/2006	14:03:24	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:03	
		Control Test	0.074	14:04	
		Air Blank	0.000	14:05	
MIAMI-DADE PD	09/27/2006	14:05:28	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:05	
		Control Test	0.077	14:06	
		Air Blank	0.000	14:07	

Control Tests

Instrument Serial #: 80-000885

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	02/21/2006	16:31:34	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:32	
		Control Test	0.078	16:32	
		Air Blank	0.000	16:32	
MIAMI-DADE PD	02/21/2006	16:33:35	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	16:34	
		ITP Check	0.078	16:34	
		Air Blank	0.000	16:34	

Control Tests

Instrument Serial #: 80-000889

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	07/06/2006	21:51:37	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:52	
		Control Test	0.080	21:52	
		Air Blank	0.000	21:52	
PINELLAS COUNTY SO	12/21/2006	19:09:47	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:08	
		Control Test	0.081	19:09	
		Air Blank	0.000	19:09	

Control Tests

Instrument Serial #: 80-000890

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	06/15/2006	03:34:03	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	03:34	
		Control Test	0.080	03:34	
		Air Blank		03:35	Ambient Fail
		Air Blank	0.000	03:35	

Control Tests

Instrument Serial #: 80-000892

Location	Date	Time	Version	Target	Source
FFWCC (PANAMA CITY)	03/15/2006	16:31:39	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:32	
		Control Test	0.079	16:32	
		Air Blank	0.000	16:32	

Control Tests

Instrument Serial #: 80-000893

Location	Date	Time	Version	Target	Source
FFWCC (LAKE CITY)	05/22/2006	13:34:27	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:34	
		Control Test	0.080	13:35	
		Air Blank	0.000	13:35	
FFWCC (LAKE CITY)	05/22/2006	13:44:51	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:45	
		Control Test	0.000	13:46	
		Air Blank	0.000	13:46	

Control Tests

Instrument Serial #: 80-000900

Location	Date	Time	Version	Target	Source
FFWCC (LAKE CITY)	05/25/2006	13:23:52	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:24	
		Control Test	0.029	13:25	
		Air Blank	0.000	13:25	
FFWCC (LAKE CITY)	05/25/2006	13:27:48	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:28	
		Control Test	0.063	13:28	
		Air Blank	0.000	13:29	
FFWCC (LAKE CITY)	05/25/2006	13:29:28	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank		13:29	Ambient Fail
		Air Blank	0.000	13:30	
FFWCC (LAKE CITY)	05/25/2006	13:30:26	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank		13:30	Ambient Fail
		Air Blank		13:31	Purge Fail
FFWCC (LAKE CITY)	05/25/2006	13:31:44	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:32	
		Control Test	0.070	13:32	
		Air Blank	0.000	13:33	
FFWCC (LAKE CITY)	05/25/2006	13:34:04	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:34	
		Control Test	0.072	13:35	
		Air Blank	0.000	13:35	
FFWCC (LAKE CITY)	05/25/2006	13:35:47	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:36	
		Control Test	0.000	13:36	
		Air Blank	0.000	13:37	

Control Tests

Instrument Serial #: 80-000901

Location	Date	Time	Version	Target	Source			
FFWCC TAMPA	05/27/2006	15:38:28	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	15:38
						Control Test	0.071	15:39
Air Blank	0.000	15:39						
FFWCC TAMPA	05/27/2006	15:40:03	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	15:40
						Control Test	0.070	15:40
Air Blank	0.000	15:41						
FFWCC TAMPA	05/27/2006	15:42:50	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	15:43
						Control Test	0.071	15:43
Air Blank	0.000	15:44						
FFWCC TAMPA	05/27/2006	15:46:16	8100.26.00	0.080	Local ITP			
						Test	Value	Time
						Air Blank	0.000	15:46
						ITP Check	0.080	15:47
Air Blank	0.000	15:47						
FFWCC TAMPA	05/27/2006	15:49:05	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	15:49
						Control Test	0.072	15:49
Air Blank	0.000	15:50						
FFWCC TAMPA	05/27/2006	16:11:51	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	16:12
						Control Test	0.075	16:12
Air Blank	0.000	16:13						
FFWCC TAMPA	05/27/2006	16:16:31	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	16:16
						Control Test	0.076	16:17
Air Blank	0.000	16:17						
FFWCC TAMPA	05/27/2006	16:36:54	8100.26.00	0.080	Local Dry gas			
						Test	Value	Time
						Air Blank	0.000	16:37
						Control Test	0.077	16:37

Control Tests

Instrument Serial #: 80-000901

Location	Date	Time	Version	Target	Source
FFWCC TAMPA	05/27/2006	16:36:54	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:38	
FFWCC TAMPA	05/28/2006	13:56:52	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:57	
		Control Test	0.079	13:57	
		Air Blank	0.000	13:58	
FFWCC TAMPA	05/28/2006	18:29:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:30	
		Control Test	0.080	18:30	
		Air Blank	0.000	18:30	
FFWCC TAMPA	05/28/2006	19:03:39	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:04	
		Control Test	0.080	19:04	
		Air Blank	0.000	19:04	
FFWCC TAMPA	05/28/2006	22:33:17	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:33	
		Control Test	0.080	22:34	
		Air Blank	0.000	22:34	
FFWCC TAMPA	06/14/2006	17:34:51	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:35	
		Control Test	0.079	17:35	
		Air Blank	0.000	17:36	
FFWCC TAMPA	09/29/2006	13:22:39	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:23	
		Control Test	0.079	13:23	
		Air Blank	0.000	13:23	
FFWCC TAMPA	10/30/2006	11:49:51	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:50	
		Control Test	0.078	11:50	
		Air Blank	0.000	11:51	

Control Tests

Instrument Serial #: 80-000901

Location	Date	Time	Version	Target	Source
FFWCC TAMPA	11/03/2006	09:58:30	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:58	
		Control Test	0.079	09:59	
		Air Blank	0.000	09:59	
FFWCC TAMPA	12/08/2006	11:02:44	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:01	
		Control Test	0.081	11:02	
		Air Blank	0.000	11:02	

Control Tests

Instrument Serial #: 80-000902

Location	Date	Time	Version	Target	Source
FFWCC	05/27/2006	16:04:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:04	
		Control Test	0.074	16:05	
		Air Blank	0.000	16:05	
FFWCC	05/27/2006	16:14:24	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:14	
		Control Test	0.076	16:15	
		Air Blank	0.000	16:15	
FFWCC	05/27/2006	16:16:09	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:16	
		Control Test	0.075	16:16	
		Air Blank	0.000	16:17	
FFWCC	05/28/2006	13:58:46	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:59	
		Control Test	0.080	13:59	
		Air Blank	0.000	14:00	
FFWCC	05/28/2006	18:30:45	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:31	
		Control Test	0.079	18:31	
		Air Blank	0.000	18:32	
FFWCC	06/03/2006	09:37:51	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:38	
		Control Test	0.078	09:38	
		Air Blank	0.000	09:39	
FFWCC	06/03/2006	09:39:26	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:39	
		Control Test	0.079	09:40	
		Air Blank	0.000	09:40	
FFWCC	06/03/2006	16:22:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:22	
		Control Test	0.079	16:23	

Control Tests

Instrument Serial #: 80-000902

Location	Date	Time	Version	Target	Source
FFWCC	06/03/2006	16:22:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:23	
FFWCC	06/04/2006	13:58:08	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:58	
		Control Test	0.079	13:58	
		Air Blank	0.000	13:59	
FFWCC	06/14/2006	17:33:14	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:33	
		Control Test	0.079	17:34	
		Air Blank	0.000	17:34	
FFWCC	08/10/2006	10:58:47	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:59	
		Control Test	0.080	10:59	
		Air Blank	0.000	11:00	
FFWCC	09/29/2006	13:21:27	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:21	
		Control Test	0.081	13:22	
		Air Blank	0.000	13:22	
FFWCC	10/30/2006	11:44:39	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:45	
		Control Test	0.081	11:45	
		Air Blank	0.000	11:45	
FFWCC	11/03/2006	09:59:23	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:59	
		Control Test	0.081	10:00	
		Air Blank	0.000	10:00	
FFWCC	12/08/2006	11:10:27	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:09	
		Control Test	0.081	11:09	
		Air Blank	0.000	11:10	

Control Tests

Instrument Serial #: 80-000906

Location	Date	Time	Version	Target	Source
FISH & WILDLIFE CC	07/19/2006	14:53:51	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:54	
		Control Test	0.079	14:54	
		Air Blank	0.000	14:55	

Control Tests

Instrument Serial #: 80-000907

Location	Date	Time	Version	Target	Source
MANATEE COUNTY SO	03/27/2006	12:12:02	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:12	
		Control Test	0.078	12:12	
		Air Blank	0.000	12:13	
MANATEE COUNTY SO	08/11/2006	18:13:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:13	
		Control Test	0.079	18:13	
		Air Blank	0.000	18:14	

Control Tests

Instrument Serial #: 80-000909

Location	Date	Time	Version	Target	Source
PINECREST PD	03/09/2006	17:18:52	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	17:19	
		ITP Check	0.078	17:19	
		Air Blank	0.000	17:20	
PINECREST PD	08/18/2006	13:09:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:09	
		Control Test	0.080	13:10	
		Air Blank	0.000	13:10	
PINECREST PD	08/18/2006	13:23:14	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:23	
		Control Test	0.309	13:24	
		Air Blank	0.000	13:24	

Control Tests

Instrument Serial #: 80-000912

Location	Date	Time	Version	Target	Source
MANATEE COUNTY SO	07/03/2006	01:04:57	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:05	
		Control Test	0.079	01:05	
		Air Blank	0.000	01:06	

Control Tests

Instrument Serial #: 80-000913

Location	Date	Time	Version	Target	Source
MANATEE COUNTY SO	10/20/2006	23:37:19	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:37	
		Control Test	0.080	23:38	
		Air Blank	0.000	23:38	

Control Tests

Instrument Serial #: 80-000914

Location	Date	Time	Version	Target	Source
FORT MYERS PD	06/06/2006	08:53:44	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:54	
		Control Test	0.077	08:54	
		Air Blank	0.000	08:55	

Control Tests

Instrument Serial #: 80-000937

Location	Date	Time	Version	Target	Source
CMI INC	01/18/2006	15:17:45	8100.26.00	0.000	Local Dry gas
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
<hr/>					
SANIBEL PD	02/16/2006	16:13:46	8100.26.00	0.080	Local Dry gas
		<u>Test</u>	<u>Value</u>	<u>Time</u>	
		Air Blank	0.000	16:14	
		Control Test	0.077	16:14	
		Air Blank	0.000	16:15	

Control Tests

Instrument Serial #: 80-000942

Location	Date	Time	Version	Target	Source
COLLIER COUNTY SO	10/05/2006	22:05:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:05	
		Control Test	0.078	22:05	
		Air Blank	0.000	22:06	
COLLIER COUNTY SO	10/16/2006	00:43:00	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:43	
		Control Test	0.078	00:43	
		Air Blank	0.000	00:44	
COLLIER COUNTY SO	10/16/2006	00:44:21	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:44	
		Control Test	0.078	00:45	
		Air Blank	0.000	00:45	

Control Tests

Instrument Serial #: 80-000943

Location	Date	Time	Version	Target	Source
COLLIER COUNTY SO	05/24/2006	11:16:42	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:17	
		Control Test	0.079	11:17	
		Air Blank	0.000	11:17	
COLLIER COUNTY SO	05/24/2006	16:47:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:47	
		Control Test	0.080	16:48	
		Air Blank	0.000	16:48	

Control Tests

Instrument Serial #: 80-000944

Location	Date	Time	Version	Target	Source
COLLIER COUNTY SO	08/08/2006	04:33:33	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:34	
		Control Test	0.075	04:34	
		Air Blank	0.000	04:34	
COLLIER COUNTY SO	08/08/2006	04:35:12	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:35	
		Control Test	0.077	04:35	
		Air Blank	0.000	04:36	
COLLIER COUNTY SO	08/08/2006	04:36:37	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:37	
		Control Test	0.077	04:37	
		Air Blank	0.000	04:37	
COLLIER COUNTY SO	08/08/2006	04:37:57	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:38	
		Control Test	0.077	04:38	
		Air Blank	0.000	04:39	
COLLIER COUNTY SO	08/08/2006	04:39:16	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:39	
		Control Test	0.077	04:40	
		Air Blank	0.000	04:40	
COLLIER COUNTY SO	11/13/2006	14:00:08	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:58	
		Control Test	0.074	13:59	
		Air Blank	0.000	14:00	

Control Tests

Instrument Serial #: 80-000945

Location	Date	Time	Version	Target	Source
CHARLOTTE COUNTY SO	06/09/2006	14:45:43	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:46	
		Control Test	0.079	14:46	
		Air Blank	0.000	14:47	
CHARLOTTE COUNTY SO	06/09/2006	14:47:03	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:47	
		Control Test	0.079	14:47	
		Air Blank	0.000	14:48	
CHARLOTTE COUNTY SO	06/09/2006	14:48:20	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		14:48	Ambient Fail
		Air Blank		14:49	Purge Fail
CHARLOTTE COUNTY SO	07/07/2006	06:02:32	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	06:02	
		Control Test	0.080	06:03	
		Air Blank	0.000	06:03	
CHARLOTTE COUNTY SO	07/07/2006	06:03:57	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	06:04	
		Control Test	0.080	06:04	
		Air Blank	0.000	06:05	

Control Tests

Instrument Serial #: 80-000946

Location	Date	Time	Version	Target	Source
CHARLOTTE COUNTY SO	11/10/2006	01:24:35	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:23	
		Control Test	0.080	01:24	
		Air Blank	0.000	01:24	
CHARLOTTE COUNTY SO	11/18/2006	23:00:48	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:59	
		Control Test	0.080	23:00	
		Air Blank	0.000	23:00	
CHARLOTTE COUNTY SO	11/18/2006	23:08:54	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	23:08	
		ITP Check	0.085	23:08	
		Air Blank	0.000	23:08	
CHARLOTTE COUNTY SO	11/18/2006	23:57:51	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:57	
		Control Test	0.082	23:57	
		Air Blank	0.000	23:57	
CHARLOTTE COUNTY SO	11/19/2006	00:04:03	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	00:03	
		ITP Check	0.080	00:03	
		Air Blank	0.000	00:04	
CHARLOTTE COUNTY SO	11/19/2006	00:05:32	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	00:04	
		ITP Check	0.079	00:05	
		Air Blank	0.000	00:05	
CHARLOTTE COUNTY SO	11/19/2006	00:07:40	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	00:06	
		ITP Check	0.079	00:07	
		Air Blank	0.000	00:07	
CHARLOTTE COUNTY SO	11/19/2006	00:09:47	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank		00:09	RFI Detect
		Air Blank	0.000	00:09	

Control Tests

Instrument Serial #: 80-000947

Location	Date	Time	Version	Target	Source
GLADES COUNTY SO	09/12/2006	18:18:09	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	18:18	
		ITP Check	0.079	18:19	
		Air Blank	0.000	18:19	

Control Tests

Instrument Serial #: 80-000949

Location	Date	Time	Version	Target	Source
TALLAHASSEE PD	10/26/2006	14:03:09	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:03	
		Control Test	0.082	14:03	
		Air Blank	0.000	14:04	

Control Tests

Instrument Serial #: 80-000951

Location	Date	Time	Version	Target	Source
HENDRY COUNTY SO	03/07/2006	17:15:56	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:16	
		Control Test	0.079	17:16	
		Air Blank	0.000	17:17	

Control Tests

Instrument Serial #: 80-000952

Location	Date	Time	Version	Target	Source
FRANKLIN COUNTY SO	03/30/2006	13:49:09	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:49	
		Control Test	0.081	13:49	
		Air Blank	0.000	13:50	

Control Tests

Instrument Serial #: 80-000954

Location	Date	Time	Version	Target	Source
JEFFERSON COUNTY SO	04/18/2006	09:45:13	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:45	
		Control Test	0.078	09:45	
		Air Blank	0.000	09:46	
JEFFERSON COUNTY SO	04/18/2006	09:47:12	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:47	
		Control Test	0.079	09:47	
		Air Blank	0.000	09:48	
JEFFERSON COUNTY SO	04/18/2006	09:48:46	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:49	
		Control Test	0.079	09:49	
		Air Blank	0.000	09:50	
JEFFERSON COUNTY SO	04/18/2006	09:51:47	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:52	
		Control Test	0.156	09:52	
		Air Blank	0.000	09:53	
JEFFERSON COUNTY SO	04/18/2006	09:53:31	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:53	
		Control Test	0.164	09:54	
		Air Blank	0.000	09:55	
JEFFERSON COUNTY SO	04/18/2006	09:57:03	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:57	
		Control Test	0.173	09:58	
		Air Blank	0.000	09:58	
JEFFERSON COUNTY SO	04/18/2006	09:59:11	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:59	
		Control Test	0.178	10:00	
		Air Blank	0.000	10:00	
JEFFERSON COUNTY SO	04/18/2006	10:00:54	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:01	
		Control Test	0.181	10:02	

Control Tests

Instrument Serial #: 80-000954

Location	Date	Time	Version	Target	Source
JEFFERSON COUNTY SO	04/18/2006	10:00:54	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:02	
JEFFERSON COUNTY SO	04/18/2006	10:03:35	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:04	
		Control Test	0.151	10:04	
		Air Blank		10:05	RFI Detect
		Air Blank	0.000	10:05	
JEFFERSON COUNTY SO	04/18/2006	10:05:47	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:06	
		Control Test	0.103	10:06	
		Air Blank	0.000	10:07	
JEFFERSON COUNTY SO	04/18/2006	10:07:43	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:08	
		Control Test	0.097	10:08	
		Air Blank	0.000	10:09	
JEFFERSON COUNTY SO	04/18/2006	10:09:46	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:10	
		Control Test	0.097	10:10	
		Air Blank	0.000	10:11	
JEFFERSON COUNTY SO	04/18/2006	10:13:06	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:13	
		Control Test	0.050	10:14	
		Air Blank	0.000	10:14	
JEFFERSON COUNTY SO	04/18/2006	10:15:26	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:15	
		Control Test	0.095	10:16	
		Air Blank	0.000	10:17	
JEFFERSON COUNTY SO	04/18/2006	10:17:21	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:17	
		Control Test	0.180	10:18	
		Air Blank	0.000	10:19	

Control Tests

Instrument Serial #: 80-000955

Location	Date	Time	Version	Target	Source
CLEWISTON PD	03/07/2006	17:15:44	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	17:16	
		ITP Check	0.079	17:16	
		Air Blank	0.000	17:17	

Control Tests

Instrument Serial #: 80-000957

Location	Date	Time	Version	Target	Source
LEON COUNTY SO	03/27/2006	17:21:50	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:22	
		Control Test	0.080	17:22	
		Air Blank	0.000	17:23	

Control Tests

Instrument Serial #: 80-000958

Location	Date	Time	Version	Target	Source
LEON COUNTY SO	03/27/2006	17:21:45	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:22	
		Control Test	0.080	17:22	
		Air Blank	0.000	17:23	

Control Tests

Instrument Serial #: 80-000959

Location	Date	Time	Version	Target	Source
GADSDEN COUNTY SO	03/31/2006	11:38:20	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:38	
		Control Test	0.078	11:39	
		Air Blank	0.000	11:39	

Control Tests

Instrument Serial #: 80-000960

Location	Date	Time	Version	Target	Source
LIBERTY COUNTY SO	03/29/2006	13:31:01	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:31	
		Control Test	0.080	13:31	
		Air Blank	0.000	13:32	

Control Tests

Instrument Serial #: 80-000961

Location	Date	Time	Version	Target	Source
WAKULLA COUNTY SO	03/30/2006	09:55:35	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:56	
		Control Test	0.077	09:56	
		Air Blank	0.000	09:56	
WAKULLA COUNTY SO	04/03/2006	10:34:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:34	
		Control Test	0.076	10:35	
		Air Blank	0.000	10:35	
WAKULLA COUNTY SO	04/03/2006	17:59:19	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:59	
		Control Test	0.077	18:00	
		Air Blank	0.000	18:00	
WAKULLA COUNTY SO	06/05/2006	11:36:32	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:36	
		Control Test	0.048	11:37	
		Air Blank		11:37	Sequence Aborted
		Air Blank	0.000	11:38	

Control Tests

Instrument Serial #: 80-000998

Location	Date	Time	Version	Target	Source
SEMINOLE SO	05/06/2006	00:22:44	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:23	
		Control Test	0.078	00:23	
		Air Blank	0.000	00:24	
SEMINOLE SO	05/06/2006	02:06:10	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	02:06	
		ITP Check	0.081	02:07	
		Air Blank	0.000	02:07	
SEMINOLE SO	05/26/2006	21:44:20	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:44	
		Control Test	0.078	21:45	
		Air Blank	0.000	21:45	
SEMINOLE SO	05/26/2006	21:46:02	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	21:46	
		ITP Check	0.081	21:46	
		Air Blank	0.000	21:47	

Control Tests

Instrument Serial #: 80-000999

Location	Date	Time	Version	Target	Source
SEMINOLE SO	10/15/2006	00:24:40	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:25	
		Control Test	0.081	00:25	
		Air Blank		00:25	RFI Detect
		Air Blank	0.000	00:26	
SEMINOLE SO	10/15/2006	00:26:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:26	
		Control Test	0.081	00:27	
		Air Blank	0.000	00:27	
SEMINOLE S.O.	11/07/2006	21:22:28	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:21	
		Control Test	0.081	21:21	
		Air Blank	0.000	21:22	
SEMINOLE S.O.	11/15/2006	23:33:14	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	23:32	
		ITP Check	0.080	23:32	
		Air Blank	0.000	23:33	
SEMINOLE S.O.	11/15/2006	23:42:40	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	23:41	
		ITP Check	0.081	23:42	
		Air Blank	0.000	23:42	
SEMINOLE S.O.	12/29/2006	22:17:42	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:16	
		Control Test	0.079	22:17	
		Air Blank	0.000	22:17	

Control Tests

Instrument Serial #: 80-001005

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	07/29/2006	22:43:18	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:43	
		Control Test	0.079	22:44	
		Air Blank	0.000	22:44	
PINELLAS COUNTY SO	07/29/2006	22:44:49	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	22:45	
		ITP Check	0.079	22:45	
		Air Blank	0.000	22:46	

Control Tests

Instrument Serial #: 80-001006

Location	Date	Time	Version	Target	Source
INDIALANTIC PD	03/21/2006	13:42:35	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:43	
		Control Test	0.079	13:43	
		Air Blank	0.000	13:43	
INDIALANTIC PD	12/07/2006	07:01:05	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	07:00	
		Control Test	0.079	07:00	
		Air Blank	0.000	07:01	

Control Tests

Instrument Serial #: 80-001007

Location	Date	Time	Version	Target	Source
HAINES CITY PD	10/12/2006	10:29:17	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:29	
		Control Test	0.000	10:30	
		Air Blank	0.000	10:30	

Control Tests

Instrument Serial #: 80-001018

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	03/27/2006	15:53:07	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:53	
		Control Test	0.078	15:53	
		Air Blank	0.000	15:54	

Control Tests

Instrument Serial #: 80-001043

Location	Date	Time	Version	Target	Source
HIGHLANDS COUNTY SO	03/16/2006	17:25:28	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:25	
		Control Test	0.079	17:26	
		Air Blank	0.000	17:26	

Control Tests

Instrument Serial #: 80-001045

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	05/12/2006	07:00:31	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	07:00	
		Control Test	0.079	07:01	
		Air Blank	0.000	07:01	
JACKSONVILLE SO	07/19/2006	12:36:52	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:37	
		Control Test	0.080	12:37	
		Air Blank	0.000	12:38	

Control Tests

Instrument Serial #: 80-001048

Location	Date	Time	Version	Target	Source
LAKE ALFRED PD	11/08/2006	13:10:42	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:09	
		Control Test	0.080	13:10	
		Air Blank	0.000	13:10	

Control Tests

Instrument Serial #: 80-001051

Location	Date	Time	Version	Target	Source
ST PETERSBURG PD	11/01/2006	16:55:39	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:54	
		Control Test	0.195	16:55	
		Air Blank	0.000	16:55	

Control Tests

Instrument Serial #: 80-001052

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	03/27/2006	15:52:22	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:52	
		Control Test	0.079	15:53	
		Air Blank	0.000	15:53	
JACKSONVILLE SO	03/30/2006	06:10:45	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	06:11	
		Control Test	0.078	06:11	
		Air Blank	0.000	06:12	

Control Tests

Instrument Serial #: 80-001053

Location	Date	Time	Version	Target	Source
NASSAU COUNTY SO	04/18/2006	14:45:21	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:45	
		Control Test	0.095	14:46	
		Air Blank	0.000	14:46	
CMI INC	04/19/2006	13:53:55	8100.26.00	0.000	Local ITP
		Test	Value	Time	
		Air Blank	0.000	13:54	
		ITP Check	0.080	13:54	
		Air Blank	0.000	13:55	

Control Tests

Instrument Serial #: 80-001056

Location	Date	Time	Version	Target	Source
DAVIE PD	04/24/2006	18:38:31	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	18:38	
		ITP Check	0.079	18:39	
		Air Blank	0.000	18:39	
DAVIE PD	04/25/2006	14:55:55	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:56	
		Control Test	0.075	14:57	
		Air Blank	0.000	14:57	
DAVIE PD	04/25/2006	15:17:12	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:17	
		Control Test	0.077	15:18	
		Air Blank	0.000	15:18	
DAVIE PD	05/24/2006	15:35:23	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:35	
		Control Test	0.198	15:36	
		Air Blank	0.000	15:37	
DAVIE PD	11/22/2006	17:42:11	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:41	
		Control Test	0.080	17:41	
		Air Blank	0.000	17:42	

Control Tests

Instrument Serial #: 80-001057

Location	Date	Time	Version	Target	Source
LAKELAND PD	08/31/2006	09:26:42	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:27	
		Control Test	0.083	09:27	
		Air Blank	0.000	09:27	
LAKELAND PD	08/31/2006	09:29:28	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	09:29	
		ITP Check	0.081	09:30	
		Air Blank	0.000	09:30	
LAKELAND PD	08/31/2006	09:36:04	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:36	
		Control Test	0.079	09:37	
		Air Blank	0.000	09:37	
LAKELAND PD	08/31/2006	09:38:40	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:39	
		Control Test	0.199	09:39	
		Air Blank	0.000	09:40	

Control Tests

Instrument Serial #: 80-001058

Location	Date	Time	Version	Target	Source
LAKE WALES PD	07/18/2006	09:27:10	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:27	
		Control Test	0.050	09:28	
		Air Blank	0.000	09:28	

Control Tests

Instrument Serial #: 80-001064

Location	Date	Time	Version	Target	Source
NEW PORT RICHEY PD	09/20/2006	14:22:59	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:23	
		Control Test	0.000	14:24	
		Air Blank	0.000	14:24	

Control Tests

Instrument Serial #: 80-001065

Location	Date	Time	Version	Target	Source
TARPON SPRINGS PD	03/07/2006	14:10:15	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	14:10	
		ITP Check	0.080	14:11	
		Air Blank	0.000	14:11	
TARPON SPRINGS PD	03/07/2006	14:14:31	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:14	
		Control Test	0.082	14:15	
		Air Blank	0.000	14:15	

Control Tests

Instrument Serial #: 80-001066

Location	Date	Time	Version	Target	Source
ST PETE BEACH PD	03/02/2006	10:53:48	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	10:54	
		Control Test	0.08	10:54	
		Air Blank		10:55	Purge Fail
		Air Blank	0.00	10:55	
ST PETE BEACH PD	03/02/2006	10:56:58	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	10:57	
		Control Test	0.08	10:57	
		Air Blank	0.00	10:58	
ST PETE BEACH PD	03/02/2006	10:59:21	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	10:59	
		Control Test	0.08	11:00	
		Air Blank	0.00	11:00	
ST PETE BEACH PD	03/02/2006	11:01:17	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:01	
		Control Test	0.08	11:02	
		Air Blank	0.00	11:02	
ST PETE BEACH PD	03/02/2006	11:06:46	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:07	
		Control Test	0.08	11:07	
		Air Blank	0.00	11:08	
ST PETE BEACH PD	03/02/2006	11:08:11	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:08	
		Control Test	0.08	11:08	
		Air Blank	0.00	11:09	
ST PETE BEACH PD	03/02/2006	11:10:30	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:10	
		Control Test	0.08	11:11	
		Air Blank	0.00	11:11	
ST PETE BEACH PD	03/02/2006	11:13:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:13	

Control Tests

Instrument Serial #: 80-001066

Location	Date	Time	Version	Target	Source
ST PETE BEACH PD	03/02/2006	11:13:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Control Test	0.08	11:13	
		Air Blank	0.00	11:14	
ST PETE BEACH PD	03/02/2006	11:14:42	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:15	
		Control Test	0.08	11:15	
		Air Blank	0.00	11:15	
ST PETE BEACH PD	03/02/2006	11:16:08	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.00	11:16	
		Control Test	0.08	11:16	
		Air Blank	0.00	11:17	
CMI INC	03/20/2006	10:54:08	8100.26.00	0.000	Local ITP
		Test	Value	Time	
		Air Blank	0.000	10:54	
		ITP Check	0.080	10:55	
		Air Blank	0.000	10:55	
CMI INC	03/28/2006	11:48:33	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	11:49	
		ITP Check	0.080	11:49	
		Air Blank	0.000	11:49	

Control Tests

Instrument Serial #: 80-001067

Location	Date	Time	Version	Target	Source
BROWARD COUNTY SO	04/22/2006	01:27:12	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:27	
		Control Test	0.080	01:27	
		Air Blank	0.000	01:28	
BROWARD COUNTY SO	04/22/2006	01:34:16	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	01:34	
		ITP Check	0.079	01:35	
		Air Blank	0.000	01:35	
BROWARD COUNTY SO	06/30/2006	11:57:57	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:58	
		Control Test	0.081	11:58	
		Air Blank	0.000	11:59	
BROWARD COUNTY SO	07/05/2006	10:13:02	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:13	
		Control Test	0.081	10:13	
		Air Blank	0.000	10:14	

Control Tests

Instrument Serial #: 80-001068

Location	Date	Time	Version	Target	Source
CLEARWATER PD	05/03/2006	11:58:37	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	11:59	
		ITP Check	0.080	11:59	
		Air Blank	0.000	11:59	
CLEARWATER PD	05/03/2006	12:17:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:17	
		Control Test	0.081	12:18	
		Air Blank	0.000	12:18	

Control Tests

Instrument Serial #: 80-001069

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	03/27/2006	16:45:37	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:46	
		Control Test	0.079	16:46	
		Air Blank	0.000	16:46	

Control Tests

Instrument Serial #: 80-001071

Location	Date	Time	Version	Target	Source
CLAY COUNTY SO	05/06/2006	11:08:11	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:08	
		Control Test	0.079	11:08	
		Air Blank	0.000	11:09	
CLAY COUNTY SO	06/03/2006	03:30:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	03:30	
		Control Test	0.078	03:31	
		Air Blank	0.000	03:31	
CLAY COUNTY SO	07/10/2006	21:54:11	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:54	
		Control Test	0.079	21:54	
		Air Blank	0.000	21:55	
CLAY COUNTY SO	11/18/2006	21:57:09	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:56	
		Control Test	0.080	21:56	
		Air Blank	0.000	21:57	

Control Tests

Instrument Serial #: 80-001074

Location	Date	Time	Version	Target	Source
PASCO COUNTY SO	04/06/2006	17:54:29	8100.26.00	3.809	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:54	
		Control Test	0.078	17:55	
		Air Blank	0.000	17:55	

Control Tests

Instrument Serial #: 80-001079

Location	Date	Time	Version	Target	Source
LARGO PD	08/31/2006	18:18:46	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	18:19	
		ITP Check	0.077	18:19	
		Air Blank	0.000	18:20	

Control Tests

Instrument Serial #: 80-001082

Location	Date	Time	Version	Target	Source
PLANTATION PD	04/06/2006	22:05:10	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:05	
		Control Test	0.075	22:05	
		Air Blank	0.000	22:06	
PLANTATION PD	04/06/2006	22:08:46	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:09	
		Control Test	0.076	22:09	
		Air Blank	0.000	22:10	
PLANTATION PD	04/06/2006	22:10:47	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:11	
		Control Test	0.076	22:11	
		Air Blank	0.000	22:12	
PLANTATION PD	04/06/2006	22:12:09	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:12	
		Control Test	0.076	22:12	
		Air Blank	0.000	22:13	
PLANTATION PD	04/06/2006	22:13:34	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:14	
		Control Test	0.078	22:14	
		Air Blank	0.000	22:14	
PLANTATION PD	04/06/2006	22:14:54	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		22:15	Ambient Fail
		Air Blank	0.000	22:15	
PLANTATION PD	04/06/2006	22:15:52	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		22:16	Ambient Fail
		Air Blank		22:17	Purge Fail
PLANTATION PD	04/06/2006	22:17:11	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:17	
		Control Test	0.076	22:17	
		Air Blank	0.000	22:18	

Control Tests

Instrument Serial #: 80-001082

Location	Date	Time	Version	Target	Source
PLANTATION PD	04/06/2006	22:18:49	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:19	
		Control Test	0.076	22:19	
		Air Blank	0.000	22:20	
PLANTATION PD	04/06/2006	22:20:11	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:20	
		Control Test	0.076	22:20	
		Air Blank	0.000	22:21	
PLANTATION PD	11/08/2006	21:29:48	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:28	
		Control Test	0.080	21:29	
		Air Blank	0.000	21:29	

Control Tests

Instrument Serial #: 80-001083

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	05/15/2006	08:38:58	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:39	
		Control Test	0.075	08:39	
		Air Blank	0.000	08:40	
JACKSONVILLE SO	07/19/2006	12:34:53	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:35	
		Control Test	0.082	12:35	
		Air Blank	0.000	12:36	

Control Tests

Instrument Serial #: 80-001111

Location	Date	Time	Version	Target	Source
FHP TROOP A	04/28/2006	14:26:33	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:27	
		Control Test	0.047	14:27	
		Air Blank	0.000	14:28	

Control Tests

Instrument Serial #: 80-001112

Location	Date	Time	Version	Target	Source
FL HIGHWAY PATROL	03/24/2006	10:25:18	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:25	
		Control Test	0.000	10:26	Interferent Detect
		Air Blank	0.000	10:26	
FL HIGHWAY PATROL	06/02/2006	12:41:09	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:41	
		Control Test	0.080	12:41	
		Air Blank	0.000	12:42	
FL HIGHWAY PATROL	06/02/2006	12:43:48	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:44	
		Control Test	0.079	12:44	
		Air Blank	0.000	12:45	

Control Tests

Instrument Serial #: 80-001115

Location	Date	Time	Version	Target	Source
FHP TROOP H	12/28/2006	14:27:16	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:26	
		Control Test	0.067	14:26	
		Air Blank	0.000	14:27	

Control Tests

Instrument Serial #: 80-001118

Location	Date	Time	Version	Target	Source
FL HIGHWAY PATROL	08/30/2006	11:54:18	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:54	
		Control Test	0.199	11:55	
		Air Blank	0.000	11:55	
FL HIGHWAY PATROL	08/30/2006	11:58:10	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:58	
		Control Test	0.080	11:59	
		Air Blank	0.000	11:59	

Control Tests

Instrument Serial #: 80-001123

Location	Date	Time	Version	Target	Source
FHP TROOP F FT MYERS	03/20/2006	15:12:27	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:12	
		Control Test	0.077	15:13	
		Air Blank	0.000	15:13	
FHP TROOP F FT MYERS	03/20/2006	15:13:55	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	15:14	
		ITP Check	0.078	15:14	
		Air Blank	0.000	15:15	

Control Tests

Instrument Serial #: 80-001139

Location	Date	Time	Version	Target	Source
BROWARD COUNTY SO	06/30/2006	11:59:33	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:00	
		Control Test	0.079	12:00	
		Air Blank	0.000	12:00	
BROWARD COUNTY SO	06/30/2006	12:01:23	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:01	
		Control Test	0.079	12:02	
		Air Blank	0.000	12:02	
BROWARD COUNTY SO	09/07/2006	09:14:25	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:14	
		Control Test	0.080	09:15	
		Air Blank	0.000	09:15	
BROWARD COUNTY SO	09/07/2006	09:16:01	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:16	
		Control Test	0.080	09:16	
		Air Blank	0.000	09:17	
BROWARD COUNTY SO	11/07/2006	11:17:57	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:17	
		Control Test	0.080	11:17	
		Air Blank	0.000	11:17	

Control Tests

Instrument Serial #: 80-001150

Location	Date	Time	Version	Target	Source
MANATEE COUNTY SO	07/29/2006	00:25:06	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:25	
		Control Test	0.076	00:25	
		Air Blank	0.000	00:26	
MANATEE COUNTY SO	07/29/2006	00:26:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:26	
		Control Test	0.076	00:27	
		Air Blank	0.000	00:27	
MANATEE COUNTY SO	12/30/2006	01:18:03	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:17	
		Control Test	0.076	01:17	
		Air Blank	0.000	01:18	

Control Tests

Instrument Serial #: 80-001157

Location	Date	Time	Version	Target	Source
LAKE WALES PD	07/18/2006	09:15:24	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:15	
		Control Test	0.047	09:16	
		Air Blank	0.000	09:17	

Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	11/06/2006	10:48:38	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:47	
		Control Test	0.000	10:48	
		Air Blank	0.000	10:48	
PONCE INLET P.D.	11/17/2006	20:13:02	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:12	
		Control Test	0.081	20:12	
		Air Blank	0.000	20:13	
PONCE INLET P.D.	11/17/2006	20:14:38	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	20:13	
		ITP Check	0.077	20:14	
		Air Blank	0.000	20:14	
PONCE INLET P.D.	12/23/2006	16:21:04	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:20	
		Control Test	0.080	16:20	
		Air Blank	0.000	16:21	

Control Tests

Instrument Serial #: 80-001174

Location	Date	Time	Version	Target	Source
CLERMONT PD	05/05/2006	17:31:42	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	17:32	
		ITP Check	0.079	17:32	
		Air Blank	0.000	17:33	

Control Tests

Instrument Serial #: 80-001177

Location	Date	Time	Version	Target	Source
CMI INC	03/30/2006	13:48:50	8100.26.00	0.000	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:49	
		Control Test	0.000	13:49	
		Air Blank		13:49	Sequence Aborted
		Air Blank	0.000	13:50	

Control Tests

Instrument Serial #: 80-001178

Location	Date	Time	Version	Target	Source
MIAMI PD	05/16/2006	19:57:40	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	19:58	
		Control Test	0.046	19:58	
		Air Blank	0.000	19:59	

Control Tests

Instrument Serial #: 80-001179

Location	Date	Time	Version	Target	Source
MIAMI PD	04/05/2006	21:27:28	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	21:27	
		ITP Check	0.100	21:28	
		Air Blank	0.000	21:28	

Control Tests

Instrument Serial #: 80-001202

Location	Date	Time	Version	Target	Source
COLLIER COUNTY SO	03/30/2006	10:03:23	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:03	
		Control Test	0.079	10:04	
		Air Blank	0.000	10:04	
COLLIER COUNTY SO	09/05/2006	18:31:00	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:31	
		Control Test	0.076	18:31	
		Air Blank	0.000	18:32	
COLLIER COUNTY SO	09/05/2006	18:32:33	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:33	
		Control Test	0.077	18:33	
		Air Blank	0.000	18:33	
COLLIER COUNTY SO	09/05/2006	18:34:08	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:34	
		Control Test	0.077	18:34	
		Air Blank	0.000	18:35	
COLLIER COUNTY SO	09/05/2006	18:35:44	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:36	
		Control Test	0.077	18:36	
		Air Blank	0.000	18:37	

Control Tests

Instrument Serial #: 80-001233

Location	Date	Time	Version	Target	Source
PALM BEACH SO BG	04/11/2006	13:40:11	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:40	
		Control Test	0.046	13:41	
		Air Blank	0.000	13:41	

Control Tests

Instrument Serial #: 80-001245

Location	Date	Time	Version	Target	Source
PORT ORANGE PD	08/10/2006	11:58:26	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:58	
		Control Test	0.077	11:59	
		Air Blank	0.000	12:00	

Control Tests

Instrument Serial #: 80-001247

Location	Date	Time	Version	Target	Source
BROWARD COUNTY SO	11/17/2006	12:32:58	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:32	
		Control Test	0.080	12:32	
		Air Blank	0.000	12:32	

Control Tests

Instrument Serial #: 80-001248

Location	Date	Time	Version	Target	Source
BROWARD COUNTY SO	04/06/2006	15:02:18	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	15:02	
		ITP Check	0.079	15:03	
		Air Blank	0.000	15:03	
BROWARD COUNTY SO	04/19/2006	20:53:55	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	20:54	
		ITP Check	0.079	20:54	
		Air Blank	0.000	20:55	
BROWARD COUNTY SO	04/22/2006	01:31:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:31	
		Control Test	0.079	01:32	
		Air Blank	0.000	01:32	
BROWARD COUNTY SO	09/25/2006	17:33:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:33	
		Control Test	0.080	17:34	
		Air Blank	0.000	17:34	

Control Tests

Instrument Serial #: 80-001256

Location	Date	Time	Version	Target	Source
ORANGE COUNTY SO	04/06/2006	09:25:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:25	
		Control Test	0.079	09:26	
		Air Blank	0.000	09:26	

Control Tests

Instrument Serial #: 80-001259

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	04/15/2006	01:03:00	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:03	
		Control Test	0.075	01:03	
PINELLAS COUNTY SO	06/16/2006	04:09:35	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	04:10	
		Control Test	0.075	04:10	
PINELLAS COUNTY SO	06/16/2006	04:11:03	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	04:11	
		ITP Check	0.079	04:11	
PINELLAS COUNTY SO	08/26/2006	20:12:32	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:12	
		Control Test	0.076	20:13	
PINELLAS COUNTY SO	08/30/2006	14:32:16	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:32	
		Control Test	0.076	14:33	
PINELLAS COUNTY SO	12/03/2006	01:02:29	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:01	
		Control Test	0.077	01:02	
PINELLAS COUNTY SO	12/03/2006	01:36:42	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	01:35	
		ITP Check	0.079	01:36	
PINELLAS COUNTY SO	12/03/2006	01:36:42	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	01:36	

Control Tests

Instrument Serial #: 80-001264

Location	Date	Time	Version	Target	Source
CMI INC	03/07/2006	10:43:02	8100.26.00	0.000	Local ITP
		Test	Value	Time	
		Air Blank	0.000	10:43	
		ITP Check	0.080	10:43	
		Air Blank	0.000	10:44	
MELBOURNE BEACH PD	10/11/2006	13:55:39	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:56	
		Control Test	0.075	13:56	
		Air Blank	0.000	13:56	

Control Tests

Instrument Serial #: 80-001271

Location	Date	Time	Version	Target	Source
SEMINOLE SO	06/23/2006	12:49:26	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:49	
		Control Test	0.075	12:50	
		Air Blank	0.000	12:50	
SEMINOLE SO	07/28/2006	22:19:34	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	22:20	
		Control Test	0.077	22:20	
		Air Blank	0.000	22:20	
SEMINOLE SO	08/08/2006	09:52:34	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:53	
		Control Test	0.077	09:53	
		Air Blank	0.000	09:53	

Control Tests

Instrument Serial #: 80-001272

Location	Date	Time	Version	Target	Source
SEMINOLE SO	04/01/2006	20:51:25	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:51	
		Control Test	0.078	20:52	
		Air Blank	0.000	20:52	
SEMINOLE SO	04/15/2006	21:21:28	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:21	
		Control Test	0.079	21:22	
		Air Blank	0.000	21:22	
SEMINOLE SO	04/15/2006	21:23:35	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	21:24	
		ITP Check	0.079	21:24	
		Air Blank	0.000	21:24	
SEMINOLE SO	06/25/2006	02:32:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	02:33	
		Control Test	0.079	02:33	
		Air Blank	0.000	02:33	
SEMINOLE SO	07/08/2006	00:57:51	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:58	
		Control Test	0.079	00:58	
		Air Blank	0.000	00:59	
SEMINOLE S.O.	11/14/2006	13:24:47	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:23	
		Control Test	0.080	13:24	
		Air Blank	0.000	13:24	

Control Tests

Instrument Serial #: 80-001273

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	04/15/2006	01:38:40	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:39	
		Control Test	0.078	01:39	
		Air Blank	0.000	01:39	
PINELLAS COUNTY SO	05/12/2006	19:38:52	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:39	
		Control Test	0.077	19:39	
		Air Blank	0.000	19:40	
PINELLAS COUNTY SO	08/30/2006	18:57:16	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:57	
		Control Test	0.079	18:58	
		Air Blank	0.000	18:58	
PINELLAS COUNTY SO	09/01/2006	20:52:42	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	20:53	
		Control Test	0.080	20:53	
		Air Blank	0.000	20:54	

Control Tests

Instrument Serial #: 80-001274

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	04/19/2006	18:35:49	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:36	
		Control Test	0.078	18:36	
		Air Blank	0.000	18:37	
PINELLAS COUNTY SO	05/12/2006	19:34:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	19:34	
		Control Test	0.078	19:35	
		Air Blank	0.000	19:35	
PINELLAS COUNTY SO	08/30/2006	14:31:40	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:32	
		Control Test	0.078	14:32	
		Air Blank	0.000	14:32	

Control Tests

Instrument Serial #: 80-001275

Location	Date	Time	Version	Target	Source
HIALEAH GARDENS PD	12/20/2006	09:20:37	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:19	
		Control Test	0.069	09:20	
		Air Blank	0.000	09:20	
HIALEAH GARDENS PD	12/21/2006	11:25:24	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:24	
		Control Test	0.079	11:24	
		Air Blank	0.000	11:25	

Control Tests

Instrument Serial #: 80-001276

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	12/20/2006	08:20:08	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:19	
		Control Test	0.072	08:19	
		Air Blank	0.000	08:20	
MIAMI-DADE PD	12/20/2006	08:21:30	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:20	
		Control Test	0.073	08:20	
		Air Blank	0.000	08:21	
MIAMI-DADE PD	12/20/2006	08:39:46	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:38	
		Control Test	0.079	08:39	
		Air Blank	0.000	08:39	
MIAMI-DADE PD	12/20/2006	08:41:10	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:40	
		Control Test	0.079	08:40	
		Air Blank	0.000	08:41	
MIAMI-DADE PD	12/21/2006	11:55:55	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:55	
		Control Test	0.078	11:55	
		Air Blank	0.000	11:55	
MIAMI-DADE PD	12/23/2006	23:13:26	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:12	
		Control Test	0.080	23:12	
		Air Blank	0.000	23:13	
MIAMI-DADE PD	12/30/2006	23:26:52	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:26	
		Control Test	0.079	23:26	
		Air Blank	0.000	23:26	
MIAMI-DADE PD	12/31/2006	23:09:12	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:08	
		Control Test	0.079	23:08	

Control Tests

Instrument Serial #: 80-001276

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	12/31/2006	23:09:12	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:09	

Control Tests

Instrument Serial #: 80-001277

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	04/01/2006	00:46:00	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:46	
		Control Test	0.077	00:46	
		Air Blank	0.000	00:47	
MIAMI-DADE PD	04/01/2006	01:52:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:52	
		Control Test	0.078	01:53	
		Air Blank	0.000	01:53	
MIAMI-DADE PD	06/09/2006	23:25:53	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	23:26	
		Control Test	0.078	23:26	
		Air Blank	0.000	23:27	
MIAMI-DADE PD	06/10/2006	01:55:52	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:56	
		Control Test	0.077	01:56	
		Air Blank	0.000	01:57	
MIAMI-DADE PD	12/20/2006	08:21:22	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:20	
		Control Test	0.072	08:20	
		Air Blank	0.000	08:21	
MIAMI-DADE PD	12/20/2006	08:22:49	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:21	
		Control Test	0.073	08:22	
		Air Blank	0.000	08:22	
MIAMI-DADE PD	12/20/2006	08:40:13	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:39	
		Control Test	0.078	08:39	
		Air Blank	0.000	08:40	
MIAMI-DADE PD	12/20/2006	08:41:39	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:40	
		Control Test	0.078	08:41	

Control Tests

Instrument Serial #: 80-001277

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	12/20/2006	08:41:39	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:41	
MIAMI-DADE PD	12/21/2006	11:58:02	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:57	
		Control Test	0.077	11:57	
		Air Blank	0.000	11:58	

Control Tests

Instrument Serial #: 80-001278

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	04/07/2006	11:20:09	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:20	
		Control Test	0.068	11:21	
		Air Blank	0.000	11:21	
JACKSONVILLE SO	04/07/2006	11:29:11	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:29	
		Control Test	0.075	11:30	
		Air Blank	0.000	11:30	

Control Tests

Instrument Serial #: 80-001279

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	03/27/2006	15:57:14	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	15:57	
		Control Test	0.078	15:58	
		Air Blank	0.000	15:58	
JACKSONVILLE SO	04/17/2006	13:06:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:06	
		Control Test	0.078	13:07	
		Air Blank	0.000	13:07	
JACKSONVILLE SO	07/19/2006	12:35:48	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:36	
		Control Test	0.080	12:36	
		Air Blank	0.000	12:37	

Control Tests

Instrument Serial #: 80-001280

Location	Date	Time	Version	Target	Source
JACKSONVILLE SO	03/30/2006	05:55:04	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	05:55	
		Control Test	0.077	05:55	
		Air Blank	0.000	05:56	
JACKSONVILLE SO	03/30/2006	07:26:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	07:26	
		Control Test	0.077	07:26	
		Air Blank	0.000	07:27	
JACKSONVILLE SO	09/25/2006	05:15:51	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	05:16	
		Control Test	0.079	05:16	
		Air Blank	0.000	05:17	
JACKSONVILLE SO	09/25/2006	05:17:56	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	05:18	
		Control Test	0.079	05:18	
		Air Blank	0.000	05:19	
JACKSONVILLE SO	09/25/2006	05:19:49	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	05:20	
		ITP Check	0.081	05:20	
		Air Blank	0.000	05:21	
JACKSONVILLE SO	09/27/2006	14:22:01	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:22	
		Control Test	0.079	14:22	
		Air Blank	0.000	14:23	

Control Tests

Instrument Serial #: 80-001284

Location	Date	Time	Version	Target	Source
GILCHRIST COUNTY SO	04/11/2006	13:59:55	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:00	
		Control Test	0.079	14:00	
		Air Blank	0.000	14:01	
GILCHRIST COUNTY SO	04/13/2006	13:21:36	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:22	
		Control Test	0.077	13:22	
		Air Blank	0.000	13:22	
GILCHRIST COUNTY SO	04/14/2006	12:11:14	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:11	
		Control Test	0.079	12:12	
		Air Blank	0.000	12:12	
GILCHRIST COUNTY SO	04/14/2006	12:16:25	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:16	
		Control Test	0.079	12:17	
		Air Blank	0.000	12:17	
GILCHRIST COUNTY SO	12/15/2006	09:16:10	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:14	
		Control Test	0.045	09:15	
		Air Blank	0.000	09:16	
GILCHRIST COUNTY SO	12/15/2006	09:18:02	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:16	
		Control Test	0.046	09:17	
		Air Blank	0.000	09:18	
GILCHRIST COUNTY SO	12/15/2006	09:19:57	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:18	
		Control Test	0.076	09:19	
		Air Blank	0.000	09:19	
GILCHRIST COUNTY SO	12/15/2006	09:21:40	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:20	
		Control Test	0.193	09:21	

Control Tests

Instrument Serial #: 80-001284

Location	Date	Time	Version	Target	Source
GILCHRIST COUNTY SO	12/15/2006	09:21:40	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:21	
GILCHRIST COUNTY SO	12/15/2006	09:23:04	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	09:22	
		Control Test	0.079	09:22	
		Air Blank	0.000	09:23	

Control Tests

Instrument Serial #: 80-001285

Location	Date	Time	Version	Target	Source
GILCHRIST COUNTY SO	12/15/2006	10:29:13	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:28	
		Control Test	0.048	10:28	
		Air Blank	0.000	10:29	
GILCHRIST COUNTY SO	12/15/2006	10:31:20	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:30	
		Control Test	0.078	10:30	
		Air Blank	0.000	10:31	
GILCHRIST COUNTY SO	12/15/2006	10:33:04	8100.27.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:31	
		Control Test	0.198	10:32	
		Air Blank	0.000	10:33	

Control Tests

Instrument Serial #: 80-001293

Location	Date	Time	Version	Target	Source
WILLISTON PD	04/05/2006	01:03:10	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:03	
		Control Test	0.078	01:03	
		Air Blank	0.000	01:04	
WILLISTON PD	04/05/2006	01:05:43	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:06	
		Control Test	0.078	01:06	
		Air Blank	0.000	01:07	
WILLISTON PD	06/05/2006	01:17:32	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:17	
		Control Test	0.077	01:18	
		Air Blank	0.000	01:18	
WILLISTON PD	06/07/2006	11:31:30	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	11:31	
		Control Test	0.078	11:32	
		Air Blank	0.000	11:32	
WILLISTON PD	06/24/2006	16:14:50	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:15	
		Control Test	0.078	16:15	
		Air Blank	0.000	16:16	

Control Tests

Instrument Serial #: 80-001305

Location	Date	Time	Version	Target	Source
MARIANNA PD	05/28/2006	10:02:55	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:03	
		Control Test	0.067	10:04	
		Air Blank	0.000	10:04	

Control Tests

Instrument Serial #: 80-001309

Location	Date	Time	Version	Target	Source
BAY COUNTY SO	12/07/2006	16:28:14	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:27	
		Control Test	0.083	16:27	
		Air Blank	0.000	16:28	
BAY COUNTY SO	12/07/2006	16:29:55	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:29	
		Control Test	0.082	16:29	
		Air Blank	0.000	16:29	

Control Tests

Instrument Serial #: 80-001312

Location	Date	Time	Version	Target	Source
WALTON COUNTY SO	10/20/2006	10:01:56	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:02	
		Control Test	0.078	10:02	
		Air Blank	0.000	10:03	

Control Tests

Instrument Serial #: 80-001322

Location	Date	Time	Version	Target	Source
OKEECHOBEE PD	11/12/2006	09:17:58	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	09:17	
		ITP Check	0.081	09:17	
		Air Blank	0.000	09:17	

Control Tests

Instrument Serial #: 80-001323

Location	Date	Time	Version	Target	Source
CMI INC	11/29/2006	10:41:50	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		10:41	Sequence Aborted
		Air Blank	0.000	10:41	

Control Tests

Instrument Serial #: 80-001329

Location	Date	Time	Version	Target	Source
INDIAN RIVER CO SO	08/07/2006	13:50:20	8100.26.00	0.100	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:50	
		Control Test	0.044	13:51	
		Air Blank	0.000	13:51	
INDIAN RIVER CO SO	08/22/2006	16:39:46	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:40	
		Control Test	0.000	16:40	
		Air Blank	0.000	16:41	

Control Tests

Instrument Serial #: 80-001344

Location	Date	Time	Version	Target	Source
SARASOTA COUNTY SO	03/02/2006	16:40:28	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	16:40	
		Control Test	0.080	16:41	
		Air Blank	0.000	16:41	

Control Tests

Instrument Serial #: 80-001345

Location	Date	Time	Version	Target	Source
EUSTIS P.D.	10/27/2006	03:32:44	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	03:31	
		Control Test	0.000	03:32	
		Air Blank	0.000	03:32	

Control Tests

Instrument Serial #: 80-001350

Location	Date	Time	Version	Target	Source
FRUITLAND PARK PD	04/21/2006	13:14:29	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:14	
		Control Test	0.081	13:15	
		Air Blank	0.000	13:15	
FRUITLAND PARK PD	05/02/2006	17:09:37	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	17:10	
		ITP Check	0.080	17:10	
		Air Blank	0.000	17:10	

Control Tests

Instrument Serial #: 80-001360

Location	Date	Time	Version	Target	Source
EVERGLADES NAT. PARK	03/10/2006	13:03:59	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	13:04	
		ITP Check	0.079	13:04	
		Air Blank	0.000	13:05	
EVERGLADES NAT. PARK	05/31/2006	13:34:50	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank		13:35	Sequence Aborted
		Air Blank	0.000	13:35	
EVERGLADES NAT. PARK	05/31/2006	13:36:37	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:37	
		Control Test	0.073	13:37	
		Air Blank	0.000	13:38	
EVERGLADES NAT. PARK	05/31/2006	13:38:22	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:38	
		Control Test	0.078	13:39	
		Air Blank	0.000	13:39	
EVERGLADES NAT. PARK	05/31/2006	13:55:55	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	13:56	
		ITP Check	0.080	13:56	
		Air Blank	0.000	13:57	
EVERGLADES NAT. PARK	05/31/2006	13:57:59	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	13:58	
		Control Test	0.047	13:59	
		Air Blank	0.000	13:59	
EVERGLADES NAT. PARK	05/31/2006	13:59:42	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank		14:00	Ambient Fail
		Air Blank	0.000	14:00	
EVERGLADES NAT. PARK	05/31/2006	14:00:40	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:01	
		Control Test	0.075	14:01	
		Air Blank	0.000	14:02	

Control Tests

Instrument Serial #: 80-001360

Location	Date	Time	Version	Target	Source
EVERGLADES NAT. PARK	05/31/2006	14:03:04	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:03	
		Control Test	0.189	14:04	
		Air Blank	0.000	14:04	
EVERGLADES NAT. PARK	05/31/2006	14:05:06	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:05	
		Control Test	0.076	14:06	
		Air Blank	0.000	14:06	
EVERGLADES NAT. PARK	05/31/2006	14:14:45	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:15	
		Control Test	0.080	14:15	
		Air Blank	0.000	14:16	
EVERGLADES NAT. PARK	05/31/2006	14:16:08	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	14:16	
		ITP Check	0.079	14:17	
		Air Blank	0.000	14:17	
EVERGLADES NAT. PARK	05/31/2006	14:17:34	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:18	
		Control Test	0.046	14:18	
		Air Blank	0.000	14:19	
EVERGLADES NAT. PARK	05/31/2006	14:19:38	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:20	
		Control Test	0.075	14:20	
		Air Blank	0.000	14:21	
EVERGLADES NAT. PARK	05/31/2006	14:21:20	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank		14:21	Ambient Fail
		Air Blank	0.000	14:22	
EVERGLADES NAT. PARK	05/31/2006	14:22:19	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:22	
		Control Test	0.192	14:23	
		Air Blank	0.000	14:24	

Control Tests

Instrument Serial #: 80-001360

Location	Date	Time	Version	Target	Source
EVERGLADES NAT. PARK	05/31/2006	14:24:05	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:24	
		Control Test	0.080	14:24	
		Air Blank	0.000	14:25	
EVERGLADES NAT. PARK	05/31/2006	14:27:22	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:27	
		Control Test	0.077	14:28	
		Air Blank	0.000	14:29	
EVERGLADES NAT. PARK	05/31/2006	14:29:52	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:30	
		Control Test	0.193	14:31	
		Air Blank	0.000	14:31	
EVERGLADES NAT. PARK	05/31/2006	14:31:47	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:32	
		Control Test	0.046	14:32	
		Air Blank	0.000	14:33	
EVERGLADES NAT. PARK	05/31/2006	14:44:24	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:44	
		Control Test	0.196	14:45	
		Air Blank	0.000	14:46	
EVERGLADES NAT. PARK	05/31/2006	14:46:08	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	14:46	
		Control Test	0.195	14:47	
		Air Blank	0.000	14:47	

Control Tests

Instrument Serial #: 80-001361

Location	Date	Time	Version	Target	Source
FL HIGHWAY PATROL	06/02/2006	12:40:50	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:41	
		Control Test	0.077	12:41	
		Air Blank	0.000	12:42	
FL HIGHWAY PATROL	06/02/2006	12:44:04	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:44	
		Control Test	0.079	12:44	
		Air Blank	0.000	12:45	

Control Tests

Instrument Serial #: 80-001364

Location	Date	Time	Version	Target	Source
GAINESVILLE PD	11/30/2006	06:49:12	8100.27.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	06:48	
		Control Test	0.081	06:48	
		Air Blank	0.000	06:49	
GAINESVILLE PD	11/30/2006	06:50:42	8100.27.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	06:49	
		ITP Check	0.081	06:50	
		Air Blank	0.000	06:50	

Control Tests

Instrument Serial #: 80-001367

Location	Date	Time	Version	Target	Source
PINELLAS COUNTY SO	04/29/2006	21:13:08	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:13	
		Control Test	0.078	21:13	
		Air Blank	0.000	21:14	
PINELLAS COUNTY SO	05/18/2006	01:51:35	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	01:52	
		Control Test	0.079	01:52	
		Air Blank	0.000	01:52	
PINELLAS COUNTY SO	07/12/2006	17:55:42	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	17:56	
		Control Test	0.080	17:56	
		Air Blank	0.000	17:56	

Control Tests

Instrument Serial #: 80-001420

Location	Date	Time	Version	Target	Source
ORANGE COUNTY SO	03/14/2006	12:59:16	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	12:59	
		Control Test	0.080	13:00	
		Air Blank	0.000	13:00	
ORANGE COUNTY SO	03/14/2006	13:05:07	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	13:05	
		ITP Check	0.079	13:06	
		Air Blank	0.000	13:06	
ORANGE COUNTY SO	04/11/2006	10:24:12	8100.26.00	0.080	Remote Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:24	
		Control Test	0.079	10:24	
		Air Blank	0.000	10:25	

Control Tests

Instrument Serial #: 80-001441

Location	Date	Time	Version	Target	Source
KENNEDY SPACE CENTER	05/05/2006	09:49:16	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:49	
		Control Test	0.078	09:50	
		Air Blank	0.000	09:50	

Control Tests

Instrument Serial #: 80-001443

Location	Date	Time	Version	Target	Source
CLAY COUNTY SO	06/03/2006	03:29:50	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	03:30	
		Control Test	0.081	03:30	
		Air Blank	0.000	03:31	
CLAY COUNTY SO	07/26/2006	00:20:13	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	00:20	
		Control Test	0.076	00:20	
		Air Blank	0.000	00:21	
CLAY COUNTY SO	09/14/2006	21:38:32	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	21:38	
		Control Test	0.080	21:39	
		Air Blank	0.000	21:39	
CLAY COUNTY SO	09/14/2006	21:40:28	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	21:40	
		ITP Check	0.079	21:41	
		Air Blank	0.000	21:41	

Control Tests

Instrument Serial #: 80-001469

Location	Date	Time	Version	Target	Source
HURLBURT FIELD PD	04/24/2006	15:55:32	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:55	
		Control Test	0.045	15:56	
		Air Blank	0.000	15:57	
HURLBURT FIELD PD	04/24/2006	15:58:58	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	15:59	
		Control Test	0.046	16:00	
		Air Blank	0.000	16:00	
HURLBURT FIELD PD	04/24/2006	16:00:56	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:01	
		Control Test	0.047	16:02	
		Air Blank	0.000	16:02	
HURLBURT FIELD PD	04/24/2006	16:03:26	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:03	
		Control Test	0.045	16:04	
		Air Blank	0.000	16:05	
HURLBURT FIELD PD	04/24/2006	16:07:50	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	16:08	
		Control Test	0.172	16:08	
		Air Blank	0.000	16:09	
HURLBURT FIELD PD	08/25/2006	13:18:01	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:18	
		Control Test	0.080	13:18	
		Air Blank	0.000	13:19	

Control Tests

Instrument Serial #: 80-001471

Location	Date	Time	Version	Target	Source
PATRICK AFB	05/31/2006	10:04:15	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:04	
		Control Test	0.076	10:05	
		Air Blank	0.000	10:05	
PATRICK AFB	05/31/2006	10:06:13	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	10:06	
		ITP Check	0.079	10:07	
		Air Blank	0.000	10:07	
PATRICK AFB	05/31/2006	10:08:19	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	10:08	
		Control Test	0.077	10:09	
		Air Blank	0.000	10:09	
PATRICK AFB	05/31/2006	10:09:58	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:10	
		Control Test	0.075	10:11	
		Air Blank	0.000	10:11	
PATRICK AFB	09/20/2006	10:33:38	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:34	
		Control Test	0.000	10:34	Interferent Detect
		Air Blank	0.000	10:35	
PATRICK AFB	09/20/2006	10:35:35	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	10:36	
		Control Test	0.000	10:36	
		Air Blank	0.000	10:37	

Control Tests

Instrument Serial #: 80-001652

Location	Date	Time	Version	Target	Source
CMI INC	08/15/2006	13:45:24	8100.26.00	0.000	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:45	
		Control Test	0.081	13:46	Interferent Detect
		Air Blank	0.000	13:47	
CMI INC	08/15/2006	13:47:11	8100.26.00	0.000	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	13:47	
		Control Test	0.073	13:47	Interferent Detect
		Air Blank	0.000	13:48	
CMI INC	08/15/2006	14:57:18	8100.26.00	0.000	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	14:57	
		Control Test	0.091	14:58	
		Air Blank	0.000	14:58	

Control Tests

Instrument Serial #: 80-001653

Location	Date	Time	Version	Target	Source
CMI INC	07/21/2006	14:10:42	8100.26.00	0.000	Local ITP
		Test	Value	Time	
		Air Blank	0.000	14:11	
		ITP Check	0.079	14:11	
		Air Blank	0.000	14:12	

Control Tests

Instrument Serial #: 80-001731

Location	Date	Time	Version	Target	Source
HILLSBOROUGH SO	05/17/2006	08:39:19	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	08:39	
		Control Test	0.000	08:40	
		Air Blank	0.000	08:40	
HILLSBOROUGH SO	05/17/2006	08:42:52	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	08:43	
		Control Test	0.078	08:44	
		Air Blank	0.000	08:44	
HILLSBOROUGH SO	05/17/2006	08:45:30	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	08:45	
		Control Test	0.046	08:46	
		Air Blank	0.000	08:47	
HILLSBOROUGH SO	05/17/2006	08:50:20	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	08:50	
		Control Test	0.000	08:51	
		Air Blank	0.000	08:51	
HILLSBOROUGH SO	05/17/2006	08:54:56	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	08:55	
		Control Test	0.187	08:56	
		Air Blank	0.000	08:56	
HILLSBOROUGH SO	05/17/2006	08:56:37	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	08:57	
		Control Test	0.115	08:57	
		Air Blank	0.000	08:58	
HILLSBOROUGH SO	05/17/2006	09:01:55	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:02	
		Control Test	0.185	09:03	
		Air Blank	0.000	09:03	
HILLSBOROUGH SO	05/17/2006	09:08:58	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:09	
		Control Test	0.194	09:10	

Control Tests

Instrument Serial #: 80-001731

Location	Date	Time	Version	Target	Source
HILLSBOROUGH SO	05/17/2006	09:08:58	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:10	
HILLSBOROUGH SO	07/18/2006	09:55:05	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	09:55	
		Control Test	0.195	09:56	
		Air Blank	0.000	09:56	

Control Tests

Instrument Serial #: 80-001739

Location	Date	Time	Version	Target	Source
CHARLOTTE COUNTY SO	05/26/2006	18:49:06	8100.26.00	0.080	Local Dry gas
		Test	Value	Time	
		Air Blank	0.000	18:49	
		Control Test	0.078	18:49	
		Air Blank	0.000	18:50	

Control Tests

Instrument Serial #: 80-001742

Location	Date	Time	Version	Target	Source
	03/17/2006	13:47:20	8100.26.00	0.000	Local ITP
		Test	Value	Time	
		Air Blank	0.000	13:47	
		ITP Check	0.079	13:48	
		Air Blank	0.000	13:48	

Control Tests

Instrument Serial #: 80-001866

Location	Date	Time	Version	Target	Source
US AIR FORCE MACDILL	05/17/2006	11:19:19	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:19	
		Control Test	0.047	11:20	
		Air Blank	0.000	11:20	
US AIR FORCE MACDILL	05/17/2006	11:21:03	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:21	
		Control Test	0.047	11:22	
		Air Blank	0.000	11:22	
US AIR FORCE MACDILL	05/17/2006	11:22:48	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:23	
		Control Test	0.048	11:23	
		Air Blank	0.000	11:24	
US AIR FORCE MACDILL	05/17/2006	11:24:50	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:25	
		Control Test	0.048	11:25	
		Air Blank	0.000	11:26	
US AIR FORCE MACDILL	05/17/2006	11:26:34	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:27	
		Control Test	0.049	11:27	
		Air Blank	0.000	11:28	
US AIR FORCE MACDILL	05/17/2006	11:31:11	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:31	
		Control Test	0.072	11:32	
		Air Blank	0.000	11:32	
US AIR FORCE MACDILL	05/17/2006	11:33:06	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:33	
		Control Test	0.076	11:34	
		Air Blank	0.000	11:34	
US AIR FORCE MACDILL	05/17/2006	11:34:49	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:35	
		Control Test	0.077	11:35	

Control Tests

Instrument Serial #: 80-001866

Location	Date	Time	Version	Target	Source
US AIR FORCE MACDILL	05/17/2006	11:34:49	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:36	
US AIR FORCE MACDILL	05/17/2006	11:36:31	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:36	
		Control Test	0.078	11:37	
		Air Blank	0.000	11:38	
US AIR FORCE MACDILL	05/17/2006	11:38:12	8100.26.00	0.080	Local Wet bath
		Test	Value	Time	
		Air Blank	0.000	11:38	
		Control Test	0.078	11:39	
		Air Blank	0.000	11:39	

Control Tests

Instrument Serial #: 80-001962

Location	Date	Time	Version	Target	Source
	05/18/2006	10:35:25	8100.26.00	0.080	Local ITP
		Test	Value	Time	
		Air Blank	0.000	10:35	
		ITP Check	0.078	10:36	
		Air Blank	0.000	10:36	