Florida Department of Law Enforcement

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
VOLUSIA SO SKIPPER,ROGER,G	01/31/2006	09:29:16	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT

Printed: 13 June 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO LECATES,RICHARD,A	02/24/2006	13:43:04	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101L
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO LECATES,RICHARD,A	03/15/2006	13:42:29	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101L
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO LECATES,RICHARD,A	04/11/2006	14:21:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.210
0.20 g/210L Test #1 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.212
0.20 g/210L Test #2 (Retry)	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.207
0.20 g/210L Test #3 (Retry)	0.204
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

.20 TUBES CLEANED AND RECHECKED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO IHNKEN,MATTHEW,D	05/15/2006	11:16:59	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.118
0.20 g/210L Test #1 (Retry)	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.116
0.20 g/210L Test #2 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.117
0.20 g/210L Test #3 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance,THERMOMETER LOOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO IHNKEN,MATTHEW,D	06/08/2006	17:11:21	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.205
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.205
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO IHNKEN,MATTHEW,D	07/05/2006	13:48:33	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	429101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.20 g/210L Test #3 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

00: RFI Detect, RADIO USAGE, RADIO TURNED OFF. 20: Control Outside Tolerance, UNKNOWN REASON FOR LOW TEST .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
VOLUSIA SO IHNKEN,MATTHEW,D	08/07/2006	22:48:02	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.206
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.207
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

A F / M A:UNKNOWN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO IHNKEN,MATTHEW,D	09/19/2006	14:51:50	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA SO IHNKEN,MATTHEW,D	10/03/2006	17:42:07	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
VOLUSIA S.O. SKIPPER,ROGER,G	11/02/2006	22:49:38	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.075
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
VOLUSIA S.O. IHNKEN,MATTHEW,D	11/28/2006	16:59:50	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	12/27/2006	17:35:39	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	02/10/2006	19:28:20	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	520201I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	07/26/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/27/2006	08:29:58	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

INTERFERENT NOT DETECTED. INTERFERENT NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/27/2006	08:47:41	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	04/11/2006		

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.078
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.199
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/27/2006	09:30:19	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	04/30/2006	08:23:01	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	05/31/2006	15:04:35	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	08/22/2006	13:55:36	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.204
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	08/30/2006	16:19:23	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMTOHY,B	09/30/2006	18:40:34	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	10/18/2006	17:47:20	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6 (Retry)	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7 (Retry)	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8 (Retry)	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Int Det: RFI Detect RADIO / CELL PHONE USE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	10/31/2006	12:46:12	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	11/30/2006	09:10:45	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	12/23/2006	08:44:58	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	02/10/2006	19:15:01	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	525202I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

MOUTH ALCOHOL TEST 1: TOO MUCH M/A SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/27/2006	06:43:01	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/27/2006	08:10:38	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Test #2	0.000
Interferent Detect Test #2 (Retry)	0.000
Alcohol Free Test #3	0.000
Interferent Detect Test #3 (Retry)	0.000

Remarks:

Int Det: RFI Detect, . INTERFERENT NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/27/2006	09:17:40	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	04/30/2006	08:21:42	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.043
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.044
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	05/25/2006	11:44:47	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	06/29/2006	16:03:43	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	05040601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	07/13/2006	16:09:00	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006		

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.189
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.189
0.20 g/210L Test #2 (Retry)	0.190
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.190
0.20 g/210L Test #3 (Retry)	0.190
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	07/20/2006	13:51:25	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.189
0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.189
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	08/30/2006	15:57:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.188
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.188
0.08 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.189
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Sequence Aborted. Non-compliance:DID NOT PASS .20 SIMULATION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	08/30/2006	16:32:20	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.075
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	08/31/2006	14:32:30	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504061I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	09/30/2006	18:40:45	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	10/18/2006	17:35:17	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4 (Retry)	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5 (Retry)	0.047
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6 (Retry)	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7 (Retry)	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.196

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8 (Retry)	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9 (Retry)	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10 (Retry)	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

05: RFI Detect / RADIO / CELL PHONE USE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	10/31/2006	12:51:57	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	11/30/2006	09:01:17	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	12/23/2006	08:56:44	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	0504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source
DAYTONA BEACH PD SKIPPER,ROGER,G	02/27/2006	13:02:42	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/09/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	03/27/2006	00:04:38	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4156011
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.183
0.20 g/210L Test #1 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.20 g/210L Test #2 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.20 g/210L Test #3 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	04/30/2006	21:44:47	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BREATH TUBENon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	04/30/2006	21:47:03	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	05/16/2006	19:22:23	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4156011
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

APRIL 2006 AGENCY INSPECTION FAILED BREATH TUBE REPLACED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	06/10/2006	19:25:06	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4156011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/10/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	07/26/2006	19:26:24	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	08/29/2006	23:00:44	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	09/30/2006	19:30:41	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH PD WARD,JAMES,A	10/30/2006	04:04:59	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source
DAYTONA BEACH P.D. SKIPPER,ROGER,G	11/07/2006	16:53:57	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.08 g/210L Test #1 (Retry)	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.08 g/210L Test #3 (Retry)	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.08 g/210L Test #4 (Retry)	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.08 g/210L Test #5 (Retry)	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6 (Retry)	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7 (Retry)	0.080
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8 (Retry)	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9 (Retry)	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10 (Retry)	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: RFI Detect CELL PHONE IN AREA..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	11/12/2006	23:47:00	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200504	**0.08 Gas Lot#**	615802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	04/21/2007	**Expiration**	06/09/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.186
0.20 g/210L Test #1 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	12/29/2006	00:49:10	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	615802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.178
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND PD SKIPPER,ROGER,G	03/02/2006	10:13:39	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	0.000
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	0.000
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	0.000
Interferent Detect Test #6 (Retry)	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.081
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	0.000
			Interferent Detect Test #7 (Retry)	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	0.000
			Interferent Detect Test #8 (Retry)	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	0.000
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	0.000
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Int Det: .00 SIM USED. INT SIM CONNECTED FOR 2ND TEST. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND PD LECATES,RICHARD,A	03/15/2006	09:21:56	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101L
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND PD LECATES,RICHARD,A	04/12/2006	09:08:36	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND PD IHNKEN,MATTHEW,D	05/15/2006	14:34:00	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND PD IHNKEN,MATTHEW,D	06/21/2006	19:58:06	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND PD PETRELLA,TRINITY,J	07/16/2006	10:45:06	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND PD PETRELLA,TRINITY,J	08/28/2006	18:16:22	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A:TOO SOON FROM OTHER INTOXYLIZER MA TEST . TOO SOON FROM OTHER INTOXYLIZER MA TEST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND PD PETRELLA,TRINITY,J	09/30/2006	11:54:08	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	431501I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/12/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

Time-Date changed.2 MIN DIFFERENCE IN INSTRUMENTS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND PD PETRELLA,TRINITY,J	10/29/2006	08:12:18	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	431501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/12/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND P.D. SKIPPER,ROGER,G	11/01/2006	17:00:45	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,T,J	11/29/2006	14:54:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	608801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	12/30/2006	17:53:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source
FLAGLER COUNTY SO SKIPPER,ROGER,G	02/21/2006	10:12:42	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	05/01/2006	08:00:22	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.192
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.074
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.197
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.076
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.198
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:POOR SEAL ON SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	05/01/2006	09:08:02	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.196
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	06/29/2006	22:01:35	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	07/25/2006	13:45:10	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	07/25/2006	13:50:54	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	08/28/2006	08:53:07	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	09/30/2006	19:19:30	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source
FLAGLER COUNTY S.O. SKIPPER,ROGER,G	10/30/2006	18:45:28	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BISS,ADAM,B	10/31/2006	10:19:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

A F / M A: Interferent Detect NO MOUTH ALCOHOL INTRODUCED . 2ND ALC.FREE TEST DUE TO INTER. DETECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	11/19/2006	11:12:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	12/30/2006	11:20:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source
FLAGLER COUNTY SO SKIPPER,ROGER,G	02/21/2006	09:35:36	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	05/01/2006	10:24:01	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	06/29/2006	20:26:24	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:POOR SEAL ON SIMULATOR CONNECTOR AT CONTROL HEAD . Non-compliance:WOULDN'T DETECT INTERFERENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	06/29/2006	20:49:37	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

REMIXED INTERFERENT IN NEWER SIMULATOR W/PROPER RESULTS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	07/25/2006	14:41:39	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	08/28/2006	08:01:35	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance:RANGE EXCEEDED FOR M/A. START OVER.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	08/28/2006	08:07:05	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	09/30/2006	19:56:08	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source
FLAGLER COUNTY S.O. SKIPPER,ROGER,G	10/30/2006	18:28:39	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	10/31/2006	09:23:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	11/19/2006	11:49:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	12/30/2006	10:40:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	01/12/2006	12:21:00	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	03/23/2006	15:10:11	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	422301I
Expiration	11/04/2006	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	03/23/2006	15:50:17	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	04/19/2006	22:35:04	8100.26	Agency	Inspection

0.05 Lot#	200502A	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	422301I
Expiration	02/14/2007	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	05/18/2006	19:55:54	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/26/2006	09:31:13	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

PATRICIA NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	06/08/2006	14:18:16	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.083
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.205
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT

Remarks:

FRANCA DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	06/16/2006	09:29:59	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

PATRICIA NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/22/2006	15:31:50	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431501I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	11/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	08/09/2006	09:56:24	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.078
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.197
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO DICIENZO,FRANCA,R	08/09/2006	10:28:45	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	08/24/2006	08:52:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

F. DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	09/22/2006	09:17:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	10/18/2006	16:48:56	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	10/20/2006	10:44:14	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

F. DICIENZO WAS PRESENT AND OBSERVED THE INTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	11/17/2006	10:36:17	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.084
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	12/21/2006	10:23:19	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
LAKE HELEN PD SKIPPER,ROGER,G	02/07/2006	12:20:13	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN PD BELLESFIELD,SCOTT,J	04/09/2006	23:38:42	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
LAKE HELEN PD BELLESFEILD,SCOTT,J	05/02/2006	01:22:33	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.184
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.197
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.198
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.DID NOT GIVE ENOUGH TIME TO WARM UP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
LAKE HELEN PD BELLESFIELD,SCOTT,J	06/07/2006	20:50:47	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.185
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN PD BELLESFIELD,SCOTT,J	07/19/2006	20:43:12	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN PD BELLESFIELD,SCOTT,J	08/12/2006	23:31:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
LAKE HELEN PD BELLESFIELD,SCOTT,J	09/10/2006	13:30:37	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.SIMULATOR HEATED FOR 25 MIN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN PD BELLESFIELD,SCOTT,J	09/10/2006	14:32:28	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN PD BELLESFIELD,SCOTT,J	10/08/2006	13:14:39	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 2 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.067
0.20 g/210L Test #1 (Retry)	0.188
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.019
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.110
0.20 g/210L Test #2 (Retry)	0.194
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.117
0.20 g/210L Test #3 (Retry)	0.195
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance . 20: Control Outside ToleranceTAKEN OUT OF SERVICE/CONTROL OUTNon-compliance:SIMULATORS NEED SERVICE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN PD BELLESFIELD,SCOTT,J	10/11/2006	13:27:49	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT PLACED BACK IN SERVICE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
LAKE HELEN P.D. SKIPPER,ROGER,G	10/25/2006	16:58:16	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	11/05/2006	13:27:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	12/22/2006	01:28:56	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
MANATEE COUNTY SO VENTURI,GEORGE,L	01/24/2006	16:52:22	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	02/13/2006	16:28:09	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO RILEY,JASON,W	03/27/2006	17:11:23	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

DEP. ROBERT ANDREWS PRESENT DURING INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	04/24/2006	07:38:24	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
MANATEE COUNTY SO RILEY,JASON,W	05/24/2006	09:28:02	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	422301I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	08/12/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	06/21/2006	09:39:45	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	606002I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

Time-Date changed. AI DEP JASON RILEY PRESENT DURING INSPEC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,N	07/12/2006	13:05:42	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

AI DEP.J.RILEY PRESENT FOR INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO RILEY,JASON,W	08/23/2006	15:10:28	8100.26	Agency	Inspection

0.05 Lot#	200509D	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

DEP. ROBERT ANDREWS PRESENT DURING INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
MANATEE COUNTY SO VENTURI,GEORGE,L	09/14/2006	13:48:25	8100.26	Department Inspection

 # Sims 4 **Avg Std Dev** **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Test #2	0.000
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	0.000
Alcohol Free Test #3	0.000
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	0.000
Alcohol Free Test #4	0.000
Interferent Detect Test #4	0.000
Interferent Detect Test #4 (Retry)	0.000
Alcohol Free Test #5	0.000
Interferent Detect Test #5	0.000
Interferent Detect Test #5 (Retry)	0.000
Alcohol Free Test #6	0.000
Interferent Detect Test #6	0.000
Interferent Detect Test #6 (Retry)	0.000
Alcohol Free Test #7	0.000
Interferent Detect Test #7	0.000
Interferent Detect Test #7 (Retry)	0.000
Alcohol Free Test #8	0.000
Interferent Detect Test #8	0.000
Interferent Detect Test #8 (Retry)	0.000
Alcohol Free Test #9	0.000
Interferent Detect Test #9	0.000
Interferent Detect Test #9 (Retry)	0.000
Alcohol Free Test #10	0.000
Interferent Detect Test #10	0.000
Interferent Detect Test #10 (Retry)	0.000

Remarks:

Int Det:INTERFERENT NOT DETECTED. Non-compliance:DID NOT DETECT INTERFERENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
MANATEE COUNTY SO VENTURI,GEORGE,L	11/06/2006	15:48:23	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

AI DEPUTY R ANDREWS PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	11/06/2006	16:22:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO RILEY,JASON,W	12/06/2006	11:28:54	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

AI DEP. ROBERT ANDREWS PRESENT DURING INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD SUERETH,DONALD,P	02/23/2006	18:50:46	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0009 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Ambient Fail.ROOM AIRED OUT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	03/23/2006	11:20:53	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.162
0.20 g/210L Test #1 (Retry)	0.162
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.163
0.20 g/210L Test #2 (Retry)	0.163
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.164
0.20 g/210L Test #3 (Retry)	0.163
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:NOT ALL 3. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	03/23/2006	14:11:10	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Test #2	0.000
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	0.000
Alcohol Free Test #3	0.000
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	0.000

Remarks:

INTERF NOT DETECTEDNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	03/24/2006	15:08:53	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	431401L
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	11/11/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	04/28/2006	14:59:22	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	04/28/2006	15:32:57	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	No
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	INT
Alcohol Free Test #1 (Retry)	INT

Remarks:

00: Interferent Detect, Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	04/28/2006	15:43:16	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.060
0.08 g/210L Test #1 (Retry)	0.058
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.060
0.08 g/210L Test #2 (Retry)	0.059
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.061
0.08 g/210L Test #3 (Retry)	0.058
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/28/2006	18:04:23	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.036
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.037
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.037
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.037
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.037
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.037
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/28/2006	18:35:35	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.036
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.037
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.037
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.037
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.037
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect. 05: RFI Detect, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/28/2006	19:02:57	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	No
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #1 (Retry)	

Remarks:

00: RFI Detect, RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/28/2006	19:13:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.042
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.174
0.20 g/210L Test #1 (Retry)	0.188
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.043
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.178
0.20 g/210L Test #2 (Retry)	0.188
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.044
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.180
0.20 g/210L Test #3 (Retry)	0.188
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	04/28/2006	22:16:40	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.179
0.20 g/210L Test #1 (Retry)	0.181
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.182
0.20 g/210L Test #2 (Retry)	0.181
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.182
0.20 g/210L Test #3 (Retry)	0.182
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/29/2006	05:52:18	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.074
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/29/2006	06:39:27	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/29/2006	06:50:47	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/29/2006	06:55:59	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/29/2006	07:00:04	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.177
0.20 g/210L Test #1 (Retry)	0.176
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.177
0.20 g/210L Test #2 (Retry)	0.177
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.178
0.20 g/210L Test #3 (Retry)	0.177
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	05/01/2006	11:33:39	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	05/01/2006	11:41:10	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.033
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.033
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.044
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	0.000
0.05 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.033
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	05/01/2006	14:03:10	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.071
0.08 g/210L Test #1 (Retry)	0.071
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.071
0.08 g/210L Test #2 (Retry)	0.073
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.072
0.08 g/210L Test #3 (Retry)	0.073
Interferent Detect Test #3	INT

Remarks:

A F / M A: Improper Sample. 05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	05/02/2006	11:52:31	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.071
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.224
0.20 g/210L Test #1 (Retry)	0.193
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.072
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.232
0.20 g/210L Test #2 (Retry)	0.191
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.073
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.230
0.20 g/210L Test #3 (Retry)	0.188
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	05/02/2006	17:19:47	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	04/21/2007		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.177
0.20 g/210L Test #1 (Retry)	0.204
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.182
0.20 g/210L Test #2 (Retry)	0.212
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.187
0.20 g/210L Test #3 (Retry)	0.215
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD STEWART,ANGELA,R	05/03/2006	09:09:45	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401L
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.159
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.163
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.165
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
CMI INC WRTIO,SDFGH,A	08/14/2006	08:36:47	8100.26	Department Inspection

 # Sims 0 **Avg Std Dev** **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
CMI INC WERYUI,WERTYU,A	08/14/2006	08:38:03	8100.26	Department Inspection

 # Sims 0 **Avg Std Dev** **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
CMI INC WERT,SXDFG,W	08/15/2006	08:38:25	8100.26	Department Inspection

 # Sims 0 **Avg Std Dev** **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD SUERETH,DONALD,P	09/05/2006	13:33:15	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	09/06/2006	10:32:41	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	502101G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	10/11/2006	13:54:21	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	10/11/2006	13:59:12	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	10/11/2006	14:03:04	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	10/11/2006	14:06:31	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.046
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	10/11/2006	14:37:43	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502101G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	01/24/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.043
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.043
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD SUERETH,DONALD,P	11/06/2006	10:39:03	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	
0.08 g/210L Test #4 (Retry)	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5 (Retry)	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6 (Retry)	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7 (Retry)	0.079

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8 (Retry)	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9 (Retry)	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10 (Retry)	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: RFI Detect.OFFICER ENTERED AREA WITH PHONE..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD STEWART,ANGELA,R	11/06/2006	15:28:12	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	502101G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/24/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD STEWART,ANGELA,R	12/08/2006	14:16:26	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	502101G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/24/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD SKIPPER,ROGER,G	02/27/2006	15:46:57	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.083
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.083
0.20 g/210L Test #5	0.205
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.083
0.20 g/210L Test #6	0.204
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.083
0.20 g/210L Test #7	0.204
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.205
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.205
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.205
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	03/26/2006	21:48:12	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4167011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	04/29/2006	15:20:13	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4167011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	05/24/2006	10:22:42	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4167011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	06/17/2006	14:54:54	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	416701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	07/31/2006	17:49:33	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	08/30/2006	10:32:51	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	09/27/2006	09:54:34	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD SKIPPER,ROGER,G	10/17/2006	10:47:33	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.204
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.204
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	10/20/2006	11:42:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	11/17/2006	12:12:00	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	12/20/2006	10:45:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source
LEON COUNTY SO MALHIOT,MATTHEW,E	02/02/2006	07:13:02	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	428101I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	10/11/2006

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,G	03/22/2006	14:22:29	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	502101G
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO DAWSON,ANDREW,D	04/06/2006	13:47:10	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	501101G
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO DAWSON,ANDREW,D	05/31/2006	12:31:59	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	502101G
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	06/01/2006	08:14:11	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	502101G
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	07/25/2006	14:05:59	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	502101G
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	08/29/2006	15:56:24	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	502101G
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	09/29/2006	14:49:25	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	606002I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	10/03/2006	11:56:13	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	606002I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source
LEON COUNTY SO MALHIOT,MATTHEW,E	10/31/2006	17:07:21	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	606002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	11/21/2006	07:35:18	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Volume Not Met (0.%03ld - Breath Sample Not. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	11/21/2006	07:36:09	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	11/21/2006	08:27:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	12/26/2006	09:57:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source
OAK HILL PD SKIPPER,ROGER,G	02/23/2006	09:31:39	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	0.000
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.205
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	0.000
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.083
0.20 g/210L Test #6	0.206
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	0.000
Interferent Detect Test #6 (Retry)	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.082
			0.20 g/210L Test #7	0.206
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	0.000
			Interferent Detect Test #7 (Retry)	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.205
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	0.000
			Interferent Detect Test #8 (Retry)	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.206
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	0.000
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.206
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	0.000
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Int Det: NO ACETONE SOLUTION ADDED FOR 1ST TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	03/29/2006	22:07:54	8100.26	Agency	Inspection

0.05 Lot# 200502A
Expiration 02/14/2007

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.033
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.041
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.047
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	03/29/2006	22:56:48	8100.26	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	03/29/2006	23:06:00	8100.26	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Interferent Detect Test #1	0.000
Alcohol Free Test #2	0.000
Interferent Detect Test #2	0.000
Alcohol Free Test #3	0.000
Interferent Detect Test #3	0.000

Remarks:

A F / M A: . Int Det:NOT DETECTED, Ambient Fail. Non-compliance:AMB FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	04/03/2006	19:41:26	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4252021
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

PREVIOUS INSPECTIONS FAILED DUE TO AIR LEAK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	05/12/2006	00:07:31	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4252021
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	INT
0.05 g/210L Test #1 (Retry)	0.048
0.08 g/210L Test #1	0.018
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test #2	0.027
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.027
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect. 05: Interferent Detect, TUBING CHANGED. 08: Control Outside Tolerance,SIMULATOR SEAL NOT TIGHT .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	06/22/2006	20:44:15	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4252021
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	07/30/2006	01:37:18	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4252021
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/10/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. 08: Control Outside Tolerance GAS LINE NOT HOOKED UP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD GRASSO,GUY,M	08/26/2006	21:30:15	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD IHNKEN,MATTHEW,D	09/19/2006	23:10:22	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.205
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.205
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD IHNKEN,MATTHEW,D	09/20/2006	00:31:55	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL PD IHNKEN,MATTHEW,D	10/04/2006	01:49:17	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.205
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.206
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source
OAK HILL P.D. SKIPPER,ROGER,G	11/02/2006	22:13:03	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	11/28/2006	23:00:38	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.085
0.20 g/210L Test #1	0.215
0.20 g/210L Test #1 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.084
0.20 g/210L Test #2	0.210
0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.053
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.206
0.20 g/210L Test #3 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

TEMP TOO LOW. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	12/27/2006	21:55:02	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AMBIENT FAIL, UNKNOWN, Ambient Fail. Non-compliance:AMBIENT FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	12/28/2006	16:58:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.206
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.206
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001155

Agency Operator	Date	Time	Software	Source
EGLIN AFB PD MALHIOT,MATTHEW,E	06/12/2006	08:51:23	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200411G	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	210201G
Expiration	09/22/2007	**Expiration**	11/18/2006	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.204
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001155

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001155

Agency Operator	Date	Time	Software	Source
EGLIN AFB PD WATSON,HEATHER,L	08/11/2006	17:26:48	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001155

Agency Operator	Date	Time	Software	Source
EGLIN AFB PD WATSON,HEATHER,L	09/28/2006	13:22:48	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.077
0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
ORMOND BEACH PD SKIPPER,ROGER,G	02/21/2006	14:10:03	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	03/27/2006	14:50:29	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007		

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.077
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.200
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.201
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	03/27/2006	16:28:39	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	03/27/2006	16:29:29	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	04/30/2006	15:13:09	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	05/23/2006	18:05:09	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	05/23/2006	18:10:40	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

3ML OF ACETONE ADDED TO ALCOHOL FREE SOLUTION AND DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	06/30/2006	16:10:10	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

FOR GOT TO ADD THE ACETONE IN THE FIRST TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	07/31/2006	14:17:20	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

3ML OF ACETONE ADDED AND DETECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH PD HALLER,ROBERT,M	08/31/2006	14:57:54	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
ORMOND BEACH PD HALLER,ROBERT,M	09/29/2006	14:56:24	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4281011
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

3ML OF ACETONE ADDED AND DETECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
ORMOND BEACH PD HALLER,ROBERT,M	10/31/2006	15:47:36	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

3ML OF ACETONE ADDED AND DETECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
ORMOND BEACH P.D. SKIPPER,ROGER,G	11/01/2006	12:09:57	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	11/30/2006	14:21:41	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	12/30/2006	11:28:56	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
LAKE WALES PD SUERETH,DONALD,P	02/23/2006	12:56:16	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	03/30/2006	10:51:23	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502101G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	01/24/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	04/30/2006	11:13:28	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	05/30/2006	09:01:27	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD YODONIS,MICHAEL,C	07/06/2006	21:05:34	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.185
0.20 g/210L Test #1 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.188
0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.188
0.20 g/210L Test #3 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	07/18/2006	10:24:59	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	08/14/2006	13:44:51	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	10/06/2006	14:14:54	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD YODONIS,MICHAEL,C	11/08/2006	10:44:18	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail I HAD THE SOLUTION CONTAINERS TO CLOSE . 08: I DID NOT HAVE THE DRY GAS AIR HOSE HOOKED UP PROPERLY .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
LAKE WALES PD SUERETH,DONALD,P	11/08/2006	13:40:48	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
LAKE WALES PD YODONIS,MICHAEL,C	11/08/2006	14:43:58	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	12/29/2006	09:29:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
PONCE INLET PD SKIPPER,ROGER,G	02/22/2006	10:30:07	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD DOMINY,RAYMOND,F	03/27/2006	01:04:30	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4281021
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.167
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: AIR LEAK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
PONCE INLET PD SCOTT,ALLEN,J	04/03/2006	23:37:35	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4281021
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

05: AIR LEAK AT SIMULATOR. 08: LEAK AT SIMULATOR . 20: AIR LEAKS AT SIMULATOR .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	05/03/2006	01:16:39	8100.26	Agency	Inspection

0.05 Lot# 200509A **0.08 Lot#** 200509B **0.20 Lot#**
Expiration 09/22/2007 **Expiration** 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
Interferent Detect Test #3	INT

Remarks:

05: AIR LEAK AT CONNECTION. 08: PROBLEM WITH CONNECTIONS W/AIR LEAKS . Non-compliance:
.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	05/03/2006	02:13:53	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	05/03/2006	02:19:13	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509A	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4281021
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

SECOND INSPECTION CONDUCTED DUT TO AIR LEAKS W/CONNECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	06/07/2006	18:04:58	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	04/21/2007		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.072
0.08 g/210L Test #1 (Retry)	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.074
Interferent Detect Test #3	INT

Remarks:

08: AIR LEAKS AT HOSE WILL CHANGE . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	06/07/2006	18:48:33	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	04/21/2007		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.074
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.075
Interferent Detect Test #3	INT

Remarks:

05: PROBLEM WITH AIR LEAKS. 08: PROBLEM WITH HOSE LEAKS, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	06/07/2006	22:09:00	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428102I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: AIR LEAK AT TUBE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	07/11/2006	01:42:32	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
PONCE INLET PD DOMINY,RAYMOND,F	08/29/2006	00:36:06	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

FORGOT TO ADD ACETONE TO SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD SCOTT,ALLEN,J	09/06/2006	01:16:13	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD DOMINY,RAYMOND,F	10/05/2006	23:27:14	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET PD DOMINY,RAYMOND,F	10/05/2006	23:43:10	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	626903I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

LOW SAMPLE 1 READING 0.044.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
PONCE INLET P.D. SKIPPER,ROGER,G	11/06/2006	12:34:09	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.05 g/210L Test #4 (Retry)	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.05 g/210L Test #5 (Retry)	0.047
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.05 g/210L Test #6 (Retry)	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.048
			0.05 g/210L Test #7 (Retry)	0.048
			0.08 g/210L Test #7	0.076
			0.20 g/210L Test #7	0.195
			0.08 g/210L Dry Gas Std Test #7	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.05 g/210L Test #8 (Retry)	0.048
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.05 g/210L Test #9 (Retry)	0.048
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.05 g/210L Test #10 (Retry)	0.048
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0015
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

05: Control Outside Tolerance LOOSE CONNECTION TO SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	11/08/2006	21:57:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	12/04/2006	02:13:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.182
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.187
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

20: AIR LEAK AT INLET TUBE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY PD SKIPPER,ROGER,G	02/27/2006	09:42:33	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0009 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0014
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	03/24/2006	10:06:27	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	03/24/2006	10:13:30	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. 05: Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREG,E	03/24/2006	11:19:57	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.077
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.046
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY PD MELVIN,GREGORY,E	03/24/2006	12:01:02	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION TO BRING INTOXILYZER 8000 IN SERVICE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	04/25/2006	08:39:14	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.042
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:AIR LEAK W/ SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	04/25/2006	09:04:36	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance:AI ERROR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY PD MELVIN,GREGORY,E	04/25/2006	09:40:58	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.040
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.AIR LEAK W/ CONNECTION HOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	05/24/2006	09:16:20	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.08 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: RFI Detect, Control Outside Tolerance. Non-compliance:RI AND THEN SIMULATOR FAILURE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY PD MELVIN,GREGORY,E	05/25/2006	08:55:24	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

CLEAN AND DRY.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY PD MELVIN,GREGORY,E	06/26/2006	09:58:02	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance AIR LEAK IN CONNECTOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	07/24/2006	13:28:22	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY PD MELVIN,GREGORY,E	08/17/2006	14:06:18	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	09/11/2006	14:44:24	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

INTOXILYZER AREA CLEAN AND DRY.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY PD MELVIN,GREGORY,E	10/06/2006	09:31:04	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT AREA CLEAN AND DRY.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY P.D. SKIPPER,ROGER,G	10/31/2006	17:51:53	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0010 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0028
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. MELVIN,GREGORY,E	11/17/2006	11:59:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	0.023
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	0.027
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	0.027
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:AIR LEAK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	12/11/2006	13:36:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

INSTRUMENT AREA CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
BREVARD S0 SKIPPER,ROGER,G	02/13/2006	13:09:41	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 THRUSH,DANIEL,C	03/20/2006	09:25:50	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.057
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.055
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.057
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.057
0.08 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: RFI Detect, Control Outside Tolerance. Non-compliance:SIM FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 THRUSH,DANIEL,C	03/20/2006	10:29:25	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.057
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.057
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.058
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #3 (Retry)	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Control Outside Tolerance. Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 THRUSH,DANIEL,C	03/20/2006	10:51:21	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.048
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.048
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. SIM AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
BREVARD S0 THRUSH,DANIEL,C	03/20/2006	11:37:01	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	411803I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/30/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	04/17/2006	10:55:34	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	04/17/2006	11:03:35	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	425202I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A:AF .008.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
BREVARD S0 WAGNER,ROBERT,F	05/22/2006	09:59:43	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	425202I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

GOOD

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	06/26/2006	11:28:39	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SLOPE NOT MET.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	06/26/2006	11:34:25	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	425202G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

M/A ON OPERATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
BREVARD S0 WAGNER,ROBERT,F	07/24/2006	10:25:06	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	425202I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	08/28/2006	08:17:22	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	425202I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,	09/18/2006	08:53:38	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	10/16/2006	10:13:41	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
BREVARD S.O. SKIPPER,ROGER,G	10/17/2006	17:17:55	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.079
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0332
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Control Outside Tolerance HOSE DISCONNECTED. REDO..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	11/13/2006	10:16:03	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,OBERT,F	12/19/2006	17:16:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
CLERMONT PD SKIPPER,ROGER,G	02/02/2006	14:35:22	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/20/2006	17:06:36	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 2 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.075
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. Non-compliance:.08 OUTSIDE TOLERANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/26/2006	13:04:34	8100.26	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result

Remarks:

A F / M A: Ambient Fail. 00: Ambient Fail, Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/26/2006	13:40:25	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Interferent Detect, Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/26/2006	14:05:31	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.047
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	INT
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/26/2006	14:33:51	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	INT
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.070
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2 (Retry)	0.071
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.20 g/210L Test: OK	No	0.08 g/210L Test #3 (Retry)	0.072
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT

Remarks:

08: Interferent Detect, Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/26/2006	14:58:53	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.068
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.073
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	INT
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.071
0.08 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.072
		Interferent Detect Test #3	INT

Remarks:

05: Interferent Detect. 08: Control Outside Tolerance, Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	03/26/2006	15:25:49	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.159
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.180
0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.188
0.20 g/210L Test #3 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside ToleranCe.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	14:23:01	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.185
0.20 g/210L Test #1 (Retry)	0.188
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.187
0.20 g/210L Test #2 (Retry)	0.189
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.188
0.20 g/210L Test #3 (Retry)	0.189
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	15:27:09	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.20 g/210L Test: OK	No	0.08 g/210L Test #3	0.074
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	15:55:19	8100.26	Agency	Inspection

0.05 Lot# 200509A **Expiration** 09/22/2007
0.08 Lot# 200509B **Expiration** 09/22/2007
0.20 Lot#

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.05 g/210L Test #1 (Retry)	0.048
0.08 g/210L Test #1	0.072
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.073
Interferent Detect Test #3	INT

Remarks:

A F / M A: . 05: Improper Sample. 08: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	16:31:29	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail. 08: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	16:57:14	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.071
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.072
		0.08 g/210L Test #3 (Retry)	0.064
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	17:21:33	8100.26	Agency	Inspection

0.05 Lot# 200509A **Expiration** 09/22/2007
0.08 Lot# 200509B **Expiration** 09/22/2007
0.20 Lot#

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.071
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.072
0.08 g/210L Test #2 (Retry)	0.074
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.073
0.08 g/210L Test #3 (Retry)	0.073
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 08: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/28/2006	17:47:32	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.147
0.20 g/210L Test #1 (Retry)	0.164
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.182
0.20 g/210L Test #2 (Retry)	0.164
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.172
0.20 g/210L Test #3 (Retry)	0.163
Interferent Detect Test #3	INT

Remarks:

A F / M A: . 05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	12:30:00	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.042
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	13:10:51	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.066
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.070
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.072
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 08: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	13:38:11	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2 (Retry)	0.047
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.072
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.074
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. 05: Ambient Fail. 08: Control Outside Tolerance, Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	14:01:59	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2 (Retry)	0.046
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.074
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.074
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. 05: Ambient Fail. 08: Control Outside Tolerance, Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	14:25:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.176
0.20 g/210L Test #1 (Retry)	0.177
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.180
0.20 g/210L Test #2 (Retry)	0.178
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.182
0.20 g/210L Test #3 (Retry)	0.180
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance.
Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN',P	04/30/2006	15:24:25	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #3	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #3	INT
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	15:38:42	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1 (Retry)	0.000
Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test #3 (Retry)	0.000

Remarks:

A F / M A: . 00: Ambient Fail. Int Det: Ambient Fail, Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	15:51:33	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

A F / M A: . Int Det: Ambient Fail, Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	04/30/2006	16:03:14	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200509B **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 09/22/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.072
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.073
0.08 g/210L Test #3 (Retry)	0.072
Interferent Detect Test #3	INT

Remarks:

A F / M A: . CONTROL OUTSIDE. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	05/05/2006	17:43:50	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.193
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.198
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.073
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.199
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	05/05/2006	18:40:02	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.078
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.079
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.079
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance WRONG SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
CLERMONT PD LEE,JOHN,P	06/15/2006	15:06:31	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	07/19/2006	14:41:10	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	]
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
CLERMONT PD LEE,JOHN,P	08/18/2006	12:29:05	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
CLERMONT PD LEE,JOHN,P	09/20/2006	14:36:53	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.000
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.000
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.190
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Sequence Aborted. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT PD LEE,JOHN,P	10/12/2006	14:46:05	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2006	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
CLERMONT P.D. SKIPPER,ROGER,G	11/09/2006	10:39:49	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	11/16/2006	15:14:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605B	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/04/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.183
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.186
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

INTERFERENT NOT DETECTED. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	12/15/2006	09:30:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source
LADY LAKE PD SKIPPER,ROGER,G	04/10/2006	11:44:59	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	415601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE PD OLSEN,WILLIAM,E	04/14/2006	21:16:49	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE PD OLSEN,WILLIAM,E	06/11/2006	23:41:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	
0.08 g/210L Test #1 (Retry)	0.082
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2 (Retry)	0.082
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3 (Retry)	0.082
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE PD OLSEN,WILLIAM,E	07/31/2006	23:06:15	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE PD OLSEN,WILLIAM,E	08/30/2006	04:57:26	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE PD OLSEN,WILLIAM,E	09/30/2006	23:43:36	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.053
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.205
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source
LADY LAKE P.D. SKIPPER,ROGER,G	10/26/2006	11:30:25	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.054
0.08 g/210L Test #1	0.085
0.08 g/210L Test #1 (Retry)	0.084
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.053
0.08 g/210L Test #2	0.085
0.08 g/210L Test #2 (Retry)	0.083
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.053
0.08 g/210L Test #3	0.085
0.08 g/210L Test #3 (Retry)	0.084
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.054
0.08 g/210L Test #4	0.084
0.08 g/210L Test #4 (Retry)	0.084
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.054
0.08 g/210L Test #5	0.085
0.08 g/210L Test #5 (Retry)	0.082
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.054
0.08 g/210L Test #6	0.084
0.08 g/210L Test #6 (Retry)	0.084
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.055
			0.08 g/210L Test #7	0.087
			0.08 g/210L Test #7 (Retry)	0.082
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.055
			0.08 g/210L Test #8	0.085
			0.08 g/210L Test #8 (Retry)	0.083
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.055
			0.08 g/210L Test #9	0.085
			0.08 g/210L Test #9 (Retry)	0.083
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.055
			0.08 g/210L Test #10	0.085
			0.08 g/210L Test #10 (Retry)	0.083
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: Control Outside Tolerance SIMM TOO HOT . ELECTRIC OULET OVERLOADED. CHANGED TO DIFFERENT OUTLET.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	10/30/2006	05:04:09	8100.27	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
Interferent Detect Test #3	INT

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	10/30/2006	05:37:16	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	11/30/2006	23:44:54	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A:OPORATOR ERROR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source
MIAMI PD VENTURI,GEORGE,L	02/21/2006	19:21:03	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.047
			0.08 g/210L Test #7	0.077
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

08: Control Outside Tolerance :FAILED TO CONNECT D/GAS HOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	03/22/2006	23:07:49	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

UNABLE TO OBTAIN SNM BLOW TO FAST AND STOPPED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	03/22/2006	23:21:23	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	04/10/2006	13:03:01	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD ANDREWS,KIMBERLY,	05/16/2006	20:08:26	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:OPERATOR ERROR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD ANDREWS,KIMBERLY,	05/16/2006	20:21:56	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509A	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	06/21/2006	19:48:22	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	07/25/2006	20:27:34	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	08/25/2006	19:14:34	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD ANDREWS,KIMBERLY,	09/26/2006	19:32:55	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	10/24/2006	19:34:08	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	11/02/2006	06:12:45	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source
MIAMI PD VENTURI,GEORGE,L	11/02/2006	15:09:17	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	11/02/2006	17:45:18	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD ANDREWS,KIMBERLY,	12/19/2006	18:01:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
MIAMI PD VENTURI,GEORGE,L	02/21/2006	19:28:01	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.075
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	03/22/2006	23:08:04	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	04/10/2006	12:59:39	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
MIAMI PD VENTURI,GEORGE,L	07/11/2006	13:51:20	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	525202I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	07/25/2006	20:27:13	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

A F / M A:OPERATOR ERROR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	08/25/2006	19:14:16	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	518802I	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	07/12/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.206
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.206
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.207
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD ANDREWS,KIMBERLY,	09/26/2006	19:30:06	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	10/24/2006	19:33:03	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.206
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.207
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.208
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	11/02/2006	06:09:25	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
MIAMI PD VENTURI,GEORGE,L	11/02/2006	15:35:35	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	11/02/2006	17:52:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD ANDREWS,KIMBERLY,	12/19/2006	18:05:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON P.D. HACKNEY,DWITE,N	01/11/2006	12:17:10	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON P.D. BROCK,JOHN,R	03/15/2006	10:56:34	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON P.D. BROCK,JOHN,R	03/15/2006	12:10:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	11/19/2006	**Expiration**	04/11/2006	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.042
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.043
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.044
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.191
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD HACKNEY,DWITE,N	03/23/2006	14:18:54	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.046
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD BROCK,JOHN,R	03/26/2006	18:00:35	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Purge Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD BROCK,JOHN,R	04/19/2006	13:15:22	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

FIRST SAMPLE OUT OF TOLERANCE. SIM CONNECTION RE-SECURED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD BROCK,JOHN,R	05/16/2006	08:40:40	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

NEXTEL BEING SHUT OFF CAUSED RFI, PROB CORRECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	06/16/2006	14:34:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	07/20/2006	09:47:11	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.186
0.20 g/210L Test #1 (Retry)	0.189
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.188
0.20 g/210L Test #2 (Retry)	0.190
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.190
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: ON .20 SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD BROCK,JOHN,R	07/20/2006	13:14:05	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.190
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08 AND .20 RE-RUN, BOTH PASSED. e..08 AND .20 BOTH RE-RUN, AND PASSED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD BROCK,JOHN,R	08/08/2006	12:57:08	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.191
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	09/14/2006	07:57:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	10/17/2006	08:56:07	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD HACKNEY,DWITE,N	11/02/2006	13:35:35	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.075
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD BROCK,JOHN,R	11/02/2006	16:46:20	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.187
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.190
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

20: 1ST TEST OUT OF TOL, SIM CONN RE-SECURED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	12/05/2006	10:04:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.186
0.20 g/210L Test #1 (Retry)	0.189
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.188
0.20 g/210L Test #2 (Retry)	0.190
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.189
0.20 g/210L Test #3 (Retry)	0.190
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:..20 OUT OF TOLERANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HACKNEY,DWITE,N	02/22/2006	09:50:26	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	03/26/2006	10:26:11	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

PASSED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO DECKARD,JOE,B	04/27/2006	09:43:08	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	05/30/2006	11:08:13	8100.26	Agency	Inspection

0.05 Lot#	20059A	**0.08 Lot#**	20059B	**0.20 Lot#**	20059C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance O-RING FAILURE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	06/27/2006	10:04:01	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HACKNEY,DWITE,N	09/05/2006	14:55:00	8100.26	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.204
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	09/26/2006	10:24:04	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HACKNEY,DWITE,N	10/25/2006	13:03:13	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.081
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.080
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.051
			0.08 g/210L Test #7	0.084
			0.20 g/210L Test #7	0.200
			0.08 g/210L Dry Gas Std Test #7	0.080
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.081
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0414
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

08: TUBE NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	10/25/2006	14:16:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	11/29/2006	19:54:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518601I
Expiration	05/04/2006	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	12/28/2006	20:00:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD HACKNEY,DWITE,N	01/12/2006	11:50:29	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD LANE,MICHAEL,A	03/27/2006	11:51:48	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

N

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	04/28/2006	11:11:30	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:INT. NOT DETECTED, . Non-compliance:INT. NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	04/28/2006	11:33:59	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.075
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance,OUTSIDE RANGENon-compliance:OUTSIDE RANGE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD LANE,MICHAEL,A	05/12/2006	13:46:25	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION MAY 2006

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	06/19/2006	12:45:09	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR JUNE 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD LANE,MICHAEL,A	07/14/2006	17:06:52	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	08/11/2006	10:05:22	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.047
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	08/11/2006	10:32:43	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION AUGUST 2006

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD LANE,MICHAEL,A	09/08/2006	11:41:04	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

MONTHLY AGENCY INSPECTION FOR SEPTEMBER 2006

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD LANE,MICHAEL,A	10/27/2006	09:39:24	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD HACKNEY,DWITE,N	11/02/2006	12:29:15	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.046
			0.08 g/210L Test #7	0.077
			0.20 g/210L Test #7	0.195
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0384
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0015

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
PANAMA CITY BEACH PD MALHIOT,MATTHEW,E	02/22/2006	09:32:51	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	525202I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.044
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.044
0.05 g/210L Test #4 (Retry)	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.045
0.05 g/210L Test #5 (Retry)	0.047
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.044
0.05 g/210L Test #6 (Retry)	0.047
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.044
			0.05 g/210L Test #7 (Retry)	0.047
			0.08 g/210L Test #7	0.076
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.076
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.045
			0.05 g/210L Test #8 (Retry)	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.075
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.045
			0.05 g/210L Test #9 (Retry)	0.047
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.075
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.045
			0.05 g/210L Test #10 (Retry)	0.047
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

05: Control Outside Tolerance. LEAK IN SIMULATOR HOSE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	03/27/2006	18:53:11	8100.26	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	03/27/2006	18:59:22	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.185
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.188
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.190
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	04/29/2006	06:47:10	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	05/30/2006	06:24:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	06/16/2006	06:53:45	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	07/21/2006	17:56:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	08/27/2006	22:24:39	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.070
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.071
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.071
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:FORGOT TO CLOSE PORT INTERFERENT INTRODUCED INTO. 08:CONTROL OUTSIDE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	09/26/2006	15:26:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
PANAMA CITY BEACH PD MALHIOT,MATTHEW,E	10/26/2006	12:04:52	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	10/30/2006	11:26:41	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	11/17/2006	16:25:14	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

AF/MA:MA OLD SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	12/01/2006	18:24:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

AF/MA:TOO MUCH SOLUTION IN MOUTH.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source
COLLIER COUNTY SO VENTURI,GEORGE,L	01/24/2006	12:08:55	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	03/22/2006	10:07:07	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	04/21/2007		

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.059
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.067
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	03/22/2006	11:10:34	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	03/22/2006	11:12:18	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

USED DIFFERENT SIMULATOR THAN ON THE 10:07 INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	04/16/2006	16:39:12	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	05/25/2006	09:52:20	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	06/21/2006	19:26:44	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	07/12/2006	10:02:39	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	07/12/2006	10:14:15	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.046
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.046
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

PRESSED ENTER KEY AT WRONG TIME.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	07/12/2006	10:36:10	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	08/02/2006	13:57:55	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	09/13/2006	08:38:08	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

AT IJC B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	10/05/2006	15:35:18	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	10/05/2006	15:37:17	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC B MOVED TO P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	10/23/2006	09:51:24	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

TESTED AT PDC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source
COLLIER COUNTY SO VENTURI,GEORGE,L	10/23/2006	14:53:30	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	10/23/2006	16:10:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

TESTED AT PDC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	11/13/2006	16:53:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/11/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	12/12/2006	16:16:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/11/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.071
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance SIM WARMUP TIME.TESTED AT IJC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	12/12/2006	16:50:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC MOVED TO BACKUP

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source
COLLIER COUNTY SO VENTURI,GEORGE,L	01/30/2006	14:15:56	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	425202I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	03/01/2006	11:56:29	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4252021
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	04/16/2006	17:40:49	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

BAT 2

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	05/24/2006	16:21:02	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	06/21/2006	12:53:19	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	07/12/2006	15:20:50	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

TESTED SPEC OPS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	08/02/2006	20:39:02	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

TESTED AT SPEC OPS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	09/12/2006	15:40:52	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512505I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

AT SPEC OPS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	10/05/2006	11:35:43	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

TESTED AT SPEC OPS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	10/22/2006	15:16:03	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance:HAD TOO MUCH MOUTH ALCOHOL STILL IN MOUTH .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	10/22/2006	16:27:08	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

TESTED AT SPECIAL OPS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source
COLLIER COUNTY SO VENTURI,GEORGE,L	10/23/2006	14:07:32	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	10/23/2006	15:08:00	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

TESTED AT PDC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	11/13/2006	10:36:36	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

TESTED AT SPEC OPS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001202

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	12/07/2006	14:01:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512502I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

THIRD MA TEST ON ONE SAMPLE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S HACKNEY,DWITE,N	01/12/2006	08:51:44	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	03/24/2006	14:02:21	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	04/30/2006	09:50:40	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	04/21/2007		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.076
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.077
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	04/30/2006	10:28:16	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.074
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

00: Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	05/01/2006	08:05:02	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	05/01/2006	08:12:29	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200503C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	05/28/2006	22:06:25	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	06/30/2006	17:51:15	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	07/31/2006	20:24:09	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	08/24/2006	06:41:58	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	09/25/2006	06:51:24	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	10/31/2006	09:06:17	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	No
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

00: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	10/31/2006	09:26:45	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	10/31/2006	09:29:32	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	10/31/2006	09:46:55	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S HACKNEY,DWITE,N	11/02/2006	13:20:14	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.076
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source	
GREENACRES P S PIPER,KEVIN,N	11/04/2006	08:37:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001203

Agency Operator	Date	Time	Software	Source
GREENACRES P S PIPER,KEVIN,N	12/22/2006	08:14:33	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001204

Agency Operator	Date	Time	Software	Source
PANAMA CITY PD MALHIOT,MATTHEW,E	02/22/2006	08:59:58	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	525202I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.045
0.05 g/210L Test #4 (Retry)	0.046
0.08 g/210L Test #4	0.075
0.08 g/210L Test #4 (Retry)	0.076
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.045
0.05 g/210L Test #5 (Retry)	0.047
0.08 g/210L Test #5	0.075
0.08 g/210L Test #5 (Retry)	0.076
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.05 g/210L Test #6 (Retry)	0.046

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001204

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #6	0.075
			0.08 g/210L Test #6 (Retry)	0.076
			0.20 g/210L Test #6	0.197
			0.08 g/210L Dry Gas Std Test #6	0.078
			Interferent Detect Test #6	INT
			Alcohol Free Test #7	0.000
			0.05 g/210L Test #7	0.046
			0.05 g/210L Test #7 (Retry)	0.047
			0.08 g/210L Test #7	0.075
			0.08 g/210L Test #7 (Retry)	0.076
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.046
			0.05 g/210L Test #8 (Retry)	0.046
			0.08 g/210L Test #8	0.075
			0.08 g/210L Test #8 (Retry)	0.076
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.05 g/210L Test #9 (Retry)	0.047
			0.08 g/210L Test #9	0.076
			0.08 g/210L Test #9 (Retry)	0.076
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.05 g/210L Test #10 (Retry)	0.047
			0.08 g/210L Test #10	0.076
			0.08 g/210L Test #10 (Retry)	0.076
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

05 AND 08 RUNNING LOW, LEAK IN SIMULATOR HOSE l Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001204

Agency Operator	Date	Time	Software	Source	
PANAMA CITY PD MALHIOT,MATTHEW,E	11/03/2006	08:08:37	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: NEVER IN USE AS OF THIS DATE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	02/20/2006	08:41:58	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA NEFF,STEPHEN,	04/11/2006	08:10:42	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.039
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.034
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.040
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.035
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.042
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	07/24/2006	14:43:48	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA WISE,GREGORY,A	07/25/2006	14:38:25	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA RHODES,GRADY,C	08/20/2006	23:57:30	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.068
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA WISE,GREGORY,A	09/18/2006	19:20:24	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA RHODES,GRADY,C	10/23/2006	00:05:47	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	10/23/2006	16:52:50	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Interferent Detect Test #6 (Retry)	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.051
			0.08 g/210L Test #7	0.082
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Interferent Detect Test #7 (Retry)	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Interferent Detect Test #8 (Retry)	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	0.000
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Int Det: HOSE LEAK ON LAST SAMPLE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA RHODES,GRADY,C	11/28/2006	00:18:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504502I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001205

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA WISE,GREGORY,A	12/19/2006	10:25:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504502I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
HENDRY COUNTY SO VENTURI,GEORGE,L	01/19/2006	12:03:08	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	434502I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	05/05/2007	**Expiration**	12/14/2006

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source	
HENDRY COUNTY SO LEWIS,ANDY,	05/11/2006	12:31:15	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4345021
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source	
HENDRY COUNTY SO LEWIS,ANDY,	06/29/2006	13:43:18	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4345021
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
HENDRY COUNTY SO LEWIS,ANDY,	07/28/2006	13:18:46	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.052
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.052
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside LOSE SEAL ON SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source	
HENDRY COUNTY SO LEWIS,ANDY,V	08/29/2006	12:41:48	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

05: Control Outside, FAILED TO CONNECT SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source	
HENDRY COUNTY SO LEWIS,ANDY,	09/21/2006	12:42:42	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source	
HENDRY COUNTY SO LEWIS,ANDY,	10/30/2006	14:29:27	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
HENDRY COUNTY SO LEWIS,ANDY,	11/08/2006	10:19:31	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
HENDRY COUNTY SO VENTURI,GEORGE,L	11/08/2006	15:16:20	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	434502I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/14/2006

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

AI CAPT A LEWIS PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source
HENDRY COUNTY SO LEWIS,ANDY,	11/08/2006	16:46:21	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001207

Agency Operator	Date	Time	Software	Source	
HENDRY COUNTY SO LEWIS,ANDY,	12/28/2006	11:22:52	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

Non-compliance:EBS GAS CYLINDER EXPIRED 12/14/2006. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO VENTURI,GEORGE,L	01/24/2006	13:30:59	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504502I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	03/02/2006	10:08:24	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504502I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	04/11/2006	13:39:43	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	05/12/2006	10:46:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200411G	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	11/18/2006	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO CLAFLIN,DAVID,S	06/08/2006	14:47:42	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	07/14/2006	12:20:12	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	08/06/2006	12:04:29	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	09/01/2006	11:22:10	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

UNKOWNNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO VENTURI,GEORGE,L	11/21/2006	16:52:43	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	626101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001208

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source
TYNDALL AFB MALHIOT,MATTHEW,E	02/22/2006	13:28:59	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MARTIN,GORDON,T	03/30/2006	09:57:51	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.076
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:*CONTOL OUTSIDE TOLERANCE
.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source
TYNDALL AFB MARTIN,GORDON,T	03/30/2006	12:55:59	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	5045021
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	11/04/2006	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

OBSERVED ALL TEST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MALHIOT,MATTHEW,E	10/26/2006	11:06:40	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source
TYNDALL AFB MALHIOT,MATTHEW,E	10/26/2006	13:59:26	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

A F / M A: NO M/A USES ON SECOND SAMPLE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MELILLO,GARY,W	11/02/2006	09:47:24	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	04/11/2006		

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.165
0.20 g/210L Test #1 (Retry)	0.189
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.180
0.20 g/210L Test #2 (Retry)	0.190
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.185
0.20 g/210L Test #3 (Retry)	0.192
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MELILLO,GARY,W	12/09/2006	10:48:33	8100.27	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.054
0.08 g/210L Test #1 (Retry)	0.062
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.060
0.08 g/210L Test #2 (Retry)	0.063
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.066
0.08 g/210L Test #3 (Retry)	0.065
Interferent Detect Test #3	INT

Remarks:

08: BAD SIMULATOR. Non-compliance:BAD SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MELILLO,GARY,W	12/09/2006	11:46:41	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	200605F
Expiration	04/21/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT

Remarks:

A F / M A: . 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MELILLO JR.,GARY,W	12/09/2006	11:52:16	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001230

Agency Operator	Date	Time	Software	Source	
TYNDALL AFB MELILLO JR.,GARY,W	12/09/2006	11:52:40	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
AUBURNDALE PD SUERETH,DONALD,P	02/02/2006	10:07:08	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	03/29/2006	14:00:50	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	04/25/2006	09:48:06	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: FAILED TO HOOK UP SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	05/19/2006	09:08:32	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
AUBURNDALE PD FLORA,MICHAEL,O	06/19/2006	15:56:17	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	07/17/2006	09:22:34	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
AUBURNDALE PD FLORA,MICHAEL,O	08/23/2006	09:01:28	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	09/18/2006	13:48:06	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
AUBURNDALE PD FLORA,MICHAEL,O	10/26/2006	15:09:03	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	431401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/11/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
AUBURNDALE PD SUERETH,DONALD,P	11/02/2006	15:15:52	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0015 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.204
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0038
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	11/02/2006	19:32:36	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001231

Agency Operator	Date	Time	Software	Source	
AUBURNDALE PD FLORA,MICHAEL,O	12/25/2006	03:23:29	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

INTERFERENT NOT ADDED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	02/20/2006	09:15:35	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.045
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA CRUTCHFIELD,GARY,L	03/28/2006	11:40:05	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	
Expiration	11/04/2006	**Expiration**	11/18/2006		

 # Sims 2 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.062
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.069
0.08 g/210L Test #1 (Retry)	0.071
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.061
0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test #2	0.071
0.08 g/210L Test #2 (Retry)	0.072
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.057
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.072
0.08 g/210L Test #3 (Retry)	0.073
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: CHANGED SIM, Control Outside Tolerance. SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA NEFF,STEPHEN,	04/11/2006	09:18:22	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	411803I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/30/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.043
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance LEAK IN SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	07/24/2006	14:26:52	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA RHODES,GRADY,C	08/21/2006	01:08:11	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source	
NAS PENSACOLA MALHIOT,MATTHEW,E	10/23/2006	14:26:27	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI FOR UPLAOD ONLY, AI NOT COMPLETENon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	10/23/2006	15:25:33	8100.27	Department Inspection

 # Sims 0 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RESTART AFTER MIN SAMPLE (BLEW TOO HARD)Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
NAS PENSACOLA MALHIOT,MATTHEW,E	10/23/2006	16:59:06	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	INT
0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.198
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
Interferent Detect Test #6 (Retry)	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
Interferent Detect Test #7 (Retry)	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050
0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.199
Interferent Detect Test #8 (Retry)	INT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001232

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0012

Remarks:

Int Det: Sequence Aborted. 08: Interferent Detect, DRY GAS TEST. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO BG HACKNEY,DWITE,N	01/12/2006	08:51:42	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO BG MCKINSTRY,DAVID,E	03/18/2006	12:31:17	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	
Expiration	11/04/2006	**Expiration**	11/18/2006		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.046
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.074
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.074
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO BG MCKINSTRY,DAVID,E	03/18/2006	14:18:35	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 2 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.068
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.071
0.08 g/210L Test #2 (Retry)	0.074
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.073
0.08 g/210L Test #3 (Retry)	0.075
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	04/11/2006	13:53:08	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	05/22/2006	09:53:09	8100.26	Agency Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	510902I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.186
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: LOOSE CONECTION. 20: LOOSE CONECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	06/21/2006	10:32:19	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	06/21/2006	11:31:02	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	06/21/2006	11:32:14	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	07/31/2006	09:13:45	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.073
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08:SIMULATOR AIR LEAK, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	07/31/2006	12:48:55	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	08/21/2006	09:46:30	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	09/23/2006	10:02:56	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG HACKNEY,DWITE,N	10/16/2006	21:49:03	8100.27	Department Inspection

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: No Sample Provided. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG HACKNEY,DWITE,N	10/16/2006	23:34:55	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	10/17/2006	01:45:05	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	11/10/2006	12:42:06	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001233

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	12/08/2006	13:30:10	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG HACKNEY,DWITE,N	01/17/2006	08:32:57	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	03/18/2006	12:49:16	8100.26	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Y.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	03/18/2006	13:09:55	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	
Expiration	11/04/2006	**Expiration**	11/18/2006		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.065
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.070
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.069
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.072
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.071
		0.08 g/210L Test #3 (Retry)	0.072
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	04/11/2006	13:29:40	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

20: LOOSE CONNECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	05/22/2006	09:50:51	8100.26	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	05/22/2006	10:07:26	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	518702I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.185
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

20: LOOSE CONECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	06/21/2006	10:32:05	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.083
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	07/31/2006	09:29:57	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	1234
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.077
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.198
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG MCKINSTRY,DAVID,E	07/31/2006	12:19:13	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

20: LOOSE CONECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG MCKINSTRY,DAVID,E	08/21/2006	09:46:12	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG CROUCHER,GREGORY,P	09/23/2006	10:56:46	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

INT NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO BG CROUCHER,GREGORY,P	09/23/2006	11:59:13	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

NO TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG HACKNEY,DWITE,N	10/16/2006	21:50:21	8100.27	Department Inspection

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: No Sample Provided. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG HACKNEY,DWITE,N	10/17/2006	00:20:10	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.082
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	10/17/2006	01:52:22	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	11/10/2006	11:47:14	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001234

Agency Operator	Date	Time	Software	Source
PALM BEACH SO BG CROUCHER,GREGORY,P	12/08/2006	13:35:56	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

FIRST MOUTH ALCOHOL DISIPATED BEFORE ABLE TO GIVE SAMPLE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD HACKNEY,DWITE,N	01/12/2006	08:51:04	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MIKE,L	03/22/2006	08:22:01	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MIKE,K	03/22/2006	08:22:50	8100.26	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	03/22/2006	10:49:51	8100.26	Agency	Inspection

0.05 Lot#	200403E	**0.08 Lot#**	200403F	**0.20 Lot#**	
Expiration	03/26/2006	**Expiration**	03/26/2006		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.064
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.071
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.067
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.073
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.070
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	03/22/2006	11:28:02	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	04/22/2006	06:41:38	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.042
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside SIMULATOR LEAKING AIR BOTH TIMES. Non-compliance: OUT SIDE 05 TOLLERANCE WILL RETEST AFTER FI.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	04/22/2006	07:08:53	8100.26	Agency	Inspection

0.05 Lot# 200509A **0.08 Lot#** 200509B **0.20 Lot#**
Expiration 09/22/2007 **Expiration** 09/22/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.072
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.073
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:NOT DETECTED FIRST TEST. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	04/22/2006	08:05:56	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.176
0.20 g/210L Test #1 (Retry)	0.186
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.179
0.20 g/210L Test #2 (Retry)	0.187
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.180
0.20 g/210L Test #3 (Retry)	0.187
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	04/24/2006	10:41:19	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.05 g/210L Test #1 (Retry)	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.180
0.20 g/210L Test #1 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.184
0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.188
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Improper Sample.BLEW INTO MOUTH PIECE WHEN SHOULD NOT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	05/24/2006	14:00:38	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	06/03/2006	06:53:15	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

N/A

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	07/28/2006	06:59:26	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	08/12/2006	07:07:48	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	09/04/2006	12:28:30	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	10/16/2006	10:10:57	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.056
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside SIMULATOR NOT SEALED HAD AIR LEAK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD HACKNEY,DWITE,N	11/02/2006	12:03:27	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	11/09/2006	10:32:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.031
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.071
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.075
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:SIMULATOR LEAKING AIR. WILL REPAIR AND REINS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	11/09/2006	10:41:37	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	11/09/2006	11:16:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source	
WEST PALM BEACH PD WIGGS,MICHAEL,K	12/10/2006	06:49:10	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , Sequence Aborted. Non-compliance:INTER NOT ADDED IN TIME AND NEED TO BEGIN AG.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001235

Agency Operator	Date	Time	Software	Source
WEST PALM BEACH PD WIGGS,MICHAEL,K	12/10/2006	07:24:42	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:SIM JAR BROKEN NOT SEALED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source
LAKE WORTH PD HACKNEY,DWITE,N	01/11/2006	12:17:50	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	03/30/2006	15:01:11	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source
LAKE WORTH PD FRANKIEWICZ,RICHARD,	03/30/2006	15:50:18	8100.26	Agency Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	5109021
Expiration	04/11/2006	**Expiration**	11/19/2006	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	04/28/2006	22:52:39	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	05/16/2006	22:58:39	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Test #2	0.000
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	0.000
Alcohol Free Test #3	0.000
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	0.000

Remarks:

Int Det:NO,YES.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	05/16/2006	23:18:11	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.007
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.012
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	05/18/2006	19:58:06	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.042
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.042
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance. Non-compliance:.050 OUTSIDE OF TOLERANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	05/19/2006	10:15:53	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	06/23/2006	17:45:33	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	07/31/2006	16:37:22	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509A	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	08/31/2006	21:40:17	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.054
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.053
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	09/29/2006	19:41:25	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	10/30/2006	17:10:05	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source
LAKE WORTH PD HACKNEY,DWITE,N	11/02/2006	12:40:19	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.075
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.075
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.075
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.046
0.08 g/210L Test #7	0.076
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.075
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.075
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	11/30/2006	20:49:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001236

Agency Operator	Date	Time	Software	Source	
LAKE WORTH PD FRANKIEWICZ,RICHARD,	12/29/2006	10:54:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	02/02/2006	12:48:10	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	431401I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/11/2006

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.076
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	INT
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.202

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8 (Retry)	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: RFI Detect. 08: Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	04/08/2006	04:37:41	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	415602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO TARTE,PHILLIP,A	05/03/2006	11:15:54	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	415602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	06/15/2006	10:40:11	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.008
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

00: LOOSE CONNECTION. 05: LOOSE SEAL. 08: LOOSE HOSE SEAL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	07/08/2006	11:14:36	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	08/10/2006	09:11:37	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	09/07/2006	09:57:36	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	619101I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.007
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. 08: LOOSE HOSE SEAL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	10/12/2006	10:05:23	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
Interferent Detect Test #3	INT

Remarks:

08: INLET HOSE LOOSE SEAL, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	10/12/2006	10:36:44	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	619101I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	10/24/2006	16:45:58	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.08 g/210L Test #4 (Retry)	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.075
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.08 g/210L Test #5 (Retry)	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.075
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.08 g/210L Test #6 (Retry)	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.075
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.078
			0.08 g/210L Test #7 (Retry)	0.079
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.075
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.08 g/210L Test #8 (Retry)	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.075
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.08 g/210L Test #9 (Retry)	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.075
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.08 g/210L Test #10 (Retry)	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0017
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	10/24/2006	19:30:00	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612201G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	11/10/2006	11:09:27	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	619101I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001237

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	12/13/2006	09:09:49	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND PD SKIPPER,ROGER,G	03/16/2006	10:12:04	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND PD BARR,RICHARD,S	03/27/2006	10:39:53	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND PD BARR,RICHARD,S	04/21/2006	09:29:16	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND PD BARR,RICHARD,S	05/25/2006	15:20:28	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND PD BARR,RICHARD,S	06/21/2006	11:05:29	8100.26	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	INT
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect, Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND PD BARR,RICHARD,S	06/21/2006	11:56:11	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DIAGNOSTIC TEST FAILEDDIAGNOSTIC TEST FAILURE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND PD BARR,RICHARD,S	06/22/2006	14:39:28	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND PD BARR,RICHARD,S	07/19/2006	10:32:27	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND PD BARR,RICHARD,S	08/17/2006	10:17:53	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND PD BARR,RICHARD,S	09/15/2006	12:46:57	8100.26	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DID NOTPICK UP INTERFERRENT. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND PD BARR,RICHARD,S	09/15/2006	13:16:59	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND PD BARR,RICHARD,S	10/27/2006	11:39:40	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	6270021
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/29/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
GROVELAND P.D. SKIPPER,ROGER,G	10/27/2006	17:01:53	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND P.D. BARR,RICHARD,S	11/20/2006	12:06:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627002I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/29/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001238

Agency Operator	Date	Time	Software	Source	
GROVELAND P.D. BARR,RICHARD,S	12/08/2006	11:11:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627002I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/29/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001242

Agency Operator	Date	Time	Software	Source
DAYTONA BEACH SHORES SKIPPER,ROGER,G	03/16/2006	10:13:15	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	INT
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001242

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.048
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.202
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

08: Interferent Detect CYLINDER CHANGED. RESTARTED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001242

Agency Operator	Date	Time	Software	Source
DAYTONA BEACH SHORES SKIPPER,ROGER,G	10/23/2006	18:05:22	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	618801I	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001242

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.075
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.075
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

LOT #/EXPIRATION FOR 0.20 SHOULD BE 200509C/09-22-2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD SKIPPER,ROGER,G	02/24/2006	08:50:25	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 4 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	INT
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #8 (Retry)	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

00: Interferent Detect REASON UNKNOWN RETEST. TUBE CHANGE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD MCELROY,GARY,T	03/30/2006	09:13:24	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	415601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD MCELROY,GARY,T	04/27/2006	08:18:09	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	415601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/10/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

NEXTEL PHONE IN AREA. LOOSE INTAKE HOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD ROOS,CHRISTOPHER,R	06/08/2006	13:23:31	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	415601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/10/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

WRONG SIMULATOR HOOKED UP AT START OF TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source	
EDGEWATER PD ROOS,CHRISTOPHER,R	07/31/2006	07:09:09	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.071
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.073
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

SIMULATOR WAS TURNED OFF AND UNDER TEMP..08 FAILED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD ROOS,CHRISTOPHER,R	07/31/2006	07:53:26	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

CHECKED AND TIGHTENED SEALS ON .05 SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD ROOS,CHRISTOPHER,R	08/30/2006	12:01:41	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

STOPPED BLOWING DURING MOUTH ALCOHOL TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD ROOS,CHRISTOPHER,R	09/27/2006	11:16:33	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER PD ROOS,CHRISTOPHER,R	10/29/2006	10:13:57	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER P.D. SKIPPER,ROGER,G	11/07/2006	11:09:58	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER P.D. ROOS,CHRISTOPHER,R	11/17/2006	12:22:10	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001243

Agency Operator	Date	Time	Software	Source
EDGEWATER P.D. ROOS,CHRISTOPHER,R	12/20/2006	10:36:17	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source
PUTNAM COUNTY SO SKIPPER,ROGER,G	02/16/2006	11:36:18	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

A F / M A: IMPROPER SAMPLE PROVIDED FIRST TIME..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	03/27/2006	17:27:35	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200505C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	03/29/2006	20:37:34	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	04/04/2006	20:25:12	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	05/07/2006	19:53:56	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	05/07/2006	20:45:52	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	06/04/2006	20:50:44	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	07/06/2006	20:41:37	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	08/01/2006	19:31:19	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	09/21/2006	19:30:41	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	10/19/2006	14:11:08	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source
PUTNAM COUNTY S.O. SKIPPER,ROGER,G	10/30/2006	11:28:16	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618802I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.075
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	11/27/2006	11:06:53	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:TOO MUCHALCOHOL USED. Non-compliance:TOO MUCH ALCOHOL USED IN MOUTH ALCOHOL TEST .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	11/27/2006	11:53:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001244

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	12/19/2006	17:27:16	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

05: Control Outside ToleranceAIR LEAK IN SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source
PORT ORANGE PD SKIPPER,ROGER,G	02/16/2006	16:09:53	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD ZUBER,RAYMOND,T	03/22/2006	21:50:47	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	416701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD COX,MICHAEL,G	03/31/2006	16:08:44	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	416701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD COX,MICHAEL,G	04/30/2006	23:53:04	8100.26	Agency	Inspection

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD COX,MICHAEL,G	05/01/2006	00:20:24	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	416701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD COX,MICHAEL,G	06/10/2006	10:28:16	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	416701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/17/2006

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

THE SIMULATOR RETURN TUBE SLIPPED OFF.. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD ZUBER,RAYMOND,T	07/09/2006	08:18:19	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: . 05: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD ZUBER,RAYMOND,T	08/10/2006	09:54:42	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD ZUBER,RAYMOND,T	09/25/2006	19:51:10	8100.26	Agency	Inspection

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:INSUFICIENT SLOPE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD ZUBER,RAYMOND,T	09/25/2006	19:58:29	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE PD COX,MICHAEL,G	10/29/2006	19:32:10	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source
PORT ORANGE P.D. SKIPPER,ROGER,G	11/07/2006	14:03:54	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

00: Ambient Fail TOO MUCH M/A SOLUTION USED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE P.D. ZUBER,RAYMOND,T	11/15/2006	10:05:18	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.191
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE P.D. COX,MICHAEL,G	12/19/2006	09:39:40	8100.27	Agency	Inspection

0.05 Lot# 200509A **0.08 Lot#** 200509B **0.20 Lot#**
Expiration 09/22/2007 **Expiration** 09/22/2007

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.074
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.074
Interferent Detect Test #3	INT

Remarks:

08: BAD O-RING, Control Outside Tolerance. Non-compliance:0.080G/210L TEST OUT OF RANGE, POS. AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001245

Agency Operator	Date	Time	Software	Source	
PORT ORANGE P.D. COX,MICHAEL,G	12/19/2006	10:43:03	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
SOUTH DAYTONA PD SKIPPER,ROGER,G	02/22/2006	13:20:17	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.078
0.08 g/210L Test #4 (Retry)	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.077
0.08 g/210L Test #5 (Retry)	0.077
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6 (Retry)	0.077
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7 (Retry)	0.077
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8 (Retry)	0.077
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9 (Retry)	0.077
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10 (Retry)	0.077
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: RFI Detect CELL PHONE USED IN AREA.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	03/29/2006	00:30:49	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Purge Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	04/30/2006	04:10:17	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	04/30/2006	20:31:01	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance:ACCIDENTALY ENTERED N TO RESTART TEST .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	04/30/2006	20:41:18	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	05/31/2006	18:27:43	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	4364011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	06/30/2006	17:35:21	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
SOUTH DAYTONA PD RIZZUTO,GEORGE,E	07/27/2006	21:14:00	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	08/29/2006	22:03:17	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	09/27/2006	00:04:42	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	
0.20 g/210L Test #1 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA PD LABOMBARD,ANTHONY,J	10/31/2006	12:22:30	8100.26	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.044
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
SOUTH DAYTONA P.D. SKIPPER,ROGER,G	11/06/2006	15:19:58	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA P.D. LABOMBARD,ANTHONY,J	11/30/2006	13:15:59	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA P.D. LABOMBARD,ANTHONY,J	11/30/2006	15:37:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.073
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001246

Agency Operator	Date	Time	Software	Source	
SOUTH DAYTONA P.D. LABOMBARD,ANTHONY,J	12/31/2006	15:02:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	436401I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/04/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	01/17/2006	11:34:09	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	03/22/2006	16:39:14	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512501I
Expiration	11/04/2006	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/24/2006	22:07:59	8100.26	Agency	Inspection

0.05 Lot#	200502A	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512501I
Expiration	02/14/2007	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/18/2006	21:48:26	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	06/16/2006	10:26:20	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

PATRICIA NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/22/2006	18:25:11	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	08/09/2006	11:50:37	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

F. DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	09/22/2006	10:11:40	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	10/18/2006	17:18:03	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	10/20/2006	11:45:30	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

F. DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	11/09/2006	15:57:16	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

F.DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	11/17/2006	09:46:16	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	12/15/2006	10:05:50	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001247

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	12/21/2006	09:40:52	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

BUTTON PUSHED AT WRONG TIME. PAT NANZ WAS PRESENT FOR ENTIR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	01/17/2006	08:32:54	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO DICIENZO,FRANCA,R	03/22/2006	15:51:33	8100.26	Agency Inspection

0.05 Lot#	411803I	**0.08 Lot#**	411803I	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	411803I
Expiration	04/30/2006	**Expiration**	04/30/2006	**Expiration**	11/04/2006	**Expiration**	04/30/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

ENTERED INCORRECT LOT # FOR .05 AND .08 SOLUTIONS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO VENTURI,GEORGE,L	04/06/2006	17:26:35	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	411803I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/30/2006

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

AI F. DICIENZO PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	04/19/2006	21:10:53	8100.26	Agency	Inspection

0.05 Lot#	200502A	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	411803I
Expiration	02/14/2007	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	04/30/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/19/2006	00:33:56	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	06/08/2006	15:17:23	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/22/2006	17:53:47	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	08/09/2006	12:34:18	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

F. DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	09/22/2006	11:02:36	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	10/18/2006	14:54:38	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.080
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.080
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.047
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.194
			0.08 g/210L Dry Gas Std Test #7	0.080
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0383
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: TUBE NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	10/20/2006	09:51:08	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.072
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

LOOSE CONNECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	11/17/2006	10:27:57	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

F.DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001248

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	12/15/2006	11:00:46	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	01/17/2006	15:23:57	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	03/23/2006	13:58:53	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512402I
Expiration	11/04/2006	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/24/2006	23:04:05	8100.26	Agency	Inspection

0.05 Lot#	200502A	**0.08 Lot#**	200411G	**0.20 Lot#**	512501I	**0.08 Gas Lot#**	512402I
Expiration	02/14/2007	**Expiration**	11/19/2006	**Expiration**	05/10/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,N	05/18/2006	23:07:45	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

FRANCA DICIENZO WAS PRESENT FOR THE ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO DICIENZO,FRANCA,R	06/08/2006	16:34:06	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

N

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/22/2006	21:28:32	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.084
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	08/09/2006	10:50:49	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

F. DICIENZO WAS PRESENT AND OBSERVED ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,	09/22/2006	10:14:56	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	10/18/2006	15:18:03	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.190
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.075
0.20 g/210L Test #4	0.190
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.075
0.20 g/210L Test #5	0.191
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.075
0.20 g/210L Test #6	0.192
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.046
0.08 g/210L Test #7	0.075
0.20 g/210L Test #7	0.191
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.191
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.192
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.192
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	10/20/2006	12:33:11	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	11/17/2006	12:23:10	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

F.DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001249

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	12/15/2006	11:52:41	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	01/17/2006	11:53:19	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result	Test	Result

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	03/23/2006	13:54:04	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512501I
Expiration	11/04/2006	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.201
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.20 g/210L Test #3 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	04/19/2006	23:48:51	8100.26	Agency	Inspection

0.05 Lot#	200502A	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512501I
Expiration	02/14/2007	**Expiration**	11/19/2006	**Expiration**	11/04/2006	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/18/2006	23:05:04	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

PATRICIA NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	06/08/2006	16:23:48	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	11/19/2006	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

N

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/22/2006	20:09:06	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	08/09/2006	11:07:06	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

PATRICIA NANZ WAS PRESENT AND OBSERVED INSPCTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,	09/22/2006	11:09:40	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200505C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO HACKNEY,DWITE,N	10/18/2006	17:14:34	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.048
			0.08 g/210L Test #7	0.076
			0.20 g/210L Test #7	0.196
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: GAS CONTAINER NOT ATTACHED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	10/20/2006	11:43:50	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	11/09/2006	16:12:34	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

DID NOT BLOW HARD ENOUGH. P.NANZ WAS PRESENT AND OBSERVED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	11/17/2006	09:39:19	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

F.DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001250

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO DICIENZO,FRANCA,R	12/15/2006	09:57:04	8100.27	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

ACETONE NEEDED TO BE ADDED.. 05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S0 SKIPPER,ROGER,G	02/13/2006	12:35:02	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S0 THRUSH,DANIEL,C	03/20/2006	12:33:10	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	20504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	411803I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/30/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	04/17/2006	12:35:11	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.009
Date and/or Time Adjusted	Yes	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.047
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.048
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. 00: Control Outside Tolerance. 05: RFI Detect, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S0 WAGNER,ROBERT,F	04/17/2006	13:01:08	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	45202I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	09/10/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	05/22/2006	10:41:28	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S0 WAGNER,ROBERT,F	06/26/2006	12:17:35	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	06/26/2006	13:02:01	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S0 WAGNER,ROBERT,F	07/24/2006	11:02:46	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

PRINTER HOOK UPNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	07/24/2006	11:38:43	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	07/24/2006	11:40:21	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	07/24/2006	12:16:29	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance:TEMP REG FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	07/24/2006	12:44:42	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

QUIT BLOWING 2 SOON.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	08/28/2006	08:14:44	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	09/18/2006	08:40:36	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S0 WAGNER,ROBERT,F	10/16/2006	10:12:20	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S.O. SKIPPER,ROGER,G	10/17/2006	16:34:17	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.200
			0.08 g/210L Dry Gas Std Test #7	0.079
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0407
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: Control Outside Tolerance. GAS HOSE NOT CONNECTED. REDO. Non-compliance: DIAG FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
BREVARD S.O. SKIPPER,ROGER,G	10/17/2006	20:00:59	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.054
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.053
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	10/18/2006	18:20:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	11/13/2006	10:07:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001253

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	12/19/2006	17:08:17	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO HACKNEY,DWITE,N	01/11/2006	09:15:57	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
MONROE COUNTY SO HACKNEY,DWITE,N	01/11/2006	09:18:37	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.080
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.079
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.203
			0.08 g/210L Dry Gas Std Test #7	0.079
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.079
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0382
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: TUBING NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	03/22/2006	14:12:59	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededAmbient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	03/22/2006	14:30:17	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	4191021
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	04/17/2006	11:07:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	419102I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
MONROE COUNTY SO RABINOWITZ,ADAM,	05/23/2006	10:47:18	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	419102I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/14/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
Interferent Detect Test #3	INT

Remarks:

08: Purge Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	06/26/2006	14:41:36	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612202G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	07/10/2006	12:25:03	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612202G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
MONROE COUNTY SO RABINOWITZ,ADAM,	08/01/2006	10:28:21	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612202G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	09/20/2006	10:34:36	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

OLD MOUTH ALCOHOL. IN COMPLETE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	09/20/2006	10:44:58	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	10/17/2006	14:45:27	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
MONROE COUNTY SO HACKNEY,DWITE,N	10/19/2006	12:29:20	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
MONROE COUNTY SO RABINOWITZ,ADAM,	10/19/2006	13:26:43	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

CP AI AFTER SOFTWARE UPDATE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO RABINOWITZ,ADAM,	11/30/2006	10:36:31	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001254

Agency Operator	Date	Time	Software	Source
MONROE COUNTY SO BLASCO,LUIS,	12/27/2006	14:51:24	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	62202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COMMAND POST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	01/24/2006	09:59:28	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: Control Outside Tolerance TUBE DISCONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/24/2006	10:51:52	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	04/11/2006	14:10:54	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	05/17/2006	10:42:02	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	06/21/2006	10:20:31	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	07/19/2006	12:40:36	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	08/16/2006	09:04:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	09/28/2006	09:42:55	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	10/16/2006	16:25:42	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	10/18/2006	12:41:05	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	11/15/2006	12:02:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001255

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	12/13/2006	13:11:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	01/24/2006	10:00:25	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/24/2006	10:50:09	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	04/11/2006	12:43:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	05/17/2006	08:34:10	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	428101I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	05/23/2006	11:10:39	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source
ORANGE S.O. SKIPPER,ROGER,G	08/02/2006	13:09:57	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

INSPECTION AFTER REPAIR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE S.O. MELVILLE,KELLY,M	08/16/2006	09:12:53	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE S.O. MELVILLE,KELLY,M	09/28/2006	09:32:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source
ORANGE S.O. SKIPPER,ROGER,G	10/16/2006	15:06:51	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001256

Agency Operator	Date	Time	Software	Source	
ORANGE S.O. MELVILLE,KELLY,M	10/18/2006	12:52:07	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	01/25/2006	12:29:51	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/24/2006	10:53:40	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	04/11/2006	14:10:44	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	05/17/2006	10:41:59	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	06/21/2006	10:20:32	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512402I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	07/19/2006	10:27:46	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	08/16/2006	09:29:58	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	09/28/2006	09:32:16	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	10/16/2006	13:21:08	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	10/18/2006	12:44:32	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	11/15/2006	12:06:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001257

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	12/13/2006	12:42:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	01/24/2006	09:58:09	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.000
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.20 g/210L Test #4 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.20 g/210L Test #5 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.20 g/210L Test #6 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.198
			0.20 g/210L Test #7 (Retry)	0.198
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.20 g/210L Test #8 (Retry)	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.20 g/210L Test #9 (Retry)	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.20 g/210L Test #10 (Retry)	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0622
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

Time-Date changed. 20: Control Outside Tolerance SIM TUBE NOTDISCONECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/24/2006	12:44:24	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518801I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	04/11/2006	14:11:03	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518801I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	05/17/2006	10:42:04	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518801I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	06/21/2006	12:38:34	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	07/19/2006	10:27:45	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT

Remarks:

08: AI FAILED TO CONNECT DGS CYLINDER HOSE ON 1ST CONT TEST. .08/AI DID NOT CONNECT DGS SUPPLY HOSE ON 1ST CONT TEST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	08/16/2006	09:03:33	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	09/28/2006	09:32:04	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	10/16/2006	15:51:38	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	10/18/2006	12:41:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	11/15/2006	12:01:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001258

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	12/13/2006	13:15:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	02/08/2006	10:24:08	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0020
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	03/22/2006	23:46:18	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	434501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	12/14/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	04/20/2006	15:05:07	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	05/17/2006	17:58:23	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	06/08/2006	15:10:21	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	INT
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
Interferent Detect Test #3	INT

Remarks:

08: Interferent Detect,UNKNOWN SOURCENon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	06/08/2006	17:24:45	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/11/2006	23:53:20	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

Printed: 13 June 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	08/01/2006	15:13:02	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance SIMULATOR NOT SEALED TIGHT .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	09/18/2006	20:43:29	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	10/17/2006	14:48:56	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.195
0.20 g/210L Test #4 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.195
0.20 g/210L Test #5 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.20 g/210L Test #6 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.196
			0.20 g/210L Test #7 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #7	0.076
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.195
			0.20 g/210L Test #8 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.20 g/210L Test #9 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.196
			0.20 g/210L Test #10 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0022
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

20: Control Outside Tolerance. HOSE CONNECTION TIGHTENED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	10/17/2006	16:11:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	11/21/2006	21:12:49	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001259

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	12/22/2006	18:08:00	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency	Date	Time	Software	Source
COCOA PD	01/17/2006	08:31:41	8100.26	Department Inspection
HACKNEY,DWITE,N				

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.204
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source	
COCOA PD FRAZIER,JAMES,H	03/21/2006	10:25:13	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.007
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.055
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.055
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.055
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. 08: Interferent Detect, Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source	
COCOA PD FRAZIER,JAMES,H	04/11/2006	08:52:29	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance:FIRST MOUTH ALCOHOL TEST DIDNOT SHOW SLOPE N.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	04/11/2006	09:26:54	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.055
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.055
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.205
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.054
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.206
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	05/03/2006	12:47:19	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

N

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source	
COCOA PD FRAZIER,JAMES,H	06/01/2006	07:55:27	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source	
COCOA PD FRAZIER,JAMES,H	07/11/2006	08:51:12	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	08/09/2006	07:48:44	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	09/05/2006	08:33:22	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	10/04/2006	07:50:58	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	10/20/2006	08:03:30	8100.26	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD HACKNEY,DWITE,N	10/20/2006	11:31:07	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.197
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.20 g/210L Test #4 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.20 g/210L Test #5 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9 (Retry)	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0016
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source
COCOA PD FRAZIER,JAMES,H	10/20/2006	12:14:26	8100.27	Agency Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source	
COCOA PD FRAZIER,JAMES,H	11/07/2006	09:35:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001260

Agency Operator	Date	Time	Software	Source	
COCOA PD FRAZIER,JAMES,H	12/12/2006	10:27:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source
ROCKLEDGE PD HACKNEY,DWITE,N	01/17/2006	10:55:56	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	03/26/2006	20:52:42	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.078
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.041
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.021
0.05 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.08 g/210L Test: OK	No	Interferent Detect Test #3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Control Outside Tolerance. Int Det: , . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	03/26/2006	21:40:05	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.070
0.08 g/210L Test #1 (Retry)	0.103
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.110
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.110
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 05: Control Outside Tolerance. 08: Control OutsideSOLUTION OUT OF RANGENon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	03/26/2006	22:40:58	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	03/26/2006	22:41:39	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.011
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.026
Barometric Pressure Sensor Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.040
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.044
0.05 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.042
		0.05 g/210L Test #3 (Retry)	0.044
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. SOLUTION NOT WITH IN RANGE. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	05/23/2006	01:29:17	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

Printed: 13 June 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	05/23/2006	01:38:42	8100.26	Agency	Inspection

0.05 Lot# 200504C **0.08 Lot#** 200504D **0.20 Lot#**
Expiration 04/21/2007 **Expiration** 04/21/2007

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.073
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.074
Interferent Detect Test #3	INT

Remarks:

A F / M A: . 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	05/27/2006	01:03:21	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	06/29/2006	22:28:08	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

Printed: 13 June 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	06/29/2006	22:38:48	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	06/29/2006	22:43:02	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . N.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	06/29/2006	23:24:33	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.039
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.038
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	06/30/2006	03:11:59	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	07/28/2006	03:12:54	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	09/01/2006	12:51:47	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Interferent Detect, Interferent Detect. INTERFERENT DETECTED WITH .05 SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	09/01/2006	13:13:52	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	09/01/2006	13:44:51	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	INT
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect, Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD HACKNEY,DWITE,N	10/20/2006	08:27:13	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source
ROCKLEDGE PD HACKNEY,DWITE,N	10/20/2006	11:13:17	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001261

Agency Operator	Date	Time	Software	Source	
ROCKLEDGE PD GOMEZ,IVETTE,	12/22/2006	07:26:43	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source
MELBOURNE PD HACKNEY,DWITE,N	02/23/2006	09:04:39	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	03/27/2006	21:45:22	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	04/20/2006	05:02:15	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.062
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.070
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.067
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.073
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.068
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: CONTROL OUTSIDE OF ACCEPTABLE PARAMETERS . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source
MELBOURNE PD WHITRIGHT,MARK,D	04/20/2006	06:15:54	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.045
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.195
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 05: Control Outside Tolerance. 08: POOR SEALS ON TUBES. 20: POOR SEALS ON TUBES.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	05/29/2006	04:15:03	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.179
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.186
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.190
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

Time-Date changed. 00: Ambient Fail. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	06/20/2006	23:46:59	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	07/31/2006	20:00:29	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	08/16/2006	04:07:28	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	09/30/2006	20:09:16	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source
MELBOURNE PD HACKNEY,DWITE,N	10/27/2006	14:47:40	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.192
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.192
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	10/30/2006	15:40:26	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/29/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

Printed: 13 June 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	11/26/2006	05:53:37	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/29/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	12/29/2006	05:36:07	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. Int Det: , . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001262

Agency Operator	Date	Time	Software	Source	
MELBOURNE PD WHITRIGHT,MARK,D	12/29/2006	20:27:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/29/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source
INDIAN HARBOUR BEACH HACKNEY,DWITE,N	02/21/2006	09:04:13	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	03/27/2006	09:12:31	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	04/20/2006	15:01:55	8100.26	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	No
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	INT
Alcohol Free Test #1 (Retry)	0.000
Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test #3 (Retry)	0.000

Remarks:

INTERFERENT DETECTED,RFI DETECTEDNon-compliance:DID TEST IN WRONG ORDER,RFI DETECTED ALSO
.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	04/20/2006	15:29:10	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	05/18/2006	22:53:37	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	06/11/2006	00:46:48	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	07/21/2006	21:42:52	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	08/23/2006	14:51:17	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	09/20/2006	11:52:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516OO1I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	10/25/2006	03:57:30	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source
INDIAN HARBOUR BEACH HACKNEY,DWITE,N	10/27/2006	14:25:59	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.192
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.191
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.192
0.08 g/210L Dry Gas Std Test #6	0.079
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.192
			0.08 g/210L Dry Gas Std Test #7	0.078
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.192
			0.08 g/210L Dry Gas Std Test #8	0.078
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.079
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0406
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: TUBE NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	11/02/2006	00:10:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001263

Agency Operator	Date	Time	Software	Source	
INDIAN HARBOUR BEACH OAKS,KEVIN,L	12/22/2006	21:59:41	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	516001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source
MELBOURNE BEACH PD HACKNEY,DWITE,N	03/20/2006	09:48:48	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,M,S	04/07/2006	20:56:23	8100.26	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	No
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	

Remarks:

00: Range Exceeded. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source
MELBOURNE BEACH PD HACKNEY,DWITE,N	08/22/2006	14:31:35	8100.26	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	08/23/2006	15:31:04	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007		

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
Interferent Detect Test #3	INT

Remarks:

DRY GAS NOT CONNECTEDNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	08/23/2006	16:18:33	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	09/11/2006	10:56:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.087
0.08 g/210L Test #1 (Retry)	0.082
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.084
0.08 g/210L Test #2 (Retry)	0.081
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.084
0.08 g/210L Test #3 (Retry)	0.082
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	10/11/2006	13:58:51	8100.26	Agency	Inspection

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	10/11/2006	14:05:38	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source
MELBOURNE BEACH PD HACKNEY,DWITE,N	10/27/2006	15:30:26	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.190
0.20 g/210L Test #4 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.190
0.20 g/210L Test #5 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source
			0.20 g/210L Test #6	0.190
			0.20 g/210L Test #6 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #6	0.000
			0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
			Interferent Detect Test #6	INT
			Alcohol Free Test #7	0.000
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.191
			0.20 g/210L Test #7 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.190
			0.20 g/210L Test #8 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.191
			0.20 g/210L Test #9 (Retry)	0.201
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.190
			0.20 g/210L Test #10 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

20: SIMULATOR AIR LEAK. 08: GAS LINE NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	10/30/2006	10:56:41	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	11/23/2006	18:41:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.074
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.076
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001264

Agency Operator	Date	Time	Software	Source	
MELBOURNE BEACH PD GRISWOLD,MELANIE,S	12/25/2006	13:54:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.048
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.206
0.08 g/210L Dry Gas Std Test #1	0.074
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.048
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.074
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.048
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.074
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.076
Interferent Detect Test #3	INT

Remarks:

00: WRONG SIM ATTACHED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	02/14/2006	11:19:10	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	03/23/2006	22:18:04	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

MONTHLY INSP

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	04/06/2006	23:37:34	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:MOUTH PEICE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	04/06/2006	23:45:39	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	05/04/2006	22:39:54	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	06/30/2006	22:11:12	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	07/07/2006	03:43:27	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	08/11/2006	00:27:11	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.074
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
Interferent Detect Test #3	INT

Remarks:

REPLACED BAD HOSE. BAD HOSE REPLSED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	08/11/2006	01:21:38	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	09/18/2006	12:08:00	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	10/26/2006	23:22:06	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.041
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.042
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.042
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BAD MONTH PIECE. HOSE KINKED. TOLERENCE OPER ERR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	10/27/2006	00:25:16	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	10/27/2006	12:03:46	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	11/29/2006	19:27:17	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	5004601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001265

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	12/22/2006	19:57:51	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.181
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.189
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.000
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

BAD MOUTH PC. BAD SEAL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	02/14/2006	11:20:24	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result	Test	Result

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	03/23/2006	22:18:04	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

MONTHLY INSP

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	04/06/2006	23:36:52	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

A F / M A:MOUTH PEICE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	05/04/2006	22:41:02	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	06/30/2006	22:11:09	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	07/07/2006	03:43:24	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	08/11/2006	00:28:08	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

BAD TUDE AND IT WAS REPLACED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source
PALM BAY PD MORGAN,DARIN,P	09/18/2006	12:35:02	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	10/26/2006	23:21:29	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.040
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.041
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.043
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BAD SEAL ON SIM. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	10/27/2006	00:22:25	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	10/27/2006	12:26:59	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	11/29/2006	19:32:45	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001266

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	12/22/2006	20:05:33	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

BAD MOUTH PC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	01/12/2006	11:50:00	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	03/23/2006	12:48:41	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	04/17/2006	12:06:18	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	05/15/2006	10:51:07	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	06/23/2006	14:10:19	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	07/25/2006	13:29:51	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	08/24/2006	12:17:53	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	08/24/2006	12:39:19	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	09/23/2006	10:58:46	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	10/17/2006	00:51:33	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	10/17/2006	02:39:51	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	11/10/2006	10:35:03	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001267

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	12/08/2006	11:17:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	01/12/2006	08:51:32	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	03/23/2006	11:15:56	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.046
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO WEINBERG,DAVID,M	03/23/2006	11:52:16	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.010
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.180
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.196
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. 05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	04/17/2006	09:43:02	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	
Expiration	04/21/2007	**Expiration**	04/21/2007		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.071
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO WEINBERG,DAVID,M	04/17/2006	11:03:55	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	05/15/2006	09:50:38	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	06/23/2006	12:43:20	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO ASD,S,S	08/23/2006	15:40:47	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	09/11/2006	10:40:18	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	09/23/2006	13:06:30	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	10/16/2006	22:54:10	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	10/17/2006	00:48:32	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO CROUCHER,GREGORY,P	11/10/2006	11:54:30	8100.27	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001268

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	12/08/2006	10:21:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	01/11/2006	12:17:45	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO WEINBERG,DAVID,M	03/23/2006	13:46:54	8100.26	Agency Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	04/17/2006	11:04:00	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.075
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	05/15/2006	10:49:09	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	06/21/2006	12:28:45	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.192
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.076
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	06/23/2006	12:44:50	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	07/25/2006	12:19:35	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO WEINBERG,DAVID,M	08/24/2006	13:32:29	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	09/23/2006	14:13:22	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	10/16/2006	23:07:06	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	10/17/2006	00:57:11	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	11/10/2006	13:21:49	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001269

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	12/08/2006	10:13:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	01/12/2006	08:51:36	8100.26	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	03/23/2006	13:06:36	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	04/17/2006	09:42:59	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #2	0.000
Alcohol Free Test #3	0.000

Remarks:

Non-compliance:N.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	04/17/2006	09:55:38	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.147
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.150
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.138
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	05/15/2006	09:53:44	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO WEINBERG,DAVID,M	06/23/2006	13:39:49	8100.26	Agency Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	07/25/2006	13:32:25	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO WEINBERG,DAVID,M	08/24/2006	13:34:17	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	09/23/2006	13:00:35	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

FIRST MOUTH ALCOHOL SAMPLE GIVEN TO SOON.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	10/17/2006	01:41:09	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.076
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	10/17/2006	02:46:29	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	11/10/2006	10:41:43	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001270

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	12/08/2006	11:25:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

SAMPLE DISSIPATED IN MOUTH BEFORE SAMPLE GIVEN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source
SEMINOLE SO SKIPPER,ROGER,G	02/28/2006	09:37:45	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	03/27/2006	00:16:26	8100.26	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	03/27/2006	00:18:03	8100.26	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	03/27/2006	00:23:14	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006		

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.027
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.000
0.20 g/210L Test #1 (Retry)	0.000
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.026
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.000
0.20 g/210L Test #2 (Retry)	0.000
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.022
0.08 g/210L Test #3 (Retry)	0.080
0.20 g/210L Test #3	0.000
0.20 g/210L Test #3 (Retry)	0.000
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance,SIMULATOR MALFUNCTION AIR NOT BEING DRAWN IN BY INTOXILYZERSIMULATOR FAILED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	03/27/2006	01:17:06	8100.26	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , Ambient Fail. FAULTY SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	03/27/2006	01:33:01	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	4281011
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	10/11/2006

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.079
0.08 g/210L Test #3 (Retry)	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	04/28/2006	00:35:07	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

R.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	04/28/2006	00:52:48	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	428101I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	05/18/2006	23:45:50	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	428101I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.054
0.08 g/210L Test #1	0.086
0.08 g/210L Test #1 (Retry)	0.085
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.053
0.08 g/210L Test #2	0.085
0.08 g/210L Test #2 (Retry)	0.085
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.075
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.053
0.08 g/210L Test #3	0.084
0.08 g/210L Test #3 (Retry)	0.084
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	06/17/2006	21:22:23	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.075
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	07/28/2006	21:17:27	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	428101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/11/2006

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.074
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.074
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.074
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.074
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.075
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance UNHOOKED TUBE, CHECKED O GASK. Non-compliance:DRY GAS STANDARD OUT OF TOLERANCE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	07/28/2006	22:22:36	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

NEW DRY GAS STANDARD WITH NEW O-RING

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	08/23/2006	22:05:37	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	09/26/2006	23:55:27	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source
SEMINOLE S.O. SKIPPER,ROGER,G	10/24/2006	23:17:48	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	INT
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.076
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.076
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #8 (Retry)	0.000
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

00: Interferent Detect WRONG SIMULATOR ATTACHED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETHAM,KEITH,D	10/25/2006	00:51:59	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETHAM,KEITH,D	11/28/2006	16:35:55	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001271

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETHAM,KEITH,D	12/28/2006	15:00:30	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source
SEMINOLE SO SKIPPER,ROGER,G	02/28/2006	09:33:55	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518701I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.20 g/210L Test #4 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.20 g/210L Test #5 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source
			0.20 g/210L Test #6	0.197
			0.20 g/210L Test #6 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #6	0.000
			0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
			Interferent Detect Test #6	INT
			Alcohol Free Test #7	0.000
			0.05 g/210L Test #7	0.048
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.197
			0.20 g/210L Test #7 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.20 g/210L Test #8 (Retry)	0.198
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.20 g/210L Test #9 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

20: RFI Detect CELL PHONE IN AREA.. 08: Control Outside Tolerance GAS TUBE NOT CONNECTED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	03/27/2006	02:52:38	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512402I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	05/06/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.075
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2 (Retry)	0.082
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

00: Ambient Fail. Int Det: . 08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	04/28/2006	00:34:13	8100.26	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Subject Test Refused. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	04/28/2006	00:42:10	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	04/28/2006	00:44:24	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512402I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	05/18/2006	23:42:01	8100.26	Agency	Inspection

0.05 Lot#	200411C	**0.08 Lot#**	200411G	**0.20 Lot#**	200411D	**0.08 Gas Lot#**	512402I
Expiration	11/04/2006	**Expiration**	11/18/2006	**Expiration**	11/04/2006	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.083
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	06/17/2006	21:19:47	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.047
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.077
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: Control Outside Tolerance. Non-compliance:.08 OUT SIDE EXCEPTED TOLOERANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	06/17/2006	21:56:51	8100.26	Agency	Inspection

0.05 Lot#	200509C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512402I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

WRONG LOT# FOR .05, CORRECT LOT#200509A

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	07/28/2006	21:18:12	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:ACETONE NOT ADDED, ACETONE ADDED AND RERUN..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	08/23/2006	22:05:14	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE SO BETHAM,KEITH,D	09/26/2006	23:51:57	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source
SEMINOLE S.O. SKIPPER,ROGER,G	10/24/2006	23:26:59	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509B	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETHAM,KEITH,D	10/25/2006	02:08:43	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETAHM,KEITH,D	11/27/2006	23:43:56	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

NEED TO RE-RUN DIAGNOSTIC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETHAM,KEITH,D	11/28/2006	00:43:26	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001272

Agency Operator	Date	Time	Software	Source	
SEMINOLE S.O. BETHAM,KEITH,D	12/27/2006	22:44:32	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	02/08/2006	10:21:31	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	03/23/2006	21:33:10	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	04/19/2006	18:29:13	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VLIKCY,L	05/17/2006	17:43:05	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	434501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/14/2006

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.190
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	06/07/2006	16:29:57	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

Printed: 13 June 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/12/2006	00:38:40	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	08/01/2006	18:51:11	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611081G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	09/17/2006	20:55:07	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	10/17/2006	15:57:09	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.193
0.20 g/210L Test #4 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.194
0.20 g/210L Test #5 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.194
0.20 g/210L Test #6 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.077
			0.20 g/210L Test #7	0.195
			0.20 g/210L Test #7 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.194
			0.20 g/210L Test #8 (Retry)	0.195
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.195
			0.20 g/210L Test #9 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.195
			0.20 g/210L Test #10 (Retry)	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0019
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

20: Control Outside Tolerance. SOLUTION CHANGED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	10/17/2006	17:02:11	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	11/21/2006	19:17:06	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	12/22/2006	16:28:22	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001273

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	12/22/2006	17:29:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	02/08/2006	10:24:04	8100.26	Department Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518602I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0009 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.085
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.085
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.085
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.085
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.085
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.085
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.085
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.085
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.085
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0015

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	03/22/2006	22:05:09	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	04/19/2006	16:40:03	8100.26	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

NOY MET.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	04/19/2006	17:46:48	8100.26	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512501I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	05/17/2006	18:19:08	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.068
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

GASKET MISSING ON SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	06/08/2006	14:13:06	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/12/2006	03:12:25	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

Non-compliance:VOLTAGE/CURRENT FAILURE ON DIAGNOSTIC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	07/12/2006	15:20:02	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO FAIRCHILD,VICKY,L	08/01/2006	16:05:35	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	09/18/2006	23:08:16	8100.26	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	10/17/2006	13:01:39	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001274

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO ADYE,VICKY,L	10/17/2006	16:11:07	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.