

# Florida Department of Law Enforcement

## Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
CMI INC 3456780-=[OIYR,';JHGDWQ3456790,-	08/17/2006	08:07:01	8100.26	Department Inspection

# Sims 0 Avg Std Dev Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
CMI INC 234567890;LKJH,,MNBVFDSAQ2345,.	08/17/2006	08:07:58	8100.26	Agency Inspection

# Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	08/22/2006	14:31:20	8100.26	Department Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5187021 <b>Expiration</b> 07/08/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	09/18/2006	13:51:44	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	10/26/2006	23:23:04	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

Time-Date changed.

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	10/27/2006	13:37:13	8100.27	Department Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5187021 <b>Expiration</b> 07/08/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0003	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.194
		0.08 g/210L Dry Gas Std Test #4	0.000
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.000
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.000
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.048
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.194
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

## Remarks:

08: TUBE NOT CONNECTED.



# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	11/29/2006	19:21:02	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

**Remarks:**

Time-Date changed. CONNECTION ERROR.

# Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	12/22/2006	20:10:14	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies Yes</b>				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.192
		0.08 g/210L Dry Gas Std Test #2	0.079
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT

**Remarks:**

BAD SEAL. FAILED TO CONN GAS.

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO VENTURI,GEORGE,L	05/16/2006	13:32:05	8100.26	Department Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO VENTURI,GEORGE,L	06/22/2006	13:08:57	8100.26	Department Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.045
		0.05 g/210L Test #4 (Retry)	0.050
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.044
		0.05 g/210L Test #5 (Retry)	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.044
		0.05 g/210L Test #6 (Retry)	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.044
			0.05 g/210L Test #7 (Retry)	0.050
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.045
			0.05 g/210L Test #8 (Retry)	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.044
			0.05 g/210L Test #9 (Retry)	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.045
			0.05 g/210L Test #10 (Retry)	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

05: Control Outside Tolerance-TIGHTENED SIM. CONNECTIONS .

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	07/14/2006	09:38:59	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200905C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO CLAFLIN,DAVID,S	08/06/2006	13:35:05	8100.26	Agency Inspection

<b>0.05 Lot#</b> 200509A	<b>0.08 Lot#</b> 200509B	<b>0.20 Lot#</b>
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	

**# Sims** 5 **Complies No**

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.199
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.049
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.200
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.199
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

08: Control Outside Tolerance, Sequence Aborted. Non-compliance: .



# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	08/06/2006	14:09:34	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

A F / M A: No Sample Provided.

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	09/01/2006	11:39:49	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	10/10/2006	10:57:54	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	10/25/2006	10:28:09	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO VENTURI,GEORGE,L	10/25/2006	16:21:20	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0005	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN, DAVID, S	10/25/2006	17:16:55	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	11/21/2006	14:08:21	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/02/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5046011 <b>Expiration</b> 02/17/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**



# Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	12/20/2006	17:32:18	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 612202G <b>Expiration</b> 05/05/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO SUERETH,DONALD,P	05/16/2006	09:34:02	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0008	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.000
		0.08 g/210L Test #4 (Retry)	0.077
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.000
		0.08 g/210L Test #5 (Retry)	0.076
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.000
		0.08 g/210L Test #6 (Retry)	0.078
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.048
			0.08 g/210L Test #7	0.000
			0.08 g/210L Test #7 (Retry)	0.077
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.000
			0.08 g/210L Test #8 (Retry)	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.000
			0.08 g/210L Test #9 (Retry)	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.000
			0.08 g/210L Test #10 (Retry)	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

## Remarks:

08: Control Outside Tol. SIMULATOR JAR CRACKED AND REPLACED. .08 SIMULATOR JAR CRACKED DURING ANALYSES.

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO GRAY,BEVERLY,A	05/16/2006	13:16:32	8100.26	Agency Inspection

# Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO GRAY,BEVERLY,A	05/17/2006	09:11:50	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

## Remarks:

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	06/09/2006	11:13:50	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/14/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	07/18/2006	09:57:31	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO GRAY,BEVERLY,A	08/11/2006	12:08:49	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

**Remarks:**



# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	09/15/2006	12:43:56	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 831505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

LOOSE CONNECTION.

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO SUERETH,DONALD,P	10/18/2006	15:31:34	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0006	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.046
		0.08 g/210L Test #4	0.076
		0.08 g/210L Test #4 (Retry)	0.076
		0.20 g/210L Test #4	0.193
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.076
		0.08 g/210L Test #5 (Retry)	0.076
		0.20 g/210L Test #5	0.193
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.046
		0.08 g/210L Test #6	0.076
		0.08 g/210L Test #6 (Retry)	0.077
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.047
			0.08 g/210L Test #7	0.076
			0.08 g/210L Test #7 (Retry)	0.076
			0.20 g/210L Test #7	0.193
			0.08 g/210L Dry Gas Std Test #7	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.08 g/210L Test #8 (Retry)	0.076
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.08 g/210L Test #9 (Retry)	0.076
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.08 g/210L Test #10 (Retry)	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

## Remarks:

08: Control Outside Tol. ALL HOSE CONECTIONS TIGHTENED. .

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO OWENS, EDWARD, W	10/18/2006	16:20:17	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

20: TOL REMOVED CONDISION FROM TUBING

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO SUERETH,DONALD,P	11/09/2006	10:49:29	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0006	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	11/10/2006	11:23:42	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	12/21/2006	08:50:00	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

**Remarks:**



# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
CMI INC PUTMAN,BRIAN,A	05/24/2006	22:54:12	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
CHARLOTTE COUNTY SO VENTURI,GEORGE,L	05/25/2006	16:38:03	8100.26	Department Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

## Remarks:

AI CPL. B. PUTMAN PRESENT

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	05/25/2006	18:05:11	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

00: Interferent Detect - CONNECTED WRONG SIMULATOR.

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	06/15/2006	10:20:00	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

INSTRUMENT BEING SENT TO CMI FOR MEMORY UPGRADE

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
CHARLOTTE COUNTY SO VENTURI,GEORGE,L	07/11/2006	13:48:58	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5252021 <b>Expiration</b> 09/14/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.194
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	07/28/2006	19:33:59	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

INSPECTION UPON RETURN FROM SERVICE FROM CMI



# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	08/03/2006	01:28:54	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

AUGUST INSPECTION

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	09/09/2006	20:00:01	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

SEPTEMBER INSPECTION

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	10/27/2006	01:58:46	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

OCTOBER INSPECTION

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
CHARLOTTE COUNTY SO VENTURI,GEORGE,L	11/07/2006	14:13:28	8100.27	Department Inspection

<b>0.05 Lot#</b> 200509A	<b>0.08 Lot#</b> 200509B	<b>0.20 Lot#</b> 200509C	<b>0.08 Gas Lot#</b> 531505G
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 11/15/2007
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0005	<b>Complies</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.193
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	11/07/2006	15:32:00	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	12/11/2006	20:37:24	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 531505G <b>Expiration</b> 11/15/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

DECEMBER INSPECTION

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
CLEARWATER PD SUERETH,DONALD,P	08/29/2006	12:03:04	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079



# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
			Interferent Detect Test #7	INT
			Alcohol Free Test #8 (Retry)	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

00: RFI Detect. CELL PX REMOVED FROM AREA..

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER, THEODORE, J	08/30/2006	19:20:15	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 200509B <b>Expiration</b> 09/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.162
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

20: Control Outside Tolerance.

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER, THEODORE, J	09/22/2006	15:55:58	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6153011 <b>Expiration</b> 06/05/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

ACETONE LOT # 2005-C MOUTH ALCOHOL LOT #2005-A

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER, THEODORE, J	10/26/2006	17:02:09	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6153011 <b>Expiration</b> 06/05/2008		
<b># Sims</b> 4	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

ACETONE STOCK SOLUTION LOT# 2005-C A

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
CLEARWATER PD SUERETH,DONALD,P	11/07/2006	15:12:59	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER, THEODORE, J	11/07/2006	16:17:35	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6153011 <b>Expiration</b> 06/05/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

ACETONE LOT #2005-D MOUTH ALCOHOL LOT #2005-A

# Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
CLEARWATER PD MILNE, THOMAS, P	12/15/2006	06:27:39	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6153011 <b>Expiration</b> 06/05/2008	
<b># Sims</b> 4	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**



# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH SO HACKNEY,DWITE,N	07/17/2006	10:56:38	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5187021 <b>Expiration</b> 07/08/2007	
<b># Sims</b> 4	<b>Avg Std Dev</b> 0.0007	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO WEINBERG,DAVID,M	07/25/2006	12:18:28	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5125011 <b>Expiration</b> 05/10/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH SO WEINBERG,DAVID,M	08/24/2006	12:19:20	8100.26	Agency Inspection

# Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Non-compliance: .

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH SO WEINBERG,DAVID,M	08/24/2006	12:20:07	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5125011 <b>Expiration</b> 05/10/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH SO CROUCHER,GREGORY,P	09/23/2006	14:14:05	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 610001G <b>Expiration</b> 04/12/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	10/16/2006	22:37:30	8100.27	Department Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5187021 <b>Expiration</b> 07/08/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0005	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**



# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO CROUCHER,GREGORY,P	10/16/2006	23:26:27	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 610001G <b>Expiration</b> 04/12/2008	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO CROUCHER,GREGORY,P	11/10/2006	12:36:35	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 610001G <b>Expiration</b> 04/12/2008	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	12/08/2006	10:18:36	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200605C <b>Expiration</b> 05/04/2008	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 610001G <b>Expiration</b> 04/12/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
CMI INC ANDREWS,ROBERT,P	04/24/2006	07:54:23	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies No</b>				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

05: Control Outside Tolerance. Non-compliance: .

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD VENTURI,GEORGE,L	05/17/2006	16:07:37	8100.26	Department Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 4364011 <b>Expiration</b> 01/04/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0003	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.194
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.195
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.192
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	05/24/2006	09:26:01	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

**Remarks:**

A.I. DEP. JASON RILEY PRESENT AND REVIEWED INSPECTION

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD VENTURI,GEORGE,L	06/22/2006	13:08:47	8100.26	Department Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 4252021 <b>Expiration</b> 09/10/2006	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0003	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.076
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.076
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000



# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	07/12/2006	13:03:36	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

AI DEP. J. RILEY PRESENT FOR INSPECTION

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD RILEY, JASON, W	08/25/2006	15:23:17	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD ANDREWS,ROBERT,P	09/22/2006	23:00:30	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD RILEY, JASON, W	10/18/2006	11:57:23	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

DEP. ROBERT ANDREWS PRESENT DURING INSPECTION

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD ANDREWS,ROBERT,P	11/06/2006	10:02:51	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD VENTURI,GEORGE,L	11/06/2006	16:20:25	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.194
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

## Remarks:

AI DEPUTY R ANDREWS PRESENT



# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD ANDREWS,ROBERT,P	11/06/2006	17:02:28	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD RILEY, JASON, W	12/06/2006	11:25:04	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200605C <b>Expiration</b> 05/04/2008	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

AI DEP. ROBERT ANDREWS PRESENT DURING INSPECTION

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD SUERETH,DONALD,P	05/24/2006	09:19:07	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 4	<b>Avg Std Dev</b> 0.0007	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	05/24/2006	10:43:47	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200504C <b>Expiration</b> 04/21/2007	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 431501 <b>Expiration</b> 11/12/2006		
<b># Sims</b> 5	<b>Complies Yes</b>				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD SUERETH,DONALD,P	08/31/2006	08:59:18	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0008	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	10/11/2006	09:47:48	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 6191011 <b>Expiration</b> 07/12/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**



# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD SUERETH,DONALD,P	10/25/2006	10:43:07	8100.27	Department Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD KNOBLOCH,WILLIAM,T	10/26/2006	09:39:38	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 6191011 <b>Expiration</b> 07/12/2008	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD KNOBLOCH,WILLIAM,T	11/27/2006	10:24:36	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6191011 <b>Expiration</b> 07/12/2008	
<b># Sims</b> 4	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD KNOBLOCH,WILLIAM,T	12/28/2006	11:22:54	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6191011 <b>Expiration</b> 07/12/2008	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

## Remarks:

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL CHAPA, ORLANDO,	05/04/2006	11:10:38	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.051
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.05 g/210L Test #3 (Retry)	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

05: Control Outside Tolerance.CHECK VALVE STUCK.

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SUERETH,DONALD,P	05/04/2006	11:58:54	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0005	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**



# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	06/27/2006	08:50:28	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200505C <b>Expiration</b> 05/05/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 2	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.159
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.190
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.190
		0.20 g/210L Test #2 (Retry)	0.202
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

08: Control Outside Tolerance. 20: Control Outside Tolerance.

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	07/30/2006	06:54:16	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b>		
<b># Sims</b> 4	<b>Complies No</b>				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.157
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.186
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	No	0.08 g/210L Test #2 (Retry)	0.080
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2	0.189
		0.20 g/210L Test #2 (Retry)	0.200
		Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.200
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

**Remarks:**

Int Det: . 08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	07/30/2006	08:36:05	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 2	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.051
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.044
Barometric Pressure Sensor Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1	0.069
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1	0.177
Diagnostic Check (Post-Inspection): OK	Yes	0.20 g/210L Test #1 (Retry)	0.193
Alcohol Free Test: OK	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
0.05 g/210L Test: OK	Yes	Interferent Detect Test #1	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #2	0.015
0.20 g/210L Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.05 g/210L Test #2	0.047
		0.05 g/210L Test #2 (Retry)	0.048
		0.08 g/210L Test #2	0.079
		0.08 g/210L Test #2 (Retry)	0.079
		0.20 g/210L Test #2	0.189
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.080
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

00: Control Outside Tolerance. 05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	09/01/2006	08:16:17	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 2	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.155
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #1 (Retry)	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
		0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

**Remarks:**

Int Det: . 08: Control Outside Tolerance. 20: Control Outside Tolerance.

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SMITH,ELMER,H	10/27/2006	08:54:04	8100.26	Agency Inspection

# Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: .

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SMITH,ELMER,H	10/27/2006	08:58:04	8100.26	Agency Inspection

# Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

SLOPE NOT MET.

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SMITH,ELMER,H	10/27/2006	09:02:39	8100.26	Agency Inspection

<b>0.05 Lot#</b> 200509A	<b>0.08 Lot#</b> 200509B	<b>0.20 Lot#</b>
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	

**# Sims** 4 **Complies No**

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.025
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.000
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: INSTRUMENT DID NOT MEET TOLERANCE.

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SMITH,ELMER,H	10/29/2006	19:00:50	8100.26	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007	
<b># Sims</b> 2	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**



# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SUERETH,DONALD,P	11/09/2006	13:59:51	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5186021 <b>Expiration</b> 07/07/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0009	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0013
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	11/09/2006	16:07:01	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 4	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL SMITH,ELMER,H	12/28/2006	17:05:23	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5354021 <b>Expiration</b> 12/22/2007		
<b># Sims</b> 4	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

Time-Date changed.

# Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source
CMI INC HACKNEY,DWITE,N	10/26/2006	10:14:18	8100.26	Agency Inspection

# Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Non-compliance: .

# Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source
PORT ST LUCIE PD HACKNEY,DWITE,N	10/26/2006	13:57:13	8100.27	Department Inspection

<b>0.05 Lot#</b> 200605B	<b>0.08 Lot#</b> 200509B	<b>0.20 Lot#</b> 200509C	<b>0.08 Gas Lot#</b>
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	

<b># Sims</b> 5	<b>Avg Std Dev</b>	<b>Complies No</b>
-----------------	--------------------	--------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.194
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.177
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.194
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.179
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.179
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050
		0.08 g/210L Test #8	0.080
		0.20 g/210L Test #8	0.179
		Interferent Detect Test #8	INT
		Alcohol Free Test #9	0.000
		0.05 g/210L Test #9	0.050
		0.08 g/210L Test #9	0.079

# Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source
			0.20 g/210L Test #9	0.178
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.178
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0074

## Remarks:

20: INTERNAL AIR LEAK. Non-compliance: .

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source
NAS JACKSONVILLE MALHIOT,MATTHEW,E	09/26/2006	11:16:41	8100.26	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 612501G <b>Expiration</b> 05/08/2008	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0005	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000



# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	09/26/2006	12:53:02	8100.26	Agency	Inspection
<b>0.05 Lot#</b> 200411C <b>Expiration</b> 11/04/2006	<b>0.08 Lot#</b> 200504D <b>Expiration</b> 04/21/2007	<b>0.20 Lot#</b> 200411D <b>Expiration</b> 11/04/2006	<b>0.08 Gas Lot#</b> 612501G <b>Expiration</b> 05/08/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

A F / M A: Range Exceeded. TOO MUCH M/A USED

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source
NAS JACKSONVILLE MALHIOT,MATTHEW,E	10/19/2006	15:09:34	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 612501G <b>Expiration</b> 05/08/2008	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.204
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.204
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.204
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.204
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.204
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	10/20/2006	08:31:52	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200411C <b>Expiration</b> 11/04/2006	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 612501G <b>Expiration</b> 05/08/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

A F / M A: M/A NOT DETECTED

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	11/20/2006	09:28:04	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 612501G <b>Expiration</b> 05/08/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.085
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	12/20/2006	08:56:49	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 612501G <b>Expiration</b> 05/08/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

MA/NOT DETECTED.

# Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source
NAS KEY WEST HACKNEY,DWITE,N	09/13/2006	16:06:00	8100.26	Department Inspection

**0.05 Lot#** 200605B  
**Expiration** 05/04/2008

**# Sims** 5                      **Avg Std Dev**                      **Complies No**

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.050
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.049
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No	Alcohol Free Test #4 (Retry)	0.000
0.08 g/210L Test: OK	No	Interferent Detect Test #4	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #5 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		Interferent Detect Test #7	INT
		Alcohol Free Test #8 (Retry)	0.000
		Interferent Detect Test #8	INT
		Alcohol Free Test #9 (Retry)	0.000
		Interferent Detect Test #9	INT
		Alcohol Free Test #10 (Retry)	0.000
		Interferent Detect Test #10	INT

**Remarks:**

00: Ambient Fail. 05: RFI Detect, RFI Detect. Non-compliance: .



# Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source
NAS KEY WEST HACKNEY,DWITE,N	09/13/2006	16:47:31	8100.26	Department Inspection

# Sims            0                            Avg Std Dev                            Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

RFI FAILURENon-compliance: .

# Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source
NAS KEY WEST HACKNEY,DWITE,N	11/06/2006	11:35:52	8100.27	Department Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 5187021 <b>Expiration</b> 07/08/2007	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0002	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

# Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

# Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source
CMI INC MALHIOT,MATTHEW,E	10/31/2006	12:30:25	8100.26	Agency Inspection

# Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Non-compliance: TAKEN OUT OF DISABLE MODE FOR UPGRADE 27 .

# Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source
FSU POLICE MALHIOT,MATTHEW,E	10/31/2006	16:49:48	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6060011 <b>Expiration</b> 03/03/2008	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8 (Retry)	0.000

# Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

## Remarks:

00: Ambient Fail.

# Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	11/03/2006	14:30:02	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200605C <b>Expiration</b> 05/04/2008	<b>0.20 Lot#</b> 200605F <b>Expiration</b> 05/10/2008	<b>0.08 Gas Lot#</b> 6172021 <b>Expiration</b> 06/23/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD SKIPPER,ROGER,G	10/17/2006	11:02:49	8100.27	Department Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6188011 <b>Expiration</b> 07/10/2008	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0004	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.083
		0.20 g/210L Test #4	0.204
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.083
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.083
		0.20 g/210L Test #6	0.203
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.203
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8 (Retry)	0.000



# Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

## Remarks:

00: RFI Detect CELL PHONE USED NEARBY..

# Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	10/20/2006	11:48:24	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6209011 <b>Expiration</b> 08/02/2008	
<b># Sims</b> 4	<b>Complies No</b>			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

## Remarks:

TEMP REG TEST FAILED.

# Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	10/29/2006	07:38:59	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6209011 <b>Expiration</b> 08/02/2008	
<b># Sims</b> 4	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

## Remarks:

# Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	11/17/2006	12:20:59	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6200911 <b>Expiration</b> 08/02/2008		
<b># Sims</b> 4	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	12/20/2006	10:41:02	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200605B <b>Expiration</b> 05/04/2008	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6209011 <b>Expiration</b> 08/02/2008		
<b># Sims</b> 4	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
CMI INC JONES,JEFF,T	10/12/2006	15:06:27	8100.27	Department Inspection

# Sims	0	Avg Std Dev	Complies	Yes
--------	---	-------------	----------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
CMI INC UNGER,FELIX,C	10/12/2006	15:07:35	8100.27	Agency Inspection

# Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
MIAMI BEACH PD VENTURI,GEORGE,L	10/27/2006	12:44:47	8100.27	Department Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6228021 <b>Expiration</b> 08/18/2008	
<b># Sims</b> 5	<b>Avg Std Dev</b> 0.0005	<b>Complies</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000



# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

AI OFF. K. MILLAN PRESENT

# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	10/27/2006	23:32:38	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6228021 <b>Expiration</b> 08/18/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**

# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	11/29/2006	14:07:06	8100.27	Agency	Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6228021 <b>Expiration</b> 08/18/2008		
<b># Sims</b> 5	<b>Complies</b> Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

**Remarks:**

A F / M A:DIDN'T BLOW ENOUGH.

# Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
MIAMI BEACH PD MILLAN,KEVIN,	12/20/2006	07:22:46	8100.27	Agency Inspection
<b>0.05 Lot#</b> 200509A <b>Expiration</b> 09/22/2007	<b>0.08 Lot#</b> 200509B <b>Expiration</b> 09/22/2007	<b>0.20 Lot#</b> 200509C <b>Expiration</b> 09/22/2007	<b>0.08 Gas Lot#</b> 6228021 <b>Expiration</b> 08/18/2008	
<b># Sims</b> 5	<b>Complies</b> Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

**Remarks:**