

Florida Department of Law Enforcement

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	02/15/2006	13:36:27	8100.26	Agency	Inspection

# Sims	0	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP BARBUTO,JOSEPH,L	02/15/2006	14:43:36	8100.26	Agency	Inspection

# Sims	0	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	02/20/2006	09:29:56	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP OAKS,KEVIN,L	02/21/2006	09:45:50	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

UNKNOWN. Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HOLSTINE,SCOTT,W	02/21/2006	12:03:14	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP GRISWOLD,MELANIE,S	02/21/2006	13:45:52	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP DECKARD,JOE,B	02/22/2006	10:32:35	8100.26	Agency	Inspection

# Sims	0	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	02/22/2006	13:43:58	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	03/06/2006	10:18:32	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	03/08/2006	10:40:06	8100.26	Agency	Inspection
# Sims	0	Complies			
		No			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	03/14/2006	08:35:17	8100.26	Agency	Inspection
# Sims	0	Complies			
		No			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator		Date	Time	Software		Source	
FDLE/ATP HACKNEY,DWITE,N		03/14/2006	08:47:13	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	411802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP HACKNEY,DWITE,N	03/14/2006	09:32:50	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator		Date	Time	Software	Source	
FDLE/ATP MELBOURNE HACKNEY,DWITE,N		03/20/2006	10:02:52	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot# 518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Test (Std Dev)	0.0012
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE SKIPPER,ROGER,G	05/09/2006	11:31:08	8100.26	Agency	Inspection
# Sims	0	Complies No			
Check or Test		Result	Test	Result	

Remarks:

RESET AI DATE ONLYNon-compliance: .

Inspection Tests

Instrument Serial #80-000200

Agency Operator		Date	Time	Software		Source	
FDLE/MELBOURNE HACKNEY,DWITE,N		11/06/2006	11:44:57	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000200

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software	Source	
FDLE/ATP PROC MALHIOT,MATTHEW,E		05/25/2006	10:29:14	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 525202I
Expiration	09/22/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 09/14/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC		06/01/2006	12:21:03	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610001G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC NEFF,STEPHEN,		07/12/2006	14:46:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:AIR LEAK IN SIMULATOR ON 1ST SAMPLE.

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC NEFF,STEPHEN,		08/18/2006	10:15:57	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

OFFICER TATRO OBSERVED AGENCY INSPECTION.

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC NEFF,STEPHEN,		08/30/2006	08:37:20	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC SMITH,JAMES,C		08/31/2006	15:07:48	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.155
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.162
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.166
0.08 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.164
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, RFI Detect. .20 OUT OF TOLERANCE.

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC SMITH,JAMES,C		08/31/2006	19:14:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.038
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.070
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.106
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.160
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.20 g/210L Test: OK	No	0.08 g/210L Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2 (Retry)	0.076
		0.20 g/210L Test #2	0.145
		0.20 g/210L Test #2 (Retry)	0.159
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3 (Retry)	0.159
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: RFI Detect, Control Outside Tolerance.
Non-compliance: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source	
FDLE/ATP PROC SMITH,JAMES,C	09/02/2006	09:43:24	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC SMITH,JAMES,C		09/02/2006	09:47:54	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.162
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.186
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.178
0.20 g/210L Test: OK	No	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.176
		0.20 g/210L Test #3 (Retry)	0.192
		Interferent Detect Test #3	INT

Remarks:

00: Interferent Detect. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source	
FDLE/ATP PROC SMITH,JAMES,C	09/02/2006	10:32:58	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.031
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.036
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.039
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.038
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.041
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.040
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source	
FDLE/ATP PROC SMITH,JAMES,C	09/02/2006	14:48:24	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.032
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.040
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.046
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.048
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC SMITH,JAMES,C		09/02/2006	15:15:17	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	20509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.029
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.068
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.146
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.173
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.042
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	No	0.08 g/210L Test #2	0.074
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2 (Retry)	0.080
		0.20 g/210L Test #2	0.166
		0.20 g/210L Test #2 (Retry)	0.177
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.082
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.172
		0.20 g/210L Test #3 (Retry)	0.176
		Interferent Detect Test #3	INT

Remarks:

THIS MACHINE SUCKSNon-compliance: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software	Source	
FDLE/ATP PROC MALHIOT,MATTHEW,E		10/11/2006	16:30:37	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 620901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 08/02/2008
# Sims	5	Avg Std Dev	0.0007	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software	Source	
FDLE/ATP PROC PERBTANI,AAMIR,A		10/23/2006	16:37:42	8100.27	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#		
Expiration	09/22/2007	Expiration	09/22/2007			
# Sims	2	Complies Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.073
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.073
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC PERBTANI,AAMIR,A		10/23/2006	18:19:15	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	620901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	08/02/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.071
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.188
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.193
Diagnostic Check (Post-Inspection): OK	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Alcohol Free Test: OK	Yes	Interferent Detect Test #1	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2	0.075
		0.08 g/210L Test #2 (Retry)	0.077
		0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC PERBTANI,AAMIR,A		11/29/2006	19:54:19	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	620901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	08/02/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source
FDLE/ATP PROC PERBTANI,AAMIR,A	12/30/2006	11:36:48	8100.27	Agency Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot#		
# Sims 2		Complies No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000206

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC PERBTANI,AAMIR,A		12/30/2006	13:31:53	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	620901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	08/02/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.057
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.054
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.065
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.064
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000219

Agency Operator		Date	Time	Software	Source	
FDLE FMROC VENTURI,GEORGE,L		01/30/2006	14:10:33	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 422301I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 08/12/2006
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.076
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.076
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.076
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.076
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000219

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000219

Agency Operator		Date	Time	Software	Source	
FDLE FMROC VENTURI,GEORGE,L		05/16/2006	13:31:31	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 415602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 06/10/2006
# Sims	5	Avg Std Dev	0.0001	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000219

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

Inspection Tests

Instrument Serial #80-000219

Agency Operator		Date	Time	Software		Source	
FDLE FMROC WORTH,DUSTIN,L		09/01/2006	23:47:03	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	619101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/12/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000219

Agency Operator		Date	Time	Software		Source	
FDLE FMROC WORTH,DUSTIN,L		10/03/2006	21:40:02	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	619101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/12/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000219

Agency Operator		Date	Time	Software	Source	
FDLE FMROC VENTURI,GEORGE,L		10/18/2006	19:11:23	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot# 626101I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration 09/20/2008
# Sims	5	Avg Std Dev	0.0003	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000219

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000220

Agency Operator	Date	Time	Software	Source
FL. DEPT. OF LAW ENF SUERETH,DONALD,P	05/30/2006	13:49:34	8100.26	Department Inspection

# Sims	0	Avg Std Dev	Complies	No
--------	---	-------------	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000220

Agency Operator		Date	Time	Software		Source	
FDLE SUERETH,DONALD,P		07/11/2006	09:50:19	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000220

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000220

Agency Operator		Date	Time	Software		Source	
FDLE NICHOLS,KENNETH,D		07/12/2006	17:53:55	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	431401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/11/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000220

Agency Operator		Date	Time	Software		Source	
FDLE NICHOLS,KENNETH,D		08/22/2006	11:12:20	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	431401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/11/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000220

Agency Operator		Date	Time	Software		Source	
FDLE NICHOLS,KENNETH,D		09/13/2006	11:43:03	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	431401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000221

Agency Operator		Date	Time	Software		Source	
FDLE FMROC VENTURI,GEORGE,L		05/16/2006	13:29:40	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	428102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.076
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.076
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.076
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000221

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

Inspection Tests

Instrument Serial #80-000221

Agency Operator		Date	Time	Software	Source	
FDLE FMROC VENTURI,GEORGE,L		10/18/2006	18:27:15	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot# 626101I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration 09/20/2008
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.045
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.046
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000221

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source
FDLE/ATP MELBOURNE HACKNEY,DWITE,N	03/20/2006	10:02:39	8100.26	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 518702I Expiration 07/08/2007	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.073
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.073
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.073
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.073
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.074
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.074
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.073
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0031
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE RIGGINS,JERRY,	04/14/2006	13:15:22	8100.26	Agency	Inspection
# Sims	5	Complies No			
Check or Test		Result	Test	Result	

Remarks:

00: Sequence Aborted. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software		Source	
FDLE/ATP MELBOURNE RIGGINS,JERRY,		04/14/2006	13:22:27	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	504502I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	5	Complies		Yes			

Check or Test	Result	Test	Result
---------------	--------	------	--------

Remarks:

RISIDUAL FROM PRIOR TEST. 20: Control Outside Tolerance COLD SIMULATOR.

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software		Source	
FDLE/ATP MELBOURNE HUGHES,HAROLD,W		04/14/2006	14:34:55	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	504502I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	5	Complies		Yes			

Check or Test	Result	Test	Result
---------------	--------	------	--------

Remarks:

Int Det:WRONG SIM.

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software	Source	
FDLE/ATP MELBOURNE HACKNEY,DWITE,N		05/31/2006	22:18:51	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot# 518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.046
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software	Source	
FDLE/ATP MELBOURNE DIMINO,LEANETTE,M		06/01/2006	08:25:23	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#		
Expiration	09/22/2007	Expiration	09/22/2007			
# Sims	4	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.068
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.045
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.046
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Ambient Fail, Purge Fail. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software		Source	
FDLE/ATP MELBOURNE DIMINO, LEANETTE, M		06/01/2006	11:04:27	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	515901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.190
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
		0.20 g/210L Test #2	0.195
		0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE DIMINO,LEANETTE,M	06/30/2006	08:13:46	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , Ambient Fail. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE DIMINO,LEANETTE,M	06/30/2006	08:47:12	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.038
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.041
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.043
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.041
0.08 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Ambient Fail. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE STRICKLAND,SELBY,J	06/30/2006	21:55:44	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.013
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.017
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.020
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE STRICKLAND,SELBY,J	06/30/2006	22:56:33	8100.26	Agency	Inspection
0.05 Lot#	200509A				
Expiration	09/22/2007				
# Sims	4		Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.040
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.043
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.044
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE STRICKLAND,SELBY,J	06/30/2006	23:31:36	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software		Source	
FDLE/ATP MELBOURNE STRICKLAND,SELBY,J		07/01/2006	00:02:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	515901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.040
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.071
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
		0.20 g/210L Test #2	0.195
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE DIMINO,LEANETTE,M	08/29/2006	06:54:51	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE DIMINO,LEANETTE,M	08/29/2006	07:37:40	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.043
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.044
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	0.000
0.05 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE DIMINO, LEANETTE, M	08/29/2006	13:03:00	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.042
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: SIMULATOR AIR LEAK, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000223

Agency Operator		Date	Time	Software		Source	
FDLE/ATP MELBOURNE DIMINO, LEANETTE, M		08/29/2006	13:59:34	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	515901I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMULATOR AIR LEAK.

Inspection Tests

Instrument Serial #80-000223

Agency Operator	Date	Time	Software	Source	
FDLE/ATP MELBOURNE HACKNEY,DWITE,N	11/06/2006	08:45:44	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software	Source	
FDLE/ATP JROC MALHIOT,MATTHEW,E		05/25/2006	10:16:55	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 419101I
Expiration	09/22/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 07/14/2006
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000224

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP JROC MALHIOT,MATTHEW,E		06/02/2006	17:43:28	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	610201G
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration	04/17/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

0.08 SIMULATOR RUNNING A BIT HOT, HIGH RESULTS WITHIN RANGE

Inspection Tests

Instrument Serial #80-000224

Agency Operator	Date	Time	Software	Source	
FDLE/ATP JROC M,M,E	07/14/2006	15:30:26	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP JROC VAN DYKE,KEITH,T		08/10/2006	19:07:08	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000224

Agency Operator	Date	Time	Software	Source
FDLE/ATP JROC VANDYKE,KEITH,T	09/28/2006	14:30:05	8100.26	Agency Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#
Expiration	05/04/2008	Expiration	05/04/2008	
# Sims	4	Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.074
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.072
0.08 g/210L Test: OK	No	0.08 g/210L Test #2 (Retry)	0.074
0.20 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.073
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect. 05: AIR LEAK AT THERMOMETER. 08: AIR LEAK . Non-compliance: .

Inspection Tests

Instrument Serial #80-000224

Agency Operator	Date	Time	Software	Source	
FDLE/ATP JROC VANDYKE,KEITH,T	09/28/2006	15:30:43	8100.26	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.037
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.040
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.049
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.051
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	INT
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: AIR LEAK AT O-RING, Interferent Detect. Non-compliance: .

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP JROC VANDYKE,KEITH,T		09/28/2006	16:43:37	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

RFI.

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC MALHIOT,MATTHEW,E		10/24/2006	18:49:49	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	5	Avg Std Dev	0.0008	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.05 g/210L Test #4 (Retry)	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.195
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.05 g/210L Test #5 (Retry)	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.195
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.046
		0.05 g/210L Test #6 (Retry)	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000224

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.046
			0.05 g/210L Test #7 (Retry)	0.049
			0.08 g/210L Test #7	0.077
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	INT
			0.05 g/210L Test #8 (Retry)	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9 (Retry)	0.049
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10 (Retry)	0.049
			0.08 g/210L Test #10	0.075
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

05: Interferent Detect, CHANGED HOSES.

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC VANDYKE,KEITH,T		10/31/2006	16:01:20	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC VAN DYKE,KEITH,T		11/29/2006	12:40:56	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: LOW SIMULATOR TEMP.

Inspection Tests

Instrument Serial #80-000224

Agency Operator		Date	Time	Software		Source	
FDLE/ATP PROC VANDYKE,KEITH,T		12/31/2006	13:52:26	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000227

Agency Operator		Date	Time	Software		Source	
FDLE OROC SKIPPER,ROGER,G		06/12/2006	10:51:27	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	4	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000227

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000227

Agency Operator		Date	Time	Software		Source	
FDLE OROC SKIPPER,ROGER,G		10/23/2006	17:48:32	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	618801I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/10/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000227

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000228

Agency Operator		Date	Time	Software	Source	
FDLE MROC VENTURI,GEORGE,L		01/30/2006	14:06:25	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 411802I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 04/30/2006
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.076
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000228

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000228

Agency Operator	Date	Time	Software	Source	
FDLE MROC SHHAJ,ANKDN,Z	09/07/2006	11:05:52	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000228

Agency Operator		Date	Time	Software	Source	
FDLE MROC VENTURI,GEORGE,L		09/19/2006	15:02:12	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 611501G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 04/27/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.201
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.05 g/210L Test #4 (Retry)	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.203
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.05 g/210L Test #5 (Retry)	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.05 g/210L Test #6 (Retry)	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.203
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000228

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.051
			0.05 g/210L Test #7 (Retry)	0.051
			0.08 g/210L Test #7	0.081
			0.20 g/210L Test #7	0.203
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	
			0.05 g/210L Test #8 (Retry)	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9 (Retry)	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10 (Retry)	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

05: RFI Detect: CELL PHONE PASSING IN HALL.

Inspection Tests

Instrument Serial #80-000228

Agency Operator		Date	Time	Software	Source	
FDLE MROC VENTURI,GEORGE,L		10/18/2006	18:11:28	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot# 520102I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration 07/22/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1	INT
Barometric Pressure Sensor Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.051
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.202
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3 (Retry)	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4 (Retry)	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5 (Retry)	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6 (Retry)	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7 (Retry)	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078

Inspection Tests

Instrument Serial #80-000228

Agency Operator	Date	Time	Software	Source
			Interferent Detect Test #7	INT
			Alcohol Free Test #8 (Retry)	0.000
			0.05 g/210L Test #8 (Retry)	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9 (Retry)	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10 (Retry)	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Interferent Detect SIMULATOR CONTAMINATION. 05: Interferent Detect SIMULATOR CONTAMINATION.

Inspection Tests

Instrument Serial #80-000229

Agency Operator	Date	Time	Software	Source
FDLE TBROC SUERETH,DONALD,P	04/11/2006	10:17:07	8100.26	Department Inspection

# Sims	5	Avg Std Dev	Complies	No
--------	---	-------------	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

INTENTIONALLY FAILED TO CHECK FORM REVISION NBRNon-compliance: .

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE TBROC SUERETH,DONALD,P		04/11/2006	10:24:16	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000229

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE TBROC CHAPA,ORLANDO,		04/11/2006	12:49:21	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE CHAPA,ORLANDO,		05/04/2006	13:32:21	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.197
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.007
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.SIM.JAR CRACKED DURING ANALYS. 20: RFI Detect.CELL PHONE REMOVED FROM AREA.

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE BALDWIN,DEAN,		06/29/2006	09:59:28	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

ACETONE STOCK SOLUTION LOT # 2003-D

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE BALDWIN,DEAN,		07/21/2006	00:29:05	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	612202G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

ACETONE STOCK SOLUTION LOT #2003-D

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE BALDWIN,DEAN,		08/31/2006	15:47:13	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

ACETONE STOCK SOLUTION LOT # 2003-D

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE BALDWIN,DEAN,		09/29/2006	16:47:56	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

ACETONE STOCK SOLUTION LOT#2003-D

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE BALDWIN,DEAN,		10/04/2006	17:18:10	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

ACETONE STOCK SOLUTION LOT #2003-D

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE BALDWIN,DEAN,		10/11/2006	16:13:08	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000229

Agency Operator		Date	Time	Software		Source	
FDLE SUERETH,DONALD,P		11/15/2006	16:14:09	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000229

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000230

Agency Operator		Date	Time	Software	Source	
FDLE - MROC VENTURI,GEORGE,L		03/10/2006	12:19:47	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 422301I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 08/12/2006
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000230

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

Inspection Tests

Instrument Serial #80-000230

Agency Operator		Date	Time	Software		Source	
FDLE - MROC VELAZQUEZ,KIMBERLY,		08/01/2006	11:22:57	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610001G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA WAS PRESENT FOR AI

Inspection Tests

Instrument Serial #80-000230

Agency Operator		Date	Time	Software		Source	
FDLE - MROC VELAZQUEZ,KIMBERLY,		09/27/2006	12:19:34	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	610001G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA

Inspection Tests

Instrument Serial #80-000230

Agency Operator		Date	Time	Software		Source	
FDLE - MROC VELAZQUEZ,KIMBERLY,		10/02/2006	10:27:25	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610001G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/12/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000230

Agency Operator		Date	Time	Software		Source	
FDLE - MROC VELAZQUEZ,KIMBERLY,		10/27/2006	09:23:20	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	610001G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/12/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT

Inspection Tests

Instrument Serial #80-000230

Agency Operator		Date	Time	Software	Source	
FDLE - MROC VENTURI,GEORGE,L		11/21/2006	16:12:03	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot# 626101I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration 09/20/2008
# Sims	5	Avg Std Dev	0.0002	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000230

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software	Source	
ALACHUA COUNTY SO SUERETH,DONALD,P		03/27/2006	09:12:44	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	4	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000737

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

A F / M A: Range Exceeded.TOO MUCH ALCOHOL USED FOR M/A T.

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO LEVY,JAYSON,I		03/27/2006	11:14:59	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	411902I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator	Date	Time	Software	Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R	04/29/2006	18:51:02	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	No	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000737

Agency Operator	Date	Time	Software	Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R	04/29/2006	19:10:20	8100.26	Agency	Inspection

# Sims	4	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:FIRST REPEAT ADDED WRONG INT. (MOUTH ALCOHOL). Non-compliance: .

Inspection Tests

Instrument Serial #80-000737

Agency Operator	Date	Time	Software	Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R	04/29/2006	19:34:37	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: UNKNOWN PROBLEM. Non-compliance: NOT DETECTING INTERFERENT.

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		04/30/2006	16:46:33	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.163
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.187
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance/ SIM. PROBLEMS. Non-compliance: NO AGREEMENT WITH 0.20 SIM.

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		05/04/2006	16:11:55	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	504602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software	Source	
ALACHUA COUNTY SO SUERETH,DONALD,P		05/08/2006	09:47:49	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot# 518602I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.083
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.083
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.083
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.083
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000737

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.083
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.083
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		06/07/2006	17:41:44	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	504602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.085
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		07/20/2006	20:44:06	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	504602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.086
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.085
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.086
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.085
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.085
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.085
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance:LOOSE CYL. CONNECTION.

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		08/27/2006	21:14:26	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		09/22/2006	17:23:43	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

FORGOT TO USE MOUTH ALCOHOL.

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO SUERETH,DONALD,P		10/20/2006	12:47:55	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Avg Std Dev	0.0008	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000737

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0019
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		10/21/2006	01:41:30	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		11/27/2006	23:16:18	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	606002I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	03/03/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000737

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		12/25/2006	22:39:07	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software	Source	
FT. WALTON BEACH PD MALHIOT,MATTHEW,E		02/21/2006	11:26:37	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 520201I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/26/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000738

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator	Date	Time	Software	Source	
FT. WALTON BEACH PD HULL,R,S	03/27/2006	14:39:56	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence AbortedSequence Aborted. Non-compliance: .

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software	Source	
FT. WALTON BEACH PD HULL,R,S		03/27/2006	14:50:42	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#		
Expiration	09/22/2007	Expiration	04/21/2007			
# Sims	2	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.067
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.047
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.073
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.047
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.075
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RFI. Non-compliance: .

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		03/27/2006	16:23:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	411902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.179
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.187
		0.20 g/210L Test #2 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Ambient Fail. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		04/24/2006	11:35:23	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	411902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		05/30/2006	14:40:51	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		06/26/2006	11:16:36	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		07/29/2006	12:58:43	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		08/22/2006	17:04:34	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		09/29/2006	10:11:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator	Date	Time	Software	Source
FT. WALTON BEACH PD MALHIOT,MATTHEW,E	10/24/2006	15:52:09	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 609301G Expiration 04/05/2008	
# Sims 5	Avg Std Dev 0.0007	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.195
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000738

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		10/26/2006	13:07:10	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		11/22/2006	14:03:36	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000738

Agency Operator		Date	Time	Software		Source	
FT. WALTON BEACH PD HULL,R,S		12/29/2006	20:45:20	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO MALHIOT,MATTHEW,E		03/27/2006	12:54:50	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	411901I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		03/27/2006	20:11:22	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	2005C	0.08 Gas Lot#	4118021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: No Sample Provided.

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software	Source	
OKALOOSA COUNTY SO MALHIOT,MATTHEW,E		06/12/2006	08:09:25	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot# 610201G
Expiration	09/22/2007	Expiration	11/18/2006	Expiration	09/22/2007	Expiration 04/17/2008
# Sims	5	Avg Std Dev	0.0003	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		06/14/2006	02:31:39	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	611801G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W	06/17/2006	00:08:54	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Subject Test Refused.

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W	07/09/2006	21:30:46	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIW,W	07/09/2006	21:36:34	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

NY.

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W	07/09/2006	21:40:08	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W	07/11/2006	22:07:12	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W	07/11/2006	22:09:14	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W	07/11/2006	22:25:35	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		07/11/2006	22:30:44	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	611801G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		08/22/2006	09:16:08	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	N
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	03/03/2003
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		08/22/2006	10:08:15	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	611801G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		09/28/2006	09:12:44	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO MALHIOT,MATTHEW,E		10/25/2006	15:20:48	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	611801G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/01/2008
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000739

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		10/31/2006	02:56:48	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		11/29/2006	02:01:30	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000739

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO BOZEMAN,CHARLIE,W		12/27/2006	02:42:41	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software	Source	
VALPARAISO PD MALHIOT,MATTHEW,E		02/21/2006	12:41:30	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 525202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 09/14/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000742

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software		Source	
VALPARAISO PD RATELIFF,ALAN,W		05/10/2006	18:55:59	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.183
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software		Source	
VALPARAISO PD RATELIFF,ALAN,W		06/03/2006	15:36:42	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.168
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance LEAK IN SIMULATOR .

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software		Source	
VALPARAISO PD RATELIFF,ALAN,W		07/27/2006	23:16:44	8100.26		Agency	Inspection
0.05 Lot# Expiration	200509A 09/22/2007	0.08 Lot# Expiration	200509C 09/22/2007	0.20 Lot# Expiration	200509C 09/22/2007	0.08 Gas Lot#	
# Sims	3	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.162
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.227
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.160
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3 (Retry)	0.154
		Interferent Detect Test #3	INT

Remarks:

20: Purge Fail, Control Outside Tolerance. Non-compliance: PURGE FAILURE AND CONTROL TEST OUT OF TOLE.

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software	Source	
VALPARAISO PD RATELIFF,ALAN,W		08/24/2006	00:15:06	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#		
Expiration	09/22/2007	Expiration	09/22/2007			
# Sims	3	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.039
Barometric Pressure Sensor Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1	0.072
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Test #1 (Retry)	0.068
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.042
0.08 g/210L Test: OK	No	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	No	0.08 g/210L Test #2	0.074
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2 (Retry)	0.069
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.011
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.068
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. 05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software		Source	
VALPARASIO PD MALHIOT,MATTHEW,E		10/24/2006	16:07:49	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000742

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software		Source	
VALPARASIO PD RATELIFF,ALAN,W		11/01/2006	21:11:34	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance DUE TO AIR LEAK IN SIMULATOR .

Inspection Tests

Instrument Serial #80-000742

Agency Operator		Date	Time	Software		Source	
VALPARASIO PD RATELIFF,ALAN,W		12/09/2006	16:57:13	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software	Source	
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E		02/20/2006	07:36:43	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 520201I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 07/26/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000744

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		03/27/2006	15:11:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	5252021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	09/14/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		04/30/2006	15:44:51	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	525202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	09/14/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.058
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.075
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.194
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
		0.20 g/210L Test #2	0.194
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: SIM TUBE LEAK. 08: SIM TUBE LEAK.

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		05/27/2006	08:31:00	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	525202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	09/14/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		06/24/2006	12:21:03	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	525202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	09/14/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.043
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.043
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: SIM LEAK

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		07/31/2006	07:54:31	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
		0.20 g/210L Test #2	0.192
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect. 05: Control Outside Tolerance/LOW SIM TEMP. 08: Control Outside Tolerance/LOW SIM TEMP.

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		08/30/2006	10:18:20	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000744

Agency Operator	Date	Time	Software	Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E	09/29/2006	10:27:33	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:SIM LEAK, . Non-compliance:SIM MALFUNCTION/LEAK.

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		09/29/2006	12:07:49	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

08: Purge Fail/BLOCKED BREATH TUBE.

Inspection Tests

Instrument Serial #80-000744

Agency Operator	Date	Time	Software	Source
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E	10/23/2006	11:43:05	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 611801G Expiration 05/01/2008	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.082
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.082
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.082
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000744

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		10/31/2006	09:41:15	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	611801G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/01/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.190
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A:APPLIED MOUTH SOULUTION PRIOR TO A/F TEST.

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		11/29/2006	11:14:46	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	620901I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	08/02/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.LOW SIM TEMP

Inspection Tests

Instrument Serial #80-000744

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		12/22/2006	14:55:29	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.188
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.08 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.189
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, RFI Detect. Non-compliance:CONTROL OUTSIDE TOL. ON .200 G/210L.

Inspection Tests

Instrument Serial #80-000749

Agency Operator	Date	Time	Software	Source
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E	02/20/2006	08:05:30	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 520201I Expiration 07/26/2007	
# Sims 5	Avg Std Dev 0.0004	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.046
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.046
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.046
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000749

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.045
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		03/27/2006	16:58:05	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.075
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
		0.20 g/210L Test #2	0.194
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance SIM LEAK. 08: Control Outside Tolerance SIM LEAK.

Inspection Tests

Instrument Serial #80-000749

Agency Operator	Date	Time	Software	Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E	04/30/2006	17:23:29	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	
Expiration	04/21/2007	Expiration	04/21/2007		
# Sims	4	Complies	No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.045
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.073
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:SIM LEAK.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		04/30/2006	18:13:08	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: SIM LEAK.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		05/27/2006	09:33:06	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		06/24/2006	13:28:57	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.038
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.196
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.039
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
		0.20 g/210L Test #2	0.197
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.039
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05:SIM LEAK. 08:SIM LEAK.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		07/31/2006	08:46:49	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		08/30/2006	11:05:20	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: LOW SIM TEMP.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		09/29/2006	12:49:53	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

A F / M A:LOW BREATH SAMPLE.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software	Source	
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E		10/23/2006	12:16:41	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 619101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/12/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.082
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #2	0.082
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.082
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.082
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.20 g/210L Test #4 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #4	0.081
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.083
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.082
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6 (Retry)	0.198

Inspection Tests

Instrument Serial #80-000749

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Dry Gas Std Test #6 (Retry)	0.082
			Interferent Detect Test #6	INT
			Alcohol Free Test #7	0.000
			0.05 g/210L Test #7	0.047
			0.08 g/210L Test #7	0.077
			0.20 g/210L Test #7 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.082
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9 (Retry)	0.198
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10 (Retry)	0.197
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

20: RFI Detect / CELL PHONE IN AREA. 08: RFI Detect.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		10/31/2006	10:23:55	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		11/29/2006	12:15:41	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.191
		0.20 g/210L Test #2 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.LOW SIM TEMP. 20: Control Outside Tolerance.LOW SIM TEMP.

Inspection Tests

Instrument Serial #80-000749

Agency Operator		Date	Time	Software		Source	
ESCAMBIA COUNTY SO ISENBERG,KESLEY,E		12/22/2006	15:41:45	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000751

Agency Operator	Date	Time	Software	Source
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E	02/20/2006	08:21:09	8100.26	Department Inspection

# Sims	5	Avg Std Dev	Complies	No
--------	---	-------------	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

PURGE FAIL AFTER MOUTH ALCOHOL TESTNon-compliance: .

Inspection Tests

Instrument Serial #80-000751

Agency Operator		Date	Time	Software	Source	
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E		02/20/2006	11:07:06	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 520201I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 07/26/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000751

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

A F / M A: TOO MUCH M A SOLUTION.

Inspection Tests

Instrument Serial #80-000751

Agency Operator	Date	Time	Software	Source	
ESCAMBIA COUNTY SO MALHIOT,MATTHEW,E	10/23/2006	14:23:35	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI FOR UPLOAD ONLY, INSPECTION NOT COMPLETENon-compliance: .

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SUERETH,DONALD,P		03/21/2006	09:17:36	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0008	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000752

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0021
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		03/21/2006	17:45:16	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

RECHECKED SEALS/CHANGED GLASSWARE.

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		04/04/2006	04:25:36	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000752

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO SAVARESE,K,P	05/28/2006	05:49:31	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000752

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO SAVARESE,K,P	05/28/2006	06:00:04	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DUI BROUGHT IN HAD TO EXIT INSPECTION

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		05/28/2006	07:00:27	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		06/25/2006	11:17:12	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SECKEL,ERIC,J		07/24/2006	17:33:54	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515901I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SECKEL,ERIC,J		08/17/2006	11:37:54	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		09/26/2006	09:21:17	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		11/30/2006	14:13:17	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200504D	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	04/21/2007	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000752

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,M,F		12/16/2006	03:30:08	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software	Source	
ALACHUA COUNTY SO MALHIOT,MATTHEW,E		01/25/2006	13:23:39	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot# 411901I
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	04/11/2006	Expiration 04/30/2006
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000760

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		03/27/2006	11:37:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator	Date	Time	Software	Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R	04/29/2006	18:49:04	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ACETONE ADDED PRIOR TO RUNNING OF SIM.

Inspection Tests

Instrument Serial #80-000760

Agency Operator	Date	Time	Software	Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R	04/29/2006	19:17:24	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000760

Agency Operator	Date	Time	Software	Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R	04/30/2006	16:09:58	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1 (Retry)	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #2 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No	Interferent Detect Test #3 (Retry)	INT
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:LOOSE SEAL ON SIM.. 05: RFI DECT.. Non-compliance:RFI.

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		04/30/2006	16:37:07	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512501I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance/LOOSE SIM. HOSE.

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		05/31/2006	09:38:26	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		06/07/2006	16:42:02	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		07/20/2006	20:52:45	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		08/27/2006	21:18:17	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	434502I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	12/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		09/22/2006	17:23:30	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

FORGOT TO USE MOUTH ALCOHOL.

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO SUERETH,DONALD,P		10/20/2006	12:58:29	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000760

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO PATTEN,WILLIAM,R		10/21/2006	01:47:50	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail/ MOUTHPIECE ON WHILE IN AIR BLANK .

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		11/27/2006	23:05:07	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000760

Agency Operator		Date	Time	Software		Source	
ALACHUA COUNTY SO ULRICH,SCOTT,C		12/25/2006	22:28:29	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO MALHIOT,MATTHEW,E		02/21/2006	12:05:37	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	520201I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/26/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000763

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0013
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO VAN DYKE,KEITH,T		03/27/2006	19:40:51	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4118021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/30/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.068
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

LOOSE CONNECTION TO GAS BOTTLE.

Inspection Tests

Instrument Serial #80-000763

Agency Operator	Date	Time	Software	Source	
OKALOOSA COUNTY SO VAN DYKE,KEITH,T	04/22/2006	07:06:04	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO VAN DYKE,KEITH,T		04/22/2006	07:10:27	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO VAN DYKE,KEITH,T		05/30/2006	23:51:36	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	609301G
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.084
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.083
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside LOOSE TOP.

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO VAN DYKE,KEITH,T		06/30/2006	18:58:36	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.045
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.197
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.045
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
		0.20 g/210L Test #2	0.198
		0.08 g/210L Dry Gas Std Test #2	0.082
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.046
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

00: USED .05 SOLUTION IN ERROR. 08: AIR LEAK.

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO VANDYKE,KEITH,T		07/28/2006	18:39:03	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200504C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A:BAD STOCK SOLUTION.

Inspection Tests

Instrument Serial #80-000763

Agency Operator		Date	Time	Software		Source	
OKALOOSA COUNTY SO VANDYKE,KEITH,T		08/10/2006	17:42:55	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software	Source	
SANTA ROSA COUNTY SO MALHIOT,MATTHEW,E		02/20/2006	10:21:08	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 520201I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 07/26/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8 (Retry)	0.000

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

A F / M A: NO M A USED. 00: Ambient Fail, ON WATER.

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/27/2006	03:12:23	8100.26	Agency	Inspection

# Sims	5	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.011
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/27/2006	03:20:15	8100.26	Agency	Inspection

Sims 5 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/27/2006	03:23:23	8100.26	Agency	Inspection
0.05 Lot#	200411C				
Expiration	11/04/2006				
# Sims	5		Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.048
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.052
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/27/2006	03:43:37	8100.26	Agency	Inspection

Sims 5 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/27/2006	03:48:19	8100.26	Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200411G	0.20 Lot#	200411D
Expiration	11/04/2006	Expiration	11/19/2006	Expiration	11/04/2006
				0.08 Gas Lot#	515902I
				Expiration	06/13/2007
# Sims	5	Complies Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.204
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.085
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.206
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.204
		0.20 g/210L Test #3 (Retry)	0.206
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		04/15/2006	00:23:59	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	20411G	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	05/08/2006	00:37:08	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007
				0.08 Gas Lot#	515902I
				Expiration	06/13/2007
# Sims	5	Complies Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		06/03/2006	01:02:47	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	07/08/2006	03:17:02	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.039
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.039
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.039
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA CO SO MALHIOT,MATTHEW,E		08/28/2006	11:10:04	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.083
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA CO SO BRINKERHOFF,CRAIG,O		09/07/2006	10:16:29	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	518802I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA CO SO BRINKERHOFF,CRAIG,O		10/13/2006	00:11:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software	Source	
SANTA ROSA CO SO MALHIOT,MATTHEW,E		10/24/2006	10:49:10	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/12/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.083
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.083
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000765

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA CO SO BRINKERHOFF,CRAIG,O		11/15/2006	23:00:38	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.209
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

Inspection Tests

Instrument Serial #80-000765

Agency Operator		Date	Time	Software		Source	
SANTA ROSA CO SO BRINKERHOFF,CRAIG,O		12/07/2006	01:55:45	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software	Source	
SUWANNEE COUNTY SO MALHIOT,MATTHEW,E		03/02/2006	10:29:02	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 512501I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 05/10/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000768

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator	Date	Time	Software	Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G	03/25/2006	16:58:08	8100.26	Agency	Inspection

Sims 5 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		03/27/2006	15:37:20	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419101I
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator	Date	Time	Software	Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G	04/27/2006	09:42:59	8100.26	Agency	Inspection

# Sims	4	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		04/27/2006	10:01:41	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419101I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		05/30/2006	21:40:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419101L
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		06/28/2006	21:04:33	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419101L
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		07/31/2006	14:26:18	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	617202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.076
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.077
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		08/31/2006	21:05:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	617202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		09/30/2006	20:55:12	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	617202I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000768

Agency Operator	Date	Time	Software	Source	
SUWANNEE COUNTY SO MALHIOT,MATTHEW,E	10/17/2006	08:08:58	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000770

Agency Operator	Date	Time	Software	Source
HAMILTON COUNTY S.O. MALHIOT,MATTHEW,E	01/26/2006	10:07:23	8100.26	Department Inspection
0.05 Lot# 200411C Expiration 11/04/2006	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200411D Expiration 04/11/2006	0.08 Gas Lot# 419101I Expiration 07/14/2006	
# Sims 4	Avg Std Dev 0.0007	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000770

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		03/22/2006	09:49:20	8100.26		Agency	Inspection
0.05 Lot#	200509B	0.08 Lot#	200509B	0.20 Lot#	200411D	0.08 Gas Lot#	419101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/04/2006	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R	04/24/2006	10:03:29	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		04/24/2006	10:23:35	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200509B	0.20 Lot#	200411D	0.08 Gas Lot#	419101I
Expiration	11/04/2006	Expiration	09/22/2007	Expiration	11/04/2006	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R	05/25/2006	09:47:36	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. MALHIOT,MATTHEW,E		08/28/2006	13:06:40	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000770

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		09/18/2006	14:09:31	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	617202I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. MALHIOT,MATTHEW,E		10/17/2006	13:55:53	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	617202I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/23/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000770

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		10/25/2006	10:07:03	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	617202I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		11/28/2006	09:08:04	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000770

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		12/27/2006	15:59:37	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software	Source	
PERRY PD MALHIOT,MATTHEW,E		02/02/2006	09:48:36	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot# 428101I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	11/04/2006	Expiration 10/11/2006
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000771

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator	Date	Time	Software	Source	
PERRY PD PARKER,C,	03/28/2006	10:39:06	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N. Non-compliance: .

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		03/28/2006	10:44:32	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,CLA,		04/27/2006	09:49:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.185
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.196
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Interferent Detect. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,CLA,		05/30/2006	08:53:33	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		06/30/2006	08:56:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

INT NOT DETECTED.

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		07/28/2006	09:14:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD GRAY,LIN,		08/16/2006	05:35:31	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	2006C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.188
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.197
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.044
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.196
		0.20 g/210L Test #2 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #2	0.082
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect. 00: Ambient Fail. 05: Control Outside Tolerance. 20: RFI Detect.

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		09/28/2006	09:02:29	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000771

Agency Operator	Date	Time	Software	Source
PERRY PD MALHIOT,MATTHEW,E	10/30/2006	13:43:14	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 612201G Expiration 05/05/2008	
# Sims 5	Avg Std Dev 0.0006	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.083
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.082
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.082
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000771

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0010
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator	Date	Time	Software	Source	
PERRY PD PARKER,C,	10/31/2006	10:11:16	8100.27	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200509C
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007
0.08 Gas Lot#	612201G	Expiration	05/05/2008		
# Sims	2	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator		Date	Time	Software		Source	
PERRY PD GRAY,LIN,		11/27/2006	03:56:21	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	612201G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000771

Agency Operator	Date	Time	Software	Source	
PERRY PD PARKER,C,	12/27/2006	14:10:32	8100.27	Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008
0.08 Gas Lot#	612201G	Expiration	05/05/2008		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source
SANTA ROSA COUNTY SO MALHIOT,MATTHEW,E	02/20/2006	10:01:51	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 520201I Expiration 07/26/2007	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.082
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.082
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/27/2006	02:15:49	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		03/27/2006	02:16:30	8100.26		Agency	Inspection
0.05 Lot#	200511C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot#	518802I
Expiration	11/04/2006	Expiration	11/19/2006	Expiration	11/04/2006	Expiration	07/21/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		04/14/2006	22:48:34	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	518802I
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	07/21/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	05/07/2006	22:57:09	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007
0.08 Gas Lot#	518802I	Expiration	07/12/2007		
# Sims	5	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		06/03/2006	00:19:05	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518802I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.207
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		07/08/2006	02:07:48	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200505C	0.20 Lot#	200505C	0.08 Gas Lot#	200504D
Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2007	Expiration	04/21/2007
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.209
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.206
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.210
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.206
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Dry Gas Std Test #2	0.083
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.211
		0.20 g/210L Test #3 (Retry)	0.207
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance. 08: Control Outside Tolerance. Non-compliance: INCORRECT LOT NUMBERS ENTERED.

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	07/08/2006	03:19:15	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		07/08/2006	03:25:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518802I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	07/10/2006	02:27:07	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007
				0.08 Gas Lot#	518802I
				Expiration	07/12/2007
# Sims	5	Complies Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.209
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.210
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.085
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		08/05/2006	03:19:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518802I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		09/07/2006	09:27:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	518802I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000772

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	10/12/2006	22:45:54	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000772

Agency Operator		Date	Time	Software		Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O		10/12/2006	22:48:45	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/12/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software	Source	
LAFAYETTE COUNTY SO SUERETH,DONALD,P		03/20/2006	10:27:42	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.05 g/210L Test #4 (Retry)	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.05 g/210L Test #5 (Retry)	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.05 g/210L Test #6 (Retry)	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000773

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.05 g/210L Test #7 (Retry)	0.048
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.05 g/210L Test #8 (Retry)	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	
			0.05 g/210L Test #9 (Retry)	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10 (Retry)	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

05: RFI Detect.CELL PHONE REMOVED FROM AREA.

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MALHIOT,MATTHEW,E		08/28/2006	12:18:23	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000773

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000773

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	09/28/2006	12:14:59	8100.26	Agency	Inspection

# Sims	2	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance: .

Inspection Tests

Instrument Serial #80-000773

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	09/28/2006	12:32:16	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DID NOT MEET SLOP. Non-compliance: .

Inspection Tests

Instrument Serial #80-000773

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	09/28/2006	12:58:07	8100.26	Agency	Inspection

# Sims	2	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance: .

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		09/28/2006	13:14:29	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008		
# Sims	2	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.174
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.183
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.187
		0.20 g/210L Test #3 (Retry)	0.193
		Interferent Detect Test #3	INT

Remarks:

LOW TEMP. N.

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		10/20/2006	11:10:29	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	52502I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/14/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.010
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.184
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.192
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.189
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

RANGE EX. REPLACED HOSE. AIR LEAK.

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO SUERETH,DONALD,P		10/31/2006	09:59:21	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000773

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		11/01/2006	09:10:18	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	52502I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/14/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.184
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		11/28/2006	15:43:52	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	52502I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/14/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK.

Inspection Tests

Instrument Serial #80-000773

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		12/28/2006	10:12:59	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	627001I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/29/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software	Source	
COLUMBIA COUNTY SO MALHIOT,MATTHEW,E		02/14/2006	10:27:42	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot# 436401I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	11/04/2006	Expiration 01/04/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.083
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.083
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000775

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		03/17/2006	08:05:23	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NO REASON

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		04/16/2006	19:45:55	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		05/14/2006	18:11:11	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.203
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		06/11/2006	13:28:45	8100.26		Agency	Inspection
0.05 Lot# Expiration	200509A 09/22/2007	0.08 Lot# Expiration	200509B 09/22/2007	0.20 Lot# Expiration	200509C 09/22/2007	0.08 Gas Lot#	
# Sims	2	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.074
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.08 g/210L Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: RFI Detect, RFI Detect. Non-compliance: .

Inspection Tests

Instrument Serial #80-000775

Agency Operator	Date	Time	Software	Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A	06/11/2006	14:51:34	8100.26	Agency	Inspection

# Sims	2	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI DetectRFI DetectRFI DetectRFI Detect. Non-compliance: .

Inspection Tests

Instrument Serial #80-000775

Agency Operator	Date	Time	Software	Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A	06/11/2006	15:14:31	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		06/16/2006	14:10:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.023
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.051
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.203
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.025
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		07/28/2006	21:25:53	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		08/27/2006	08:34:51	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range ExceededAmbient Fail.

Inspection Tests

Instrument Serial #80-000775

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		09/29/2006	09:05:34	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	43601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000775

Agency Operator	Date	Time	Software	Source
COLUMBIA COUNTY SO MALHIOT,MATTHEW,E	10/20/2006	12:46:18	8100.27	Department Inspection

# Sims	0	Avg Std Dev	Complies	No
--------	---	-------------	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BARO SENSOR CHECK FAILED 988/1007 2Non-compliance: MORE THE 1.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE CO SO SUERETH,DONALD,P		03/20/2006	11:34:48	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.076
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.076
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0017
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		04/21/2006	10:59:08	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	2	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.182
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.187
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.20 g/210L Test: OK	No	0.08 g/210L Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2 (Retry)	0.077
		0.20 g/210L Test #2	0.185
		0.20 g/210L Test #2 (Retry)	0.191
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.193
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		04/28/2006	14:38:11	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	4314011
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/11/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.186
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.194
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Alcohol Free Test: OK	Yes	Interferent Detect Test #1	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2	0.077
		0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.078
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.Y

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		05/24/2006	08:36:49	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	4314011
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/11/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	05/24/2006	10:13:46	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	05/24/2006	10:15:24	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	05/24/2006	10:17:08	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	05/24/2006	10:22:11	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		06/23/2006	10:28:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	07/27/2006	09:05:33	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	
Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	2	Complies No			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. AIR LEAK.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		07/27/2006	10:06:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.187
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.194
		0.20 g/210L Test #2 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK. AIR LEAK.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		08/24/2006	09:47:56	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		09/28/2006	10:22:46	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615802I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		10/20/2006	09:16:48	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	52502I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/14/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.185
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.194
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.044
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.045
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

TEMP BELOW 33.8. TEMP BELOW 33.8.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO SUERETH,DONALD,P		10/31/2006	09:47:37	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	11/01/2006	10:13:16	8100.27	Agency	Inspection
0.05 Lot#	200606B	0.08 Lot#	200605C	0.20 Lot#	
Expiration	05/04/2008	Expiration	05/04/2008		
# Sims	2	Complies No			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.072
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.073
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

TEMP BELOW 33.8 , Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000776

Agency Operator	Date	Time	Software	Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,	11/01/2006	11:15:58	8100.27	Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	
Expiration	05/04/2008	Expiration	05/04/2008		
# Sims	2	Complies	No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.072
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

TEMP BELOW 33.8, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		11/01/2006	12:45:00	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	52502I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/14/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.073
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		11/28/2006	14:25:01	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	52502I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/14/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.184
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.194
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK. AIR LEAK.

Inspection Tests

Instrument Serial #80-000776

Agency Operator		Date	Time	Software		Source	
LAFAYETTE COUNTY SO MANDOLA,MATTHEW,		12/28/2006	11:50:54	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	627001I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/29/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.188
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK. AIR LEAK.

Inspection Tests

Instrument Serial #80-000777

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO MALHIOT,MATTHEW,E		04/04/2006	11:11:15	8100.26		Department Inspection	
0.05 Lot#	200411C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	434502I
Expiration	11/04/2006	Expiration	11/18/2006	Expiration	05/05/2007	Expiration	12/14/2006
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000777

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000777

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		05/14/2006	20:10:37	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/04/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000777

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		06/25/2006	08:01:47	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000777

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		07/25/2006	10:52:38	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000777

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		08/29/2006	09:49:45	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000777

Agency Operator		Date	Time	Software		Source	
COLUMBIA COUNTY SO PERBTANI,AAMIR,A		09/28/2006	14:26:23	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD MALHIOT,MATTHEW,E		02/09/2006	10:45:11	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000778

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

MOUTH ALCOHOL TEST: TOO MUCH M/A SOLUTION

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		03/27/2006	11:08:39	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	4119021
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	04/30/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: AIR LEAK.

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		04/24/2006	10:30:31	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	411902I
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	04/30/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #2	0.000
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

20: USED WRONG SIMULATOR. 08: HOSE NOT CONNECT.

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		05/23/2006	11:20:29	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	416701I
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	06/17/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		06/22/2006	10:50:42	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	612202G
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		07/21/2006	08:25:23	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	612202G
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,	08/18/2006	09:33:33	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000778

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,	08/18/2006	09:41:06	8100.26	Agency	Inspection

# Sims	5	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RANGE EXCEEDEDRange Exceeded. Non-compliance: .

Inspection Tests

Instrument Serial #80-000778

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,	08/18/2006	09:59:43	8100.26	Agency	Inspection

# Sims	5	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance: .

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD		08/18/2006	10:04:41	8100.26		Agency	Inspection
HEMPHILL,CLOTE,							
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	612202G
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		09/26/2006	13:01:35	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	612202G
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD MALHIOT,MATTHEW,E		10/19/2006	15:21:45	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000778

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		10/20/2006	08:38:43	8100.27		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	11/04/2006	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:M/A NOT DETECTED

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		11/20/2006	09:44:40	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000778

Agency Operator		Date	Time	Software		Source	
NAS JACKSONVILLE PD HEMPHILL,CLOTE,		12/20/2006	09:09:40	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000779

Agency Operator		Date	Time	Software	Source	
MADISON COUNTY SO MALHIOT,MATTHEW,E		03/13/2006	09:24:34	8100.26	Department Inspection	
0.05 Lot#	200411C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot# 434502I
Expiration	11/04/2006	Expiration	11/18/2006	Expiration	11/04/2006	Expiration 12/14/2006
# Sims	5	Avg Std Dev	0.0002	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000779

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000779

Agency Operator		Date	Time	Software		Source	
MADISON COUNTY SO STRICKLAND,JAMES,M		03/16/2006	13:03:38	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	419101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000779

Agency Operator		Date	Time	Software		Source	
MADISON COUNTY SO STRICKLAND,JAMES,M		08/15/2006	13:32:08	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

NONE/BACK-UP UNIT

Inspection Tests

Instrument Serial #80-000779

Agency Operator		Date	Time	Software		Source	
MADISON COUNTY SO STRICKLAND,JAMES,M		09/11/2006	11:48:38	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

BACK-UP INSTRUMENT

Inspection Tests

Instrument Serial #80-000779

Agency Operator		Date	Time	Software		Source	
MADISON COUNTY SO STRICKLAND,JAMES,M		10/05/2006	13:40:16	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	428101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

NONE-SPARE UNIT

Inspection Tests

Instrument Serial #80-000780

Agency Operator		Date	Time	Software	Source	
SUWANNEE COUNTY SO MALHIOT,MATTHEW,E		03/02/2006	10:42:05	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 512501I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 05/10/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.203
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.076
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000780

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000780

Agency Operator	Date	Time	Software	Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G	03/25/2006	16:53:41	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

PASSED.

Inspection Tests

Instrument Serial #80-000780

Agency Operator	Date	Time	Software	Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G	03/25/2006	17:12:19	8100.26	Agency	Inspection

Sims 5 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000780

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		03/27/2006	14:47:48	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419101I
Expiration	11/04/2006	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000780

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		10/06/2006	19:11:00	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	6172021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000780

Agency Operator		Date	Time	Software	Source	
SUWANNEE COUNTY SO MALHIOT,MATTHEW,E		10/17/2006	10:52:52	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 617202I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 06/23/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000780

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000780

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		11/30/2006	11:53:57	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000780

Agency Operator		Date	Time	Software		Source	
SUWANNEE COUNTY SO FALGOUT,DAVID,G		12/30/2006	12:40:08	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000781

Agency Operator	Date	Time	Software	Source
WASHINGTON COUNTY SO MALHIOT,MATTHEW,E	03/08/2006	09:27:55	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 411803I Expiration 04/30/2006	
# Sims 5	Avg Std Dev 0.0006	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000781

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0011

Remarks:

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		03/27/2006	10:00:40	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504C	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Purge Fail.

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		04/30/2006	08:26:35	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		05/29/2006	22:43:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		07/04/2006	18:54:23	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.046
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.046
Barometric Pressure Sensor Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.192
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.046
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.194
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.047
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance PICKED UP THE WRONG SIMULATOR. 05: RFI Detect.

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		07/23/2006	15:02:53	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		08/30/2006	10:50:45	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:WEAK STOCK SOULTION.

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		09/30/2006	20:47:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000781

Agency Operator	Date	Time	Software	Source
WASHINGTON COUNTY SO MALHIOT,MATTHEW,E	10/25/2006	13:54:06	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 609301G Expiration 04/05/2008	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000781

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

A F / M A, LEFT OVER ALCOHOL FROM OTHER TEST.

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		10/31/2006	23:27:30	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		11/30/2006	21:38:00	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.230
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.227
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.224
		0.20 g/210L Test #3 (Retry)	0.190
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000781

Agency Operator	Date	Time	Software	Source	
WASHINGTON COUNTY SO BROCK,BILLY,J	12/31/2006	14:21:17	8100.27	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	INT
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ALCOHOL IN MOUTH . 00: Interferent Detect, Interferent Detect. Non-compliance:SIMULATOR CONTAMINATED
WILL WASH AND REPEAT .

Inspection Tests

Instrument Serial #80-000781

Agency Operator	Date	Time	Software	Source	
WASHINGTON COUNTY SO BROCK,BILLY,J	12/31/2006	16:55:48	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Interferent Detect SIMULATOR CONTAMINATED, Interferent . Non-compliance:CONTAMINATED SIMULATOR.

Inspection Tests

Instrument Serial #80-000781

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		12/31/2006	19:01:04	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator	Date	Time	Software	Source
WASHINGTON COUNTY SO MALHIOT,MATTHEW,E	03/08/2006	09:34:21	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 411803I Expiration 04/30/2006	
# Sims 5	Avg Std Dev 0.0006	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.087
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.083
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.085
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.083
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.084
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.083
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.083
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.083
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.083
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.082
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.083
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.082
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000782

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.083
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.082
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.082
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.082
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.082
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: Control Outside Tolerance. DRY GAS RUNNING HIGH

Inspection Tests

Instrument Serial #80-000782

Agency Operator	Date	Time	Software	Source
WASHINGTON COUNTY SO BROCK,BILLY,J	03/27/2006	08:41:40	8100.26	Agency Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200504D Expiration 04/21/2007	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 200504D Expiration 04/21/2007	
# Sims 4	Complies No			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.189
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.184
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #1	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2	0.000
0.08 g/210L Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2	0.078
		0.20 g/210L Test #2	0.195
		0.20 g/210L Test #2 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #2	0.000
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.000
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

Control Outside Tolerance. 08: Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		03/27/2006	09:47:08	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	416701I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/17/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		04/30/2006	08:23:55	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	416701I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/17/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: . 05: RFI Detect.

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		05/29/2006	22:34:30	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		07/04/2006	18:48:22	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		07/23/2006	14:53:14	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		08/30/2006	10:49:12	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:WEAK STOCK SOULTION.

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		09/30/2006	20:44:46	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO MALHIOT,MATTHEW,E		10/25/2006	15:27:56	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000782

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		10/31/2006	23:31:59	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator	Date	Time	Software	Source	
WASHINGTON COUNTY SO BROCK,BILLY,J	11/30/2006	19:59:24	8100.27	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:WEEK ACETONE, . Non-compliance:FAILED ACETONE TEST.

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		11/30/2006	21:29:56	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software	Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		12/31/2006	19:03:45	8100.27	Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#		
Expiration	05/04/2008	Expiration	05/04/2008			
# Sims	4	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	INT
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.20 g/210L Test: OK	No	0.08 g/210L Test #3 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT

Remarks:

08: Interferent Detect CONTAMINATED SIMULATOR, Control Out . SIMULATOR CONTAMINATED NO O RING IN SIMULATORNon-compliance:SIMULATOR CONTAMINATED NO O RING IN SIMULATO.

Inspection Tests

Instrument Serial #80-000782

Agency Operator		Date	Time	Software		Source	
WASHINGTON COUNTY SO BROCK,BILLY,J		12/31/2006	20:26:53	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	609301G
Expiration	05/04/2008	Expiration	05/04/2007	Expiration	05/10/2008	Expiration	04/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Improper Sample. 05: Interferent Detect.

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software	Source	
HOLMES COUNTY SO MALHIOT,MATTHEW,E		03/08/2006	08:42:41	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 411803I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 04/30/2006
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000784

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY, MICHAEL,		03/27/2006	09:35:08	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

WAIT TO LONG ON 1ST M/A, NO M/A ON 2ND

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		04/26/2006	15:03:35	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		05/04/2006	15:03:03	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		06/13/2006	17:00:44	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator	Date	Time	Software	Source	
HOLMES COUNTY SO RALEY,MICHAEL,	07/13/2006	14:23:19	8100.26	Agency	Inspection

Sims 2 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		07/13/2006	14:28:40	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		07/20/2006	15:49:13	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		08/11/2006	12:42:26	8100.26		Agency	Inspection
0.05 Lot#	200506B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

HOSE NOT HOOKED UP.

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		09/20/2006	15:49:59	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	618101I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		10/24/2006	20:32:44	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	606002I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	03/03/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO MALHIOT,MATTHEW,E		10/25/2006	12:55:18	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000784

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		11/15/2006	22:16:04	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	618101I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000784

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		12/27/2006	21:27:56	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	606002I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	03/03/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software	Source	
HOLMES COUNTY SO MALHIOT,MATTHEW,E		03/08/2006	08:35:59	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 411803I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 04/30/2006
# Sims	5	Avg Std Dev	0.0010	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000785

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		03/27/2006	09:33:25	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		04/26/2006	15:08:12	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.058
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.066
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.191
		0.20 g/210L Test #2 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.TEMPERATURE OUT OF RANGE

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		05/04/2006	15:02:51	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	415601I
Expiration	09/20/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		06/13/2006	17:01:13	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000785

Agency Operator	Date	Time	Software	Source	
HOLMES COUNTY SO RALEY,MICHAEL,	07/13/2006	14:23:05	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	
Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	2	Complies	No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		07/20/2006	15:53:07	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		08/11/2006	12:45:26	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

LEAK IN TUBING.

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		09/20/2006	15:55:10	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	618101I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		10/24/2006	20:26:08	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	618101I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator	Date	Time	Software	Source
HOLMES COUNTY SO MALHIOT,MATTHEW,E	10/25/2006	12:45:13	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 6060021 Expiration 03/03/2008	
# Sims 5	Avg Std Dev 0.0007	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.195
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.076
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000785

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		11/12/2006	22:06:23	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	606002I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	03/03/2008
# Sims	3	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		Interferent Detect Test #3	INT

Remarks:

JAIL WAS PAINTED TODAYNon-compliance: .

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		11/15/2006	22:23:51	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	606002I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	03/03/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

LEAK IN TUBING.

Inspection Tests

Instrument Serial #80-000785

Agency Operator		Date	Time	Software		Source	
HOLMES COUNTY SO RALEY,MICHAEL,		12/27/2006	21:27:08	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	618101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO HACKNEY,DWITE,N		01/12/2006	12:15:20	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000786

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		03/24/2006	16:11:09	8100.26		Agency	Inspection
0.05 Lot#	200411C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot#	512402I
Expiration	11/04/2006	Expiration	11/19/2006	Expiration	11/04/2006	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO VENTURI,GEORGE,L		04/06/2006	17:26:59	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000786

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

AI F. DICIENZO PRESENT

Inspection Tests

Instrument Serial #80-000786

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/19/2006	21:14:04	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO NANZ,PATRICIA,M		04/19/2006	21:24:03	8100.26		Agency	Inspection
0.05 Lot#	200502A	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot#	512402I
Expiration	02/14/2007	Expiration	11/19/2006	Expiration	11/04/2006	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO NANZ,PATRICIA,M		05/18/2006	22:10:12	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

FRANCA DICIENZO WAS PRESENT FOR THE ENTIRE INSPECTION

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		06/08/2006	15:22:40	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		07/22/2006	20:05:53	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200411D	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	11/04/2006	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		08/09/2006	11:53:14	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

PATRICIA NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		09/22/2006	11:53:08	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO HACKNEY,DWITE,N		10/18/2006	15:04:15	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000786

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		10/20/2006	09:59:36	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		10/27/2006	10:06:07	8100.27		Agency	Inspection
0.05 Lot#	512402I	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	05/06/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

ENTERED WRONG LOT # FOR 0.050 SOLUTION.P NANZ WAS PRESENT.

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		11/17/2006	11:26:03	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R		12/15/2006	10:56:52	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512402I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/06/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

Inspection Tests

Instrument Serial #80-000786

Agency Operator		Date	Time	Software		Source	
BROWARD COUNTY SO NANZ,PATRICIA,		12/21/2006	09:33:11	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	608301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	03/27/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

Inspection Tests

Instrument Serial #80-000787

Agency Operator	Date	Time	Software	Source
GULF BREEZE PD MALHIOT,MATTHEW,E	02/20/2006	07:48:05	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 520201I Expiration 07/26/2007	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000787

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		03/24/2006	10:39:40	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	504601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD NEFF,STEPHEN,		04/21/2006	08:14:52	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	504601I
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

OBSERVED BY OFFICER TATRO

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		05/24/2006	13:13:48	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Improper Sample-INTRODUCED SAMPLE DURING PURGE . DURING .05 - AIR LEAK IN SIMULATOR

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		06/01/2006	12:15:14	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.191
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: DID NOT ADD 3ML TILL LAST SAMPLE
. 05: Control Outside Tolerance-HOOKED UP SIM LATE.

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD MALHIOT,MATTHEW,E		08/22/2006	12:36:26	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000787

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		08/30/2006	10:19:23	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Intft Det-TYGON HAD RESIDUAL ACETONE - CHANGED TYGON .

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		09/21/2006	10:30:23	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. MOUTHPIECE REMOVED LATE

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		10/05/2006	07:06:31	8100.26		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD MALHIOT,MATTHEW,E		10/23/2006	09:11:15	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000787

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		11/20/2006	07:34:50	8100.27		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000787

Agency Operator		Date	Time	Software		Source	
GULF BREEZE PD		12/13/2006	08:11:53	8100.27		Agency	Inspection
NEFF,STEPHEN,							
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	504601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/17/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO HACKNEY,DWITE,N		02/20/2006	09:49:52	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000788

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		03/22/2006	11:10:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		04/27/2006	12:26:59	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect. 20: RFI Detect.

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		05/25/2006	13:29:24	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007		
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.201
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		05/25/2006	14:02:23	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/27/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		06/19/2006	16:03:25	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

N/A

Inspection Tests

Instrument Serial #80-000788

Agency Operator	Date	Time	Software	Source	
ST LUCIE COUNTY SO SNOW,DAVID,J	07/25/2006	13:58:28	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SLOPE NOT MET LAST SAMPLE.

Inspection Tests

Instrument Serial #80-000788

Agency Operator	Date	Time	Software	Source	
ST LUCIE COUNTY SO SNOW,DAVID,J	07/25/2006	14:07:17	8100.26	Agency	Inspection

Sims 5 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance:RESIDUAL M/A.

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		07/25/2006	14:11:25	8100.26		Agency	Inspection
0.05 Lot#	20054C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		08/23/2006	18:03:59	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator	Date	Time	Software	Source	
ST LUCIE COUNTY SO SNOW,DAVID,J	09/27/2006	17:45:14	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A/F OK, M/A NEW SOLUTION.

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		09/27/2006	18:03:13	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO HACKNEY,DWITE,N		10/26/2006	13:26:10	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000788

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0016
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		10/26/2006	16:06:01	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	5354021
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	12/22/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		11/29/2006	18:57:55	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	535402I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	12/22/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000788

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		12/13/2006	18:37:15	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200605F	0.08 Gas Lot#	535402I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2008	Expiration	12/22/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD HACKNEY,DWITE,N		01/17/2006	14:23:09	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.203
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.203
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J		03/24/2006	10:18:37	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#		
Expiration	09/22/2007	Expiration	09/22/2007			
# Sims	4	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.052
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.080
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.026
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.023
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 08: Control Outside Tolerance, Purge Fail. Non-compliance:CONTOL OUTSIDE ALLOWANCE.

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	03/26/2006	04:55:47	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . INTERFERENT NOT DETECTED.

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		03/26/2006	05:25:50	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.174
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		04/30/2006	14:53:21	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.054
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.054
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J		05/31/2006	16:53:47	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#		
Expiration	09/22/2007	Expiration	09/22/2007			
# Sims	4	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.054
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.051
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.050
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Purge Fail, Ambient Fail. Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	05/31/2006	18:05:19	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. Int Det: RFI Detect. Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	05/31/2006	18:23:33	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.008
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.007
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Purge Fail, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD HARRELSON,THOMAS,M		05/31/2006	21:34:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		06/29/2006	11:58:55	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	07/31/2006	17:18:27	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		07/31/2006	17:22:55	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		08/29/2006	10:57:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		09/29/2006	10:25:05	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.168
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.177
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.183
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator		Date	Time	Software		Source	
VERO BEACH PD ADAMSKI,FRANK,J		09/29/2006	11:21:40	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	10/30/2006	17:52:04	8100.26	Agency	Inspection

# Sims	4	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	INT
Interferent Detect Test: Interferent Detect	Yes		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	10/30/2006	18:12:07	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #3	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , Range Exceeded. Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	10/30/2006	18:30:14	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N.

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ADAMSKI,FRANK,J	10/30/2006	18:31:14	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD JKL,DFS,F	12/05/2006	13:42:49	8100.27	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD GRDG,FDSFDS,T	12/11/2006	09:35:15	8100.27	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD ASDFHJKL,R4TYUIP,W	12/16/2006	08:09:57	8100.27	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD SCHOFIELD,DEBORAH,K	12/16/2006	17:27:00	8100.27	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source	
VERO BEACH PD FDS,FS,F	12/16/2006	17:27:42	8100.27	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000790

Agency Operator	Date	Time	Software	Source
VERO BEACH PD SCHOFIELD,DEB,K	12/16/2006	17:28:33	8100.27	Department Inspection

# Sims	0	Avg Std Dev	Complies	Yes
--------	---	-------------	----------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software	Source	
FT PIERCE PD HACKNEY,DWITE,N		02/20/2006	09:48:47	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot# 518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000791

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD		03/20/2006	10:03:54	8100.26		Agency	Inspection
CURRY,ROBERT,E							
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.183
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.192
		0.20 g/210L Test #2 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		04/16/2006	11:05:13	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		05/24/2006	03:16:34	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.180
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.191
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software	Source	
FT PIERCE PD CURRY,ROBERT,E		06/28/2006	02:12:20	8100.26	Agency	Inspection
0.05 Lot# Expiration	200509A 09/22/2007	0.08 Lot# Expiration	200509B 09/22/2007	0.20 Lot#		
# Sims	3	Complies Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		Interferent Detect Test #1	
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1 (Retry)	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.048
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #2 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #3	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #3	0.049
0.05 g/210L Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: RFI Detect. 08: RFI Detect, RFI Detect.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		06/28/2006	02:34:26	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.073
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.178
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.192
Diagnostic Check (Post-Inspection): OK	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Alcohol Free Test: OK	Yes	Interferent Detect Test #1	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2	0.076
		0.08 g/210L Test #2 (Retry)	0.078
		0.20 g/210L Test #2	0.186
		0.20 g/210L Test #2 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator	Date	Time	Software	Source	
FT PIERCE PD CURRY,ROBERT,E	07/25/2006	21:59:26	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD		07/25/2006	22:03:55	8100.26		Agency	Inspection
CURRY,ROBERT,E							
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A : . 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		08/20/2006	01:02:47	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		09/30/2006	01:43:52	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software	Source	
FT PIERCE PD HACKNEY,DWITE,N		10/26/2006	13:59:09	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.195
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8 (Retry)	0.000

Inspection Tests

Instrument Serial #80-000791

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		10/28/2006	01:27:54	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		11/24/2006	22:45:40	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000791

Agency Operator		Date	Time	Software		Source	
FT PIERCE PD CURRY,ROBERT,E		12/29/2006	22:40:17	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	502101G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	01/24/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO HACKNEY,DWITE,N		02/20/2006	09:49:30	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000792

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		03/22/2006	14:28:15	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		04/27/2006	13:59:51	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		05/25/2006	15:27:44	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		06/19/2006	18:21:21	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		07/25/2006	15:41:53	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		08/23/2006	18:46:02	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		09/27/2006	19:07:43	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software	Source	
ST LUCIE COUNTY SO HACKNEY,DWITE,N		10/26/2006	13:46:26	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0003	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.191
		0.08 g/210L Dry Gas Std Test #4	0.000
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.190
		0.08 g/210L Dry Gas Std Test #5	0.000
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.190
		0.08 g/210L Dry Gas Std Test #6	0.000
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000792

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.190
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.192
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.190
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.191
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

08: TUBE NOT CONNECTED.

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		10/26/2006	16:09:32	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		11/29/2006	19:05:29	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000792

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		12/13/2006	18:53:07	8100.27		Agency	Inspection
0.05 Lot#	2005O9A	0.08 Lot#	200509B	0.20 Lot#	20065F	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2008	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO HACKNEY,DWITE,N		02/20/2006	09:48:44	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000794

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		03/23/2006	09:06:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		04/27/2006	13:11:20	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO BRIGLIA,MATTHEW,K		05/25/2006	14:43:37	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		06/19/2006	15:13:58	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		07/25/2006	14:51:37	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	53542I
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	12/22/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

.200 LOW 1ST TEST; CHECKED SEAL AND CORRECTED PROBLEM

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		08/23/2006	17:25:35	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	535402I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	12/22/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		09/27/2006	18:21:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	535402I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	12/22/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO HACKNEY,DWITE,N		10/26/2006	13:15:06	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000794

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		10/26/2006	16:03:26	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		11/29/2006	18:51:19	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000794

Agency Operator		Date	Time	Software		Source	
ST LUCIE COUNTY SO SNOW,DAVID,J		12/13/2006	18:45:05	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200605F	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2008	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO MALHIOT,MATTHEW,E		03/24/2006	12:18:06	8100.26		Department Inspection	
0.05 Lot#	200411C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	504502I
Expiration	11/04/2006	Expiration	11/18/2006	Expiration	05/05/2007	Expiration	02/17/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.204
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000795

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000795

Agency Operator	Date	Time	Software	Source	
GULF COUNTY SO NEWSOME,JAMES,A	03/27/2006	15:02:37	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		03/27/2006	15:10:36	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007		
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.177
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.077
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.198
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		03/27/2006	16:19:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4281021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		04/26/2006	21:29:30	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4281021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator	Date	Time	Software	Source	
GULF COUNTY SO NEWSOME,JAMES,A	05/25/2006	22:05:57	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000795

Agency Operator	Date	Time	Software	Source	
GULF COUNTY SO NEWSOME,JAMES,A	05/25/2006	22:07:52	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Ambient Fail. Non-compliance: .

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		05/25/2006	22:25:20	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4281021
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		06/28/2006	22:08:24	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		07/27/2006	21:39:02	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	428102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		08/28/2006	23:13:02	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	428102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Time-Date changed. Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		09/30/2006	21:24:03	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	626101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/20/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

00: Ambient Fail. Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		10/29/2006	21:52:43	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	626101I
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/20/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO MALHIOT,MATTHEW,E		11/03/2006	15:42:21	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	626101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/20/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.083
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.083
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.082
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.083
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000795

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.083
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.083
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.083
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		11/30/2006	21:33:45	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	626101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/20/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000795

Agency Operator		Date	Time	Software		Source	
GULF COUNTY SO NEWSOME,JAMES,A		12/30/2006	18:51:36	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	626101I
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	09/20/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD VENTURI,GEORGE,L		03/02/2006	15:32:57	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.076
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.076
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.076
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.076
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		03/22/2006	15:58:02	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E	04/28/2006	19:37:58	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E	04/28/2006	19:47:55	8100.26	Agency	Inspection
0.05 Lot#	200509A				
Expiration	09/22/2007				
# Sims	4		Complies No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		04/28/2006	20:34:08	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.184
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.054
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.186
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.054
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.187
		0.20 g/210L Test #3 (Retry)	0.202
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E	05/23/2006	20:43:44	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

OPERATOR ERROR

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		05/23/2006	20:56:04	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.201
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.054
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.204
		0.08 g/210L Dry Gas Std Test #2	0.076
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.054
		0.08 g/210L Test #3	0.079
		0.08 g/210L Test #3 (Retry)	0.081
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.076
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 08: Control Outside Tolerance.SIMULATOR TOP NOT TIGHTENED PROPERLY

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		06/25/2006	14:18:00	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.200
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.044
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.194
		0.20 g/210L Test #2 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.20 g/210L Test #3 (Retry)	0.203
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.SIMULATOR HAD LEAK,TOP NOT TIGHTENED ENOUGH.

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E	07/17/2006	16:56:44	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		07/17/2006	17:06:14	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		08/05/2006	04:21:49	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		09/04/2006	08:04:45	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		10/23/2006	17:36:11	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	200509B
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.055
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.187
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.055
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD VENTURI,GEORGE,L		11/17/2006	19:20:17	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	512601I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/10/2007
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.195
		0.08 g/210L Dry Gas Std Test #4	0.075
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.075
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.075
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.195
		0.08 g/210L Dry Gas Std Test #7	0.075
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.075
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

--

Inspection Tests

Instrument Serial #80-000796

Agency Operator	Date	Time	Software	Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E	11/18/2006	11:05:21	8100.27	Agency	Inspection
0.05 Lot#	200509A				
Expiration	09/22/2007				
# Sims	3		Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.059
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.042
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.060
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.060
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.046
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		11/18/2006	11:56:31	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000796

Agency Operator		Date	Time	Software		Source	
LONGBOAT KEY PD SHOEMAKER,FRANK,E		12/02/2006	14:51:34	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	512601I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

HEADSPACE IN WETBATH SIMULATOR.

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source
HAMILTON COUNTY S.O. MALHIOT,MATTHEW,E	01/26/2006	09:54:50	8100.26	Department Inspection
0.05 Lot# 200411C Expiration 11/04/2006	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200411D Expiration 04/11/2006	0.08 Gas Lot# 419101I Expiration 07/14/2006	
# Sims 5	Avg Std Dev 0.0008	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

MOUTh ALC. IMPROPERLY DONE.

Inspection Tests

Instrument Serial #80-000797

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		05/25/2006	10:25:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	419101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R	06/27/2006	13:22:10	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.038
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.041
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.040
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.041
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.040
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R	06/27/2006	13:51:46	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.032
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.042
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.034
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.041
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.035
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside HOSE KINKED. Non-compliance: .

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R	06/27/2006	14:17:39	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.039
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.043
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.042
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.043
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.042
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R	06/28/2006	13:26:34	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
0.08 Gas Lot#	419101I	Expiration	07/14/2006		
# Sims	5	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.044
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: WRONG SIMULATOR.

Inspection Tests

Instrument Serial #80-000797

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		07/28/2006	08:58:31	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	6172021
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000797

Agency Operator		Date	Time	Software		Source	
HAMILTON COUNTY S.O. ALLEN,PAMELA,R		08/28/2006	10:24:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	6172021
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source	
HAMILTON COUNTY S.O. MALHIOT,MATTHEW,E	10/17/2006	12:07:09	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source
HAMILTON COUNTY S.O. MALHIOT,MATTHEW,E	10/17/2006	14:20:37	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 6172021 Expiration 06/23/2008	
# Sims 5	Avg Std Dev 0.0004	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000797

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		02/14/2006	10:47:58	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.203
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000808

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	03/22/2006	12:36:31	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		03/27/2006	14:42:15	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.182
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: LOOSE CONECTION

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		04/07/2006	10:07:42	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		05/26/2006	12:49:35	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	516002I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

TUBING DISCONNECTED AT SIMULATOR.

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		06/14/2006	10:04:02	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		07/20/2006	08:58:01	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		08/14/2006	11:05:20	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

TUBING DISCONNECTED AT SIMULATOR PORT.

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		09/19/2006	12:21:57	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		10/18/2006	11:16:38	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000808

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		10/18/2006	13:12:52	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.013
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.201
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.195
		0.20 g/210L Test #2 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tol. DIST.WATER CHANGED.. 20: LOOSE CONOutside Tolerance.

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		11/13/2006	12:30:57	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		11/22/2006	11:30:18	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000808

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		12/21/2006	09:41:54	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612210G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software	Source	
DIXIE COUNTY SO SUERETH,DONALD,P		03/06/2006	12:46:54	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000809

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.078
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	INT
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

A F / M A: RFI Detect. 08: Interferent Detect.

Inspection Tests

Instrument Serial #80-000809

Agency Operator	Date	Time	Software	Source
DIXIE COUNTY SO MANDOLA,MATTHEW,	03/29/2006	11:58:34	8100.26	Agency Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200504D Expiration 04/21/2007	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 4191011 Expiration 07/14/2006	
# Sims 2	Complies Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.050
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #2	0.079
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance. 08: Control Outside Tolerance.MADE SEVERAL MISTAKE FOR 1ST TEST

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		04/28/2006	10:42:35	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4191011
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		05/22/2006	09:10:41	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4191011
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000809

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	06/22/2006	11:15:39	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	
Expiration	09/22/2007	Expiration	04/21/2007		
# Sims	2		Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.071
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.073
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		06/22/2006	13:16:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		07/24/2006	10:53:11	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		08/21/2006	08:39:40	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		09/27/2006	09:14:32	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.184
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK.

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		10/30/2006	08:02:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NO

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software	Source	
DIXIE COUNTY SO SUERETH,DONALD,P		10/31/2006	13:57:42	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000809

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		10/31/2006	14:47:58	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		11/27/2006	10:17:23	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000809

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		12/29/2006	09:17:32	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615802I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK.

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	04/03/2006	09:44:26	8100.26	Agency	Inspection

Sims 2 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	04/03/2006	10:28:59	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

CHECKINGNon-compliance: .

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	04/03/2006	10:34:41	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		04/03/2006	11:08:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4191011
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.181
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.189
		0.20 g/210L Test #2 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.N

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		04/28/2006	08:20:32	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4191011
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		05/22/2006	11:04:52	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	4191011
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.184
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.196
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.193
		0.20 g/210L Test #2 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	06/22/2006	08:30:36	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Y. Non-compliance: .

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	06/22/2006	08:47:48	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		06/22/2006	08:59:04	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.196
		0.08 g/210L Dry Gas Std Test #2	0.000
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 08: Control Outside Tolerance.AIR LEAK

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,	06/22/2006	10:22:47	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		07/24/2006	08:52:12	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

CHECK

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		08/21/2006	10:56:41	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
		0.20 g/210L Test #2	0.196
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK. AIR LEAK.

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		09/27/2006	07:08:49	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		10/30/2006	09:53:18	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

NO

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO SUERETH,DONALD,P		10/31/2006	13:39:56	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000810

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		10/31/2006	14:37:20	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		11/27/2006	08:37:59	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000810

Agency Operator		Date	Time	Software		Source	
DIXIE COUNTY SO MANDOLA,MATTHEW,		12/29/2006	11:17:58	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615802I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/09/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK.

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source
PERRY PD MALHIOT,MATTHEW,E	02/02/2006	09:36:44	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200411G Expiration 11/18/2006	0.20 Lot# 200411D Expiration 11/04/2006	0.08 Gas Lot# 428101 Expiration 10/11/2006	
# Sims 5	Avg Std Dev 0.0007	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4 (Retry)	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5 (Retry)	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6 (Retry)	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7 (Retry)	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source
Alcohol Free Test #8				0.000
0.05 g/210L Test #8 (Retry)				0.050
0.08 g/210L Test #8				0.079
0.20 g/210L Test #8				0.199
0.08 g/210L Dry Gas Std Test #8				0.078
Interferent Detect Test #8				INT
Alcohol Free Test #9				0.000
0.05 g/210L Test #9 (Retry)				0.050
0.08 g/210L Test #9				0.079
0.20 g/210L Test #9				0.199
0.08 g/210L Dry Gas Std Test #9				0.077
Interferent Detect Test #9				INT
Alcohol Free Test #10				0.000
0.05 g/210L Test #10 (Retry)				0.050
0.08 g/210L Test #10				0.079
0.20 g/210L Test #10				0.199
0.08 g/210L Dry Gas Std Test #10				0.077
Interferent Detect Test #10				INT
0.05 g/210L Test (Std Dev)				0.0007
0.08 g/210L Test (Std Dev)				0.0006
0.20 g/210L Test (Std Dev)				0.0013
0.08 g/210L Dry Gas Std Test (Std Dev)				0.0005

Remarks:

05: RFI Detect (CELL PHONE).

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source	
PERRY PD PARKER,C,	03/27/2006	11:47:21	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source	
PERRY PD PARKER,C,	03/27/2006	12:00:26	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		03/27/2006	12:14:29	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	200509B
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD FRANKLIN,MICHAEL,E		04/30/2006	15:55:32	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	419101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/14/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,CLA,		05/30/2006	10:50:24	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		06/30/2006	13:27:04	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		07/28/2006	12:09:13	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD GRAY,LIN,		08/21/2006	00:50:01	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		09/28/2006	10:29:44	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	2006C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source
PERRY PD MALHIOT,MATTHEW,E	10/30/2006	13:56:46	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 612201G Expiration 05/05/2008	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD PARKER,C,		10/31/2006	10:19:50	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200605C	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	05/04/2008	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000811

Agency Operator		Date	Time	Software		Source	
PERRY PD GRAY,LIN,		11/27/2006	05:22:22	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	612201G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/05/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.183
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000811

Agency Operator	Date	Time	Software	Source	
PERRY PD PARKER,C,	12/28/2006	15:15:07	8100.27	Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008
0.08 Gas Lot#	612201G	Expiration	05/05/2008		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO SUERETH,DONALD,P		02/16/2006	10:02:50	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.046
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/24/2006	15:29:37	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/24/2006	15:36:35	8100.26	Agency	Inspection

0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#
Expiration	09/22/2007	Expiration	09/22/2007	

# Sims	4	Complies	No
---------------	---	-----------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.072
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.058
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.062
0.08 g/210L Test: OK	No	Interferent Detect Test #2	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.071
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/24/2006	16:13:41	8100.26	Agency	Inspection

0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#
Expiration	09/22/2007	Expiration	09/22/2007	

# Sims	4	Complies	No
---------------	---	-----------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.071
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.076
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/24/2006	16:41:48	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/24/2006	16:49:24	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.009
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.043
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/24/2006	17:28:50	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.027
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	No	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.036
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.049
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.049
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Purge Fail. 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		03/24/2006	18:17:21	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	03/26/2006	11:27:17	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	04/23/2006	18:02:09	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	04/23/2006	18:24:32	8100.26	Agency	Inspection

0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#
Expiration	09/22/2007	Expiration	09/22/2007	

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.059
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.068
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.044
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.069
0.08 g/210L Test: OK	No	0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.069
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	04/23/2006	19:15:38	8100.26	Agency	Inspection

# Sims	4	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	04/23/2006	19:31:36	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
# Sims	4	Complies No			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.135
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.164
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.151
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.172
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.158
		0.20 g/210L Test #3 (Retry)	0.174
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		04/23/2006	20:23:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	05/25/2006	12:34:38	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		05/25/2006	12:53:24	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		06/20/2006	13:56:33	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A:NOT ENOUGH ALCOHOL USED..

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		07/27/2006	02:47:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		08/01/2006	12:55:30	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		09/20/2006	15:45:41	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

A F / M A : . Int Det:TIGHTENED SEAL.

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	10/12/2006	07:51:04	8100.26	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	10/12/2006	07:57:49	8100.26	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:INT NOT DETECTED, . Non-compliance: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		10/12/2006	08:17:07	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.179
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software	Source	
HERNANDO COUNTY SO SUERETH,DONALD,P		10/27/2006	14:36:55	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518602I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.201
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.203
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.197
		0.20 g/210L Test #4 (Retry)	0.203
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.199
		0.20 g/210L Test #5 (Retry)	0.203
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.20 g/210L Test #6 (Retry)	0.204
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000812

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.199
			0.20 g/210L Test #7 (Retry)	0.204
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.20 g/210L Test #8 (Retry)	0.204
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.20 g/210L Test #9 (Retry)	0.205
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.20 g/210L Test #10 (Retry)	0.204
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0040
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

20: OUTSIDE TOL. 1ST ANALYSIS DUE TO A/C COOLING TUBING. . INSTRUMENT MOVED FROM DIRECT CONTACT WITH COOL AIR FROM A/C

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		10/27/2006	15:17:01	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		11/25/2006	05:20:17	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	609301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/05/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.053
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000812

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		12/29/2006	23:00:06	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	630702I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	11/06/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software	Source	
MARION COUNTY SO SUERETH,DONALD,P		03/21/2006	09:37:27	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.076
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	INT
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049

Inspection Tests

Instrument Serial #80-000813

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.196
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

08: Interferent Detect. RM. FANNED.

Inspection Tests

Instrument Serial #80-000813

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO FORTE,MICHAEL,F	03/21/2006	18:52:54	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007
0.08 Gas Lot#	502101G	Expiration	01/24/2007		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		04/04/2006	04:27:26	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	519901I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO SAVARESE,K,P	05/28/2006	05:46:41	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		05/28/2006	06:00:55	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

NONE

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		06/25/2006	11:18:53	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SECKEL,ERIC,J		07/24/2006	16:30:13	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SECKEL,ERIC,J		08/17/2006	11:28:10	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200605F	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/10/2008	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		09/26/2006	09:03:00	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SUERETH,DONALD,P		10/19/2006	20:37:15	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.195
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.195
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.195
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000813

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		10/19/2006	22:06:53	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200505C	0.08 Gas Lot#	619101I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/05/2007	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		11/26/2006	07:28:05	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200504D	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	05/04/2008	Expiration	04/21/2007	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000813

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,M,F		12/16/2006	03:28:32	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	619101I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO SUERETH,DONALD,P		02/15/2006	18:04:01	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000814

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		06/20/2006	13:55:02	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	07/27/2006	01:15:29	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		07/27/2006	01:23:53	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.052
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.083
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

A F / M A: . Int Det:3 ML NOT ADDED YET.

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		07/27/2006	02:09:38	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	07/27/2006	03:30:52	8100.26	Agency	Inspection

# Sims	0	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		08/01/2006	13:01:40	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		09/20/2006	16:00:28	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		10/12/2006	08:03:48	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A:SLOPE NOT MET.

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO SUERETH,DONALD,P		10/27/2006	14:57:01	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	09/22/2006	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000814

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		10/27/2006	15:33:26	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000814

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	11/25/2006	04:20:49	8100.27	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , Sequence Aborted. Non-compliance: .

Inspection Tests

Instrument Serial #80-000814

Agency Operator	Date	Time	Software	Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M	11/25/2006	04:51:37	8100.27	Agency	Inspection

0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#
Expiration	09/22/2007	Expiration	09/22/2007	

# Sims	3	Complies	No
---------------	---	-----------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.143
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.090
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.052
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2 (Retry)	0.084
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.052
0.20 g/210L Test: OK	No	0.08 g/210L Test #3 (Retry)	0.082
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT

Remarks:

08: Improper Sample, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		11/25/2006	05:30:46	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000814

Agency Operator		Date	Time	Software		Source	
HERNANDO COUNTY SO GREIFENBERGER,FRANCES,M		12/29/2006	22:25:58	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	518701I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SUERETH,DONALD,P		02/09/2006	14:06:44	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000815

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0010

Remarks:

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		03/27/2006	14:59:00	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.172
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	No	0.08 g/210L Test #2 (Retry)	0.080
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2	0.184
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.079
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.191
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance, Ambient Fail.
Non-compliance: .

Inspection Tests

Instrument Serial #80-000815

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO SEGREST,DARRELL,E	03/27/2006	16:47:25	8100.26	Agency	Inspection
0.05 Lot#	200509A				
Expiration	09/22/2007				
# Sims	4		Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.028
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.040
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.019
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2 (Retry)	0.045
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	No	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.007
0.20 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3 (Retry)	0.046
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

00: Control Outside Tolerance. Int Det: . 05: Improper Sample, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		03/28/2006	01:38:36	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0112005
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.181
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.193
		0.20 g/210L Test #2 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J		04/28/2006	13:42:12	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	0112005
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.173
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		05/10/2006	14:06:29	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	0112005
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

Inspection Tests

Instrument Serial #80-000815

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO SEGREST,DARRELL,E	06/29/2006	07:58:59	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	
Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies	No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.062
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.067
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.067
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.057
		0.08 g/210L Test #3 (Retry)	0.067
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		07/27/2006	09:01:37	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0150594
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/08/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.185
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.188
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		08/30/2006	19:48:21	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0150594
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/08/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.073
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.199
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.044
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
		0.20 g/210L Test #2	0.199
		0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.045
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		09/29/2006	09:26:29	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0150594
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/08/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		10/30/2006	21:55:09	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0150594
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/08/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SUERETH,DONALD,P		11/16/2006	18:15:27	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0009	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000815

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Test (Std Dev)	0.0012
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000815

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO SEGREST,DARRELL,E		12/25/2006	23:02:57	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	0150594
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	05/08/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000816

Agency Operator		Date	Time	Software	Source	
SUMTER COUNTY SO SUERETH,DONALD,P		02/09/2006	14:13:15	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.000
		0.20 g/210L Test #4 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.000
		0.20 g/210L Test #5 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #5	0.082
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.000
		0.20 g/210L Test #6 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.000
			0.20 g/210L Test #7 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #7	0.081
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.000
			0.20 g/210L Test #8 (Retry)	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.000
			0.20 g/210L Test #9 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.000
			0.20 g/210L Test #10 (Retry)	0.199
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

20:TOP HOSE DETACHED FROM SIMULATOR.

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J	03/27/2006	14:45:28	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J	03/27/2006	15:06:05	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.036
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.044
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.046
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . Non-compliance: .

Inspection Tests

Instrument Serial #80-000816

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J		03/27/2006	15:41:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0112005
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.161
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.087
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.084
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.188
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #2	0.087
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.085
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.085
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.085
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance. 08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J	04/28/2006	14:44:02	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.035
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.036
Barometric Pressure Sensor Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.012
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.042
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.05 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	Alcohol Free Test #3 (Retry)	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.046
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000816

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J		05/10/2006	14:21:10	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	0112005
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J	06/29/2006	08:01:20	8100.26	Agency	Inspection

Sims 5 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000816

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J		06/29/2006	08:24:25	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	0150594
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/08/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000816

Agency Operator		Date	Time	Software		Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J		07/27/2006	09:02:08	8100.26		Agency	Inspection
0.05 Lot#	2000509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.174
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.162
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.182
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.154
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.246
		0.20 g/210L Test #3 (Retry)	0.148
		Interferent Detect Test #3	INT

Remarks:

20: Purge Fail, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000816

Agency Operator		Date	Time	Software	Source	
SUMTER COUNTY SO SUERETH,DONALD,P		11/15/2006	21:19:38	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot# 518602I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000816

Agency Operator	Date	Time	Software	Source	
SUMTER COUNTY SO PREVATT,WILLIAM,J	11/16/2006	17:01:49	8100.27	Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	2000605F
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008
0.08 Gas Lot#		0150594			
Expiration		05/08/2008			
# Sims	5	Complies Yes			

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software	Source	
CITRUS COUNTY SO SUERETH,DONALD,P		02/15/2006	15:55:04	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0008	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.203
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000817

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0019
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		03/27/2006	11:54:28	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.063
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		04/27/2006	16:55:25	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.179
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.186
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:LINE OFF. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		05/29/2006	07:41:06	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		06/27/2006	06:35:17	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		07/30/2006	11:25:24	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	610301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/18/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		08/27/2006	10:26:13	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	610301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/18/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I	09/24/2006	08:20:41	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		09/24/2006	08:35:12	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007		
# Sims	3	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.176
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.187
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.181
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.190
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.183
		0.20 g/210L Test #3 (Retry)	0.190
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:HOSE WAS LOOSE ON .20 SIMILATOR.

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		09/24/2006	09:27:32	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	610301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/18/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.193
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.HOSE

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO SUERETH,DONALD,P		10/19/2006	11:03:42	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000817

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I	10/22/2006	13:09:43	8100.27	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007
0.08 Gas Lot#	610301G	Expiration	04/18/2008		
# Sims	3	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		11/27/2006	08:27:44	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	610301G
Expiration	05/04/2008	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/18/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000817

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO WASHBURN,ROBERT,I		12/30/2006	17:30:23	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200505C	0.08 Gas Lot#	610301G
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/05/2007	Expiration	04/18/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO SUERETH,DONALD,P		02/07/2006	20:24:05	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	03/29/2006	13:01:19	8100.26	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededRange ExceededAmbient Fail. Non-compliance: AMBIENT AIR FAIL - WILL VENTILATE AREA .

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	03/29/2006	13:15:03	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Interferent Detect.

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	03/29/2006	13:22:01	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	03/29/2006	13:40:58	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	03/29/2006	13:49:00	8100.26	Agency	Inspection

0.05 Lot# 200504C
Expiration 04/21/2007

Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.037
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.040
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.040
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, RFI Detect.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		03/29/2006	14:23:34	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		04/26/2006	14:20:34	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		05/24/2006	10:07:36	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007		
# Sims	3	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.048
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.183
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.185
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.048
		0.20 g/210L Test #3 (Retry)	0.188
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		05/24/2006	13:13:20	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		06/27/2006	10:45:02	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		07/18/2006	18:33:41	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	08/31/2006	08:06:37	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RINSED MOUTH W/ ALCOHOL AGAIN

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	08/31/2006	08:19:03	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ROOM AC OFF/WILL VENTILATE.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,		08/31/2006	08:56:56	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#		
Expiration	04/21/2007	Expiration	04/21/2007			
# Sims	3	Complies No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. UNKNOWN.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		08/31/2006	12:20:21	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		09/12/2006	20:20:10	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/27/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software	Source	
CITRUS COUNTY SO SUERETH,DONALD,P		10/19/2006	13:09:48	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source
--------------------	------	------	----------	--------

0.05 g/210L Test #8	0.048
0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.077
Interferent Detect Test #8	INT
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.077
Interferent Detect Test #9	INT
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.078
Interferent Detect Test #10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	10/25/2006	18:59:32	8100.27	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2 (Retry)	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3 (Retry)	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. TERMINATE INPSPECTION AND RESTART VENT RMNon-compliance: .

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	10/25/2006	19:06:04	8100.27	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. ABORTED - TANK NOT CONNECTED.

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	10/25/2006	19:16:15	8100.27	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Ambient Fail. Non-compliance:AMBIENT AIR FAILED.

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	10/30/2006	20:32:04	8100.27	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Interferent Detect. Non-compliance: UNKNOWN.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		10/30/2006	21:34:16	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.048
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.188
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: WRONG SIMULATOR - CORRECTED.

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		11/16/2006	14:45:30	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	11/16/2006	14:47:32	8100.27	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000819

Agency Operator	Date	Time	Software	Source	
CITRUS COUNTY SO KING,WAYNE,	12/29/2006	13:57:11	8100.27	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

MUST VENT ROOM AND RESTART. Non-compliance: .

Inspection Tests

Instrument Serial #80-000819

Agency Operator		Date	Time	Software		Source	
CITRUS COUNTY SO KING,WAYNE,		12/29/2006	17:44:12	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO SKIPPER,ROGER,G		02/01/2006	10:05:02	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000822

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO TAYLOR,JEFF,W		03/28/2006	13:51:13	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	05/05/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.189
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.078
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.198
		Interferent Detect Test #3	INT

Remarks:

A F / M A: . 20: RFI Detect, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO TAYLOR,JEFF,W		03/31/2006	16:20:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO TAYLOR,JEFFREY,W		04/28/2006	05:09:50	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.185
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.040
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.190
		0.20 g/210L Test #2 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO TAYLOR,JEFFREY,W		05/24/2006	15:16:38	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO MCKANE,THOMAS,D		06/21/2006	10:30:58	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	11/04/2006	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO CHESSHER,SANDRA,W		07/17/2006	12:23:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000822

Agency Operator	Date	Time	Software	Source	
LAKE SO CHESSHER,SANDRA,W	08/08/2006	15:53:17	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ALREAY CHECKED

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO CHESSHER,SANDRA,W		08/08/2006	15:54:54	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO CHESSHER,SANDRA,W		09/28/2006	12:52:36	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	515902I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION

Inspection Tests

Instrument Serial #80-000822

Agency Operator	Date	Time	Software	Source	
LAKE SO CHESSHER,SANDRA,W	10/25/2006	07:04:04	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AGENCY INSPECTIONNon-compliance: .

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE SO CHESSHER,SANDRA,W		10/25/2006	07:06:08	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software	Source	
LAKE S.O. SKIPPER,ROGER,G		10/27/2006	10:43:22	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 615802I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 06/09/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.066
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1 (Retry)	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.078
		0.08 g/210L Test #4 (Retry)	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.078
		0.08 g/210L Test #5 (Retry)	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.08 g/210L Test #6 (Retry)	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000822

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.079
			0.08 g/210L Test #7 (Retry)	0.080
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.08 g/210L Test #8 (Retry)	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.08 g/210L Test #9 (Retry)	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.08 g/210L Test #10 (Retry)	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0040
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

08: Control Outside Tolerance SIM HOSE LOOSE. REDONE..

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE S.O. CHESSHER,SANDRA,W		11/08/2006	15:17:30	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION

Inspection Tests

Instrument Serial #80-000822

Agency Operator		Date	Time	Software		Source	
LAKE S.O. MCKANE,THOMAS,D		12/27/2006	01:47:45	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515902I
Expiration	05/04/2008	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE SO SKIPPER,ROGER,G		02/01/2006	10:03:35	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0007	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000823

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE SO TAYLOR,JEFF,W		03/28/2006	13:38:04	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	01/04/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000823

Agency Operator	Date	Time	Software	Source	
LAKE SO TAYLOR,JEFFREY,W	04/28/2006	05:07:03	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1	0.020
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1 (Retry)	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #2 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #3 (Retry)	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No	Interferent Detect Test #3 (Retry)	0.019
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. Int Det: , . Non-compliance: .

Inspection Tests

Instrument Serial #80-000823

Agency Operator	Date	Time	Software	Source	
LAKE SO TAYLOR,JEFFREY,W	04/28/2006	05:47:23	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.007
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #3	0.000
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Control Outside Tolerance, Ambient Fail. Non-compliance: .

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software	Source	
LAKE S.O. SKIPPER,ROGER,G		08/02/2006	13:11:03	8100.26	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot# 618801I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration 07/18/2008
# Sims	5	Avg Std Dev	0.0003	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000823

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.200
			0.08 g/210L Dry Gas Std Test #7	0.079
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0333
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

08: DRY GAS HOSE NOT CONNECTED, REPEATED. INSP AFTER REPAIR.

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE S.O. MCKANE,THOMAS,D		08/08/2006	15:54:49	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/04/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSPECTION

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE S.O. MCKANE,THOMAS,D		09/28/2006	12:52:21	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	436401I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	01/04/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE S.O. CHESSHER,SANDRA,W		10/25/2006	07:04:53	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/04/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE S.O. SKIPPER,ROGER,G		10/27/2006	10:56:22	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000823

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE S.O. MCKANE,THOMAS,D		11/08/2006	15:22:05	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	436401I
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/04/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION

Inspection Tests

Instrument Serial #80-000823

Agency Operator		Date	Time	Software		Source	
LAKE S.O. MCKANE,THOMAS,D		12/27/2006	01:44:25	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	436401I
Expiration	05/04/2008	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/04/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD SKIPPER,ROGER,G		02/15/2006	11:39:09	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000824

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator	Date	Time	Software	Source	
LEESBURG PD GRAY,SCOTT,G	03/27/2006	12:26:22	8100.26	Agency	Inspection

# Sims	3	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

REDO.

Inspection Tests

Instrument Serial #80-000824

Agency Operator	Date	Time	Software	Source	
LEESBURG PD GRAY,SCOTT,G	03/27/2006	12:33:06	8100.26	Agency	Inspection

Sims 3 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. REDO.

Inspection Tests

Instrument Serial #80-000824

Agency Operator	Date	Time	Software	Source	
LEESBURG PD GRAY,SCOTT,G	03/27/2006	12:37:01	8100.26	Agency	Inspection

Sims 3 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		03/27/2006	12:46:07	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		04/30/2006	19:21:28	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	50201G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator	Date	Time	Software	Source	
LEESBURG PD GRAY,SCOTT,G	05/30/2006	17:36:54	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.020
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.047
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.048
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

CONT. WAS NOT LOCKED IN PLACE, RFI Detect. 1-CONT NOT CONNECTED CORRECTLY 2-RFI.

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		05/30/2006	19:52:24	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.176
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

LOOSE CONNECTION.

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		06/27/2006	09:12:34	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: SIM. WAS NOT HOOKED UP ON FIRST ATTEMPT .

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		07/28/2006	07:14:27	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509A	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		08/27/2006	17:07:39	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG PD GRAY,SCOTT,G		09/29/2006	21:30:51	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software	Source	
LEESBURG P.D. SKIPPER,ROGER,G		10/26/2006	14:38:32	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 618801I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/10/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000824

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG P.D. GRAY,SCOTT,G		10/28/2006	09:08:35	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502701G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	02/02/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PASS

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG P.D. GRAY,SCOTT,G		11/29/2006	14:04:09	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000824

Agency Operator		Date	Time	Software		Source	
LEESBURG P.D. GRAY,SCOTT,G		12/17/2006	20:12:17	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA PD SKIPPER,ROGER,G		03/16/2006	10:11:59	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000825

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA PD WADE,WILLIAM,E		06/27/2006	21:34:36	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA PD		07/25/2006	20:07:36	8100.26		Agency	Inspection
WADE,WILIAM,E							
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA PD WADE,WILLIAM,E		08/30/2006	09:33:53	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA PD WADE,WILLIAM,E		09/24/2006	10:42:40	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA PD WADE,WILLIAM,E		10/26/2006	20:07:17	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA P.D. SKIPPER,ROGER,G		10/27/2006	13:48:44	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	615802I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/09/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000825

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA P.D. WADE,WILLIAM,E		11/29/2006	04:59:19	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000825

Agency Operator		Date	Time	Software		Source	
MOUNT DORA P.D. WADE,WILLIAM,E		12/25/2006	16:55:18	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SUERETH,DONALD,P		03/21/2006	09:10:25	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		03/21/2006	17:45:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot#	512501I
Expiration	04/21/2007	Expiration	11/19/2006	Expiration	05/05/2007	Expiration	05/10/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

CHECK FOR ACCURACY

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO FORTE,MICHAEL,F	04/04/2006	04:30:01	8100.26	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007
0.08 Gas Lot#	512501I	Expiration	05/10/2007		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO SAVARESE,K,P	05/28/2006	05:47:13	8100.26	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		05/28/2006	05:54:06	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515901I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		06/25/2006	11:19:10	8100.26		Agency	Inspection
0.05 Lot# Expiration	200504C 04/21/2007	0.08 Lot# Expiration	200504D 04/21/2007	0.20 Lot# Expiration	200505C 05/05/2007	0.08 Gas Lot# Expiration	515901I 06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SECKEL,ERIC,J		07/24/2006	16:38:17	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	515901I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SECKEL,ERIC,J		08/17/2006	11:43:13	8100.26		Agency	Inspection
0.05 Lot# Expiration	200504C 04/21/2007	0.08 Lot# Expiration	200504D 04/21/2007	0.20 Lot# Expiration	200605F 05/10/2008	0.08 Gas Lot# Expiration	515901I 06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO FORTE,MICHAEL,F		09/26/2006	09:18:51	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	619101I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/12/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: LOOSE SEAL ON SIM.

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SUERETH,DONALD,P		10/19/2006	20:55:20	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO FORTE,MICHAEL,F	10/19/2006	22:13:43	8100.27	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200505C
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/05/2007
0.08 Gas Lot#	515901I	Expiration	06/13/2007		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000827

Agency Operator		Date	Time	Software		Source	
MARION COUNTY SO SAVARESE,K,P		11/26/2006	07:27:31	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200504D	0.20 Lot#	200605F	0.08 Gas Lot#	515901I
Expiration	05/04/2008	Expiration	04/21/2007	Expiration	05/10/2008	Expiration	06/13/2007
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO FORTE,MICHAEL,F	12/16/2006	01:55:29	8100.27	Agency	Inspection

Sims 4 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

Inspection Tests

Instrument Serial #80-000827

Agency Operator	Date	Time	Software	Source	
MARION COUNTY SO FORTE,M,F	12/16/2006	02:37:58	8100.27	Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008
0.08 Gas Lot#	515901I	Expiration	06/13/2007		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software	Source	
HILLSBOROUGH SO SUERETH,DONALD,P		02/14/2006	10:52:01	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot# 518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.052
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.000
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.000
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.000
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000829

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.053
			0.08 g/210L Test #7	0.082
			0.20 g/210L Test #7	0.202
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.081
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: GAS STD HOSE DETACHED FROM INSTRUMENT.

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		03/27/2006	13:25:18	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

CASE IN GOOD AND CLEAN CONDITION

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		04/04/2006	10:24:22	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		04/17/2006	10:56:24	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		04/20/2006	12:14:49	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		05/29/2006	11:07:14	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		06/14/2006	13:04:16	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		07/18/2006	11:46:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		08/14/2006	09:46:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		09/19/2006	10:57:06	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software	Source	
HILLSBOROUGH SO SUERETH,DONALD,P		10/18/2006	12:09:40	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.077
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.076
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	INT
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000829

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

08: Interferent Detect. ROOM AIRED OUT..

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		10/18/2006	14:07:23	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.198
		Interferent Detect Test #3	INT

Remarks:

20: LOOSE CONNECTION, SOLUTION CHANGED. Non-compliance: .

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		10/18/2006	14:49:14	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		11/13/2006	10:17:12	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		11/21/2006	10:22:02	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	518602I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/07/2007
# Sims	4	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.082
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.081
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000829

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		11/22/2006	10:44:42	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000829

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		12/19/2006	13:06:46	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		02/14/2006	10:28:07	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000830

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		03/22/2006	10:38:44	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

TUBING CONNECTION OFF.

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		03/27/2006	09:54:56	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007		
# Sims	4	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.206
		Interferent Detect Test #3	INT

Remarks:

SIMULATOR CONNECTION LOOSE, RFI Detect. RFI DETECTED UNABLE TO FINISH.

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		03/29/2006	12:29:28	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		04/06/2006	12:51:06	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.078
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.078
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

ATTACHED TO WRONG SIMULATOR.

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		05/12/2006	23:35:15	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT

Remarks:

TUBING BECAME DETACHED, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		05/16/2006	10:24:48	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	531505G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	11/15/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tol. SIM. RETURN HOSE LEAK DISCOVERED

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		06/09/2006	12:39:19	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518502I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		07/18/2006	12:47:42	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		08/11/2006	10:36:40	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		09/15/2006	11:41:43	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #2	0.077
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

SIMULATOR NOT SEALED TIGH. GAS CYLENDER NOT TIGHT.

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		10/18/2006	12:25:49	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0008	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.195
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000830

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

A F / M A:CONTAMINATED FROM PREVIOUS TEST.

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		10/18/2006	15:32:00	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/3007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.072
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
		0.20 g/210L Test #2	0.197
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance LOOSE CONNECTION. 08: Control Outside Tolerance LOOSE CONNECTION.

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		11/10/2006	10:18:05	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000830

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		12/19/2006	10:45:54	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HACKNEY,DWITE,N	02/22/2006	09:53:27	8100.26	Department Inspection

0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

<u>Check or Test</u>	<u>Result</u>	<u>Test</u>	<u>Result</u>
----------------------	---------------	-------------	---------------

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		03/26/2006	10:26:42	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

PASSED

Inspection Tests

Instrument Serial #80-000831

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO DECKARD,JOE,B	04/27/2006	09:40:40	8100.26	Agency Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#
Expiration	09/22/2007	Expiration	09/22/2007	
# Sims	5	Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:SIMULATOR SEAL FAILURE.

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		04/27/2006	10:42:40	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance HOSE DISCONNECTED.

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		05/30/2006	11:07:29	8100.26		Agency	Inspection
0.05 Lot#	20059A	0.08 Lot#	20059B	0.20 Lot#	20059C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		06/27/2006	10:02:29	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO HACKNEY,DWITE,N		09/05/2006	14:55:29	8100.26		Department Inspection	
0.05 Lot#	N	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.203
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.203
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.203
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000831

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		09/26/2006	10:24:36	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software	Source	
MARTIN COUNTY SO HACKNEY,DWITE,N		10/25/2006	11:51:28	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000831

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		10/25/2006	12:38:13	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000831

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		11/29/2006	22:43:40	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO HACKNEY,DWITE,N		02/22/2006	09:50:26	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		03/26/2006	10:27:40	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	502101G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

PASSED

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		04/27/2006	09:44:51	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		05/30/2006	11:07:20	8100.26		Agency	Inspection
0.05 Lot#	20059A	0.08 Lot#	20059B	0.20 Lot#	20059C	0.08 Gas Lot#	
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	5	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.188
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.187
		0.20 g/210L Test #3 (Retry)	0.192
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance POWER TO SIM OFF, Control Out. Non-compliance:SIMULATOR POWER OFF.

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		05/30/2006	12:20:09	8100.26		Agency	Inspection
0.05 Lot#	20059A	0.08 Lot#	20059B	0.20 Lot#	20059C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		06/27/2006	10:06:06	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	502101G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	01/24/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		07/28/2006	10:09:55	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		08/29/2006	09:17:25	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	07/08/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		09/26/2006	10:33:10	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HOPE,JEAN,J	10/11/2006	08:48:39	8100.26	Department Inspection

# Sims	0	Avg Std Dev	Complies	Yes
--------	---	-------------	----------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO ERTY',FGH,F	10/11/2006	08:50:35	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO QWERT,WERTY,E	10/11/2006	09:03:32	8100.26	Department Inspection

# Sims	0	Avg Std Dev	Complies	Yes
--------	---	-------------	----------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO ERWER,EWRWER,N	10/11/2006	09:04:32	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software	Source	
MARTIN COUNTY SO HACKNEY,DWITE,N		10/25/2006	12:21:37	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 07/08/2007
# Sims	5	Avg Std Dev	0.0003	Complies	No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.000
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.000
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.000
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.000
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.000
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.000
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.000
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.051
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.000
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.000
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.000
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.000
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

08: TUBE NOT CONNECTED, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO HACKNEY,DWITE,N		10/25/2006	13:48:08	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000832

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		10/25/2006	14:34:55	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		11/29/2006	19:44:18	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect RADIO.

Inspection Tests

Instrument Serial #80-000832

Agency Operator		Date	Time	Software		Source	
MARTIN COUNTY SO DECKARD,JOE,B		12/28/2006	21:52:36	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	615301I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		02/14/2006	10:33:43	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	518602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000833

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		03/24/2006	12:31:59	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415601I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

CASE IN GOOD AND CLEA CONDITION

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		04/03/2006	11:17:54	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	415602I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	06/10/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		05/26/2006	13:54:17	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		06/14/2006	11:06:10	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		07/20/2006	09:49:00	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.180
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TUBING LOOSE AT SIMULATOR PORT.

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		08/14/2006	12:25:33	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	516002I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	06/13/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TUBING DISCONNECTED AT SIMULATOR PORT.

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		09/19/2006	13:20:43	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO SUERETH,DONALD,P		10/18/2006	10:50:37	8100.27		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000833

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO OWENS,EDWARD,W		10/18/2006	11:59:56	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518602I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/07/2007
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		11/13/2006	11:17:11	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000833

Agency Operator		Date	Time	Software		Source	
HILLSBOROUGH SO GRAY,BEVERLY,A		12/19/2006	11:35:16	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA PD SKIPPER,ROGER,G		02/14/2006	10:41:13	8100.26		Department Inspection	
0.05 Lot#	2005059A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518701I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.053
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.053
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.053
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000836

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.053
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA PD SEELEY,DAVID,A		05/18/2006	10:44:03	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	510902I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/21/2007
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA PD SEELEY,DAVID,A		09/30/2006	08:13:24	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	510902I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/21/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA PD SEELEY,DAVID,A		10/15/2006	08:47:21	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007		
# Sims	3	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.188
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.187
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.188
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.188
		0.20 g/210L Test #3 (Retry)	0.189
		Interferent Detect Test #3	INT

Remarks:

20, Control Outside Tolerance. .20 NOT MEETING MIN. REQUIREMENTS.

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA PD SEELEY,DAVID,A		10/27/2006	09:12:48	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	510902I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	04/21/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

REINSPECTED INTOX 8000 DUE TO COOL A/C TEMP COOLING SIM

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA P.D. SKIPPER,ROGER,G		11/09/2006	14:23:00	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	627201I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/03/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.053
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.053
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.053
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000836

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.053
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.053
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.053
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA P.D. SEELEY,DAVID,A		11/24/2006	10:06:26	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	510902I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/21/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000836

Agency Operator		Date	Time	Software		Source	
UMATILLA P.D. SEELEY,DAVID,A		12/19/2006	11:53:48	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	510902I
Expiration	04/21/2007	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	04/21/2007
# Sims	3	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

DISREGUARD HIT BUTTON BY MISTAKE

Inspection Tests

Instrument Serial #80-000838

Agency Operator	Date	Time	Software	Source
PANAMA CITY BEACH PD MALHIOT,MATTHEW,E	02/22/2006	09:37:07	8100.26	Department Inspection
0.05 Lot# 200504C Expiration 04/21/2007	0.08 Lot# 200504D Expiration 04/21/2007	0.20 Lot# 200505C Expiration 05/05/2007	0.08 Gas Lot# 5252021 Expiration 09/14/2007	
# Sims 5	Avg Std Dev 0.0005	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.076
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.077
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.077
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.076
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.077
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000838

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		04/29/2006	07:39:04	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		05/30/2006	07:19:23	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:POSS SIM AIR LEAK, CHANGED GASKET/CHANGED SOLUTION . POSS SIM AIR LEAK, CHANGED GASKET, USED NEW SOLUTION

Inspection Tests

Instrument Serial #80-000838

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	06/16/2006	07:53:04	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	
Expiration	09/22/2007	Expiration	09/22/2007		
# Sims	2	Complies	No		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.071
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.075
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.071
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.071
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:UNKNOWN MAY BE SOLUTION,WILL REDO INSPECTION.

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		06/16/2006	08:57:34	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD		07/21/2006	18:48:13	8100.26		Agency	Inspection
GAILFOIL,MICHAEL,L							
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	08/27/2006	23:55:47	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
0.08 Gas Lot#	610201G	Expiration	04/17/2008		
# Sims	2	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		09/26/2006	16:28:03	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

Inspection Tests

Instrument Serial #80-000838

Agency Operator	Date	Time	Software	Source
PANAMA CITY BEACH PD MALHIOT,MATTHEW,E	10/26/2006	12:17:22	8100.27	Department Inspection
0.05 Lot# 200509A Expiration 09/22/2007	0.08 Lot# 200509B Expiration 09/22/2007	0.20 Lot# 200509C Expiration 09/22/2007	0.08 Gas Lot# 610201G Expiration 04/17/2008	
# Sims 5	Avg Std Dev 0.0003	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000838

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		10/30/2006	12:47:31	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

AF/MA:TRYING NEW MA SOLUTION. SOLUTION USED 1ST DISCARDED

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		11/17/2006	17:07:49	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000838

Agency Operator		Date	Time	Software		Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L		12/01/2006	19:09:16	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

AF/MA:INADEQUATE SAMPLE.

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software	Source	
LYNN HAVEN PD MALHIOT,MATTHEW,E		04/04/2006	10:19:03	8100.26	Department Inspection	
0.05 Lot#	200411C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 434502I
Expiration	11/04/2006	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 12/14/2006
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.078
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source	
LYNN HAVEN PD SMITH,JAMES,C	05/13/2006	00:11:01	8100.26	Agency	Inspection

Sims 2 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source	
LYNN HAVEN PD SMITH,JAMES,C	05/13/2006	00:58:03	8100.26	Agency	Inspection

Sims 0 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source	
LYNN HAVEN PD SMITH,JAMES,C	05/13/2006	00:59:43	8100.26	Agency	Inspection

Sims 2 Complies Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source	
LYNN HAVEN PD SMITH,JAMES,C	05/13/2006	01:13:24	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N. Non-compliance: .

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source	
LYNN HAVEN PD SMITH,JAMES,C	05/13/2006	01:31:00	8100.26	Agency	Inspection

Sims 2 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . N.

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software		Source	
LYNN HAVEN PD SMITH,JAMES,C		07/12/2006	23:10:23	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
# Sims	3	Complies No					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.177
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.073
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.072
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.073
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.073
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.073
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance. 08: Control Outside Tolerance,DRY GAS CYLINDER REMOVED AND RETIGHTENEDNon-compliance: .

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software		Source	
LYNN HAVEN PD MALHIOT,MATTHEW,E		10/02/2006	09:06:02	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610201G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/17/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.083
		0.20 g/210L Test #4	0.204
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.083
		0.20 g/210L Test #5	0.204
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.083
		0.20 g/210L Test #6	0.204
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.083
		0.20 g/210L Test #7	0.204
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.204
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software		Source	
LYNN HAVEN PD SMITH,JAMES,C		10/04/2006	09:58:13	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software	Source	
LYNN HAVEN PD MALHIOT,MATTHEW,E		10/26/2006	14:11:08	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 6172021
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 06/23/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.081
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.081
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.081
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software		Source	
LYNN HAVEN PD SMITH,JAMES,C		11/29/2006	16:24:44	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000840

Agency Operator	Date	Time	Software	Source	
LYNN HAVEN PD SMITH,JAMES,C	12/24/2006	13:28:46	8100.27	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000840

Agency Operator		Date	Time	Software		Source	
LYNN HAVEN PD SMITH,JAMES,C		12/24/2006	14:07:44	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	617202I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	06/23/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software	Source	
CAPE CORAL PD VENTURI,GEORGE,L		01/24/2006	16:48:28	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 428102I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 10/11/2006
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.076
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.077
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.076
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.076
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000859

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	03/22/2006	16:23:55	8100.26	Agency	Inspection

# Sims	5	Complies	No
--------	---	----------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		03/22/2006	16:28:35	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		04/11/2006	08:39:01	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		05/12/2006	13:22:46	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200411G	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	11/18/2006	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		06/02/2006	14:26:43	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		07/13/2006	11:32:52	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.181
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance USED .200 SIMULATOR.

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		08/03/2006	15:17:24	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	601301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/18/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		09/05/2006	08:41:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/18/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		10/06/2006	12:25:30	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/18/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000859

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	10/25/2006	10:28:41	8100.26	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.039
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.042
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.040
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.043
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.040
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.043
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software		Source	
CAPE CORAL PD CLAFLIN,DAVID,S		10/25/2006	12:03:49	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	610301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	04/18/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. REPEATED INSPECTION-SIM LEAK

Inspection Tests

Instrument Serial #80-000859

Agency Operator		Date	Time	Software	Source	
CAPE CORAL PD VENTURI,GEORGE,L		12/05/2006	15:09:18	8100.27	Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot# 626101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration 09/20/2008
# Sims	5	Avg Std Dev	0.0004	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.076
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.076
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.076
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.076
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000859

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.075
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO VENTURI,GEORGE,L		01/23/2006	16:04:20	8100.26		Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	434502I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	12/14/2006
# Sims	5	Avg Std Dev	0.0006	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.078
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000861

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		03/27/2006	01:09:40	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200411D	0.08 Gas Lot#	434502I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	04/11/2006	Expiration	12/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		04/25/2006	08:22:10	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	608301G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	03/27/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		05/16/2006	08:17:55	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	608301G
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	03/27/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		06/26/2006	08:54:58	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	608301G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	03/27/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A:DID NOT DETECT MOUTH ALCOHOL.

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		07/21/2006	07:47:18	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	608301G
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	03/27/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		08/09/2006	13:32:02	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	608301G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	03/27/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000861

Agency Operator		Date	Time	Software		Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T		09/14/2006	11:56:08	8100.26		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	608301G
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	03/27/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000861

Agency Operator	Date	Time	Software	Source
SARASOTA COUNTY SO VENTURI,GEORGE,L	12/05/2006	15:37:27	8100.27	Department Inspection
0.05 Lot# 200605B Expiration 05/04/2008	0.08 Lot# 200605C Expiration 05/04/2008	0.20 Lot# 200605F Expiration 05/10/2008	0.08 Gas Lot# 626101I Expiration 09/20/2008	
# Sims 5	Avg Std Dev 0.0003	Complies Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.075
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.074
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.075
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.075
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.075
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.075
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.075
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.075
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000

Inspection Tests

Instrument Serial #80-000861

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7	0.050
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.074
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.075
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.074
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.075
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.074
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.075
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.074
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.075
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: Control Outside Tolerance-DGS AMBIENT CONDITION

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software	Source	
BLOUNTSTOWN PD MALHIOT,MATTHEW,E		03/08/2006	10:07:39	8100.26	Department Inspection	
0.05 Lot#	200504C	0.08 Lot#	200411G	0.20 Lot#	200505C	0.08 Gas Lot# 411803I
Expiration	04/21/2007	Expiration	11/18/2006	Expiration	05/05/2007	Expiration 04/30/2006
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000862

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		03/28/2006	12:23:30	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	428102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	2	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator	Date	Time	Software	Source	
BLOUNTSTOWN PD TANNER,WARREN,L	04/26/2006	12:08:06	8100.26	Agency	Inspection

Sims 4 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		04/26/2006	12:14:08	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	428102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		05/30/2006	18:25:28	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	428102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		06/23/2006	16:26:26	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		07/20/2006	20:59:24	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	428102I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/11/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: No Sample Provided.

Inspection Tests

Instrument Serial #80-000862

Agency Operator	Date	Time	Software	Source	
BLOUNTSTOWN PD TANNER,WARREN,L	08/28/2006	07:47:31	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
0.08 Gas Lot#	428102I	Expiration	10/11/2006		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator	Date	Time	Software	Source	
BLOUNTSTOWN PD TANNER,WARREN,L	09/25/2006	08:37:37	8100.26	Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007
0.08 Gas Lot#	428102I	Expiration	10/11/2006		
# Sims	4	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.187
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		10/30/2006	08:25:36	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	627101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/03/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software	Source	
BLOUNTSTOWN PD MALHIOT,MATTHEW,E		11/02/2006	16:50:54	8100.27	Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot# 427101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration 10/03/2008
# Sims	5	Avg Std Dev	0.0005	Complies	Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.193
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.080
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.080
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.193
		0.08 g/210L Dry Gas Std Test #7	0.080
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000862

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		11/28/2006	08:12:53	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	637101I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/03/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000862

Agency Operator		Date	Time	Software		Source	
BLOUNTSTOWN PD TANNER,WARREN,L		12/19/2006	08:33:17	8100.27		Agency	Inspection
0.05 Lot#	200605B	0.08 Lot#	200605C	0.20 Lot#	200605F	0.08 Gas Lot#	627101I
Expiration	05/04/2008	Expiration	05/04/2008	Expiration	05/10/2008	Expiration	10/03/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000867

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO HACKENY,DWITE,N	01/11/2006	09:16:39	8100.26	Agency	Inspection

Sims 0 Complies No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO HACKNEY,DWITE,N		01/11/2006	09:19:11	8100.26		Department Inspection	
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200505C	0.08 Gas Lot#	518702I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0003	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000867

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		03/27/2006	12:10:45	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	419102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		04/17/2006	10:36:11	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		05/23/2006	10:17:55	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	419102I
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	07/14/2006
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MARATHON

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		06/26/2006	10:13:21	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	612202G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MARATHON STATION

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		07/10/2006	11:44:03	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	612202G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MARATHON STATION

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		08/01/2006	11:13:25	8100.26		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	612202G
Expiration	04/21/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MARATHON STATION

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		09/20/2006	10:22:25	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MARATHON STATION

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		10/17/2006	13:32:22	8100.26		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MARATHON STATION

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO HACKNEY,DWITE,N		10/19/2006	12:41:28	8100.27		Department Inspection	
0.05 Lot#	200605B	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	518702I
Expiration	05/04/2008	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	07/08/2007
# Sims	5	Avg Std Dev	0.0004	Complies	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.080
		Interferent Detect Test #4	INT
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Interferent Detect Test #5	INT
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Interferent Detect Test #6	INT
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Interferent Detect Test #7	INT
		Alcohol Free Test #8	0.000

Inspection Tests

Instrument Serial #80-000867

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		10/19/2006	13:37:04	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200504D	0.20 Lot#	200505C	0.08 Gas Lot#	612202G
Expiration	09/22/2007	Expiration	04/21/2007	Expiration	05/05/2007	Expiration	05/05/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

KV AI AFTER SOFTWARE UPDATE

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO RABINOWITZ,ADAM,		11/28/2006	10:54:07	8100.27		Agency	Inspection
0.05 Lot#	200504C	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	612202G
Expiration	04/21/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	05/05/2008
# Sims	4	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MARATHON STATION

Inspection Tests

Instrument Serial #80-000867

Agency Operator	Date	Time	Software	Source	
MONROE COUNTY SO BLASCO,LUIS,	12/27/2006	13:51:21	8100.27	Agency	Inspection

# Sims	5	Complies	Yes
--------	---	----------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Inspection Tests

Instrument Serial #80-000867

Agency Operator		Date	Time	Software		Source	
MONROE COUNTY SO BLASCO,LUIS,		12/27/2006	14:41:02	8100.27		Agency	Inspection
0.05 Lot#	200509A	0.08 Lot#	200509B	0.20 Lot#	200509C	0.08 Gas Lot#	627201I
Expiration	09/22/2007	Expiration	09/22/2007	Expiration	09/22/2007	Expiration	10/03/2008
# Sims	5	Complies Yes					

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

KV STATION