

# Florida Department of Law Enforcement

## Control Tests

Instrument Serial #: 80-000219

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
FDLE FMROC	12/04/2007	17:55:48	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:54	
		ITP Check	0.079	17:55	
		Air Blank	0.000	17:55	

# Control Tests

Instrument Serial #: 80-000227

Location	Date	Time	Version	Target	Source
FDLE OROC	01/17/2007	18:25:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:24	
		Control Test	0.079	18:24	
		Air Blank	0.000	18:25	
FDLE OROC	01/17/2007	18:28:59	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:27	
		Control Test	0.042	18:28	
		Air Blank	0.000	18:28	
FDLE OROC	01/17/2007	18:30:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:29	
		Control Test	0.045	18:30	
		Air Blank	0.000	18:30	
FDLE OROC	01/17/2007	18:36:01	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:34	
		Control Test	0.188	18:35	
		Air Blank	0.000	18:36	
FDLE OROC	01/17/2007	18:38:04	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:36	
		Control Test	0.077	18:37	
		Air Blank	0.000	18:38	
FDLE OROC	01/17/2007	18:52:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:51	
		Control Test	0.044	18:52	
		Air Blank	0.000	18:52	
FDLE OROC	03/13/2007	16:07:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:05	
		Control Test	0.047	16:06	
		Air Blank	0.000	16:07	
FDLE OROC	04/11/2007	12:20:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:19	
		Control Test	0.081	12:20	

# Control Tests

Instrument Serial #: 80-000227

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
FDLE OROC	04/11/2007	12:20:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:20	

# Control Tests

Instrument Serial #: 80-000228

Location	Date	Time	Version	Target	Source
FDLE MROC	07/26/2007	23:35:13	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:34	
		Control Test	0.080	23:34	
		Air Blank	0.000	23:35	
FDLE MROC	07/26/2007	23:36:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:35	
		Control Test	0.079	23:36	
		Air Blank	0.000	23:36	
FDLE MROC	07/26/2007	23:38:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:37	
		Control Test	0.080	23:37	
		Air Blank	0.000	23:38	

# Control Tests

Instrument Serial #: 80-000749

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ESCAMBIA COUNTY SO	05/01/2007	10:56:13	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		10:55	Sequence Aborted
		Air Blank	0.000	10:56	

# Control Tests

Instrument Serial #: 80-000752

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MARION COUNTY SO	08/17/2007	09:44:41	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:43	
		Control Test	0.080	09:44	
		Air Blank	0.000	09:44	
MARION COUNTY SO	12/06/2007	19:27:09	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:25	
		Control Test	0.199	19:26	
		Air Blank	0.000	19:27	
MARION COUNTY SO	12/06/2007	20:20:44	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:19	
		Control Test	0.080	20:20	
		Air Blank	0.000	20:20	

# Control Tests

Instrument Serial #: 80-000760

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ALACHUA COUNTY SO	10/15/2007	14:52:27	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:51	
		Control Test	0.079	14:51	
		Air Blank	0.000	14:52	
ALACHUA COUNTY SO	10/15/2007	14:54:00	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:53	
		ITP Check	0.079	14:53	
		Air Blank	0.000	14:54	

# Control Tests

Instrument Serial #: 80-000763

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
OKALOOSA COUNTY SO	04/14/2007	16:59:38	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:58	
		Control Test	0.078	16:59	
		Air Blank	0.000	16:59	

# Control Tests

Instrument Serial #: 80-000779

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MADISON COUNTY SO	01/22/2007	12:25:14	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:24	
		Control Test	0.190	12:24	
		Air Blank	0.000	12:25	

# Control Tests

Instrument Serial #: 80-000785

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HOLMES COUNTY SO	04/30/2007	13:37:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:36	
		Control Test	0.079	13:37	
		Air Blank	0.000	13:37	
HOLMES COUNTY SO	04/30/2007	13:40:33	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:39	
		ITP Check	0.078	13:40	
		Air Blank	0.000	13:40	

# Control Tests

Instrument Serial #: 80-000786

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BROWARD S.O.	06/22/2007	10:01:53	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:01	
		Control Test	0.079	10:01	
		Air Blank	0.000	10:01	

# Control Tests

Instrument Serial #: 80-000788

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ST LUCIE COUNTY SO	03/18/2007	00:02:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:02	
		Control Test	0.078	00:02	
		Air Blank		00:02	RFI Detect
		Air Blank	0.000	00:02	

# Control Tests

Instrument Serial #: 80-000790

Location	Date	Time	Version	Target	Source
VERO BEACH PD	04/30/2007	09:53:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:52	
		Control Test	0.000	09:53	
		Air Blank	0.000	09:53	
VERO BEACH PD	04/30/2007	09:56:07	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:54	
		Control Test	0.046	09:55	
		Air Blank	0.000	09:56	
VERO BEACH PD	04/30/2007	09:59:38	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:58	
		Control Test	0.000	09:59	Interferent Detect
		Air Blank	0.000	09:59	
VERO BEACH PD	04/30/2007	10:04:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:03	
		Control Test	0.073	10:04	
		Air Blank	0.000	10:04	
VERO BEACH PD	04/30/2007	10:07:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:06	
		Control Test	0.049	10:06	
		Air Blank	0.000	10:07	
VERO BEACH PD	04/30/2007	10:08:59	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:07	
		Control Test	0.048	10:08	
		Air Blank	0.000	10:08	
VERO BEACH PD	04/30/2007	10:11:28	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:10	
		Control Test	0.048	10:10	
		Air Blank	0.000	10:11	
VERO BEACH PD	04/30/2007	10:13:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:12	
		Control Test	0.076	10:12	

# Control Tests

Instrument Serial #: 80-000790

Location	Date	Time	Version	Target	Source
VERO BEACH PD	04/30/2007	10:13:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:13	
VERO BEACH PD	04/30/2007	10:15:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:13	
		Control Test	0.077	10:14	
		Air Blank	0.000	10:15	
VERO BEACH PD	05/31/2007	12:53:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:52	
		Control Test	0.000	12:52	Interferent Detect
		Air Blank	0.000	12:53	
VERO BEACH PD	05/31/2007	12:55:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:54	
		Control Test	0.047	12:55	
		Air Blank	0.000	12:55	
VERO BEACH PD	05/31/2007	12:57:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:56	
		Control Test	0.072	12:57	
		Air Blank	0.000	12:57	
VERO BEACH PD	05/31/2007	13:01:24	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:00	
		Control Test	0.000	13:00	Interferent Detect
		Air Blank	0.000	13:01	
VERO BEACH PD	05/31/2007	13:03:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:02	
		Control Test	0.047	13:02	
		Air Blank	0.000	13:03	
VERO BEACH PD	05/31/2007	13:05:12	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:04	
		Control Test	0.080	13:04	
		Air Blank	0.000	13:05	

# Control Tests

Instrument Serial #: 80-000790

Location	Date	Time	Version	Target	Source
VERO BEACH PD	05/31/2007	13:07:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:05	
		Control Test	0.080	13:06	
		Air Blank	0.000	13:07	
VERO BEACH PD	08/31/2007	10:59:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:58	
		Control Test	0.000	10:59	
		Air Blank	0.000	10:59	
VERO BEACH PD	08/31/2007	11:04:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:02	
		Control Test	0.045	11:03	
		Air Blank	0.000	11:04	
VERO BEACH PD	08/31/2007	11:08:13	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:07	
		Control Test	0.046	11:07	
		Air Blank	0.000	11:08	
VERO BEACH PD	08/31/2007	11:13:01	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:11	
		Control Test	0.074	11:12	
		Air Blank	0.000	11:13	
VERO BEACH PD	08/31/2007	11:15:43	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:14	
		Control Test	0.193	11:15	
		Air Blank	0.000	11:15	
VERO BEACH PD	09/30/2007	09:17:57	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:16	
		Control Test	0.000	09:17	
		Air Blank	0.000	09:17	
VERO BEACH PD	09/30/2007	09:19:50	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:18	
		Control Test	0.048	09:19	

# Control Tests

Instrument Serial #: 80-000790

Location	Date	Time	Version	Target	Source
VERO BEACH PD	09/30/2007	09:19:50	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:19	
VERO BEACH PD	09/30/2007	09:22:45	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:21	
		Control Test	0.074	09:22	
		Air Blank	0.000	09:22	
VERO BEACH PD	10/31/2007	12:33:34	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:32	
		Control Test	0.046	12:33	
		Air Blank	0.000	12:33	
VERO BEACH PD	10/31/2007	12:35:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:34	
		Control Test	0.074	12:34	
		Air Blank	0.000	12:35	
VERO BEACH PD	10/31/2007	12:37:33	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:36	
		Control Test	0.193	12:37	
		Air Blank	0.000	12:37	
VERO BEACH PD	11/29/2007	15:25:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:23	
		Control Test	0.047	15:24	
		Air Blank	0.000	15:25	
VERO BEACH PD	11/29/2007	15:26:52	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:25	
		Control Test	0.075	15:26	
		Air Blank	0.000	15:26	
VERO BEACH PD	11/29/2007	15:29:10	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:27	
		Control Test	0.192	15:28	
		Air Blank	0.000	15:29	

# Control Tests

Instrument Serial #: 80-000790

Location	Date	Time	Version	Target	Source
VERO BEACH PD	12/18/2007	15:08:27	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		15:07	RFI Detect
		Air Blank	0.000	15:08	
VERO BEACH PD	12/18/2007	15:10:10	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:08	
		Control Test	0.043	15:09	
		Air Blank	0.000	15:10	
VERO BEACH PD	12/18/2007	15:12:31	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:11	
		Control Test	0.044	15:12	
		Air Blank	0.000	15:12	
VERO BEACH PD	12/18/2007	15:20:45	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:19	
		Control Test	0.045	15:20	
		Air Blank	0.000	15:20	
VERO BEACH PD	12/18/2007	15:22:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:21	
		Control Test	0.073	15:22	
		Air Blank	0.000	15:22	
VERO BEACH PD	12/18/2007	15:24:45	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:23	
		Control Test	0.075	15:24	
		Air Blank	0.000	15:24	
VERO BEACH PD	12/18/2007	15:26:11	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:25	
		Control Test	0.195	15:26	RFI Detect
		Air Blank		15:26	RFI Detect
VERO BEACH PD	12/18/2007	15:29:23	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:28	
		Control Test	0.196	15:28	
		Air Blank		15:28	RFI Detect

# Control Tests

Instrument Serial #: 80-000790

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
VERO BEACH PD	12/18/2007	15:29:23	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:29	
VERO BEACH PD	12/18/2007	15:39:44	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:38	
		Control Test	0.000	15:39	
		Air Blank	0.000	15:39	
VERO BEACH PD	12/18/2007	15:41:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:40	
		Control Test	0.047	15:40	
		Air Blank	0.000	15:41	

# Control Tests

Instrument Serial #: 80-000794

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ST LUCIE COUNTY SO	01/23/2007	21:55:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:54	
		Control Test	0.081	21:54	
		Air Blank	0.000	21:55	

# Control Tests

Instrument Serial #: 80-000795

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
GULF COUNTY SO	10/31/2007	19:51:21	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:50	
		Control Test	0.000	19:50	
		Air Blank	0.000	19:51	
GULF COUNTY SO	10/31/2007	19:53:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:52	
		Control Test	0.184	19:53	
		Air Blank	0.000	19:53	
GULF COUNTY SO	10/31/2007	19:57:12	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:55	
		Control Test	0.188	19:56	
		Air Blank	0.000	19:57	
GULF COUNTY SO	10/31/2007	19:59:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:58	
		Control Test	0.191	19:59	
		Air Blank	0.000	19:59	
GULF COUNTY SO	10/31/2007	21:17:56	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:17	
				21:17	Sequence Aborted
		Air Blank	0.000	21:17	

# Control Tests

Instrument Serial #: 80-000796

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
LONGBOAT KEY PD	09/12/2007	11:43:12	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:42	
		ITP Check	0.078	11:42	
		Air Blank	0.000	11:43	

# Control Tests

Instrument Serial #: 80-000808

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HILLSBOROUGH SO	03/20/2007	10:08:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:06	
		Control Test	0.049	10:07	
		Air Blank	0.000	10:08	

## Control Tests

Instrument Serial #: 80-000812

Location	Date	Time	Version	Target	Source
HERNANDO COUNTY SO	05/10/2007	12:45:11	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		12:44	Sequence Aborted
		Air Blank	0.000	12:45	
HERNANDO COUNTY SO	05/10/2007	12:56:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:55	
		Control Test	0.078	12:55	
Air Blank	0.000	12:56			
HERNANDO COUNTY SO	05/10/2007	12:58:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:56	
		Control Test	0.077	12:57	
Air Blank	0.000	12:58			
HERNANDO COUNTY SO	05/10/2007	12:59:49	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:58	
		ITP Check	0.079	12:59	
Air Blank	0.000	12:59			
HERNANDO COUNTY SO	05/10/2007	14:16:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:15	
		Control Test	0.079	14:16	
Air Blank	0.000	14:16			
HERNANDO COUNTY SO	05/10/2007	14:18:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:17	
		Control Test	0.079	14:17	
Air Blank	0.000	14:18			
HERNANDO COUNTY SO	05/10/2007	14:19:53	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:18	
		ITP Check	0.084	14:19	
Air Blank	0.000	14:19			
HERNANDO COUNTY SO	06/06/2007	15:57:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:56	
		Control Test	0.079	15:56	
Air Blank	0.000	15:57			

# Control Tests

Instrument Serial #: 80-000812

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HERNANDO COUNTY SO	06/07/2007	16:02:21	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:01	
		Control Test	0.079	16:01	
		Air Blank	0.000	16:02	
HERNANDO COUNTY SO	07/02/2007	11:00:39	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:59	
		ITP Check	0.083	11:00	
		Air Blank	0.000	11:00	
HERNANDO COUNTY SO	07/02/2007	12:07:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:06	
		Control Test	0.080	12:07	
		Air Blank	0.000	12:07	

# Control Tests

Instrument Serial #: 80-000813

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MARION COUNTY SO	06/27/2007	20:04:53	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:04	
		Control Test	0.080	20:04	
		Air Blank	0.000	20:04	

# Control Tests

Instrument Serial #: 80-000817

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	02/27/2007	11:48:39	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:47	
		Control Test	0.000	11:48	Interferent Detect
		Air Blank	0.000	11:48	
CITRUS COUNTY SO	11/12/2007	17:01:51	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:01	
		Control Test	0.080	17:01	
		Air Blank	0.000	17:01	
CITRUS COUNTY SO	11/12/2007	17:03:30	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:02	
		ITP Check	0.077	17:03	
		Air Blank	0.000	17:03	
CITRUS COUNTY SO	11/13/2007	15:47:53	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:46	
		ITP Check	0.077	15:47	
		Air Blank	0.000	15:47	
CITRUS COUNTY SO	11/13/2007	15:49:30	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:48	
		Control Test	0.079	15:49	
		Air Blank	0.000	15:49	
CITRUS COUNTY SO	11/14/2007	09:44:31	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:43	
		ITP Check	0.077	09:44	
		Air Blank	0.000	09:44	

# Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	01/30/2007	13:40:32	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:39	
		Control Test	0.078	13:40	
		Air Blank	0.000	13:40	
CITRUS COUNTY SO	01/30/2007	13:42:38	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:41	
		Control Test	0.050	13:42	
		Air Blank	0.000	13:42	
CITRUS COUNTY SO	01/30/2007	13:44:49	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:43	
		Control Test	0.080	13:44	
		Air Blank	0.000	13:44	
CITRUS COUNTY SO	02/25/2007	12:46:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:45	
		Control Test	0.077	12:46	
		Air Blank	0.000	12:46	
CITRUS COUNTY SO	02/25/2007	12:49:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:47	
		Control Test	0.049	12:48	
		Air Blank	0.000	12:49	
CITRUS COUNTY SO	02/25/2007	12:51:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:50	
		Control Test	0.080	12:50	
		Air Blank	0.000	12:51	
CITRUS COUNTY SO	02/25/2007	12:56:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:55	
		Control Test	0.197	12:55	
		Air Blank	0.000	12:56	
CITRUS COUNTY SO	03/31/2007	11:54:02	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:53	
		Control Test	0.077	11:53	

# Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	03/31/2007	11:54:02	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:54	
CITRUS COUNTY SO	03/31/2007	11:56:31	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:55	
		Control Test	0.078	11:55	
		Air Blank	0.000	11:56	
CITRUS COUNTY SO	03/31/2007	11:58:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:57	
		Control Test	0.049	11:58	
		Air Blank	0.000	11:58	
CITRUS COUNTY SO	03/31/2007	12:22:22	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		12:21	Ambient Fail
		Air Blank	0.000	12:22	
CITRUS COUNTY SO	03/31/2007	12:24:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:23	
		Control Test	0.000	12:23	
		Air Blank		12:23	Ambient Fail
		Air Blank	0.000	12:24	
CITRUS COUNTY SO	03/31/2007	12:29:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:28	
		Control Test	0.000	12:29	
		Air Blank	0.000	12:29	
CITRUS COUNTY SO	03/31/2007	12:31:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:30	
		Control Test	0.035	12:30	
		Air Blank	0.000	12:31	
CITRUS COUNTY SO	03/31/2007	12:33:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:32	
		Control Test	0.044	12:33	
		Air Blank	0.000	12:33	

# Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	03/31/2007	12:35:37	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:34	
		Control Test	0.056	12:35	
		Air Blank	0.000	12:35	
CITRUS COUNTY SO	03/31/2007	12:59:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:58	
		Control Test	0.051	12:58	
		Air Blank	0.000	12:59	
CITRUS COUNTY SO	03/31/2007	13:09:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:08	
		Control Test	0.065	13:08	
		Air Blank	0.000	13:09	
CITRUS COUNTY SO	03/31/2007	13:26:19	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:25	
		ITP Check	0.079	13:25	
		Air Blank	0.000	13:26	
CITRUS COUNTY SO	03/31/2007	15:03:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:02	
		Control Test	0.070	15:02	
		Air Blank	0.000	15:03	
CITRUS COUNTY SO	03/31/2007	15:05:50	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:04	
		Control Test	0.074	15:05	
		Air Blank	0.000	15:05	
CITRUS COUNTY SO	03/31/2007	22:14:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:13	
		Control Test	0.052	22:13	Interferent Detect
		Air Blank	0.000	22:14	
CITRUS COUNTY SO	03/31/2007	22:18:54	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:18	
		Control Test	0.049	22:18	Interferent Detect

## Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	03/31/2007	22:18:54	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:18	
CITRUS COUNTY SO	03/31/2007	22:21:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:19	
		Control Test	0.023	22:20	Interferent Detect
		Air Blank	0.000	22:21	
CITRUS COUNTY SO	03/31/2007	22:40:30	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		22:39	Ambient Fail
		Air Blank	0.000	22:40	
CITRUS COUNTY SO	04/01/2007	17:15:31	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:14	
		ITP Check	0.079	17:15	
		Air Blank	0.000	17:15	
CITRUS COUNTY SO	04/01/2007	17:16:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:16	
		Control Test	0.076	17:16	
		Air Blank	0.000	17:16	
CITRUS COUNTY SO	04/01/2007	17:18:23	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:17	
		Control Test	0.077	17:17	
		Air Blank	0.000	17:18	
CITRUS COUNTY SO	04/01/2007	17:19:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:18	
		Control Test	0.076	17:19	
		Air Blank	0.000	17:19	
CITRUS COUNTY SO	04/24/2007	19:55:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:54	
		Control Test	0.076	19:54	
		Air Blank	0.000	19:55	
CITRUS COUNTY SO	04/24/2007	19:56:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:55	

## Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	04/24/2007	19:56:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Control Test	0.076	19:56	
		Air Blank	0.000	19:56	
CITRUS COUNTY SO	04/24/2007	19:58:56	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:58	
		ITP Check	0.080	19:58	
Air Blank	0.000	19:58			
CITRUS COUNTY SO	04/24/2007	20:01:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:59	
		Control Test	0.000	20:00	
Air Blank	0.000	20:01			
CITRUS COUNTY SO	04/24/2007	20:05:04	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:03	
		Control Test	0.050	20:04	
Air Blank	0.000	20:05			
CITRUS COUNTY SO	04/24/2007	20:06:57	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:05	
		Control Test	0.081	20:06	
Air Blank	0.000	20:06			
CITRUS COUNTY SO	05/10/2007	15:07:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		15:06	Sequence Aborted
		Air Blank	0.000	15:07	
CITRUS COUNTY SO	05/10/2007	15:12:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:11	
		Control Test	0.078	15:11	
Air Blank	0.000	15:12			
CITRUS COUNTY SO	05/10/2007	15:16:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:15	
		Control Test	0.076	15:16	
Air Blank	0.000	15:16			

# Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	05/10/2007	15:18:51	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:17	
		ITP Check	0.080	15:18	
		Air Blank	0.000	15:18	
CITRUS COUNTY SO	05/25/2007	15:18:43	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:17	
		Control Test	0.077	15:18	
		Air Blank	0.000	15:18	
CITRUS COUNTY SO	07/20/2007	12:45:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:44	
		Control Test	0.079	12:45	
		Air Blank	0.000	12:45	
CITRUS COUNTY SO	07/20/2007	12:48:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:46	
		Control Test	0.080	12:47	
		Air Blank	0.000	12:48	
CITRUS COUNTY SO	07/20/2007	12:49:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:48	
		Control Test	0.051	12:49	
		Air Blank	0.000	12:49	
CITRUS COUNTY SO	07/20/2007	12:57:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:56	
		Control Test	0.000	12:56	
		Air Blank	0.000	12:57	
CITRUS COUNTY SO	07/20/2007	12:59:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:58	
		Control Test	0.000	12:58	
		Air Blank	0.000	12:59	
CITRUS COUNTY SO	07/20/2007	13:01:21	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:00	
		ITP Check	0.081	13:00	

## Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	07/20/2007	13:01:21	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:01	
CITRUS COUNTY SO	07/20/2007	13:49:41	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		13:49	Ambient Fail
		Air Blank	0.000	13:49	
CITRUS COUNTY SO	07/20/2007	13:51:12	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:50	
		ITP Check	0.080	13:50	
		Air Blank	0.000	13:51	
CITRUS COUNTY SO	07/20/2007	13:53:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:52	
		Control Test	0.023	13:52	Interferent Detect
		Air Blank	0.000	13:53	
CITRUS COUNTY SO	07/23/2007	09:14:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:13	
		Control Test	0.077	09:13	
		Air Blank	0.000	09:14	
CITRUS COUNTY SO	07/23/2007	09:22:01	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:20	
		Control Test	0.080	09:21	
		Air Blank	0.000	09:22	
CITRUS COUNTY SO	07/23/2007	09:24:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:23	
		Control Test	0.050	09:24	
		Air Blank	0.000	09:24	
CITRUS COUNTY SO	07/23/2007	09:27:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:26	
		Control Test	0.079	09:26	
		Air Blank	0.000	09:27	
CITRUS COUNTY SO	07/23/2007	10:20:10	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:18	

## Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	07/23/2007	10:20:10	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Control Test	0.198	10:19	
		Air Blank	0.000	10:20	
CITRUS COUNTY SO	07/23/2007	10:21:59	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:20	
		Control Test	0.199	10:21	
		Air Blank	0.000	10:21	
CITRUS COUNTY SO	07/23/2007	10:23:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:22	
		Control Test	0.200	10:23	
		Air Blank	0.000	10:23	
CITRUS COUNTY SO	08/30/2007	11:08:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:07	
		Control Test	0.079	11:08	
		Air Blank	0.000	11:08	
CITRUS COUNTY SO	08/30/2007	11:17:01	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:16	
		Control Test	0.079	11:16	
		Air Blank	0.000	11:17	
CITRUS COUNTY SO	08/30/2007	11:19:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:18	
		Control Test	0.080	11:18	
		Air Blank	0.000	11:19	
CITRUS COUNTY SO	08/30/2007	11:21:25	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:20	
		Control Test	0.200	11:20	
		Air Blank	0.000	11:21	
CITRUS COUNTY SO	08/30/2007	11:22:52	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:21	
		ITP Check	0.080	11:22	
		Air Blank	0.000	11:22	

# Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	09/26/2007	08:57:49	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:56	
		ITP Check	0.080	08:57	
		Air Blank	0.000	08:57	
CITRUS COUNTY SO	09/26/2007	09:03:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:02	
		Control Test	0.195	09:02	
		Air Blank	0.000	09:03	
CITRUS COUNTY SO	09/26/2007	09:06:35	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:05	
		Control Test	0.078	09:06	
		Air Blank	0.000	09:06	
CITRUS COUNTY SO	09/26/2007	09:18:25	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:17	
		Control Test	0.080	09:17	
		Air Blank	0.000	09:18	
CITRUS COUNTY SO	09/26/2007	09:22:43	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:21	
		Control Test	0.204	09:22	
		Air Blank	0.000	09:22	
CITRUS COUNTY SO	09/26/2007	09:30:35	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:29	
		Control Test	0.204	09:30	
		Air Blank	0.000	09:30	
CITRUS COUNTY SO	11/12/2007	10:47:13	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:46	
		ITP Check	0.078	10:46	
		Air Blank	0.000	10:47	
CITRUS COUNTY SO	11/12/2007	10:51:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:50	
		Control Test	0.080	10:50	

## Control Tests

Instrument Serial #: 80-000819

Location	Date	Time	Version	Target	Source
CITRUS COUNTY SO	11/12/2007	10:51:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		10:51	Purge Fail
		Air Blank		10:51	Purge Fail
CITRUS COUNTY SO	11/12/2007	11:20:35	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:19	
		ITP Check	0.078	11:20	
		Air Blank	0.000	11:20	
CITRUS COUNTY SO	11/12/2007	11:22:33	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:21	
		Control Test	0.029	11:21	
		Air Blank	0.000	11:22	
CITRUS COUNTY SO	11/12/2007	11:27:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:26	
		Control Test	0.000	11:26	
		Air Blank	0.000	11:27	
CITRUS COUNTY SO	11/12/2007	11:36:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:36	
		Control Test	0.000	11:36	
		Air Blank	0.000	11:36	
CITRUS COUNTY SO	11/12/2007	11:38:33	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:37	
		Control Test	0.000	11:38	
		Air Blank	0.000	11:38	
CITRUS COUNTY SO	11/12/2007	12:02:22	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:01	
		Control Test	0.000	12:01	
		Air Blank	0.000	12:02	
CITRUS COUNTY SO	11/12/2007	13:21:46	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:19	
		Control Test	0.143	13:20	
		Air Blank		13:20	Purge Fail

# Control Tests

Instrument Serial #: 80-000819

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CITRUS COUNTY SO	11/12/2007	13:21:46	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:21	
CITRUS COUNTY SO	11/12/2007	13:23:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:22	
		Control Test	0.081	13:22	
		Air Blank	0.000	13:23	
CITRUS COUNTY SO	11/12/2007	13:24:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:23	
		Control Test	0.081	13:24	
		Air Blank	0.000	13:24	
CITRUS COUNTY SO	11/12/2007	13:26:27	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:25	
		Control Test	0.203	13:25	
		Air Blank	0.000	13:26	
CITRUS COUNTY SO	11/29/2007	16:05:27	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:04	
		ITP Check	0.078	16:05	
		Air Blank	0.000	16:05	

# Control Tests

Instrument Serial #: 80-000827

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MARION COUNTY SO	11/18/2007	13:04:25	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:03	
		Control Test	0.081	13:03	
		Air Blank	0.000	13:04	
MARION COUNTY SO	11/18/2007	13:06:12	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:05	
		Control Test	0.000	13:05	
		Air Blank	0.000	13:06	
MARION COUNTY SO	11/18/2007	13:07:40	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:06	
		ITP Check	0.083	13:07	
		Air Blank	0.000	13:07	

# Control Tests

Instrument Serial #: 80-000829

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HILLSBOROUGH SO	04/03/2007	09:56:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:55	
		Control Test	0.080	09:55	
		Air Blank	0.000	09:56	
HILLSBOROUGH SO	04/03/2007	09:58:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:57	
		Control Test	0.080	09:57	
		Air Blank	0.000	09:58	
HILLSBOROUGH SO	04/03/2007	10:51:58	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:51	
		ITP Check	0.079	10:51	
		Air Blank	0.000	10:51	

# Control Tests

Instrument Serial #: 80-000830

Location	Date	Time	Version	Target	Source
HILLSBOROUGH SO	01/23/2007	16:54:51	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:54	
		Control Test	0.080	16:54	
		Air Blank	0.000	16:54	
HILLSBOROUGH SO	03/08/2007	09:20:12	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:19	
		ITP Check	0.078	09:19	
		Air Blank	0.000	09:20	
HILLSBOROUGH SO	04/25/2007	13:03:11	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:02	
		Control Test	0.092	13:02	Interferent Detect
		Air Blank		13:03	Purge Fail
HILLSBOROUGH SO	04/25/2007	13:04:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:04	
		Control Test	0.079	13:04	
		Air Blank	0.000	13:04	
HILLSBOROUGH SO	05/01/2007	16:15:31	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:14	
		Control Test	0.080	16:15	
		Air Blank	0.000	16:15	
HILLSBOROUGH SO	05/01/2007	16:58:27	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:57	
		ITP Check	0.079	16:58	
		Air Blank	0.000	16:58	

# Control Tests

Instrument Serial #: 80-000833

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HILLSBOROUGH SO	06/18/2007	14:09:35	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:08	
		Control Test	0.077	14:09	
		Air Blank	0.000	14:09	

# Control Tests

Instrument Serial #: 80-000861

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SARASOTA COUNTY SO	05/25/2007	23:02:31	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:01	
		Control Test	0.078	23:02	
		Air Blank	0.000	23:02	
SARASOTA COUNTY SO	05/25/2007	23:05:28	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:04	
		ITP Check	0.079	23:05	
		Air Blank	0.000	23:05	
SARASOTA COUNTY SO	06/15/2007	10:04:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:03	
		Control Test	0.076	10:03	
		Air Blank	0.000	10:04	
SARASOTA COUNTY SO	06/15/2007	22:24:17	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:23	
		Control Test	0.077	22:23	
		Air Blank	0.000	22:24	
SARASOTA COUNTY SO	10/02/2007	11:45:32	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:44	
		ITP Check	0.081	11:44	
		Air Blank		11:45	RFI Detect
		Air Blank	0.000	11:45	

# Control Tests

Instrument Serial #: 80-000867

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MONROE CO SO	04/30/2007	13:54:51	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:52	
		Control Test	0.071	13:53	
		Air Blank		13:54	Purge Fail
		Air Blank		13:54	Purge Fail
MONROE CO SO	04/30/2007	14:09:03	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:08	
		Control Test	0.000	14:08	
		Air Blank	0.000	14:09	
MONROE CO SO	04/30/2007	14:12:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:11	
		Control Test	0.000	14:11	
		Air Blank	0.000	14:12	

# Control Tests

Instrument Serial #: 80-000873

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MIAMI PD	03/20/2007	19:42:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:41	
		Control Test	0.080	19:42	
		Air Blank	0.000	19:42	
MIAMI PD	03/20/2007	19:44:38	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:43	
		ITP Check	0.078	19:44	
		Air Blank	0.000	19:44	

# Control Tests

Instrument Serial #: 80-000876

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HIALEAH PD	09/10/2007	15:59:07	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:57	
		Control Test	0.200	15:58	
		Air Blank	0.000	15:59	
HIALEAH PD	09/11/2007	11:59:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:58	
		Control Test	0.000	11:58	
		Air Blank	0.000	11:59	
HIALEAH PD	09/11/2007	12:00:46	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:59	
		ITP Check	0.077	12:00	
		Air Blank	0.000	12:00	
HIALEAH PD	09/11/2007	14:51:13	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:50	
		Control Test	0.000	14:50	
		Air Blank	0.000	14:51	

# Control Tests

Instrument Serial #: 80-000879

Location	Date	Time	Version	Target	Source
MIAMI BEACH PD	10/17/2007	13:54:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:53	
		Control Test	0.096	13:54	
		Air Blank	0.000	13:54	
MIAMI BEACH PD	10/17/2007	20:04:42	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:03	
		Control Test	0.098	20:04	
		Air Blank	0.000	20:04	
MIAMI BEACH PD	10/17/2007	20:15:17	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:14	
		Control Test	0.000	20:14	
		Air Blank	0.000	20:15	
MIAMI BEACH PD	10/17/2007	20:17:03	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:15	
		Control Test	0.012	20:16	
		Air Blank	0.000	20:17	
MIAMI BEACH PD	10/17/2007	20:19:09	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:18	
		Control Test	0.000	20:18	
		Air Blank	0.000	20:19	
MIAMI BEACH PD	10/18/2007	10:49:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:48	
		Control Test	0.202	10:48	
		Air Blank	0.000	10:49	
MIAMI BEACH PD	10/19/2007	11:46:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:45	
		Control Test	0.079	11:45	
		Air Blank	0.000	11:46	
MIAMI BEACH PD	10/25/2007	12:18:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:17	
		Control Test	0.085	12:18	

# Control Tests

Instrument Serial #: 80-000879

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MIAMI BEACH PD	10/25/2007	12:18:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:18	
MIAMI BEACH PD	10/25/2007	15:03:52	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:02	
		Control Test	0.000	15:03	
		Air Blank	0.000	15:03	
MIAMI BEACH PD	11/12/2007	14:52:03	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:51	
		Control Test	0.079	14:51	
		Air Blank	0.000	14:52	
MIAMI BEACH PD	11/12/2007	14:55:23	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:54	
		Control Test	0.051	14:54	
		Air Blank	0.000	14:55	
MIAMI BEACH PD	11/12/2007	14:56:22	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		14:55	Sequence Aborted
		Air Blank	0.000	14:56	

# Control Tests

Instrument Serial #: 80-000884

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MIAMI-DADE PD	03/08/2007	19:02:14	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:01	
		Control Test	0.076	19:01	
		Air Blank	0.000	19:02	
MIAMI-DADE PD	03/14/2007	15:09:30	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:08	
		Control Test	0.078	15:08	
		Air Blank	0.000	15:09	

# Control Tests

Instrument Serial #: 80-000887

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CHATTAHOOCHEE PD	06/18/2007	09:56:29	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:55	
		Control Test	0.076	09:55	
		Air Blank	0.000	09:56	
CHATTAHOOCHEE PD	06/18/2007	10:05:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:04	
		Control Test	0.074	10:05	
		Air Blank	0.000	10:05	
CHATTAHOOCHEE PD	06/19/2007	16:59:36	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:58	
		Control Test	0.038	16:59	Interferent Detect
		Air Blank	0.000	16:59	
CHATTAHOOCHEE PD	06/20/2007	11:10:51	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:09	
		Control Test	0.209	11:10	
		Air Blank	0.000	11:10	

# Control Tests

Instrument Serial #: 80-000888

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	09/15/2007	01:48:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:47	
		Control Test	0.081	01:47	
		Air Blank	0.000	01:48	
PINELLAS COUNTY SO	09/15/2007	01:49:44	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:48	
		ITP Check	0.078	01:49	
		Air Blank	0.000	01:49	

# Control Tests

Instrument Serial #: 80-000889

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	09/20/2007	00:45:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:45	
		Control Test	0.080	00:45	
		Air Blank	0.000	00:45	
PINELLAS COUNTY SO	09/20/2007	00:47:43	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:46	
		ITP Check	0.077	00:47	
		Air Blank	0.000	00:47	
PINELLAS COUNTY SO	12/12/2007	22:01:52	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:01	
		Control Test	0.080	22:01	
		Air Blank	0.000	22:01	
PINELLAS COUNTY SO	12/12/2007	22:03:24	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:02	
		ITP Check	0.078	22:02	
		Air Blank	0.000	22:03	
PINELLAS COUNTY SO	12/13/2007	00:51:54	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:51	
		ITP Check	0.078	00:51	
		Air Blank	0.000	00:51	

# Control Tests

Instrument Serial #: 80-000890

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	11/23/2007	19:29:32	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:28	
		ITP Check	0.078	19:29	
		Air Blank	0.000	19:29	

# Control Tests

Instrument Serial #: 80-000901

Location	Date	Time	Version	Target	Source
FFWCC TAMPA	01/15/2007	11:25:00	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:24	
		Control Test	0.079	11:24	
		Air Blank	0.000	11:25	
FFWCC TAMPA	01/19/2007	10:02:07	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:01	
		Control Test	0.080	10:01	
		Air Blank	0.000	10:02	
FFWCC TAMPA	01/27/2007	15:07:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:06	
		Control Test	0.079	15:06	
		Air Blank	0.000	15:07	
FFWCC TAMPA	01/27/2007	16:40:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:39	
		Control Test	0.079	16:40	
		Air Blank	0.000	16:40	
FFWCC TAMPA	02/06/2007	11:25:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:24	
		Control Test	0.080	11:24	
		Air Blank	0.000	11:25	
FFWCC TAMPA	03/16/2007	13:32:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:31	
		Control Test	0.078	13:31	
		Air Blank	0.000	13:32	
FFWCC TAMPA	04/16/2007	12:58:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:57	
		Control Test	0.079	12:57	
		Air Blank	0.000	12:58	
FFWCC TAMPA	05/09/2007	14:36:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:35	
		Control Test	0.080	14:36	

## Control Tests

Instrument Serial #: 80-000901

Location	Date	Time	Version	Target	Source
FFWCC TAMPA	05/09/2007	14:36:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:36	
FFWCC TAMPA	06/22/2007	06:35:21	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:34	
		Control Test	0.079	06:34	
Air Blank	0.000	06:35			
FFWCC TAMPA	07/17/2007	08:37:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:36	
		Control Test	0.078	08:36	
Air Blank	0.000	08:37			
FFWCC TAMPA	08/22/2007	18:30:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:29	
		Control Test	0.078	18:29	
Air Blank	0.000	18:30			
FFWCC TAMPA	09/17/2007	15:29:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:28	
		Control Test	0.078	15:28	
Air Blank	0.000	15:29			
FFWCC TAMPA	10/15/2007	11:03:01	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:02	
		Control Test	0.079	11:02	
Air Blank	0.000	11:03			
FFWCC TAMPA	11/08/2007	19:29:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:29	
		Control Test	0.079	19:29	
Air Blank	0.000	19:29			
FFWCC TAMPA	12/18/2007	14:25:51	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:25	
		Control Test	0.079	14:25	
Air Blank	0.000	14:25			

## Control Tests

Instrument Serial #: 80-000902

Location	Date	Time	Version	Target	Source
FFWCC	01/15/2007	11:26:13	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:25	
		Control Test	0.080	11:25	
		Air Blank	0.000	11:26	
FFWCC	01/19/2007	10:02:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:01	
		Control Test	0.081	10:02	
		Air Blank	0.000	10:02	
FFWCC	01/19/2007	10:05:01	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:04	
		Control Test	0.081	10:04	
		Air Blank	0.000	10:05	
FFWCC	01/27/2007	15:34:41	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:33	
		Control Test	0.080	15:34	
		Air Blank	0.000	15:34	
FFWCC	01/27/2007	16:42:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:41	
		Control Test	0.080	16:42	
		Air Blank	0.000	16:42	
FFWCC	02/06/2007	11:16:29	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:15	
		Control Test	0.081	11:15	
		Air Blank	0.000	11:16	
FFWCC	03/16/2007	13:33:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:32	
		Control Test	0.080	13:32	
		Air Blank	0.000	13:33	
FFWCC	04/16/2007	12:57:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:56	
		Control Test	0.082	12:56	

## Control Tests

Instrument Serial #: 80-000902

Location	Date	Time	Version	Target	Source
FFWCC	04/16/2007	12:57:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:57	
FFWCC	06/22/2007	06:38:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:37	
		Control Test	0.080	06:38	
Air Blank	0.000	06:38			
FFWCC	07/17/2007	08:38:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:37	
		Control Test	0.080	08:37	
Air Blank	0.000	08:38			
FFWCC	08/22/2007	18:30:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:29	
		Control Test	0.080	18:29	
Air Blank	0.000	18:30			
FFWCC	09/17/2007	15:30:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:29	
		Control Test	0.080	15:29	
Air Blank	0.000	15:30			
FFWCC	10/15/2007	11:03:41	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:02	
		Control Test	0.080	11:03	
Air Blank	0.000	11:03			
FFWCC	11/08/2007	19:30:51	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:30	
		Control Test	0.080	19:30	
Air Blank	0.000	19:30			
FFWCC	12/18/2007	14:27:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:26	
		Control Test	0.081	14:26	
Air Blank	0.000	14:27			

# Control Tests

Instrument Serial #: 80-000907

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MANATEE COUNTY SO	02/14/2007	10:08:33	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:07	
		Control Test	0.077	10:08	
		Air Blank	0.000	10:08	

# Control Tests

Instrument Serial #: 80-000913

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MANATEE COUNTY SO	10/16/2007	19:21:30	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:20	
		ITP Check	0.079	19:21	
		Air Blank	0.000	19:21	
MANATEE COUNTY SO	12/11/2007	10:21:45	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:20	
		ITP Check	0.079	10:21	
		Air Blank	0.000	10:21	

# Control Tests

Instrument Serial #: 80-000942

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COLLIER COUNTY SO	05/14/2007	10:09:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:08	
		Control Test	0.075	10:09	
		Air Blank	0.000	10:09	
COLLIER COUNTY SO	05/14/2007	10:11:30	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:10	
		Control Test	0.075	10:11	
		Air Blank	0.000	10:11	
COLLIER COUNTY SO	05/14/2007	10:13:13	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:12	
		Control Test	0.076	10:12	
		Air Blank	0.000	10:13	

# Control Tests

Instrument Serial #: 80-000943

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COLLIER COUNTY SO	06/20/2007	15:24:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:23	
		Control Test	0.080	15:23	
		Air Blank	0.000	15:24	
COLLIER COUNTY SO	08/01/2007	13:47:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:46	
		Control Test	0.049	13:47	
		Air Blank	0.000	13:47	
COLLIER COUNTY SO	08/01/2007	13:49:22	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:48	
		ITP Check	0.099	13:48	
		Air Blank	0.000	13:49	
COLLIER COUNTY SO	08/01/2007	13:50:51	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:50	
		Control Test	0.079	13:50	
		Air Blank	0.000	13:50	
COLLIER COUNTY SO	11/20/2007	11:41:49	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:40	
		ITP Check	0.080	11:41	
		Air Blank	0.000	11:41	
COLLIER COUNTY SO	11/20/2007	13:36:59	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:36	
		ITP Check	0.080	13:36	
		Air Blank	0.000	13:36	

# Control Tests

Instrument Serial #: 80-000944

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COLLIER COUNTY SO	01/01/2007	12:33:28	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:32	
		Control Test	0.193	12:32	
		Air Blank	0.000	12:33	
COLLIER COUNTY SO	11/20/2007	11:40:03	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:39	
		ITP Check	0.080	11:39	
		Air Blank	0.000	11:40	
COLLIER COUNTY SO	11/20/2007	13:27:29	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:26	
		ITP Check	0.080	13:27	
		Air Blank	0.000	13:27	
COLLIER COUNTY SO	12/25/2007	09:33:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:32	
		Control Test	0.080	09:33	
		Air Blank	0.000	09:33	
COLLIER COUNTY SO	12/25/2007	09:35:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:34	
		Control Test	0.080	09:35	
		Air Blank	0.000	09:35	

# Control Tests

Instrument Serial #: 80-000945

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CHARLOTTE COUNTY SO	08/13/2007	13:30:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:29	
		Control Test	0.094	13:29	
		Air Blank	0.000	13:30	

## Control Tests

Instrument Serial #: 80-000946

Location	Date	Time	Version	Target	Source
CHARLOTTE COUNTY SO	02/10/2007	20:26:27	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:25	
		Control Test	0.000	20:25	Interferent Detect
		Air Blank	0.000	20:26	
CHARLOTTE COUNTY SO	02/10/2007	20:28:14	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:27	
		Control Test	0.000	20:27	Interferent Detect
		Air Blank	0.000	20:28	
CHARLOTTE COUNTY SO	02/10/2007	20:30:01	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:28	
		Control Test	0.000	20:29	Interferent Detect
		Air Blank	0.000	20:30	
CHARLOTTE COUNTY SO	02/10/2007	20:31:53	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:30	
		Control Test	0.000	20:31	Interferent Detect
		Air Blank	0.000	20:31	
CHARLOTTE COUNTY SO	02/10/2007	20:33:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:32	
		Control Test	0.000	20:33	Interferent Detect
		Air Blank	0.000	20:33	
CHARLOTTE COUNTY SO	02/10/2007	20:38:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:37	
		Control Test	0.000	20:37	Interferent Detect
		Air Blank	0.000	20:38	
CHARLOTTE COUNTY SO	02/10/2007	20:42:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:41	
		Control Test	0.000	20:41	Interferent Detect
		Air Blank	0.000	20:42	
CHARLOTTE COUNTY SO	02/10/2007	20:44:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:43	
		Control Test	0.000	20:43	Interferent Detect

# Control Tests

Instrument Serial #: 80-000946

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CHARLOTTE COUNTY SO	02/10/2007	20:44:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:44	
CHARLOTTE COUNTY SO	02/10/2007	20:45:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:44	
		Control Test	0.000	20:45	Interferent Detect
		Air Blank	0.000	20:45	
CHARLOTTE COUNTY SO	02/10/2007	20:47:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:46	
		Control Test	0.000	20:47	Interferent Detect
		Air Blank	0.000	20:47	
CHARLOTTE COUNTY SO	03/25/2007	21:04:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:03	
		Control Test	0.077	21:03	
		Air Blank	0.000	21:04	

# Control Tests

Instrument Serial #: 80-000950

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HAVANA POLICE DEPT	04/24/2007	15:18:44	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:17	
		Control Test	0.077	15:18	
		Air Blank	0.000	15:18	

# Control Tests

Instrument Serial #: 80-000951

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HENDRY COUNTY SO	09/27/2007	16:52:08	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:51	
		ITP Check	0.080	16:51	
		Air Blank	0.000	16:52	

# Control Tests

Instrument Serial #: 80-000955

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CLEWISTON PD	09/27/2007	11:45:50	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:44	
		ITP Check	0.078	11:45	
		Air Blank	0.000	11:45	

# Control Tests

Instrument Serial #: 80-000962

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
WAKULLA COUNTY SO	10/03/2007	14:36:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:35	
		Control Test	0.080	14:35	
		Air Blank	0.000	14:36	

# Control Tests

Instrument Serial #: 80-000963

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ORANGE COUNTY SO	10/02/2007	15:26:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:26	
		Control Test	0.082	15:26	
		Air Blank	0.000	15:26	
ORANGE COUNTY SO	10/02/2007	15:29:11	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:28	
		ITP Check	0.080	15:28	
		Air Blank	0.000	15:29	

# Control Tests

Instrument Serial #: 80-000966

Location	Date	Time	Version	Target	Source
OSCEOLA CORRECTIONS	06/18/2007	01:00:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:58	
		Control Test	0.000	00:59	
		Air Blank		01:00	Ambient Fail
		Air Blank		01:00	Purge Fail

# Control Tests

Instrument Serial #: 80-000967

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
APOPKA P.D.	12/03/2007	10:24:14	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:23	
		Control Test	0.079	10:23	
		Air Blank	0.000	10:24	

# Control Tests

Instrument Serial #: 80-000998

Location	Date	Time	Version	Target	Source
SEMINOLE S.O.	03/17/2007	21:33:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:32	
		Control Test	0.078	21:32	
		Air Blank		21:33	RFI Detect
		Air Blank	0.000	21:33	
SEMINOLE S.O.	03/25/2007	22:28:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:27	
		Control Test	0.076	22:27	
		Air Blank	0.000	22:28	
SEMINOLE S.O.	03/29/2007	23:43:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:42	
		Control Test	0.079	23:42	
		Air Blank	0.000	23:43	
SEMINOLE S.O.	09/15/2007	04:30:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	04:29	
		Control Test	0.074	04:29	
		Air Blank	0.000	04:30	
SEMINOLE S.O.	09/15/2007	04:31:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	04:31	
		Control Test	0.075	04:31	
		Air Blank	0.000	04:31	
SEMINOLE S.O.	09/15/2007	04:33:52	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	04:33	
		Control Test	0.077	04:33	
		Air Blank	0.000	04:33	

# Control Tests

Instrument Serial #: 80-000999

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SEMINOLE S.O.	02/18/2007	03:48:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:48	
		Control Test	0.081	03:48	
		Air Blank	0.000	03:48	
SEMINOLE S.O.	08/29/2007	21:57:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:56	
		Control Test	0.079	21:56	
		Air Blank	0.000	21:57	
SEMINOLE S.O.	09/15/2007	05:12:14	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:11	
		Control Test	0.077	05:11	
		Air Blank	0.000	05:12	
SEMINOLE S.O.	09/15/2007	05:13:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:12	
		Control Test	0.077	05:13	
		Air Blank	0.000	05:13	

# Control Tests

Instrument Serial #: 80-001003

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	12/26/2007	17:13:43	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:12	
		Control Test	0.079	17:13	
		Air Blank	0.000	17:13	

# Control Tests

Instrument Serial #: 80-001005

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	11/13/2007	10:29:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:28	
		Control Test	0.081	10:28	
		Air Blank	0.000	10:29	
PINELLAS COUNTY SO	11/13/2007	10:30:58	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:30	
		ITP Check	0.079	10:30	
		Air Blank	0.000	10:30	

# Control Tests

Instrument Serial #: 80-001006

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
INDIALANTIC PD	01/02/2007	10:02:00	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:01	
		Control Test	0.079	10:01	
		Air Blank	0.000	10:02	
INDIALANTIC PD	04/27/2007	07:01:35	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	07:00	
		Control Test	0.081	07:01	
		Air Blank	0.000	07:01	

# Control Tests

Instrument Serial #: 80-001007

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HAINES CITY PD	03/09/2007	13:09:55	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:08	
		Control Test	0.000	13:09	
		Air Blank	0.000	13:09	

# Control Tests

Instrument Serial #: 80-001018

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
JACKSONVILLE SO	01/10/2007	06:36:54	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:36	
		Control Test	0.079	06:36	
		Air Blank	0.000	06:36	
JACKSONVILLE SO	01/10/2007	08:43:09	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:41	
		Control Test	0.084	08:42	
		Air Blank	0.000	08:43	
JACKSONVILLE SO	09/14/2007	06:04:52	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:04	
		Control Test	0.079	06:04	
		Air Blank	0.000	06:04	

# Control Tests

Instrument Serial #: 80-001044

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COCONUT CREEK PD	02/26/2007	15:32:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:31	
		Control Test	0.077	15:32	
		Air Blank	0.000	15:32	
COCONUT CREEK PD	08/01/2007	14:54:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:53	
		Control Test	0.076	14:53	
		Air Blank	0.000	14:54	
COCONUT CREEK PD	08/13/2007	17:24:33	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:23	
		Control Test	0.078	17:24	
		Air Blank	0.000	17:24	

# Control Tests

Instrument Serial #: 80-001046

Location	Date	Time	Version	Target	Source
CLAY COUNTY SO	03/07/2007	03:18:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:17	
		Control Test	0.079	03:17	
		Air Blank	0.000	03:18	
CLAY COUNTY SO	03/07/2007	03:20:15	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:19	
		ITP Check	0.079	03:19	
		Air Blank	0.000	03:20	
CLAY COUNTY SO	03/19/2007	08:43:32	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:42	
		Control Test	0.078	08:43	
		Air Blank	0.000	08:43	
CLAY COUNTY SO	03/19/2007	08:45:38	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:44	
		ITP Check	0.078	08:45	
		Air Blank	0.000	08:45	
CLAY COUNTY SO	05/15/2007	10:06:46	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:05	
		Control Test	0.079	10:06	
		Air Blank	0.000	10:06	
CLAY COUNTY SO	07/28/2007	22:30:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:29	
		Control Test	0.079	22:29	
		Air Blank	0.000	22:30	

# Control Tests

Instrument Serial #: 80-001047

Location	Date	Time	Version	Target	Source
BROWARD COUNTY SO	01/12/2007	10:36:41	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:35	
		Control Test	0.082	10:36	
		Air Blank	0.000	10:36	
BROWARD COUNTY SO	01/12/2007	10:38:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:37	
		Control Test	0.082	10:38	
		Air Blank	0.000	10:38	
BROWARD COUNTY SO	02/07/2007	16:36:40	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:35	
		Control Test	0.081	16:36	
		Air Blank	0.000	16:36	
BROWARD COUNTY SO	03/30/2007	10:44:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:43	
		Control Test	0.079	10:43	
		Air Blank	0.000	10:44	
BROWARD S.O.	08/03/2007	13:04:12	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:03	
		Control Test	0.079	13:03	
		Air Blank	0.000	13:04	
BROWARD S.O.	08/29/2007	13:23:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:22	
		Control Test	0.078	13:22	
		Air Blank	0.000	13:23	
BROWARD S.O.	08/29/2007	15:34:14	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:33	
		Control Test	0.050	15:33	
		Air Blank	0.000	15:34	
BROWARD S.O.	08/29/2007	15:36:39	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:35	
		Control Test	0.081	15:36	

# Control Tests

Instrument Serial #: 80-001047

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BROWARD S.O.	08/29/2007	15:36:39	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:36	

# Control Tests

Instrument Serial #: 80-001052

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
JACKSONVILLE SO	02/20/2007	11:11:20	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:10	
		Control Test	0.081	11:10	
		Air Blank	0.000	11:11	
JACKSONVILLE SO	06/19/2007	05:02:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:02	
		Control Test	0.078	05:02	
		Air Blank	0.000	05:02	
JACKSONVILLE SO	08/17/2007	13:46:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:45	
		Control Test	0.078	13:45	
		Air Blank	0.000	13:46	
JACKSONVILLE SO	11/16/2007	10:16:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:15	
		Control Test	0.080	10:16	
		Air Blank	0.000	10:16	
JACKSONVILLE SO	11/16/2007	10:22:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:21	
		Control Test	0.080	10:22	
		Air Blank	0.000	10:22	

# Control Tests

Instrument Serial #: 80-001053

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
NASSAU CO SO	06/20/2007	06:14:43	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:13	
		Control Test	0.078	06:14	
		Air Blank	0.000	06:14	
NASSAU CO SO	09/11/2007	05:58:25	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:57	
		Control Test	0.079	05:57	
		Air Blank	0.000	05:58	

# Control Tests

Instrument Serial #: 80-001058

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
LAKE WALES PD	03/06/2007	12:33:36	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:32	
		ITP Check	0.080	12:33	
		Air Blank	0.000	12:33	
LAKE WALES PD	03/06/2007	12:41:21	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:40	
		Control Test	0.047	12:40	
		Air Blank	0.000	12:41	
LAKE WALES PD	07/06/2007	09:58:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:58	
		Control Test	0.079	09:58	
		Air Blank	0.000	09:58	

# Control Tests

Instrument Serial #: 80-001059

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
FERNANDINA BEACH PD	02/27/2007	08:26:17	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:25	
		ITP Check	0.080	08:25	
		Air Blank	0.000	08:26	

# Control Tests

Instrument Serial #: 80-001060

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
WINTER HAVEN PD	09/20/2007	09:59:41	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:58	
		ITP Check	0.079	09:59	
		Air Blank	0.000	09:59	
WINTER HAVEN PD	10/22/2007	10:22:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:20	
		Control Test	0.000	10:21	
		Air Blank	0.000	10:22	

# Control Tests

Instrument Serial #: 80-001062

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
TREASURE ISLAND PD	02/05/2007	09:26:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:25	
		Control Test	0.081	09:26	
		Air Blank	0.000	09:26	
TREASURE ISLAND PD	02/05/2007	09:28:15	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:27	
		ITP Check	0.079	09:27	
		Air Blank	0.000	09:28	

# Control Tests

Instrument Serial #: 80-001063

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HOLLYWOOD PD	07/23/2007	14:40:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:39	
		Control Test	0.193	14:40	
		Air Blank	0.000	14:40	
HOLLYWOOD PD	07/23/2007	14:47:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:46	
		Control Test	0.184	14:47	
		Air Blank	0.000	14:47	

# Control Tests

Instrument Serial #: 80-001064

Location	Date	Time	Version	Target	Source
NEW PORT RICHEY PD	01/03/2007	10:38:15	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:37	
		Control Test	0.079	10:37	
		Air Blank	0.000	10:38	
NEW PORT RICHEY PD	01/03/2007	10:39:48	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:38	
		ITP Check	0.078	10:39	
		Air Blank	0.000	10:39	
NEW PORT RICHEY PD	01/03/2007	14:56:47	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:55	
		ITP Check	0.402	14:56	
		Air Blank	0.000	14:56	
NEW PORT RICHEY PD	01/03/2007	14:58:28	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:57	
		ITP Check	0.403	14:58	
		Air Blank	0.000	14:58	
NEW PORT RICHEY PD	01/03/2007	15:00:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:59	
		Control Test	0.079	14:59	
		Air Blank	0.000	15:00	
NEW PORT RICHEY PD	02/27/2007	10:47:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:46	
		Control Test	0.077	10:47	
		Air Blank	0.000	10:47	
NEW PORT RICHEY PD	02/27/2007	10:49:26	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:48	
		ITP Check	0.417	10:48	
		Air Blank	0.000	10:49	
NEW PORT RICHEY PD	02/28/2007	15:08:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:07	
		Control Test	0.077	15:07	

# Control Tests

Instrument Serial #: 80-001064

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
NEW PORT RICHEY PD	02/28/2007	15:08:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:08	
NEW PORT RICHEY PD	02/28/2007	15:09:33	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:08	
		ITP Check	0.416	15:09	
		Air Blank	0.000	15:09	
NEW PORT RICHEY PD	02/28/2007	15:48:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:48	
		Control Test	0.080	15:48	
		Air Blank	0.000	15:48	
NEW PORT RICHEY PD	02/28/2007	15:51:01	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:49	
		Control Test	0.080	15:50	
		Air Blank	0.000	15:51	
NEW PORT RICHEY PD	02/28/2007	15:53:18	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:52	
		ITP Check	0.417	15:52	
		Air Blank	0.000	15:53	

# Control Tests

Instrument Serial #: 80-001067

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BROWARD COUNTY SO	03/06/2007	16:50:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:50	
		Control Test	0.082	16:50	
		Air Blank	0.000	16:50	
BROWARD COUNTY SO	03/19/2007	10:17:43	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:16	
		Control Test	0.082	10:17	
		Air Blank	0.000	10:17	

# Control Tests

Instrument Serial #: 80-001068

Location	Date	Time	Version	Target	Source
CLEARWATER PD	02/16/2007	13:33:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:32	
		Control Test	0.085	13:33	
		Air Blank	0.000	13:33	
CLEARWATER PD	02/16/2007	13:35:33	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:34	
		Control Test	0.116	13:35	
		Air Blank	0.000	13:35	
CLEARWATER PD	02/16/2007	14:21:17	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:20	
		Control Test	0.168	14:20	
		Air Blank	0.000	14:21	
CLEARWATER PD	02/16/2007	14:23:01	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:21	
		Control Test	0.167	14:22	
		Air Blank	0.000	14:23	
CLEARWATER PD	02/20/2007	11:38:08	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:37	
		ITP Check	0.080	11:37	
		Air Blank	0.000	11:38	
CLEARWATER PD	02/24/2007	22:50:54	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:49	
		Control Test	0.000	22:50	Interferent Detect
		Air Blank	0.000	22:50	
CLEARWATER PD	02/24/2007	22:52:39	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:51	
		Control Test	0.047	22:52	
		Air Blank	0.000	22:52	
CLEARWATER PD	02/24/2007	22:53:36	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		22:53	Ambient Fail
		Air Blank	0.000	22:53	

## Control Tests

Instrument Serial #: 80-001068

Location	Date	Time	Version	Target	Source
CLEARWATER PD	02/24/2007	22:55:28	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:54	
		Control Test	0.075	22:54	
		Air Blank	0.000	22:55	
CLEARWATER PD	02/24/2007	22:56:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		22:55	Ambient Fail
		Air Blank	0.000	22:56	
CLEARWATER PD	02/24/2007	22:58:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:56	
		Control Test	0.075	22:57	
		Air Blank	0.000	22:58	
CLEARWATER PD	02/24/2007	23:00:49	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:59	
		Control Test	0.196	23:00	
		Air Blank	0.000	23:00	
CLEARWATER PD	02/24/2007	23:02:36	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:01	
		Control Test	0.068	23:02	
		Air Blank	0.000	23:02	
CLEARWATER PD	02/24/2007	23:05:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:04	
		Control Test	0.072	23:05	
		Air Blank	0.000	23:05	
CLEARWATER PD	02/24/2007	23:07:38	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:06	
		Control Test	0.075	23:07	
		Air Blank	0.000	23:07	
CLEARWATER PD	02/24/2007	23:09:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:08	
		Control Test	0.000	23:08	
		Air Blank	0.000	23:09	

# Control Tests

Instrument Serial #: 80-001068

Location	Date	Time	Version	Target	Source
CLEARWATER PD	02/24/2007	23:11:17	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:10	
		Control Test	0.068	23:10	
		Air Blank	0.000	23:11	
CLEARWATER PD	02/24/2007	23:12:14	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		23:11	Ambient Fail
		Air Blank	0.000	23:12	
CLEARWATER PD	02/24/2007	23:13:12	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		23:12	Ambient Fail
		Air Blank	0.000	23:13	
CLEARWATER PD	02/24/2007	23:14:10	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		23:13	Ambient Fail
		Air Blank	0.000	23:14	
CLEARWATER PD	02/24/2007	23:15:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:14	
		Control Test	0.000	23:15	
		Air Blank	0.000	23:15	

# Control Tests

Instrument Serial #: 80-001071

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CLAY COUNTY SO	03/19/2007	08:42:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:41	
		Control Test	0.080	08:41	
		Air Blank	0.000	08:42	

# Control Tests

Instrument Serial #: 80-001074

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PASCO COUNTY SO	04/25/2007	13:30:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:29	
		Control Test	0.082	13:29	
		Air Blank	0.000	13:30	

# Control Tests

Instrument Serial #: 80-001111

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
FHP TROOP A	06/07/2007	06:55:03	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:54	
		Control Test	0.079	06:54	
		Air Blank	0.000	06:55	
FHP TROOP A	06/07/2007	06:56:27	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:55	
		Control Test	0.080	06:55	
		Air Blank	0.000	06:56	

## Control Tests

Instrument Serial #: 80-001119

Location	Date	Time	Version	Target	Source
FHP TROOP D - DELAND	08/22/2007	15:25:32	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:24	
		Control Test	0.079	15:25	
		Air Blank	0.000	15:25	
FHP TROOP D - DELAND	08/29/2007	09:27:27	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:26	
		Control Test	0.080	09:26	
		Air Blank	0.000	09:27	
FHP TROOP D - DELAND	10/01/2007	16:02:27	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:01	
		Control Test	0.033	16:01	
		Air Blank	0.000	16:02	
FHP TROOP D - DELAND	10/01/2007	16:12:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:10	
		Control Test	0.033	16:11	
		Air Blank	0.000	16:12	
FHP TROOP D - DELAND	10/01/2007	16:16:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:15	
		Control Test	0.034	16:15	
		Air Blank	0.000	16:16	
FHP TROOP D - DELAND	10/01/2007	16:20:07	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:18	
		Control Test	0.043	16:19	
		Air Blank	0.000	16:20	
FHP TROOP D - DELAND	10/01/2007	16:28:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:27	
		Control Test	0.047	16:28	
		Air Blank	0.000	16:28	
FHP TROOP D - DELAND	10/01/2007	16:41:37	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:40	
		Control Test	0.000	16:41	

# Control Tests

Instrument Serial #: 80-001119

Location	Date	Time	Version	Target	Source
FHP TROOP D - DELAND	10/01/2007	16:41:37	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:41	
FHP TROOP D - DELAND	10/01/2007	16:43:34	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:42	
		Control Test	0.000	16:43	
		Air Blank	0.000	16:43	
FHP TROOP D - DELAND	10/01/2007	16:51:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:50	
		Control Test	0.027	16:51	
		Air Blank	0.000	16:51	
FHP TROOP D - DELAND	10/01/2007	16:57:06	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:55	
		Control Test	0.028	16:56	
		Air Blank	0.000	16:57	
FHP TROOP D - DELAND	10/01/2007	17:02:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:00	
		Control Test	0.035	17:01	
		Air Blank	0.000	17:02	
FHP TROOP D - DELAND	10/01/2007	17:05:49	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:04	
		Control Test	0.049	17:05	
		Air Blank	0.000	17:05	
FHP TROOP D - DELAND	10/01/2007	17:08:33	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:07	
		Control Test	0.049	17:08	
		Air Blank	0.000	17:08	
FHP TROOP D - DELAND	10/01/2007	17:10:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:09	
		Control Test	0.049	17:09	
		Air Blank	0.000	17:10	

# Control Tests

Instrument Serial #: 80-001134

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ST JOHN'S COUNTY SO	09/25/2007	12:57:47	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:56	
		Control Test	0.097	12:57	
		Air Blank	0.000	12:57	

# Control Tests

Instrument Serial #: 80-001139

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BROWARD COUNTY SO	06/08/2007	10:07:56	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:07	
		Control Test	0.081	10:07	
		Air Blank	0.000	10:07	
BROWARD COUNTY SO	08/02/2007	16:23:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:22	
		Control Test	0.080	16:22	
		Air Blank	0.000	16:23	
BROWARD COUNTY SO	09/15/2007	20:03:52	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:03	
		Control Test	0.080	20:03	
		Air Blank	0.000	20:03	
BROWARD COUNTY SO	11/26/2007	14:33:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:32	
		Control Test	0.080	14:33	
		Air Blank	0.000	14:33	

# Control Tests

Instrument Serial #: 80-001150

Location	Date	Time	Version	Target	Source
MANATEE COUNTY SO	03/24/2007	01:23:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:22	
		Control Test	0.083	01:22	
		Air Blank	0.000	01:23	
MANATEE COUNTY SO	08/31/2007	21:05:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:05	
		Control Test	0.077	21:05	
		Air Blank	0.000	21:05	
MANATEE COUNTY SO	08/31/2007	21:07:17	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:06	
		Control Test	0.082	21:06	
		Air Blank	0.000	21:07	
MANATEE COUNTY SO	08/31/2007	21:57:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:56	
		Control Test	0.080	21:56	
		Air Blank	0.000	21:57	
MANATEE COUNTY SO	08/31/2007	23:07:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:06	
		Control Test	0.078	23:06	
		Air Blank	0.000	23:07	
MANATEE COUNTY SO	09/01/2007	01:21:01	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:20	
		Control Test	0.081	01:20	
		Air Blank	0.000	01:21	
MANATEE COUNTY SO	09/28/2007	22:35:40	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:34	
		Control Test	0.079	22:35	
		Air Blank	0.000	22:35	
MANATEE COUNTY SO	09/28/2007	23:51:11	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:50	
		Control Test	0.079	23:50	

# Control Tests

Instrument Serial #: 80-001150

Location	Date	Time	Version	Target	Source
MANATEE COUNTY SO	09/28/2007	23:51:11	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	23:51	
MANATEE COUNTY SO	09/29/2007	01:26:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:25	
		Control Test	0.077	01:26	
		Air Blank	0.000	01:26	
MANATEE COUNTY SO	09/29/2007	01:27:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:27	
		Control Test	0.078	01:27	
		Air Blank	0.000	01:27	
MANATEE COUNTY SO	09/29/2007	02:01:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:01	
		Control Test	0.078	02:01	
		Air Blank	0.000	02:01	
MANATEE COUNTY SO	10/19/2007	22:28:07	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:27	
		Control Test	0.079	22:27	
		Air Blank	0.000	22:28	
MANATEE COUNTY SO	10/19/2007	22:34:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:33	
		Control Test	0.079	22:34	
		Air Blank	0.000	22:34	
MANATEE COUNTY SO	11/06/2007	08:49:29	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:48	
		Control Test	0.080	08:49	
		Air Blank	0.000	08:49	
MANATEE COUNTY SO	11/09/2007	22:50:38	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:49	
		Control Test	0.080	22:50	
		Air Blank	0.000	22:50	

# Control Tests

Instrument Serial #: 80-001157

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
LAKE WALES PD	07/06/2007	09:58:07	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:57	
		Control Test	0.079	09:57	
		Air Blank	0.000	09:58	

## Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	03/15/2007	04:00:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:59	
		Control Test	0.080	03:59	
		Air Blank	0.000	04:00	
PONCE INLET P.D.	03/18/2007	17:54:34	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:53	
		Control Test	0.044	17:54	
		Air Blank	0.000	17:54	
PONCE INLET P.D.	03/18/2007	18:17:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:16	
		Control Test	0.047	18:16	
		Air Blank	0.000	18:17	
PONCE INLET P.D.	03/18/2007	18:19:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:18	
		Control Test	0.047	18:19	
		Air Blank	0.000	18:19	
PONCE INLET P.D.	03/18/2007	18:21:39	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:20	
		Control Test	0.048	18:21	
		Air Blank	0.000	18:21	
PONCE INLET P.D.	03/18/2007	18:23:37	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:22	
		Control Test	0.047	18:23	
		Air Blank	0.000	18:23	
PONCE INLET P.D.	03/18/2007	18:25:37	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:24	
		Control Test	0.048	18:25	
		Air Blank	0.000	18:25	
PONCE INLET P.D.	03/18/2007	18:28:11	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:26	
		Control Test	0.075	18:27	

# Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	03/18/2007	18:28:11	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:28	
PONCE INLET P.D.	03/18/2007	18:34:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:33	
		Control Test	0.077	18:34	
		Air Blank	0.000	18:34	
PONCE INLET P.D.	03/18/2007	18:37:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:36	
		Control Test	0.076	18:36	
		Air Blank	0.000	18:37	
PONCE INLET P.D.	03/18/2007	18:40:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:38	
		Control Test	0.077	18:39	
		Air Blank	0.000	18:40	
PONCE INLET P.D.	03/18/2007	18:42:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:40	
		Control Test	0.076	18:41	
		Air Blank	0.000	18:42	
PONCE INLET P.D.	03/18/2007	18:44:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:43	
		Control Test	0.047	18:43	
		Air Blank	0.000	18:44	
PONCE INLET P.D.	03/18/2007	18:46:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:45	
		Control Test	0.076	18:46	
		Air Blank	0.000	18:46	
PONCE INLET P.D.	03/18/2007	18:49:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:48	
		Control Test	0.078	18:48	
		Air Blank	0.000	18:49	

# Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	03/18/2007	18:51:49	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:50	
		Control Test	0.047	18:51	
		Air Blank	0.000	18:51	
PONCE INLET P.D.	03/18/2007	18:54:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:52	
		Control Test	0.047	18:53	
		Air Blank	0.000	18:54	
PONCE INLET P.D.	03/18/2007	18:57:04	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:55	
		Control Test	0.195	18:56	
		Air Blank	0.000	18:57	
PONCE INLET P.D.	03/18/2007	18:59:26	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:58	
		Control Test	0.196	18:58	
		Air Blank	0.000	18:59	
PONCE INLET P.D.	03/18/2007	19:02:13	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:00	
		Control Test	0.197	19:01	
		Air Blank	0.000	19:02	
PONCE INLET P.D.	03/18/2007	19:04:21	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:03	
		Control Test	0.196	19:03	
		Air Blank	0.000	19:04	
PONCE INLET P.D.	03/18/2007	19:06:33	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:05	
		Control Test	0.197	19:05	
		Air Blank	0.000	19:06	
PONCE INLET P.D.	03/18/2007	19:08:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:07	
		Control Test	0.197	19:08	

# Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	03/18/2007	19:08:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:08	
PONCE INLET P.D.	03/18/2007	19:27:13	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:26	
		Control Test	0.080	19:26	
		Air Blank	0.000	19:27	
PONCE INLET P.D.	04/05/2007	18:56:43	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:55	
		Control Test	0.044	18:56	
		Air Blank	0.000	18:56	
PONCE INLET P.D.	04/05/2007	18:58:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:57	
		Control Test	0.046	18:58	
		Air Blank	0.000	18:58	
PONCE INLET P.D.	04/05/2007	19:01:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:00	
		Control Test	0.045	19:01	
		Air Blank	0.000	19:01	
PONCE INLET P.D.	04/05/2007	19:05:33	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:04	
		Control Test	0.046	19:05	
		Air Blank	0.000	19:05	
PONCE INLET P.D.	04/05/2007	19:06:30	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		19:06	Ambient Fail
		Air Blank	0.000	19:06	
PONCE INLET P.D.	04/05/2007	19:09:06	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:07	
		Control Test	0.047	19:08	
		Air Blank	0.000	19:09	
PONCE INLET P.D.	04/05/2007	19:12:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:11	

# Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	04/05/2007	19:12:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Control Test	0.046	19:11	
		Air Blank	0.000	19:12	
PONCE INLET P.D.	04/05/2007	19:14:35	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:13	
		Control Test	0.073	19:14	
		Air Blank	0.000	19:14	
PONCE INLET P.D.	04/05/2007	19:19:53	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:18	
		Control Test	0.074	19:19	
		Air Blank	0.000	19:19	
PONCE INLET P.D.	04/05/2007	19:22:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:21	
		Control Test	0.076	19:22	
		Air Blank	0.000	19:22	
PONCE INLET P.D.	04/05/2007	19:25:25	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:24	
		Control Test	0.075	19:24	
		Air Blank	0.000	19:25	
PONCE INLET P.D.	04/05/2007	19:27:54	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:26	
		Control Test	0.075	19:27	
		Air Blank	0.000	19:27	
PONCE INLET P.D.	04/05/2007	19:29:54	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:28	
		Control Test	0.193	19:29	
		Air Blank	0.000	19:29	
PONCE INLET P.D.	04/05/2007	19:31:52	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:30	
		Control Test	0.193	19:31	
		Air Blank	0.000	19:31	

# Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	04/05/2007	19:33:37	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:32	
		Control Test	0.045	19:33	
		Air Blank	0.000	19:33	
PONCE INLET P.D.	04/05/2007	19:39:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:37	
		Control Test	0.046	19:38	
		Air Blank	0.000	19:39	
PONCE INLET P.D.	04/05/2007	19:42:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:41	
		Control Test	0.075	19:41	
		Air Blank	0.000	19:42	
PONCE INLET P.D.	04/05/2007	19:44:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:42	
		Control Test	0.075	19:43	
		Air Blank	0.000	19:44	
PONCE INLET P.D.	04/05/2007	19:46:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:44	
		Control Test	0.076	19:45	
		Air Blank	0.000	19:46	
PONCE INLET P.D.	04/05/2007	19:48:03	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:46	
		Control Test	0.046	19:47	
		Air Blank	0.000	19:48	
PONCE INLET P.D.	05/06/2007	18:56:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:54	
		Control Test	0.045	18:55	
		Air Blank	0.000	18:56	
PONCE INLET P.D.	05/06/2007	18:59:42	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:58	
		Control Test	0.046	18:59	

# Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	05/06/2007	18:59:42	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:59	
PONCE INLET P.D.	05/06/2007	19:04:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:03	
		Control Test	0.047	19:04	
		Air Blank	0.000	19:04	
PONCE INLET P.D.	05/06/2007	19:09:09	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:07	
		Control Test	0.048	19:08	
		Air Blank	0.000	19:09	
PONCE INLET P.D.	05/06/2007	19:13:31	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:12	
		Control Test	0.076	19:12	
		Air Blank	0.000	19:13	
PONCE INLET P.D.	05/06/2007	19:16:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:15	
		Control Test	0.078	19:15	
		Air Blank	0.000	19:16	
PONCE INLET P.D.	05/06/2007	19:24:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:23	
		Control Test	0.196	19:24	
		Air Blank	0.000	19:24	
PONCE INLET P.D.	06/02/2007	03:08:45	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:07	
		Control Test	0.076	03:08	
		Air Blank	0.000	03:08	
PONCE INLET P.D.	06/02/2007	03:10:51	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:09	
		Control Test	0.076	03:10	
		Air Blank	0.000	03:10	

## Control Tests

Instrument Serial #: 80-001158

Location	Date	Time	Version	Target	Source
PONCE INLET P.D.	06/02/2007	03:13:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:12	
		Control Test	0.045	03:12	
		Air Blank	0.000	03:13	
PONCE INLET P.D.	06/02/2007	03:15:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:14	
		Control Test	0.046	03:14	
		Air Blank	0.000	03:15	
PONCE INLET P.D.	09/11/2007	22:23:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:22	
		Control Test	0.044	22:23	
		Air Blank	0.000	22:23	
PONCE INLET P.D.	09/11/2007	22:31:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:30	
		Control Test	0.046	22:31	
		Air Blank	0.000	22:31	
PONCE INLET P.D.	09/11/2007	22:33:50	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:32	
		Control Test	0.075	22:33	
		Air Blank	0.000	22:33	
PONCE INLET P.D.	09/11/2007	22:35:50	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:34	
		Control Test	0.193	22:35	
		Air Blank	0.000	22:35	
PONCE INLET P.D.	12/22/2007	05:11:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		05:11	RFI Detect
		Air Blank		05:11	RFI Detect
PONCE INLET P.D.	12/22/2007	05:13:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:12	
		Control Test	0.071	05:13	
		Air Blank	0.000	05:13	

# Control Tests

Instrument Serial #: 80-001158

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PONCE INLET P.D.	12/22/2007	05:15:21	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:14	
		Control Test	0.074	05:14	
		Air Blank	0.000	05:15	
PONCE INLET P.D.	12/22/2007	05:16:54	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:16	
		Control Test	0.075	05:16	
		Air Blank	0.000	05:16	

# Control Tests

Instrument Serial #: 80-001159

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ORANGE CITY P.D.	05/23/2007	09:15:12	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:14	
		Control Test	0.000	09:14	
		Air Blank	0.000	09:15	

# Control Tests

Instrument Serial #: 80-001174

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CLERMONT P.D.	02/15/2007	14:31:57	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:31	
		ITP Check	0.080	14:31	
		Air Blank	0.000	14:31	
CLERMONT P.D.	02/15/2007	14:33:20	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:32	
		Control Test	0.070	14:32	
		Air Blank	0.000	14:33	
CLERMONT P.D.	02/15/2007	14:53:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:52	
		Control Test	0.069	14:53	Interferent Detect
		Air Blank	0.000	14:53	
CLERMONT P.D.	04/18/2007	13:00:53	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:00	
		Control Test	0.079	13:00	
		Air Blank	0.000	13:00	

# Control Tests

Instrument Serial #: 80-001179

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MIAMI P.D.	10/17/2007	19:44:24	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:43	
		Control Test	0.000	19:43	
		Air Blank	0.000	19:44	
MIAMI P.D.	10/18/2007	16:53:17	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:52	
		Control Test	0.082	16:52	
		Air Blank	0.000	16:53	
MIAMI P.D.	11/10/2007	10:56:07	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:54	
		Control Test	0.089	10:55	Interferent Detect
		Air Blank	0.000	10:56	
MIAMI P.D.	12/04/2007	10:17:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:14	
		Control Test	0.183	10:15	
		Air Blank		10:16	Purge Fail
		Air Blank		10:17	Purge Fail

# Control Tests

Instrument Serial #: 80-001201

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COLLIER COUNTY SO	09/13/2007	11:01:59	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:01	
		ITP Check	0.081	11:01	
		Air Blank	0.000	11:01	

# Control Tests

Instrument Serial #: 80-001202

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COLLIER COUNTY SO	09/13/2007	11:02:39	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:01	
		ITP Check	0.078	11:02	
		Air Blank	0.000	11:02	

# Control Tests

Instrument Serial #: 80-001203

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
GREENACRES P S	07/22/2007	08:01:14	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:00	
		Control Test	0.082	08:00	
		Air Blank	0.000	08:01	
GREENACRES P S	07/22/2007	08:02:47	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:01	
		ITP Check	0.079	08:02	
		Air Blank	0.000	08:02	
GREENACRES P S	07/24/2007	13:21:24	8100.27.00	0.080	Remote Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:20	
		Control Test	0.082	13:20	
		Air Blank	0.000	13:21	
GREENACRES P S	07/24/2007	13:30:22	8100.27.00	0.080	Remote Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:29	
		Control Test	0.081	13:29	
		Air Blank	0.000	13:30	

# Control Tests

Instrument Serial #: 80-001207

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HENDRY COUNTY SO	09/27/2007	14:10:54	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:10	
		ITP Check	0.079	14:10	
		Air Blank	0.000	14:10	

# Control Tests

Instrument Serial #: 80-001208

Location	Date	Time	Version	Target	Source
LEE COUNTY SO	02/20/2007	14:20:35	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		14:20	Sequence Aborted
		Air Blank	0.000	14:20	
LEE COUNTY SO	12/04/2007	16:34:58	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:34	
		ITP Check	0.080	16:34	
		Air Blank	0.000	16:34	
LEE COUNTY SO	12/04/2007	16:49:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:48	
		Control Test	0.079	16:48	
		Air Blank	0.000	16:49	
LEE COUNTY SO	12/04/2007	16:50:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:50	
		Control Test	0.000	16:50	
		Air Blank	0.000	16:50	
LEE COUNTY SO	12/04/2007	16:53:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:52	
		Control Test	0.000	16:52	
		Air Blank	0.000	16:53	
LEE COUNTY SO	12/04/2007	16:54:57	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:53	
		Control Test	0.199	16:54	
		Air Blank	0.000	16:54	

# Control Tests

Instrument Serial #: 80-001231

Location	Date	Time	Version	Target	Source
AUBURNDALE PD	08/19/2007	10:46:30	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:45	
		Control Test	0.000	10:46	
		Air Blank	0.000	10:46	
AUBURNDALE PD	08/19/2007	11:33:30	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:32	
		Control Test	0.000	11:33	
		Air Blank	0.000	11:33	
AUBURNDALE PD	08/19/2007	11:35:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:34	
		Control Test	0.200	11:35	
		Air Blank	0.000	11:35	
AUBURNDALE PD	08/19/2007	11:38:23	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:37	
		Control Test	0.080	11:37	
		Air Blank	0.000	11:38	
AUBURNDALE PD	08/19/2007	11:43:52	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:42	
		Control Test	0.200	11:43	
		Air Blank	0.000	11:43	
AUBURNDALE PD	08/19/2007	13:24:36	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:23	
		Control Test	0.000	13:24	
		Air Blank	0.000	13:24	
AUBURNDALE PD	08/19/2007	13:30:03	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:28	
		Control Test	0.080	13:29	
		Air Blank	0.000	13:30	
AUBURNDALE PD	08/19/2007	13:31:47	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:30	
		Control Test	0.199	13:31	

# Control Tests

Instrument Serial #: 80-001231

Location	Date	Time	Version	Target	Source
AUBURNDALE PD	08/19/2007	13:31:47	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:31	
AUBURNDALE PD	08/19/2007	14:13:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:12	
		Control Test	0.049	14:12	
		Air Blank	0.000	14:13	
AUBURNDALE PD	08/19/2007	14:16:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:15	
		Control Test	0.079	14:15	
		Air Blank	0.000	14:16	
AUBURNDALE PD	10/29/2007	02:58:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:56	
		Control Test	0.047	02:57	
		Air Blank	0.000	02:58	
AUBURNDALE PD	10/29/2007	03:00:24	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:59	
		Control Test	0.078	02:59	
		Air Blank	0.000	03:00	
AUBURNDALE PD	10/29/2007	03:02:53	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:01	
		Control Test	0.200	03:02	
		Air Blank	0.000	03:02	
AUBURNDALE PD	11/26/2007	02:21:35	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:20	
		Control Test	0.000	02:21	
		Air Blank	0.000	02:21	
AUBURNDALE PD	11/26/2007	02:23:21	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:22	
		Control Test	0.000	02:22	
		Air Blank	0.000	02:23	

# Control Tests

Instrument Serial #: 80-001231

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
AUBURNDALE PD	11/26/2007	02:25:21	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:24	
		Control Test	0.047	02:24	
		Air Blank	0.000	02:25	
AUBURNDALE PD	11/26/2007	02:27:03	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:25	
		Control Test	0.078	02:26	
		Air Blank	0.000	02:27	
AUBURNDALE PD	11/26/2007	02:29:11	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:27	
		Control Test	0.196	02:28	
		Air Blank	0.000	02:29	

# Control Tests

Instrument Serial #: 80-001245

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PORT ORANGE P.D.	04/09/2007	14:30:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:29	
		Control Test	0.195	14:29	
		Air Blank	0.000	14:30	
PORT ORANGE P.D.	04/29/2007	02:22:00	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:21	
		Control Test	0.076	02:21	
		Air Blank	0.000	02:22	
PORT ORANGE P.D.	05/02/2007	14:31:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:30	
		Control Test	0.078	14:30	
		Air Blank	0.000	14:31	
PORT ORANGE P.D.	10/27/2007	07:48:43	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	07:47	
		Control Test	0.000	07:48	Interferent Detect
		Air Blank	0.000	07:48	

# Control Tests

Instrument Serial #: 80-001246

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SOUTH DAYTONA P.D.	08/31/2007	16:56:56	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:55	
		Control Test	0.000	16:56	
		Air Blank	0.000	16:56	
SOUTH DAYTONA P.D.	08/31/2007	17:00:44	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:59	
		Control Test	0.044	17:00	
		Air Blank	0.000	17:00	
SOUTH DAYTONA P.D.	08/31/2007	17:20:54	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:19	
		Control Test	0.045	17:20	
		Air Blank	0.000	17:20	
SOUTH DAYTONA P.D.	08/31/2007	17:24:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:23	
		Control Test	0.000	17:24	Interferent Detect
		Air Blank	0.000	17:24	

# Control Tests

Instrument Serial #: 80-001248

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BROWARD COUNTY SO	03/27/2007	11:48:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:47	
		Control Test	0.079	11:47	
		Air Blank	0.000	11:48	
BROWARD COUNTY SO	09/11/2007	13:10:16	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:09	
		Control Test	0.039	13:09	
		Air Blank	0.000	13:10	
BROWARD COUNTY SO	09/17/2007	10:59:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:58	
		Control Test	0.080	10:58	
		Air Blank	0.000	10:59	
BROWARD COUNTY SO	09/17/2007	11:43:10	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:41	
		Control Test	0.040	11:42	
		Air Blank	0.000	11:43	
BROWARD COUNTY SO	09/17/2007	12:26:57	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:25	
		Control Test	0.192	12:26	
		Air Blank	0.000	12:26	
BROWARD COUNTY SO	09/17/2007	13:35:28	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:34	
		Control Test	0.201	13:34	
		Air Blank	0.000	13:35	
BROWARD COUNTY SO	10/26/2007	13:44:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:43	
		Control Test	0.082	13:44	
		Air Blank	0.000	13:44	

# Control Tests

Instrument Serial #: 80-001249

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BROWARD CO SO	08/03/2007	12:17:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:16	
		Control Test	0.075	12:17	
		Air Blank	0.000	12:17	
BROWARD CO SO	08/03/2007	12:20:52	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:20	
		Control Test	0.075	12:20	
		Air Blank	0.000	12:20	
BROWARD CO SO	08/03/2007	12:22:54	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:22	
		Control Test	0.075	12:22	
		Air Blank	0.000	12:22	
BROWARD CO SO	08/29/2007	12:14:21	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:13	
		Control Test	0.074	12:13	
		Air Blank	0.000	12:14	
BROWARD CO SO	08/29/2007	12:15:41	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:14	
		Control Test	0.074	12:15	
		Air Blank	0.000	12:15	
BROWARD CO SO	10/10/2007	16:54:53	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:53	
		Control Test	0.000	16:54	
		Air Blank	0.000	16:54	

# Control Tests

Instrument Serial #: 80-001255

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ORANGE COUNTY SO	08/31/2007	21:51:56	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:51	
		Control Test	0.079	21:51	
		Air Blank	0.000	21:51	

# Control Tests

Instrument Serial #: 80-001257

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ORANGE COUNTY SO	11/09/2007	22:42:27	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:41	
		ITP Check	0.081	22:41	
		Air Blank	0.000	22:42	

# Control Tests

Instrument Serial #: 80-001258

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ORANGE COUNTY S.O.	11/09/2007	22:13:10	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:12	
		ITP Check	0.081	22:12	
		Air Blank	0.000	22:13	

## Control Tests

Instrument Serial #: 80-001259

Location	Date	Time	Version	Target	Source			
PINELLAS COUNTY SO	02/10/2007	01:36:08	8100.27.00	0.080	Local Dry gas			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	01:35
						Control Test	0.072	01:35
PINELLAS COUNTY SO	02/10/2007	01:37:38	8100.27.00	0.080	Local ITP			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	01:36
						ITP Check	0.079	01:37
PINELLAS COUNTY SO	10/20/2007	21:36:29	8100.27.00	0.080	Local Dry gas			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	21:35
						Control Test	0.078	21:35
PINELLAS COUNTY SO	10/20/2007	21:39:40	8100.27.00	0.080	Local ITP			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	21:38
						ITP Check	0.082	21:39
PINELLAS COUNTY SO	10/20/2007	21:47:25	8100.27.00	0.080	Local Wet bath			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	21:46
						Control Test	0.000	21:46
PINELLAS COUNTY SO	11/28/2007	21:33:10	8100.27.00	0.080	Local Dry gas			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	21:32
						Control Test	0.077	21:32
PINELLAS COUNTY SO	11/28/2007	21:34:36	8100.27.00	0.080	Local Dry gas			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	21:33
						Control Test	0.077	21:34
PINELLAS COUNTY SO	12/12/2007	22:06:56	8100.27.00	0.080	Local Dry gas			
						<b>Test</b>	<b>Value</b>	<b>Time</b>
						Air Blank	0.000	22:06
						Control Test	0.077	22:06

# Control Tests

Instrument Serial #: 80-001259

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	12/12/2007	22:06:56	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:06	
PINELLAS COUNTY SO	12/12/2007	22:08:22	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:07	
		ITP Check	0.081	22:07	
		Air Blank	0.000	22:08	
PINELLAS COUNTY SO	12/29/2007	02:25:05	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:24	
		Control Test	0.076	02:24	
		Air Blank	0.000	02:25	
PINELLAS COUNTY SO	12/29/2007	02:26:40	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:25	
		ITP Check	0.081	02:26	
		Air Blank	0.000	02:26	

# Control Tests

Instrument Serial #: 80-001263

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
INDIAN HARBOUR BEACH	09/19/2007	12:16:59	8100.27.00	0.080	Remote Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:16	
		Control Test	0.078	12:16	
		Air Blank	0.000	12:16	

# Control Tests

Instrument Serial #: 80-001264

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MELBOURNE BEACH PD	01/26/2007	11:14:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:13	
		Control Test	0.075	11:14	
		Air Blank	0.000	11:14	
MELBOURNE BEACH PD	01/26/2007	11:15:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:15	
		Control Test	0.077	11:15	
		Air Blank	0.000	11:15	
MELBOURNE BEACH PD	04/24/2007	09:38:14	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:37	
		Control Test	0.079	09:37	
		Air Blank	0.000	09:38	
MELBOURNE BEACH PD	08/26/2007	04:07:46	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	04:06	
		Control Test	0.076	04:07	
		Air Blank	0.000	04:07	
MELBOURNE BEACH PD	09/08/2007	17:48:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:47	
		Control Test	0.076	17:48	
		Air Blank	0.000	17:48	

# Control Tests

Instrument Serial #: 80-001265

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PALM BAY PD	10/25/2007	00:13:39	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:12	
		Control Test	0.075	00:13	
		Air Blank	0.000	00:13	
PALM BAY PD	10/25/2007	00:15:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:14	
		Control Test	0.077	00:14	
		Air Blank	0.000	00:15	
PALM BAY PD	10/25/2007	00:16:40	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:15	
		Control Test	0.077	00:16	
		Air Blank	0.000	00:16	
PALM BAY PD	10/25/2007	00:18:16	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:17	
		ITP Check	0.081	00:17	
		Air Blank	0.000	00:18	

# Control Tests

Instrument Serial #: 80-001269

Location	Date	Time	Version	Target	Source
PALM BEACH CO SO	02/08/2007	16:57:17	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:56	
		Control Test	0.079	16:56	
		Air Blank	0.000	16:57	
PALM BEACH CO SO	06/18/2007	12:36:13	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:35	
		Control Test	0.078	12:35	
		Air Blank	0.000	12:36	
PALM BEACH CO SO	06/18/2007	12:37:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:36	
		Control Test	0.078	12:37	
		Air Blank	0.000	12:37	
PALM BEACH CO SO	06/18/2007	13:00:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:00	
		Control Test	0.078	13:00	
		Air Blank	0.000	13:00	
PALM BEACH CO SO	06/18/2007	13:02:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		13:01	Ambient Fail
		Air Blank	0.000	13:02	
PALM BEACH CO SO	06/18/2007	13:03:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:02	
		Control Test	0.079	13:03	
		Air Blank	0.000	13:03	

## Control Tests

Instrument Serial #: 80-001271

Location	Date	Time	Version	Target	Source
SEMINOLE S.O.	05/25/2007	22:29:02	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:28	
		Control Test	0.076	22:28	
		Air Blank	0.000	22:29	
SEMINOLE S.O.	05/26/2007	04:05:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	04:05	
		Control Test	0.073	04:05	
		Air Blank	0.000	04:05	
SEMINOLE S.O.	05/26/2007	04:08:03	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	04:07	
		Control Test	0.075	04:07	
		Air Blank	0.000	04:08	
SEMINOLE S.O.	07/25/2007	18:31:34	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:30	
		Control Test	0.076	18:31	
		Air Blank	0.000	18:31	
SEMINOLE S.O.	09/22/2007	02:37:16	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:36	
		Control Test	0.077	02:36	
		Air Blank	0.000	02:37	
SEMINOLE S.O.	09/22/2007	02:38:39	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:37	
		Control Test	0.077	02:38	
		Air Blank	0.000	02:38	
SEMINOLE S.O.	10/03/2007	01:28:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:27	
		Control Test	0.077	01:27	
		Air Blank	0.000	01:28	
SEMINOLE S.O.	10/03/2007	01:35:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:34	
		Control Test	0.076	01:34	

# Control Tests

Instrument Serial #: 80-001271

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SEMINOLE S.O.	10/03/2007	01:35:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:35	
SEMINOLE S.O.	10/03/2007	01:36:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:35	
		Control Test	0.078	01:36	
		Air Blank	0.000	01:36	
SEMINOLE S.O.	10/20/2007	00:32:30	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:31	
		Control Test	0.073	00:32	
		Air Blank	0.000	00:32	
SEMINOLE S.O.	10/20/2007	00:33:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:33	
		Control Test	0.075	00:33	
		Air Blank	0.000	00:33	
SEMINOLE S.O.	10/20/2007	00:35:30	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:34	
		Control Test	0.076	00:35	
		Air Blank	0.000	00:35	
SEMINOLE S.O.	11/25/2007	03:50:11	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:49	
		Control Test	0.078	03:49	
		Air Blank	0.000	03:50	

# Control Tests

Instrument Serial #: 80-001272

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SEMINOLE S.O.	05/05/2007	21:50:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:49	
		Control Test	0.078	21:49	
		Air Blank	0.000	21:50	
SEMINOLE S.O.	12/23/2007	01:11:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:11	
		Control Test	0.079	01:11	
		Air Blank	0.000	01:11	

# Control Tests

Instrument Serial #: 80-001274

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PINELLAS COUNTY SO	11/29/2007	19:30:49	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:29	
		Control Test	0.079	19:30	
		Air Blank	0.000	19:30	
PINELLAS COUNTY SO	11/29/2007	19:33:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:32	
		Control Test	0.080	19:32	
		Air Blank	0.000	19:33	
PINELLAS COUNTY SO	11/29/2007	19:35:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:33	
		Control Test	0.080	19:34	
		Air Blank	0.000	19:35	
PINELLAS COUNTY SO	11/29/2007	19:36:33	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:35	
		ITP Check	0.080	19:36	
		Air Blank	0.000	19:36	

# Control Tests

Instrument Serial #: 80-001275

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HIALEAH GARDENS P.D.	09/11/2007	14:29:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:28	
		Control Test	0.079	14:28	
		Air Blank	0.000	14:29	

# Control Tests

Instrument Serial #: 80-001276

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	06/02/2007	01:29:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:29	
		Control Test	0.079	01:29	
		Air Blank	0.000	01:29	
MIAMI-DADE PD	06/23/2007	03:30:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:29	
		Control Test	0.074	03:30	
		Air Blank	0.000	03:30	
MIAMI-DADE PD	06/23/2007	03:32:21	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:31	
		Control Test	0.076	03:31	
		Air Blank	0.000	03:32	
MIAMI-DADE PD	06/23/2007	03:37:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:36	
		Control Test	0.075	03:36	
		Air Blank	0.000	03:37	
MIAMI-DADE PD	06/23/2007	03:49:24	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:48	
		Control Test	0.076	03:48	
		Air Blank	0.000	03:49	
MIAMI-DADE PD	06/23/2007	03:54:07	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:53	
		Control Test	0.077	03:53	
		Air Blank	0.000	03:54	
MIAMI-DADE PD	06/23/2007	03:57:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:56	
		Control Test	0.077	03:57	
		Air Blank	0.000	03:57	
MIAMI-DADE PD	06/24/2007	00:11:39	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:10	
		Control Test	0.076	00:11	

# Control Tests

Instrument Serial #: 80-001276

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	06/24/2007	00:11:39	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:11	
MIAMI-DADE PD	06/24/2007	01:27:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:27	
		Control Test	0.077	01:27	
		Air Blank	0.000	01:27	
MIAMI-DADE PD	08/11/2007	02:45:25	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:44	
		Control Test	0.079	02:44	
		Air Blank	0.000	02:45	
MIAMI-DADE PD	10/12/2007	22:50:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:49	
		Control Test	0.077	22:49	
		Air Blank	0.000	22:50	
MIAMI-DADE PD	11/03/2007	01:06:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:05	
		Control Test	0.077	01:06	
		Air Blank	0.000	01:06	
MIAMI-DADE PD	11/03/2007	01:08:02	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:07	
		Control Test	0.077	01:07	
		Air Blank	0.000	01:08	
MIAMI-DADE PD	12/07/2007	22:38:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:37	
		Control Test	0.000	22:37	
		Air Blank	0.000	22:38	
MIAMI-DADE PD	12/07/2007	22:39:39	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:38	
		Control Test	0.077	22:39	
		Air Blank	0.000	22:39	

# Control Tests

Instrument Serial #: 80-001277

Location	Date	Time	Version	Target	Source
MIAMI-DADE PD	06/02/2007	00:54:22	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:53	
		Control Test	0.075	00:53	
		Air Blank	0.000	00:54	
MIAMI-DADE PD	06/02/2007	01:05:20	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:04	
		Control Test	0.076	01:04	
		Air Blank	0.000	01:05	
MIAMI-DADE PD	06/02/2007	01:36:10	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:35	
		Control Test	0.076	01:35	
		Air Blank	0.000	01:36	
MIAMI-DADE PD	06/02/2007	03:24:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	03:23	
		Control Test	0.077	03:24	
		Air Blank	0.000	03:24	
MIAMI-DADE PD	06/09/2007	22:31:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	22:30	
		Control Test	0.078	22:31	
		Air Blank	0.000	22:31	
MIAMI-DADE PD	10/06/2007	19:55:48	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	19:54	
		Control Test	0.076	19:55	
		Air Blank	0.000	19:55	
MIAMI-DADE PD	10/06/2007	20:21:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:20	
		Control Test	0.076	20:21	
		Air Blank	0.000	20:21	

# Control Tests

Instrument Serial #: 80-001278

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
JACKSONVILLE SO	08/14/2007	13:18:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:17	
		Control Test	0.081	13:17	
		Air Blank	0.000	13:18	

# Control Tests

Instrument Serial #: 80-001280

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
JACKSONVILLE SO	02/20/2007	11:11:20	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:10	
		Control Test	0.080	11:10	
		Air Blank	0.000	11:11	
JACKSONVILLE SO	09/14/2007	05:04:08	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:03	
		Control Test	0.068	05:03	
		Air Blank	0.000	05:04	
JACKSONVILLE SO	09/14/2007	05:06:54	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:06	
		Control Test	0.072	05:06	
		Air Blank	0.000	05:06	
JACKSONVILLE SO	09/14/2007	05:08:45	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:07	
		Control Test	0.073	05:08	
		Air Blank	0.000	05:08	
JACKSONVILLE SO	09/14/2007	05:10:58	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:10	
		Control Test	0.074	05:10	
		Air Blank	0.000	05:10	
JACKSONVILLE SO	09/14/2007	06:00:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:00	
		Control Test	0.073	06:00	
		Air Blank	0.000	06:00	

## Control Tests

Instrument Serial #: 80-001282

Location	Date	Time	Version	Target	Source
U OF FLORIDA PD	04/27/2007	09:56:11	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:55	
		Control Test	0.069	09:55	
		Air Blank	0.000	09:56	
U OF FLORIDA PD	05/10/2007	13:31:17	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:30	
		Control Test	0.078	13:30	
		Air Blank	0.000	13:31	
U OF FLORIDA PD	05/10/2007	13:35:44	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:34	
		Control Test	0.077	13:35	
		Air Blank	0.000	13:35	
U OF FLORIDA PD	05/10/2007	14:18:21	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:17	
		Control Test	0.078	14:17	
		Air Blank	0.000	14:18	
U OF FLORIDA PD	05/10/2007	15:59:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:57	
		Control Test	0.047	15:58	
		Air Blank	0.000	15:59	
U OF FLORIDA PD	06/23/2007	15:26:23	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:25	
		Control Test	0.193	15:25	
		Air Blank	0.000	15:26	
U OF FLORIDA PD	06/23/2007	16:05:09	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:03	
		Control Test	0.191	16:04	
		Air Blank	0.000	16:05	
U OF FLORIDA PD	06/23/2007	16:14:46	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:13	
		Control Test	0.195	16:14	

# Control Tests

Instrument Serial #: 80-001282

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
U OF FLORIDA PD	06/23/2007	16:14:46	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:14	

# Control Tests

Instrument Serial #: 80-001284

Location	Date	Time	Version	Target	Source
GILCHRIST COUNTY SO	01/17/2007	08:48:15	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:47	
		Control Test	0.044	08:47	
		Air Blank	0.000	08:48	
GILCHRIST COUNTY SO	01/17/2007	08:52:03	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:50	
		Control Test	0.045	08:51	
		Air Blank	0.000	08:52	
GILCHRIST COUNTY SO	01/17/2007	08:55:17	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:54	
		Control Test	0.000	08:54	
		Air Blank	0.000	08:55	
GILCHRIST COUNTY SO	01/17/2007	08:57:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:56	
		Control Test	0.076	08:56	
		Air Blank	0.000	08:57	
GILCHRIST COUNTY SO	01/17/2007	08:59:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:58	
		Control Test	0.194	08:58	
		Air Blank	0.000	08:59	
GILCHRIST COUNTY SO	01/17/2007	09:01:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:59	
		Control Test	0.046	09:00	
		Air Blank	0.000	09:01	
GILCHRIST COUNTY SO	03/05/2007	11:59:51	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:58	
		Control Test	0.046	11:59	
		Air Blank	0.000	11:59	
GILCHRIST COUNTY SO	05/20/2007	11:50:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:49	
		Control Test	0.080	11:50	

## Control Tests

Instrument Serial #: 80-001284

Location	Date	Time	Version	Target	Source
GILCHRIST COUNTY SO	05/20/2007	11:50:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:50	
GILCHRIST COUNTY SO	05/20/2007	11:53:00	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:51	
		Control Test	0.190	11:52	
		Air Blank	0.000	11:53	
GILCHRIST COUNTY SO	05/20/2007	13:57:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:56	
		Control Test	0.081	13:56	
		Air Blank	0.000	13:57	
GILCHRIST COUNTY SO	06/15/2007	08:21:58	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:20	
		Control Test	0.191	08:21	
		Air Blank	0.000	08:21	
GILCHRIST COUNTY SO	08/15/2007	10:35:39	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:34	
		Control Test	0.043	10:35	
		Air Blank	0.000	10:35	
GILCHRIST COUNTY SO	08/15/2007	10:40:19	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		10:39	Ambient Fail
		Air Blank	0.000	10:40	
GILCHRIST COUNTY SO	08/15/2007	10:42:14	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:41	
		Control Test	0.045	10:41	
		Air Blank	0.000	10:42	
GILCHRIST COUNTY SO	08/15/2007	10:44:11	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:42	
		Control Test	0.046	10:43	
		Air Blank	0.000	10:44	
GILCHRIST COUNTY SO	09/12/2007	08:08:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:07	

# Control Tests

Instrument Serial #: 80-001284

Location	Date	Time	Version	Target	Source
GILCHRIST COUNTY SO	09/12/2007	08:08:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Control Test	0.044	08:08	
		Air Blank	0.000	08:08	
GILCHRIST COUNTY SO	09/12/2007	08:10:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:09	
		Control Test	0.080	08:10	
		Air Blank	0.000	08:10	
GILCHRIST COUNTY SO	09/12/2007	08:40:28	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:39	
		Control Test	0.045	08:39	
		Air Blank	0.000	08:40	
GILCHRIST COUNTY SO	10/24/2007	07:12:24	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	07:11	
		ITP Check	0.080	07:11	
		Air Blank	0.000	07:12	
GILCHRIST COUNTY SO	10/24/2007	07:14:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	07:13	
		Control Test	0.189	07:14	
		Air Blank	0.000	07:14	
GILCHRIST COUNTY SO	10/24/2007	07:16:38	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	07:15	
		Control Test	0.075	07:16	
		Air Blank	0.000	07:16	
GILCHRIST COUNTY SO	10/24/2007	07:18:51	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	07:17	
		Control Test	0.045	07:18	
		Air Blank	0.000	07:18	

# Control Tests

Instrument Serial #: 80-001285

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
GILCHRIST COUNTY SO	03/05/2007	12:57:33	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:56	
		Control Test	0.080	12:57	
		Air Blank	0.000	12:57	
GILCHRIST COUNTY SO	05/20/2007	12:38:35	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:37	
		ITP Check	0.079	12:38	
		Air Blank	0.000	12:38	

# Control Tests

Instrument Serial #: 80-001291

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BRADFORD COUNTY SO	06/03/2007	20:38:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:38	
		Control Test	0.077	20:38	
		Air Blank	0.000	20:38	

# Control Tests

Instrument Serial #: 80-001292

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
LEVY COUNTY SO	05/31/2007	09:33:29	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:32	
		ITP Check	0.079	09:33	
		Air Blank	0.000	09:33	

# Control Tests

Instrument Serial #: 80-001306

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SNEADS PD	04/30/2007	09:51:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank		09:50	Sequence Aborted
		Air Blank	0.000	09:51	

# Control Tests

Instrument Serial #: 80-001308

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BAY COUNTY SO	03/11/2007	17:50:50	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:50	
		Control Test	0.079	17:50	
		Air Blank	0.000	17:50	

# Control Tests

Instrument Serial #: 80-001313

Location	Date	Time	Version	Target	Source
SPRINGFIELD PD	09/20/2007	14:33:29	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:32	
		Control Test	0.077	14:32	
		Air Blank	0.000	14:33	
SPRINGFIELD PD	09/20/2007	14:35:51	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:34	
		Control Test	0.192	14:35	
		Air Blank	0.000	14:35	
SPRINGFIELD PD	09/20/2007	15:53:12	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:51	
		Control Test	0.193	15:52	
		Air Blank	0.000	15:53	
SPRINGFIELD PD	09/20/2007	16:07:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:06	
		Control Test	0.045	16:07	
		Air Blank	0.000	16:07	
SPRINGFIELD PD	09/20/2007	16:18:20	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:17	
		Control Test	0.000	16:17	
		Air Blank	0.000	16:18	
SPRINGFIELD PD	09/20/2007	16:21:24	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:20	
		Control Test	0.000	16:20	
		Air Blank	0.000	16:21	
SPRINGFIELD PD	09/20/2007	16:23:56	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:23	
		Control Test	0.079	16:23	
		Air Blank	0.000	16:23	

# Control Tests

Instrument Serial #: 80-001341

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
DESOTO COUNTY SO	11/14/2007	13:44:53	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:43	
		ITP Check	0.080	13:44	
		Air Blank	0.000	13:44	

# Control Tests

Instrument Serial #: 80-001344

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SARASOTA COUNTY SO	03/24/2007	20:35:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:34	
		Control Test	0.081	20:34	
		Air Blank	0.000	20:35	
SARASOTA COUNTY SO	03/24/2007	20:38:41	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	20:37	
		ITP Check	0.078	20:38	
		Air Blank	0.000	20:38	
SARASOTA COUNTY SO	11/14/2007	13:43:54	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:43	
		ITP Check	0.080	13:43	
		Air Blank	0.000	13:43	

# Control Tests

Instrument Serial #: 80-001345

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
EUSTIS P.D.	07/15/2007	12:22:23	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:21	
		ITP Check	0.081	12:21	
		Air Blank	0.000	12:22	

# Control Tests

Instrument Serial #: 80-001350

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
FRUITLAND PARK P.D.	09/10/2007	09:41:23	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:40	
		ITP Check	0.081	09:40	
		Air Blank	0.000	09:41	

# Control Tests

Instrument Serial #: 80-001364

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
GAINESVILLE PD	05/02/2007	10:05:50	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:04	
		ITP Check	0.081	10:05	
		Air Blank	0.000	10:05	

# Control Tests

Instrument Serial #: 80-001365

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
LEVY COUNTY SO	05/26/2007	13:09:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:08	
		Control Test	0.079	13:08	
		Air Blank	0.000	13:09	
LEVY COUNTY SO	05/26/2007	13:10:47	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:09	
		Control Test	0.079	13:10	
		Air Blank	0.000	13:10	
LEVY COUNTY SO	05/26/2007	13:12:37	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:11	
		Control Test	0.078	13:11	
		Air Blank	0.000	13:12	

# Control Tests

Instrument Serial #: 80-001416

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ORANGE COUNTY SO	08/31/2007	21:52:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	21:52	
		Control Test	0.081	21:52	
		Air Blank	0.000	21:52	

# Control Tests

Instrument Serial #: 80-001440

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
KENNEDY SPACE CENTER	09/04/2007	11:00:13	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:59	
		Control Test	0.080	10:59	
		Air Blank	0.000	11:00	

# Control Tests

Instrument Serial #: 80-001443

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CLAY COUNTY SO	03/24/2007	12:01:49	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:00	
		Control Test	0.083	12:01	
		Air Blank	0.000	12:01	
CLAY COUNTY SO	03/24/2007	12:03:34	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:02	
		ITP Check	0.079	12:03	
		Air Blank	0.000	12:03	

# Control Tests

Instrument Serial #: 80-001469

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HURLBURT FIELD PD	03/25/2007	10:18:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:17	
		Control Test	0.081	10:17	
		Air Blank	0.000	10:18	
HURLBURT FIELD PD	08/15/2007	06:18:02	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	06:17	
		Control Test	0.080	06:17	
		Air Blank	0.000	06:18	

# Control Tests

Instrument Serial #: 80-001471

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PATRICK A.F.B.	07/31/2007	12:08:42	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	12:07	
		Control Test	0.073	12:08	
		Air Blank	0.000	12:08	

# Control Tests

Instrument Serial #: 80-001646

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
BAY COUNTY SO	03/11/2007	18:00:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:59	
		Control Test	0.079	17:59	
		Air Blank	0.000	18:00	

# Control Tests

Instrument Serial #: 80-001649

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
KENNETH CITY PD	09/19/2007	05:54:45	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:53	
		Control Test	0.048	05:54	
		Air Blank	0.000	05:54	
KENNETH CITY PD	09/19/2007	05:57:35	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:56	
		Control Test	0.078	05:56	
		Air Blank	0.000	05:57	
KENNETH CITY PD	09/19/2007	05:59:50	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:58	
		Control Test	0.198	05:59	
		Air Blank	0.000	05:59	
KENNETH CITY PD	10/10/2007	05:37:51	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:36	
		Control Test	0.000	05:37	
		Air Blank	0.000	05:37	
KENNETH CITY PD	10/10/2007	05:40:28	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	05:39	
		Control Test	0.048	05:39	
		Air Blank	0.000	05:40	

# Control Tests

Instrument Serial #: 80-001651

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
COLLIER COUNTY SO	09/13/2007	11:02:20	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:01	
		ITP Check	0.079	11:01	
		Air Blank	0.000	11:02	

# Control Tests

Instrument Serial #: 80-001652

Location	Date	Time	Version	Target	Source
WEST MIAMI PD	10/03/2007	16:10:13	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:09	
		Control Test	0.078	16:09	
		Air Blank	0.000	16:10	
WEST MIAMI PD	10/03/2007	16:17:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:15	
		Control Test	0.081	16:16	
		Air Blank	0.000	16:17	
WEST MIAMI PD	10/03/2007	16:19:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:17	
		Control Test	0.082	16:18	
		Air Blank	0.000	16:19	
WEST MIAMI PD	10/03/2007	16:21:18	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:20	
		Control Test	0.083	16:20	
		Air Blank	0.000	16:21	
WEST MIAMI PD	10/03/2007	16:23:41	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:22	
		Control Test	0.082	16:23	
		Air Blank	0.000	16:23	
WEST MIAMI PD	10/03/2007	16:27:57	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:26	
		Control Test	0.082	16:27	
		Air Blank	0.000	16:27	

# Control Tests

Instrument Serial #: 80-001653

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
ST PETERSBURG PD	09/07/2007	12:00:37	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:59	
		ITP Check	0.078	12:00	
		Air Blank	0.000	12:00	
ST PETERSBURG PD	09/10/2007	09:11:53	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	09:10	
		ITP Check	0.079	09:11	
		Air Blank	0.000	09:11	
ST PETERSBURG PD	09/10/2007	10:02:27	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	10:01	
		Control Test	0.048	10:01	
		Air Blank	0.000	10:02	

# Control Tests

Instrument Serial #: 80-001657

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MICCOSUKEE PD	04/19/2007	11:47:04	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:46	
		Control Test	0.075	11:46	
		Air Blank	0.000	11:47	
MICCOSUKEE PD	04/19/2007	11:48:28	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:47	
		Control Test	0.075	11:47	
		Air Blank	0.000	11:48	
MICCOSUKEE PD	04/19/2007	11:50:06	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:49	
		Control Test	0.075	11:49	
		Air Blank	0.000	11:50	
MICCOSUKEE PD	11/27/2007	11:38:09	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:37	
		Control Test	0.077	11:37	
		Air Blank	0.000	11:38	
MICCOSUKEE PD	11/27/2007	11:42:31	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:41	
		ITP Check	0.080	11:42	
		Air Blank	0.000	11:42	

# Control Tests

Instrument Serial #: 80-001721

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
PALM BAY PD	05/23/2007	00:04:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	00:03	
		Control Test	0.076	00:03	
		Air Blank	0.000	00:04	

# Control Tests

Instrument Serial #: 80-001723

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
VENICE PD	11/02/2007	08:28:18	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	08:27	
		Control Test	0.079	08:27	
		Air Blank	0.000	08:28	
VENICE PD	11/12/2007	14:46:19	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:45	
		Control Test	0.078	14:45	
		Air Blank	0.000	14:46	

# Control Tests

Instrument Serial #: 80-001731

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HILLSBOROUGH SO	01/06/2007	01:33:36	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:32	
		Control Test	0.077	01:33	
		Air Blank	0.000	01:33	
HILLSBOROUGH SO	01/06/2007	01:35:20	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:34	
		Control Test	0.077	01:34	
		Air Blank	0.000	01:35	
HILLSBOROUGH SO	01/06/2007	01:36:55	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	01:36	
		Control Test	0.077	01:36	
		Air Blank	0.000	01:36	

# Control Tests

Instrument Serial #: 80-001737

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
SOUTH MIAMI P.D.	11/02/2007	02:49:00	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	02:48	
		Control Test	0.076	02:48	
		Air Blank	0.000	02:49	

# Control Tests

Instrument Serial #: 80-001739

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
CHARLOTTE COUNTY SO	08/30/2007	14:27:57	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:27	
		Control Test	0.080	14:27	
		Air Blank	0.000	14:27	
CHARLOTTE COUNTY SO	08/30/2007	14:29:48	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:28	
		Control Test	0.195	14:29	
		Air Blank	0.000	14:29	
CHARLOTTE COUNTY SO	11/14/2007	13:44:20	8100.27.00	0.080	Local ITP
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	13:43	
		ITP Check	0.079	13:43	
		Air Blank	0.000	13:44	

# Control Tests

Instrument Serial #: 80-001740

Location	Date	Time	Version	Target	Source
CLEARWATER PD	04/10/2007	17:55:32	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:54	
		Control Test	0.000	17:55	
		Air Blank	0.000	17:55	
CLEARWATER PD	04/10/2007	17:58:05	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:56	
		Control Test	0.000	17:57	Interferent Detect
		Air Blank	0.000	17:58	
CLEARWATER PD	04/10/2007	18:00:02	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	17:58	
		Control Test	0.048	17:59	
		Air Blank	0.000	18:00	
CLEARWATER PD	04/10/2007	18:01:47	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:00	
		Control Test	0.079	18:01	
		Air Blank	0.000	18:01	
CLEARWATER PD	04/10/2007	18:03:31	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	18:02	
		Control Test	0.197	18:02	
		Air Blank	0.000	18:03	
CLEARWATER PD	09/10/2007	14:37:46	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	14:36	
		Control Test	0.074	14:37	
		Air Blank	0.000	14:37	
CLEARWATER PD	10/23/2007	15:54:31	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	15:53	
		Control Test	0.000	15:53	
		Air Blank		15:54	Sequence Aborted
		Air Blank	0.000	15:54	

# Control Tests

Instrument Serial #: 80-002097

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
NAS JACKSONVILLE	03/15/2007	11:31:26	8100.27.00	0.080	Local Dry gas
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:30	
		Control Test	0.081	11:30	
		Air Blank	0.000	11:31	

# Control Tests

Instrument Serial #: 80-002232

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
MIAMI BEACH PD	12/05/2007	16:23:34	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:22	
		Control Test	0.191	16:23	
		Air Blank	0.000	16:23	
MIAMI BEACH PD	12/05/2007	16:25:25	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	16:24	
		Control Test	0.193	16:24	
		Air Blank	0.000	16:25	

# Control Tests

Instrument Serial #: 80-002464

<b>Location</b>	<b>Date</b>	<b>Time</b>	<b>Version</b>	<b>Target</b>	<b>Source</b>
HIALEAH P.D.	08/02/2007	11:14:08	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:12	
		Control Test	0.047	11:13	
		Air Blank	0.000	11:14	
HIALEAH P.D.	08/02/2007	11:16:40	8100.27.00	0.080	Local Wet bath
		<b>Test</b>	<b>Value</b>	<b>Time</b>	
		Air Blank	0.000	11:15	
		Control Test	0.048	11:16	
		Air Blank	0.000	11:16	