

Inspection Tests

Instrument Serial #80-000206

Agency Operator	Date	Time	Software	Source
FDLE/ATP PROC SMITH,JAMES,C	09/02/2006	15:15:17	8100.26	Agency Inspection

0.05 Lot# 200509A	0.08 Lot# 200509B	0.20 Lot# 20509C	0.08 Gas Lot#
Expiration 09/22/2007	Expiration 09/22/2007	Expiration 09/22/2007	

Sims 5 **Complies No**

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.029
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.068
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.146
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test #1 (Retry)	0.173
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.042
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	No	0.08 g/210L Test #2	0.074
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #2 (Retry)	0.080
		0.20 g/210L Test #2	0.166
		0.20 g/210L Test #2 (Retry)	0.177
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.082
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.172
		0.20 g/210L Test #3 (Retry)	0.176
		Interferent Detect Test #3	INT

Remarks:

THIS MACHINE SUCKS Non-compliance: .