Florida Department of Law Enforcement

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	01/14/2007	16:21:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	02/18/2007	14:25:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	03/18/2007	11:58:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	04/11/2007	04:00:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	05/07/2007	18:10:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	06/08/2007	17:40:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	07/11/2007	12:27:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	08/15/2007	13:28:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	09/07/2007	12:20:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source
SATELLITE BEACH PD HACKNEY,DWITE,N	09/25/2007	11:10:12	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	10/19/2007	17:17:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	11/15/2007	18:03:45	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001001

Agency Operator	Date	Time	Software	Source	
SATELLITE BEACH PD BERRIOS,NORBERTO,	12/21/2007	14:03:47	8100.27	Agency	Inspection

0.05 Lot#	200711D	**0.08 Lot#**	200708C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	11/01/2009	**Expiration**	08/01/2009	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	01/29/2007	08:47:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.078
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.194
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	01/29/2007	09:51:05	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	02/23/2007	12:52:17	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	03/28/2007	09:01:31	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source
HARDEE COUNTY SO SUERETH,DONALD,P	04/12/2007	12:09:38	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.08 g/210L Test #4 (Retry)	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.08 g/210L Test #5 (Retry)	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.08 g/210L Test #6 (Retry)	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.079
			0.08 g/210L Test #7 (Retry)	0.079
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.08 g/210L Test #8 (Retry)	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9 (Retry)	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10 (Retry)	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: RFI Detect.OFFICER ENTERED WITH CELL PHONE.PX REMOVED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	04/27/2007	01:31:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	05/30/2007	09:44:38	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	07/01/2007	01:32:38	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	08/01/2007	09:37:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	09/22/2007	11:43:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	10/21/2007	20:34:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	11/28/2007	02:54:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001002

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	12/30/2007	08:51:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	01/26/2007	19:54:45	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	02/22/2007	21:02:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	03/12/2007	03:54:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	04/16/2007	01:38:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	04/16/2007	11:40:12	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	05/13/2007	02:24:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	06/14/2007	21:07:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: SIM NOT CONNECTED TO INSTRUMENT-RETEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/20/2007	21:41:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	08/10/2007	00:31:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	09/17/2007	03:53:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	10/21/2007	20:40:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	11/17/2007	20:58:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001003

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	12/15/2007	23:57:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	01/23/2007	16:45:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	02/21/2007	09:28:03	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	03/28/2007	15:13:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	04/04/2007	18:18:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	05/14/2007	11:25:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	06/19/2007	15:52:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/05/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	07/08/2007	14:00:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.041
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.041
		0.05 g/210L Test #3 (Retry)	0.045
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	08/20/2007	11:36:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	09/17/2007	19:41:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	10/09/2007	14:21:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source
BREVARD S.O. SKIPPER,ROGER,G	10/09/2007	17:19:30	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001004

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	11/13/2007	13:28:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	01/26/2007	22:18:29	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.067
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.072
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

08: SIMULATOR NOT AT TEMPERATURE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	02/22/2007	22:35:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: AMBIENT FAIL/INSTRUMENT UNABLE TO PURGE CHAMBER.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	03/11/2007	22:27:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	04/15/2007	23:02:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	04/16/2007	14:07:29	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	05/13/2007	23:42:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	06/13/2007	22:36:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.UNK REASON

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/20/2007	00:03:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	08/09/2007	21:13:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	09/17/2007	00:44:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	10/21/2007	22:50:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	11/17/2007	23:03:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001005

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	12/15/2007	23:17:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	01/24/2007	05:30:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	02/28/2007	03:53:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	03/25/2007	05:21:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.056
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.059
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.064
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	04/27/2007	09:43:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	05/28/2007	09:58:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	06/25/2007	09:40:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	2006605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	07/26/2007	02:48:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	08/29/2007	05:17:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	09/19/2007	04:47:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source
INDIALANTIC PD HACKNEY,DWITE,N	09/25/2007	13:33:45	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.192
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.192
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.192
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	10/27/2007	12:16:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	11/25/2007	14:56:27	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001006

Agency Operator	Date	Time	Software	Source	
INDIALANTIC PD DESCHENES,RYAN,	12/16/2007	10:31:33	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	01/19/2007	10:25:03	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	01/19/2007	10:29:37	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	01/19/2007	10:58:32	8100.27	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.048
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.048
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.043
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	01/19/2007	11:56:14	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.045
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	02/01/2007	11:44:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	03/09/2007	14:29:48	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:SIMULATOR NOT PROPERLY HEATED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	04/17/2007	14:48:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	05/30/2007	15:46:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,RODGERS,D	06/29/2007	15:20:47	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	06/29/2007	16:16:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.248
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.227
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.055
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.179
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.226
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.217
		0.20 g/210L Test #3 (Retry)	0.224
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	06/29/2007	17:03:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.183
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.189
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.TEMPERTURE PROBLEMS IN BA ROOM

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source
HAINES CITY PD SUERETH,DONALD,P	07/10/2007	11:35:33	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

K. RODGERS PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	07/10/2007	12:30:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	08/30/2007	11:18:58	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	08/30/2007	14:32:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.045
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

FORGOT TO CONNECT THE HOSES

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	09/28/2007	15:05:10	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.043
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.043
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05ROOM TEMPATURE, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	09/28/2007	15:51:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside/EXCHANGE HEADS ON SIMULATOR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	10/29/2007	11:23:19	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	11/30/2007	16:17:58	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001007

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	12/26/2007	16:59:12	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	733201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/30/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD GRENIER,D,K	01/25/2007	22:23:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD GRENIER,D,K	02/22/2007	21:38:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.000
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.188
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD GRENIER,D,K	03/23/2007	17:31:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD GRENIER,D,K	04/29/2007	08:01:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.175
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD GRENIER,D,K	05/27/2007	09:42:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.073
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	06/27/2007	18:05:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.072
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	07/19/2007	11:44:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.173
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.196
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.064
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.196
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	08/09/2007	12:19:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	09/21/2007	13:30:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source
WEST MELBOURNE PD HACKNEY,DWITE,N	10/29/2007	11:48:54	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.192
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.192
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.192
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	10/29/2007	12:18:55	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.047
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.048
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	10/29/2007	13:03:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	11/12/2007	14:03:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001008

Agency Operator	Date	Time	Software	Source	
WEST MELBOURNE PD BRYANT,JAMES,H	12/05/2007	11:16:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	01/25/2007	23:32:05	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504601I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed. HOSE UNATTACHED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	02/22/2007	23:49:23	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	03/23/2007	12:25:01	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	04/26/2007	20:47:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	535420I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/05/2007	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

BAD MOUTH P. OPER ERR FORGOT GAS LINE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	05/30/2007	19:12:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	06/12/2007	16:39:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

BAD SEAL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	07/26/2007	18:42:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	08/30/2007	19:18:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	09/20/2007	10:08:47	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	09/20/2007	12:04:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance SIMULATOR AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	10/29/2007	13:08:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,	11/22/2007	20:41:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001009

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	12/27/2007	19:33:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.176
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.111
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.085
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance WRONG SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	01/23/2007	16:51:01	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	02/21/2007	08:57:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	03/28/2007	15:05:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	04/04/2007	18:32:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	05/14/2007	10:19:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	06/19/2007	15:57:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	07/08/2007	13:53:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	08/20/2007	11:29:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	09/17/2007	20:07:32	8100.27	Agency	Inspection

0.05 Lot#	200605C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.078
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.078
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	10/09/2007	14:10:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source
BREVARD S.O. SKIPPER,ROGER,G	10/09/2007	16:51:09	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	11/13/2007	12:10:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001010

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	12/17/2007	11:23:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/16/2007	13:33:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/25/2007	09:44:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	02/16/2007	09:27:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	02/27/2007	13:04:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	03/15/2007	13:12:07	8100.27	Agency	Inspection

0.05 Lot#	200506B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

.05 LOT #200605B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	03/27/2007	13:17:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source
JACKSONVILLE SO MALHIOT,MATTHEW,E	04/12/2007	11:38:04	8100.27	Department Inspection

0.05 Lot#	200605C	**0.08 Lot#**	200605D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	04/16/2007	13:02:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	04/27/2007	13:06:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	05/01/2007	09:33:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT WAS TAKEN OUT OF SERVICE FOR BREATH TUBE REPLACE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	08/14/2007	12:34:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	08/24/2007	09:19:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	09/14/2007	09:00:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	09/27/2007	10:17:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/15/2007	12:44:24	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/31/2007	09:57:32	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/16/2007	10:10:12	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/29/2007	13:30:23	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/13/2007	13:29:00	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001018

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/31/2007	13:16:03	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	733202I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/30/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source
HIGHLANDS COUNTY SO VENTURI,GEORGE,L	01/24/2007	17:33:04	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	02/11/2007	20:01:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	03/29/2007	21:50:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	04/30/2007	08:56:37	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	05/20/2007	14:47:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	06/17/2007	08:28:57	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. MOUTH ALCOHOL PRESENT. START Non-compliance:TOO SOON BETWEEN DOING INSTRUMENTS .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	06/17/2007	13:52:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source
HIGHLANDS COUNTY SO VENTURI,GEORGE,L	07/23/2007	15:56:35	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	07/28/2007	07:56:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	08/15/2007	12:44:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	09/09/2007	11:30:52	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

UNKNOWN REASON FOR READING.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	09/09/2007	12:23:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	10/06/2007	10:19:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.074
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	11/02/2007	10:08:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	724801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001043

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	12/05/2007	22:58:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	724801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source
COCONUT CREEK PD HACKNEY,DWITE,N	01/02/2007	14:36:39	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.077
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.076
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.076
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.076
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.081
			0.20 g/210L Test #7	0.203
			0.08 g/210L Dry Gas Std Test #7	INT
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

08: Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	01/03/2007	13:17:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	02/06/2007	20:05:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	03/14/2007	19:54:25	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	04/03/2007	19:22:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	05/11/2007	19:20:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.024
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.023
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.027
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	06/07/2007	16:54:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	07/03/2007	19:41:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	08/14/2007	17:11:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source
COCONUT CREEK PD VEIGA,SANDRA,F	09/06/2007	16:32:59	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.077
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.077
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.077
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.082
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.077
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	INT
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Interferent Detect.R. MARKEN WAS PRESENT DURING INSPECT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	09/06/2007	17:30:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	10/16/2007	18:32:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	11/02/2007	16:40:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001044

Agency Operator	Date	Time	Software	Source	
COCONUT CREEK PD MARKEN,RYAN,K	12/04/2007	17:44:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance/SIM CHANGED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/16/2007	13:26:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/25/2007	09:35:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source
JACKSONVILLE SO MALHIOT,MATTHEW,E	04/12/2007	10:31:14	8100.27	Department Inspection

0.05 Lot#	200605C	**0.08 Lot#**	200605D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	05/01/2007	07:30:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	05/16/2007	09:58:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	05/30/2007	13:32:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	06/15/2007	09:20:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	06/29/2007	09:15:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	07/17/2007	10:34:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	07/30/2007	10:08:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	08/14/2007	10:45:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	08/24/2007	09:22:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	09/14/2007	08:43:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	09/27/2007	10:02:22	8100.27	Agency	Inspection

0.05 Lot#	200605	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/15/2007	12:30:56	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/31/2007	09:43:53	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/16/2007	10:18:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/29/2007	13:19:24	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/13/2007	13:04:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001045

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/31/2007	13:01:00	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

CHANGE O-RING ON SIM CONN..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	01/28/2007	11:34:14	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	02/20/2007	01:32:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	03/25/2007	08:37:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source
CLAY COUNTY SO MALHIOT,MATTHEW,E	04/19/2007	10:09:43	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	04/22/2007	00:02:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	05/20/2007	05:13:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	06/15/2007	22:53:15	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	07/05/2007	21:07:07	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	08/06/2007	22:35:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	09/22/2007	13:19:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	10/15/2007	21:59:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	10/15/2007	22:42:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	11/07/2007	11:05:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside FAILED TO CONTECT DRY GAS CYL TO INTOX .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001046

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	12/10/2007	13:27:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	01/12/2007	10:24:23	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	01/26/2007	15:30:50	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	02/09/2007	15:07:23	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	02/23/2007	09:44:04	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.207
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	03/08/2007	11:56:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	03/30/2007	09:47:40	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/20/2007	16:09:09	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/20/2007	16:44:28	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/11/2007	10:57:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,N	05/25/2007	16:34:37	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	06/08/2007	12:50:37	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source
BROWARD S.O. VEIGA,SANDRA,F	08/02/2007	15:48:05	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	79601I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

F. DICIENZO & P. NANZ WERE PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source
BROWARD S.O. VEIGA,SANDRA,F	08/02/2007	15:52:39	8100.27	Department Inspection

 # Sims 0 **Avg Std Dev** **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

TEST

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD S.O. NANZ,PATRICIA,M	08/03/2007	12:42:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD S.O. DICIENZO,FRANCA,R	09/07/2007	16:30:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZWAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD S.O. NANZ,PATRICIA,M	09/28/2007	13:15:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD S.O. DICIENZO,FRANCA,R	10/12/2007	15:52:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD S.O. NANZ,PATRICIA,M	10/26/2007	12:52:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD S.O. NANZ,PATRICIA,M	11/09/2007	16:14:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	11/30/2007	15:36:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001047

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	12/07/2007	15:53:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD FLORA,MICHAEL,O	01/24/2007	09:17:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD FLORA,MICHAEL,O	02/18/2007	08:32:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD FLORA,MICHAEL,O	03/27/2007	08:08:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD FLORA,MICHAEL,O	04/30/2007	04:03:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.018
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.019
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: PRESENCE FROM 4L01 INSPEC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD FLORA,MICHAEL,O	05/17/2007	02:38:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD FLORA,MICHAEL,O	06/25/2007	02:13:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	07/08/2007	**Expiration**	07/08/2007	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

0.05 SOL EXP 05/04/08, 0.08 EXP 05/04/08, 0.20 EXP 05/10/08

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD HOULE,DANIEL,L	07/22/2007	08:52:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD HOULE,DANIEL,L	08/23/2007	09:53:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source
LAKE ALFRED PD SUERETH,DONALD,P	08/29/2007	15:03:34	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

OFFICER D. HOULE PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD HOULE,DANIEL,L	09/19/2007	09:47:32	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD HOULE,DANIEL,L	10/19/2007	09:40:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:M A DID NOT RINSE WITH MA SOL/MA ON AF.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD HOULE,DANIEL,L	11/15/2007	10:51:06	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. TOO MUCH MA. AI SURETH NOTIFIED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001048

Agency Operator	Date	Time	Software	Source	
LAKE ALFRED PD HOULE,DANIEL,L	12/20/2007	11:13:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	01/11/2007	17:52:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	02/13/2007	19:27:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	03/13/2007	13:40:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

20:SIM TEMP. LOW SWITCHED HEADS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	04/14/2007	12:53:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source
PASCO COUNTY SO SUERETH,DONALD,P	05/16/2007	14:59:29	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.083
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.082
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

DEPUTY L. WINTERS PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	05/16/2007	16:07:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	06/19/2007	19:51:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	07/25/2007	18:47:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	08/22/2007	17:25:31	8100.27	Agency	Inspection

0.05 Lot#	200602D	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	02/23/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	09/13/2007	20:01:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	10/10/2007	12:59:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	11/07/2007	19:57:59	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001049

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	12/12/2007	21:05:41	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD BARBUTO,JOSEPH,L	01/24/2007	05:14:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.050
0.08 g/210L Test #1	
0.08 g/210L Test #1 (Retry)	0.081
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.042
0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test #2 (Retry)	0.081
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.05 g/210L Test #3 (Retry)	0.051
0.08 g/210L Test #3 (Retry)	0.082
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance AIR LEAK. 08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD BARBUTO,JOSEPH,L	01/24/2007	21:54:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	02/22/2007	15:01:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.043
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.044
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD BARBUTO,JOSEPH,L	03/30/2007	11:28:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	03/30/2007	12:07:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/18/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	04/24/2007	15:58:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	05/31/2007	09:03:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	06/28/2007	09:51:54	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. Int Det: , . Non-compliance:INT DET PROBLEM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	06/28/2007	10:45:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

SGT BARBUTO PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	07/26/2007	16:52:44	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	08/30/2007	09:11:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD BARBUTO,JOSEPH,L	09/30/2007	12:54:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source
CORAL SPRINGS PD VEIGA,SANDRA,F	10/23/2007	14:27:03	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	709601I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.081
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.078
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	INT
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Interferent Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD BARBUTO,JOSEPH,L	10/23/2007	15:06:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD WILKINSON,JON,G	11/23/2007	09:41:21	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001050

Agency Operator	Date	Time	Software	Source	
CORAL SPRINGS PD BARBUTO,JOSEPH,L	12/27/2007	09:22:22	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

OFC WILKINSON PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	01/17/2007	22:02:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	02/15/2007	20:21:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	03/21/2007	22:32:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

SIMS CLEAN AND DRY

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	04/18/2007	22:22:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ARKOVICH,ROBBIE,S	05/03/2007	20:34:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	06/13/2007	19:50:58	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	06/13/2007	19:55:06	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

IN COMPLIANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	06/13/2007	20:00:40	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ALCOHOL IN MOUTH FROM PRIOR MOUTH ALCOHOL TEST. INSPECTION NOT COMPLETED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	06/13/2007	23:06:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	07/12/2007	18:14:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	08/09/2007	19:33:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	09/26/2007	20:35:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301C
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source
ST PETERSBURG PD SUERETH,DONALD,P	10/11/2007	13:00:58	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.083
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.083
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.083
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.083
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

OFFICER M. WEISKOPF PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	10/11/2007	14:49:42	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	11/19/2007	12:12:57	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001051

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	12/20/2007	18:30:32	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source
JACKSONVILLE SO MALHIOT,MATTHEW,E	01/02/2007	10:41:47	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/25/2007	08:35:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/29/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	02/16/2007	09:33:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/29/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	02/27/2007	13:10:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/29/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	03/15/2007	13:19:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/29/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	03/27/2007	13:22:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/29/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

N/A

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	04/16/2007	13:08:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	04/27/2007	13:11:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	05/16/2007	10:08:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	05/30/2007	13:41:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	06/15/2007	09:30:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	06/29/2007	09:31:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	07/17/2007	10:46:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	07/30/2007	10:18:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	08/14/2007	10:56:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	08/24/2007	09:32:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	09/14/2007	08:52:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	09/27/2007	10:11:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/15/2007	12:40:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/31/2007	09:50:50	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/16/2007	10:06:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/29/2007	13:23:32	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/13/2007	13:23:48	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001052

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/31/2007	13:08:11	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source
NASSAU CO SO MALHIOT,MATTHEW,E	01/02/2007	12:48:04	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	01/06/2007	19:52:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	01/30/2007	14:26:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	02/27/2007	11:39:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	03/30/2007	15:56:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	04/28/2007	22:14:10	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	05/26/2007	00:08:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	06/26/2007	10:32:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612D	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/14/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	07/30/2007	09:45:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	08/27/2007	12:19:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.RAIDO TO CLOSE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	09/29/2007	11:00:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	10/31/2007	17:29:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	]630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	11/29/2007	07:11:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	7289O2I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/23/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001053

Agency Operator	Date	Time	Software	Source	
NASSAU CO SO NYE,TERRY,L	12/17/2007	09:44:51	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	728902I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/23/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	01/04/2007	10:46:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	02/10/2007	19:59:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.190
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.046
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.000
		0.20 g/210L Test #2 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance./SIMULATOR NOT HOOKED UP e/SIMULATOR NOT HOOKED UP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	03/29/2007	22:47:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.192
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.193
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.HOOKED UP WRONG SIMULATOR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	04/30/2007	09:04:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	05/20/2007	14:32:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	06/17/2007	09:33:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
		Interferent Detect Test #3	INT

Remarks:

-GAS NOT PLUGGED INTO PORT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source
HIGHLANDS COUNTY SO VENTURI,GEORGE,L	07/23/2007	16:12:41	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	07/28/2007	07:48:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	08/15/2007	12:33:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	09/09/2007	12:16:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	10/06/2007	10:11:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	11/02/2007	10:03:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance/TEMP NOT 34.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001054

Agency Operator	Date	Time	Software	Source	
HIGHLANDS COUNTY SO REBMANN,DORIS,E	12/05/2007	23:07:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	01/15/2007	09:08:56	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	01/16/2007	10:27:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT

Remarks:

20: LOOSE HOSE SEAL. 08: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	02/18/2007	09:53:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: SEAL-AIR-LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	03/21/2007	12:33:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: SEAL-AIR-LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	04/15/2007	08:28:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	05/16/2007	11:10:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.187
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08:HOSE-AIR-LEAK. 20: HOSE-AIR-LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	06/19/2007	08:24:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	07/20/2007	11:53:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: HOSE WAS DISCONNECTED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	08/08/2007	10:01:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	08/30/2007	11:35:44	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	09/21/2007	09:42:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	10/23/2007	09:55:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	11/26/2007	11:05:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	12/05/2007	16:25:25	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.206
0.08 g/210L Dry Gas Std Test #2	0.084
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.206
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.206
0.08 g/210L Dry Gas Std Test #4	0.083
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.207
0.08 g/210L Dry Gas Std Test #5	0.084
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.083
0.20 g/210L Test #6	0.207
0.08 g/210L Dry Gas Std Test #6	0.083
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.207
0.08 g/210L Dry Gas Std Test #7	0.084
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001055

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.207
			0.08 g/210L Dry Gas Std Test #8	0.084
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.207
			0.08 g/210L Dry Gas Std Test #9	0.083
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.207
			0.08 g/210L Dry Gas Std Test #10	0.083
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	01/30/2007	22:16:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	02/28/2007	22:43:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	03/31/2007	10:20:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	04/30/2007	22:49:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	05/31/2007	09:30:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	06/29/2007	18:25:44	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	07/30/2007	22:06:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	08/20/2007	17:32:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	09/29/2007	21:28:49	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source
DAVIE PD VEIGA,SANDRA,F	10/23/2007	14:51:33	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	709601I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

SGT. DALE ENGLE WAS PRESENT FOR THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	10/27/2007	15:11:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	11/30/2007	18:03:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001056

Agency Operator	Date	Time	Software	Source	
DAVIE PD ENGLE,DALE,S	12/27/2007	18:17:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	733201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/30/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	01/19/2007	13:29:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	02/13/2007	12:34:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source
LAKELAND PD SUERETH,DONALD,P	03/05/2007	15:53:28	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.082
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.082
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.083
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7 (Retry)	0.082
Interferent Detect Test #7	INT

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.052
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.083
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.053
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.053
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: RFI Detect.OFFICER W/CELL PHONE ENTERED AREA.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	03/14/2007	11:47:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	04/27/2007	11:35:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	05/18/2007	10:38:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	06/26/2007	12:23:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	07/24/2007	11:53:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	2006O5F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	08/27/2007	12:37:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	09/24/2007	12:16:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	10/23/2007	11:35:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	11/26/2007	12:04:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001057

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	12/26/2007	11:57:00	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	01/31/2007	14:23:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.201
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	02/28/2007	14:42:48	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2 (Retry)	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3 (Retry)	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	02/28/2007	14:51:51	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance DID NOT PERFORM ALL REQUIRED TESTS .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	02/28/2007	15:29:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source
LAKE WALES PD SUERETH,DONALD,P	03/06/2007	11:57:40	8100.27	Department Inspection

 # Sims 4 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail,UNKNOWN SOURCE., Ambient Fail. Non-compliance:AMBIENT FAILURES..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source
LAKE WALES PD SUERETH,DONALD,P	03/06/2007	14:17:20	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4 (Retry)	0.051
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5 (Retry)	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6 (Retry)	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7 (Retry)	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8 (Retry)	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9 (Retry)	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10 (Retry)	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

05: RFI Detect. OFFICER ENTERED ROOM W/CELL PHONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	04/30/2007	10:00:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	06/07/2007	14:40:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.051
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.082
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.040
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance DELIVERY HOSE SPLIT LEAKING . REPLACED HOSE REPEATED TEST OK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	07/06/2007	10:38:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	08/31/2007	20:01:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	09/30/2007	16:20:19	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ERROR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	09/30/2007	17:07:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	11/29/2007	05:02:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

GAS EXP NO EXT TANKNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001058

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	12/28/2007	02:27:45	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	733802I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.209
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source
FERNANDINA BEACH PD MALHIOT,MATTHEW,E	02/05/2007	11:13:06	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

# Sims	5	**Avg Std Dev**	0.0005	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	02/15/2007	15:09:32	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007		

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.047
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.047
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI DetectRFI Detect. 08: RFI Detect, RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	02/26/2007	02:52:53	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.05 .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	02/27/2007	08:44:21	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL SOL.RFI De. MOUTH ALCOHOL PROBLEMS WITH SUBJECTS NOT INSTRUMENTNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	02/27/2007	09:24:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance LOOSE SIMULATOR COVER.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	03/31/2007	23:37:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	04/26/2007	12:49:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	05/31/2007	18:45:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	06/30/2007	16:14:11	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.183
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

00: RFI Detect. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	07/31/2007	23:21:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	08/31/2007	17:49:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	09/30/2007	23:08:18	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	10/29/2007	22:00:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/21/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance..AIR LEAK TEST LINE REPAIRED . OUTSIDE OF TOLERANCE AIR LEAK..TEST LINE...REPAIRED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	11/30/2007	21:04:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001059

Agency Operator	Date	Time	Software	Source	
FERNANDINA BEACH PD WHITE,FRANK,M	12/28/2007	19:39:26	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

UNABLE TO TEST...GASKET BROKEN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	01/25/2007	10:45:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.SIMULATOR GLASS BROKEN. e.SIMULATOR GLASS BROKEN. CORRECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	02/26/2007	10:49:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	03/22/2007	11:35:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	04/26/2007	14:00:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	05/28/2007	11:04:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	06/28/2007	13:38:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	07/13/2007	15:28:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	08/23/2007	15:31:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	09/19/2007	10:40:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source
WINTER HAVEN PD SUERETH,DONALD,P	09/27/2007	10:42:41	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

OFFICER K. NICHOLS PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	10/22/2007	11:52:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	11/20/2007	10:50:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001060

Agency Operator	Date	Time	Software	Source	
WINTER HAVEN PD NICHOLS,KENNETH,D	12/17/2007	12:23:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	01/30/2007	04:22:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	02/28/2007	23:01:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	03/30/2007	21:55:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	04/24/2007	00:23:01	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	05/26/2007	19:48:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A:M.A.NOT USED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	06/24/2007	18:24:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	07/17/2007	16:58:46	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,MICHAEL,S	07/18/2007	03:46:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	08/24/2007	04:11:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	09/16/2007	18:36:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source
GULFPORT PD SUERETH,DONALD,P	09/25/2007	11:31:53	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0010
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	10/18/2007	17:16:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	11/24/2007	18:25:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,MICHAEL,S	12/23/2007	18:05:24	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:NOT DETECTED, . Non-compliance:INT NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,MICHAEL,S	12/23/2007	18:33:28	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:NOT DETECTED . Non-compliance:INT NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001061

Agency Operator	Date	Time	Software	Source	
GULFPORT PD MAROTTA,M,S	12/23/2007	19:54:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance. AIR IN SIM

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	01/11/2007	10:43:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	02/08/2007	10:09:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	03/08/2007	10:25:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	04/06/2007	09:30:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	05/06/2007	19:40:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	06/08/2007	12:11:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	07/06/2007	09:17:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	08/10/2007	09:32:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source
TREASURE ISLAND PD SUERETH,DONALD,P	08/28/2007	12:31:40	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

C. DONALDSON PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	09/07/2007	09:30:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	10/05/2007	10:16:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	11/02/2007	09:47:48	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001062

Agency Operator	Date	Time	Software	Source	
TREASURE ISLAND PD DONALDSON,CHARLES,B	12/07/2007	09:41:18	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source
HOLLYWOOD PD HACKNEY,DWITE,N	01/22/2007	12:02:26	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD BARBUTO,JOHN,E	01/30/2007	18:43:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	02/25/2007	16:02:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.067
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.073
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	03/24/2007	10:45:35	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance:RESTART INPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	03/24/2007	12:26:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.177
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	04/27/2007	13:20:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	05/24/2007	15:58:57	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. Non-compliance: PROBLEM W/MOUTH ALCOHOL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	05/24/2007	17:22:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	06/24/2007	16:44:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	07/20/2007	17:58:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.102
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.180
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	No	0.20 g/210L Test #2	0.106
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2 (Retry)	0.189
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.105
		0.20 g/210L Test #3 (Retry)	0.191
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside-CAUSED BY LOOSE HOSE. 20: Control Outside-REPLACED CRACKED SIM HOSE, Control Outs. Non-compliance .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source
HOLLYWOOD PD VEIGA,SANDRA,F	07/23/2007	14:18:21	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD BARBUTO,JOHN,E	07/31/2007	13:53:43	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.075
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	No	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.076
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.076
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:USED .08 SOL
ACCIDENTALLY SIM NOT WARMED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD BARBUTO,JOHN,E	07/31/2007	14:53:44	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	08/29/2007	16:16:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	09/28/2007	11:04:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	10/21/2007	22:52:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	11/28/2007	22:02:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001063

Agency Operator	Date	Time	Software	Source	
HOLLYWOOD PD DEMMERY,THOMAS,M	12/26/2007	18:02:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source
NEW PORT RICHEY PD SUERETH,DONALD,P	03/12/2007	13:52:11	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	03/14/2007	10:16:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	04/18/2007	19:24:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	05/22/2007	15:55:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	06/07/2007	19:31:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.045
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	07/24/2007	13:47:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630602I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,E,J	08/24/2007	17:25:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630602I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	09/23/2007	18:00:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630602I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	10/26/2007	16:12:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630602I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	11/26/2007	23:43:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724701G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001064

Agency Operator	Date	Time	Software	Source	
NEW PORT RICHEY PD JAY,ERIK,J	12/31/2007	01:27:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724701G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	01/24/2007	09:04:26	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededImproper Sample. 00: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	01/24/2007	10:38:31	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	6307011
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	02/15/2007	10:01:59	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	03/08/2007	09:39:38	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	04/16/2007	10:34:43	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	05/26/2007	10:28:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	06/06/2007	09:49:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	630701I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	07/17/2007	13:29:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source
TARPON SPRINGS PD SUERETH,DONALD,P	10/16/2007	14:14:57	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	10/17/2007	14:24:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

RETURN TO SERVICE MONTHLY INSPECTION OCTOBER 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	11/19/2007	11:30:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001065

Agency Operator	Date	Time	Software	Source	
TARPON SPRINGS PD RUGGIERO,FRANK,C	12/17/2007	10:13:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD COOK,STEPHEN,R	01/05/2007	13:05:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD COOK,STEPHEN,R	02/08/2007	21:46:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD DONALDSON,CHARLES,B	03/22/2007	10:14:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD DONALDSON,CHARLES,B	04/19/2007	10:17:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	63062I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD DONALDSON,CHARLES,B	05/22/2007	09:46:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD DONALDSON,CHARLES,B	06/22/2007	09:53:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD DONALDSON,CHARLES,B	07/20/2007	10:19:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630602I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD DONALDSON,CHARLES,B	08/17/2007	09:43:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630602I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source
ST PETE BEACH PD SUERETH,DONALD,P	09/13/2007	11:29:56	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

DETECTIVE B. JOHNSON PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD JOHNSON,BRUCE,M	09/13/2007	12:16:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD JOHNSON,BRUCE,M	10/18/2007	11:12:20	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD JOHNSON,BRUCE,M	10/18/2007	12:02:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	723501I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/26/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: MOUTH ALC STILL IN MOUTH

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD JOHNSON,BRUCE,M	11/21/2007	12:15:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008		

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.198
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001066

Agency Operator	Date	Time	Software	Source	
ST PETE BEACH PD JOHNSON,BRUCE,M	12/24/2007	11:33:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	723501I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/26/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	01/12/2007	13:43:47	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	02/09/2007	16:02:09	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	02/23/2007	11:01:26	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	609301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	03/08/2007	11:00:16	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	03/30/2007	11:39:38	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PATNA

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/20/2007	17:28:59	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/11/2007	11:44:56	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	06/08/2007	12:36:36	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/12/2007	16:02:22	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	08/03/2007	11:46:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENTED AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	08/24/2007	16:41:50	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO VEIGA,SANDRA,F	08/29/2007	15:45:52	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

P. NANZ & F. DICIENZO WERE PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	09/07/2007	16:25:57	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	09/28/2007	13:04:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	10/12/2007	14:02:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	10/26/2007	13:09:50	8100.27	Agency	Inspection

0.05 Lot#	711401G	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	04/29/2009	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION. .05 LO

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	11/09/2007	13:39:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605F	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001067

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	12/28/2007	15:24:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	01/26/2007	04:42:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	02/16/2007	12:39:50	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200606C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.181
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.190
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.193
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside MIXED UP SIM PUT .OOO ON BY MISTAKE, Co. OUT SIDE CONTROL TEST RANGENon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	02/16/2007	15:18:56	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.007
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	INT
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:TAKEN OFF LINE. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source
CLEARWATER PD SUERETH,DONALD,P	02/20/2007	13:36:20	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.083
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.082
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.083
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	02/24/2007	23:49:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

ACETONE LOT 2006-D MOUTH ALCOHOL LOT 2006-A

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	03/23/2007	07:01:39	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	04/10/2007	19:03:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance GASKET AIR LEAK SEALED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	05/25/2007	05:57:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	06/22/2007	10:04:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

A F / M A: HAD TO STOP BLOWING STARTED COUGHING.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	07/27/2007	06:42:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	08/16/2007	17:05:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MOUTH ALCOHOL LOT 2006-A ACETONE LOT 2006-D

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	09/14/2007	05:21:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	10/24/2007	18:12:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect/CELL PHONE IN ROOM ACETONE/2006A ALCH/2006D

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	11/23/2007	04:54:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001068

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	12/20/2007	16:48:47	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	615802I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

ACETONE LOT 2006-D ALCOHOL LOT 2006-A

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001069

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/16/2007	13:38:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001069

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	01/25/2007	09:48:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001069

Agency Operator	Date	Time	Software	Source
JACKSONVILLE SO MALHIOT,MATTHEW,E	04/12/2007	11:10:27	8100.27	Department Inspection

0.05 Lot#	200506C	**0.08 Lot#**	200506D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001069

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	01/08/2007	05:02:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	504601L
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	02/17/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

TEMP 80.1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	02/19/2007	04:35:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	03/28/2007	05:51:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

TEMP 74.3F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source
FT MEADE PD SUERETH,DONALD,P	04/12/2007	14:30:27	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: Ambient Fail. ROOM FANNED, UNK SOURCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	05/07/2007	19:25:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,-	06/27/2007	13:57:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail. 20: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	07/20/2007	08:40:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

ROOM TEMP 74.7 F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	08/23/2007	20:12:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	09/21/2007	05:36:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

TEMP 72.7

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	10/31/2007	09:12:13	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	11/30/2007	18:44:31	8100.27	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1 (Retry)	INT
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #3	0.000
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: Ambient Fail, RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	11/30/2007	18:49:26	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source
FT MEADE PD SUERETH,DONALD,P	12/05/2007	12:38:49	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001070

Agency Operator	Date	Time	Software	Source	
FT MEADE PD VALLE,CARLOS,	12/18/2007	16:10:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	01/28/2007	13:22:52	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	02/20/2007	00:36:01	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	03/25/2007	10:05:22	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source
CLAY COUNTY SO MALHIOT,MATTHEW,E	04/19/2007	10:35:34	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	04/22/2007	00:55:37	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	05/20/2007	05:22:43	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	06/15/2007	23:20:22	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source
CLAY COUNTY SO MALHIOT,MATTHEW,E	09/17/2007	14:18:39	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

A F / M A:FAILED TO USE M A SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	09/22/2007	13:30:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION, INSTRUMENT PLACED BACK IN SERVICE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	10/15/2007	21:59:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	11/07/2007	13:18:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008		

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.081
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.203
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	11/07/2007	13:57:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001071

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	12/10/2007	14:15:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source
DADE CITY PD SUERETH,DONALD,P	02/14/2007	13:01:52	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	02/19/2007	10:20:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	03/28/2007	16:23:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	04/26/2007	16:56:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	05/24/2007	15:42:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.050
		0.08 g/210L Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	05/24/2007	16:59:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK LOCATED IN .08 SIMULATOR DURING FIRST INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	06/26/2007	14:33:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	07/18/2007	05:53:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	08/17/2007	10:59:44	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	09/24/2007	16:38:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	10/25/2007	16:15:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	11/15/2007	16:52:33	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	612202G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001073

Agency Operator	Date	Time	Software	Source	
DADE CITY PD BALDWIN,DEAN,	12/28/2007	12:36:51	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	612202G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	01/11/2007	17:52:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/24/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	02/13/2007	19:31:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	03/13/2007	13:38:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	INT
		Interferent Detect Test #3	INT

Remarks:

08: UNK. INT?, TAKEN OUT OF SERV.FOR REPR. ONGOING INT PROB. KEEPS DETECTING INTERFERANT..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source
PASCO COUNTY SO SUERETH,DONALD,P	05/16/2007	15:54:46	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

08: CONTROL OUTSIDE TOLERANCE. ETH GAS HOSE DISCONNECTE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	05/16/2007	16:41:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	06/19/2007	19:59:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	714101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08: UNK INT. DET..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	07/25/2007	18:56:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	714101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

UNK INTERF. DET..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	08/22/2007	17:28:54	8100.27	Agency	Inspection

0.05 Lot#	200602D	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	02/23/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

UNK INTERF..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	09/13/2007	20:14:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001074

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	10/10/2007	12:55:57	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.075
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.073
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.075
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	INT
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

.073/INT R-N-R TANK CHK CONNECTIONS, Control Outside Tolera. DID NOT PASS REMOVING FROM SERVICE, CONSTANT INT. W/DG DURIREMOVED FROM SERVICE INT OM DG LAST FEW INSP. SENT FOR REP..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	01/31/2007	03:10:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	02/23/2007	04:21:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	03/29/2007	20:06:48	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	04/30/2007	17:57:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	05/23/2007	18:06:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	06/23/2007	16:45:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	07/24/2007	01:05:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source
FHP TROOP E MIAMI VEIGA,SANDRA,F	07/31/2007	12:54:20	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

RAY ADDISON WAS PRESENT DURING INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	08/31/2007	03:52:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	09/29/2007	17:01:49	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	10/29/2007	23:22:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail HOSE NEXT TO ALCO SOL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	11/23/2007	09:49:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #2	0.000
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect. 08: Control Outside HOSE DETACHED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001075

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	12/18/2007	13:40:19	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632002I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	01/12/2007	02:43:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	02/19/2007	02:09:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	03/19/2007	02:37:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	04/16/2007	01:14:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance SIM LID NOT TIGHT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	05/31/2007	01:58:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	06/24/2007	09:29:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance SIM COVER NOT TIGHT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	07/16/2007	09:39:21	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	08/27/2007	11:36:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	09/29/2007	09:16:56	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance.UNK, NO LEAKS FOUND IN SIMNon-compliance:UNK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	09/29/2007	10:42:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source
ZEPHYRHILLS PD SUERETH,DONALD,P	10/08/2007	14:24:53	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

INITIAL ANALYSES BEGAN WITH A PURGE FAILURE OF UNK. ORIGIN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD BREWER,DEREK,R	10/28/2007	07:03:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance SIM LID NOT TIGHT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD GRIMM,RANDY,K	11/08/2007	18:12:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD GRIMM,RANDY,K	12/12/2007	04:31:02	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:TO MUCH MOUTH ALCOHOL. Non-compliance:MOUTH ALCOHOL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001076

Agency Operator	Date	Time	Software	Source	
ZEPHYRHILLS PD GRIMM,RANDY,K	12/12/2007	05:06:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	01/29/2007	08:43:31	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	02/23/2007	12:47:39	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	03/28/2007	08:54:36	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source
HARDEE COUNTY SO SUERETH,DONALD,P	04/12/2007	11:41:55	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	04/27/2007	01:24:32	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	05/30/2007	09:38:28	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	07/01/2007	01:27:10	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	08/01/2007	09:26:10	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	09/22/2007	11:54:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/14/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	10/21/2007	20:39:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	11/28/2007	02:59:22	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001077

Agency Operator	Date	Time	Software	Source	
HARDEE COUNTY SO OBRYAN,DANNY,L	12/30/2007	08:31:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	01/07/2007	00:33:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	02/07/2007	15:56:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	03/08/2007	02:32:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	04/09/2007	01:54:56	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	04/09/2007	03:11:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	05/13/2007	20:31:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	06/11/2007	15:30:35	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

INT. NOT DETECTED, . INT.NOT ALWAYS DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	06/12/2007	14:35:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

REPLACE OLD ACETONE TO GIVE PROPER RESULTS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	07/10/2007	16:09:39	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DID NOT HEAR ALERT TONE., . NOT ALERTING TO ACETONE, INTERFERNT TEST..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	07/10/2007	17:07:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PREVIOUS TEST O RING PROBLEM

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	08/07/2007	14:09:06	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #3	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: TO BE SURE THAT INT. IS DECTED, Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	08/07/2007	15:09:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source
DUNDEE PD SUERETH,DONALD,P	08/29/2007	11:40:28	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	09/11/2007	21:39:36	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:NOT DETECTED, POSSIBLE BAD SEAL, . Non-compliance: POSSIBLE BAD ACETONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	09/11/2007	23:01:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	10/10/2007	15:08:27	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	11/14/2007	13:38:34	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001078

Agency Operator	Date	Time	Software	Source	
DUNDEE PD BOWMAN,RICKY,W	12/07/2007	15:32:13	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	01/25/2007	11:33:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT WAS CLEAN AND DRY; TEMPERATURE 71.6 F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	02/08/2007	11:19:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT WAS CLEAN AND DRY. TEMPERATURE 72.2 DEGREES

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	03/09/2007	14:04:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect (SHUT OFF PORTABLE RADIO).

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	04/09/2007	16:07:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 71.6 DEGREES (F)

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	05/11/2007	15:47:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 72.5 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	06/14/2007	15:29:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 72.0 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	07/24/2007	21:28:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE 72.0 DEGREES

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	08/15/2007	15:28:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 71.6 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	09/07/2007	21:31:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 72.0 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source
LARGO PD SUERETH,DONALD,P	10/02/2007	15:07:12	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

OFFICER M. FREIRE PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	10/02/2007	21:47:42	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 73.3 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	11/06/2007	21:28:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 71.8 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001079

Agency Operator	Date	Time	Software	Source	
LARGO PD FREIRE,MICHAEL,A	12/11/2007	21:48:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AMBIENT TEMPERATURE WAS 72.0 DEGREES F

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	01/16/2007	08:52:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502101G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	01/24/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	02/18/2007	06:10:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	03/21/2007	10:41:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	04/15/2007	10:02:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	05/16/2007	09:49:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	06/19/2007	09:48:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

WITNESSED BY D/D WIEGERT #4591

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	07/20/2007	10:03:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	08/08/2007	11:25:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	08/30/2007	12:24:27	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 11 April 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	09/21/2007	09:24:38	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	10/23/2007	09:42:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.177
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.184
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.183
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	11/26/2007	11:01:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.201
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: UNKNOWN RFI DETECT ALL ELECTRONICS OFF.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001080

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	12/18/2007	11:28:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.