Florida Department of Law Enforcement

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source
PASCO COUNTY SO SUERETH,DONALD,P	05/16/2007	16:04:31	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	05/16/2007	16:49:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	06/19/2007	19:56:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	07/25/2007	18:51:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	08/22/2007	17:33:25	8100.27	Agency	Inspection

0.05 Lot#	200602D	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	02/08/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	09/13/2007	20:18:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	10/10/2007	13:03:33	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	11/07/2007	19:51:44	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001081

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	12/12/2007	21:14:42	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	01/09/2007	13:02:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	02/15/2007	19:28:19	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	03/15/2007	19:13:31	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	04/18/2007	19:42:22	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	05/17/2007	20:11:22	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	06/21/2007	20:13:18	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	07/19/2007	19:01:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	08/16/2007	19:48:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source
PLANTATION PD VEIGA,SANDRA,F	09/06/2007	13:31:54	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

INSPECTOR S. CAMPBEL WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	09/20/2007	19:59:45	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

CSA R. LAHIFF IN ATTENDANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	10/24/2007	19:57:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

OBSERVED BY A.I. ROSE LAHIFF

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	11/15/2007	19:51:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

OBSERVED BY A.I. LAHIFF

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001082

Agency Operator	Date	Time	Software	Source	
PLANTATION PD CAMPBELL,STUART,	12/19/2007	20:02:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

OBSERVED BY AI R. LAHIFF

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source
JACKSONVILLE SO MALHIOT,MATTHEW,E	04/12/2007	10:15:46	8100.27	Department Inspection

0.05 Lot#	200605C	**0.08 Lot#**	200605D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/10/2007	10:16:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSTR. PLACED INTO SERVICE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/15/2007	12:26:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	10/31/2007	09:38:53	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/16/2007	09:59:15	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	N	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	11/29/2007	14:13:15	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/13/2007	12:41:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001083

Agency Operator	Date	Time	Software	Source	
JACKSONVILLE SO THOMASON,ROBERT,D	12/31/2007	13:05:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	01/03/2007	16:24:43	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	02/25/2007	22:31:37	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	03/06/2007	09:19:11	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	04/14/2007	10:34:47	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.052
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.207
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.207
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. 00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	05/11/2007	16:49:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	06/26/2007	15:59:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.082
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	07/30/2007	13:23:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

UNKNOWN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	07/31/2007	18:50:44	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	08/31/2007	11:13:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source
FL HIGHWAY PATROL SUERETH,DONALD,P	09/11/2007	10:51:06	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.053
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.053
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.053
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.053
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.053
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.053

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.053
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.053
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

TROOPER A. PALESE PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	09/22/2007	16:37:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	10/11/2007	17:03:50	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	11/27/2007	23:30:14	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: DID NOT DETECT MA. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	11/28/2007	01:06:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

A F / M A: DID NOT DETECT MA.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001109

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	12/27/2007	10:50:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	01/31/2007	03:24:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	02/23/2007	04:36:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	03/29/2007	20:33:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	04/30/2007	18:02:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	05/23/2007	18:02:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	06/23/2007	16:54:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	07/24/2007	01:14:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source
FHP TROOP E MIAMI VEIGA,SANDRA,F	07/31/2007	13:32:58	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

RAY ADDISON WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	08/31/2007	04:02:50	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	09/29/2007	17:15:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside HOSE DETACHED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	10/29/2007	23:28:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	11/23/2007	09:53:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001110

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	12/18/2007	13:48:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	01/18/2007	23:27:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200504D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	05/04/2008	**Expiration**	04/21/2007	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	02/23/2007	12:19:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200504D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	05/04/2008	**Expiration**	04/21/2007	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	03/28/2007	13:19:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200504D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	05/04/2008	**Expiration**	04/21/2007	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	04/29/2007	10:28:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance CONNECTED INCORRECT SIM 0.05.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	05/18/2007	18:21:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	06/22/2007	09:40:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	07/30/2007	16:59:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612D	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/14/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.029
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.029
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	0.029
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:ACC ADDED 3ML OF ALC INST OF ACE,.20 LOT# 200612A .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	08/03/2007	11:27:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source
FHP TROOP A GEDDINGS,MARGARET,M	09/10/2007	12:16:57	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	09/10/2007	20:04:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	10/14/2007	14:16:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	11/23/2007	02:30:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.064
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001111

Agency Operator	Date	Time	Software	Source	
FHP TROOP A RIVERA,NICK,	12/27/2007	01:02:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	01/07/2007	10:44:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	02/04/2007	15:48:27	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. A F / M A:UNKNOWNAmbient FailOPERATOR ERRORRange ExceedUNK .
Non-compliance:UNKNOWN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	02/05/2007	11:23:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.201
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: SIMULATOR AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	03/09/2007	12:37:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	04/05/2007	13:54:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:MA REASON UNKNOWN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	05/14/2007	13:12:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	06/04/2007	13:40:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A:UNKNOWN. 08: Control Outside Toler. SIM LEAK AKSIM AIR LEAK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	07/11/2007	14:45:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	08/06/2007	15:51:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A:MA UNKNOWN REASON.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source
FL HIGHWAY PATROL HACKNEY,DWITE,N	09/06/2007	14:55:00	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.076
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.076
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.076
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	09/06/2007	16:09:44	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:UNK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	10/09/2007	12:28:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	11/02/2007	15:01:41	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001112

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	12/02/2007	11:25:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	01/03/2007	16:17:05	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	02/25/2007	22:27:37	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200404D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	03/06/2007	09:08:35	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	04/14/2007	10:23:20	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	630702I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.209
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	05/11/2007	16:40:38	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	06/26/2007	15:56:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1 (Retry)	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3 (Retry)	0.082
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. 08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	07/31/2007	18:44:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

UNKNOWN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	08/31/2007	11:09:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

UNKNOWN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source
FL HIGHWAY PATROL SUERETH,DONALD,P	09/11/2007	11:08:52	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.202
0.20 g/210L Test #1 (Retry)	0.202
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2 (Retry)	0.203
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3 (Retry)	0.203
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4 (Retry)	0.203
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5 (Retry)	0.204
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6 (Retry)	0.204
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7 (Retry)	0.203
0.08 g/210L Dry Gas Std Test #7	0.076
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8 (Retry)	0.203
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9 (Retry)	0.203
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10 (Retry)	0.203
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

20: RFI Detect.OFFICER IN CLOSE PROXIMITY WITH CELL PHONE. . TROOPER A. PALESE PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	09/22/2007	16:29:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	10/11/2007	17:00:06	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	11/28/2007	01:00:40	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001113

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL PALESE,ANTHONY,W	12/27/2007	10:45:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	01/30/2007	19:57:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.178
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance SIM JAR NOT TIGHT ENOUGH . INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	02/28/2007	22:08:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:MEASURE ACETONE STOCK TO WRONG LEVEL . 05: Ambient Fail.SOLUTION IN TUBE FROM SIM PRESSURE CHECK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. HARRELSON,BOBBY,C	03/25/2007	22:08:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NOT ENOUGH M/A.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	04/27/2007	10:10:21	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

STILL HAD ALC IN MOUTH DURING AF TEST FDLE NOTIFIED Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	04/27/2007	10:38:37	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance,FDLE NOTIFIED Non-compliance: .05 TEST OUTSIDE TOLERANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source
DELAND P.D. SKIPPER,ROGER,G	04/30/2007	11:40:49	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	04/30/2007	12:55:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.071
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.187
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.197
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance AIR LEAK AT SIM. 20: Control Outside Tolerance AIR LEAK AT SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	05/31/2007	05:38:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	06/30/2007	11:20:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	07/30/2007	16:15:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	08/30/2007	07:48:38	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. HARRELSON,BOBBY,C	09/28/2007	05:30:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INTOXILYZER OPERATING CORRECTLY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	10/31/2007	04:22:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.095
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.088
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.089
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.087
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.085
		0.08 g/210L Test #3 (Retry)	0.087
		Interferent Detect Test #3	INT

Remarks:

08: PADDLE GOT STUCK ON SIM:RESOLVED, Control Outside Toler. Non-compliance: OUT OF TOLERANCE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	10/31/2007	14:44:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724701G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	11/30/2007	19:18:15	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	724701G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: COUGHED WHILE BLOWING, PRODUCED READING . INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001114

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	12/31/2007	16:15:58	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	724701G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	01/04/2007	10:08:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	627101I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	02/21/2007	10:11:40	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	606002I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.210
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.085
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.210
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.085
		0.20 g/210L Test #3	0.210
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.RECIRCULATED ROOM AIR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	03/22/2007	13:59:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.RECIRCULATED ROOM AIR. 08: RFI Detect.CELPX ON..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	04/03/2007	10:50:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	05/16/2007	15:07:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	06/18/2007	11:20:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.209
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	07/09/2007	15:30:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2006	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.RECIRCULATED ROOM AIR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	08/15/2007	09:42:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Improper Sample.INSP ERROR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	09/17/2007	11:23:39	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2 (Retry)	
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: No Sample Provided. 00: Ambient Fail, RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	09/17/2007	12:00:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	10/03/2007	13:30:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.LEFT D/GAS LINE OFF.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source
FHP TROOP H GEDDINGS,MARGARET,M	10/18/2007	14:58:28	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.205
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.204
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.205
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.205
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.205
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.206
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	11/27/2007	09:48:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001115

Agency Operator	Date	Time	Software	Source	
FHP TROOP H STRICKLAND,JAMES,M	12/11/2007	10:59:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source
EGLIN AFB POLICE MALHIOT,MATTHEW,E	02/05/2007	10:58:51	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.082
			0.20 g/210L Test #7	0.200
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0252
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Control Outside Tolerance: FAILED TO HOOK UP D/G HOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE WATSON,HEATHER,L	04/09/2007	22:23:23	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	706101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Purge Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE WATSON,HEATHER,L	05/31/2007	03:25:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	706101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE WATSON,HEATHER,M	06/25/2007	14:12:19	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	706101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/02/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE KNIGHT,GARRETT,M	07/31/2007	09:51:38	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE KNIGHT,GARRETT,M	07/31/2007	10:50:51	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	706101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.182
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.187
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.076
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:BAD SEAL ON SIMULATOR. 08: Improper Sample. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE KNIGHT,GARRETT,M	08/28/2007	14:21:09	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	706101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE KNIGHT,GARRETT,M	09/27/2007	11:43:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE KNIGHT,GARRETT,M	10/15/2007	17:14:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE KNIGHT,GARRETT,M	11/30/2007	09:33:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001116

Agency Operator	Date	Time	Software	Source	
EGLIN AFB POLICE LACHAPELLE,MARK,W	12/31/2007	16:03:42	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/03/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside: BAD AIR SEAL; FIXED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	01/11/2007	18:03:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	02/13/2007	19:24:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	03/13/2007	13:22:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	04/14/2007	12:50:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source
PASCO COUNTY SO SUERETH,DONALD,P	05/16/2007	15:24:41	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.195
0.20 g/210L Test #4 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.20 g/210L Test #5 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.196
0.20 g/210L Test #6 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.196
			0.20 g/210L Test #7 (Retry)	0.195
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.20 g/210L Test #8 (Retry)	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.20 g/210L Test #9 (Retry)	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.20 g/210L Test #10 (Retry)	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0023
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

20: Control Outside Tolerance. PWR OUTLET ACC. TURNED OFF. . DEPUTY L. WINTERS PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	05/16/2007	16:11:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	06/19/2007	19:50:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	07/25/2007	18:54:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	08/22/2007	17:30:06	8100.27	Agency	Inspection

0.05 Lot#	200602D	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	02/23/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	09/13/2007	20:06:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	713101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	10/10/2007	12:55:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	11/07/2007	19:54:47	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001117

Agency Operator	Date	Time	Software	Source	
PASCO COUNTY SO WINTERS,LANE,A	12/12/2007	21:15:24	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001118

Agency Operator	Date	Time	Software	Source
FL HIGHWAY PATROL SUERETH,DONALD,P	09/11/2007	11:22:13	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001118

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

TROOPER A. PALESE PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	01/17/2007	11:21:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	02/20/2007	17:46:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	03/22/2007	11:15:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.031
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.191
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.031
0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.179
0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.029
0.05 g/210L Test #3 (Retry)	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.181
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.SIMULATOR SD1118 ON .05 AND.20 VALVE TO INLET FOUND LOOSE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	04/18/2007	18:44:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.069
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.046
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.044
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:MIXED NEW SIMULATOR WITH ACETONE. 08: Control OutsideSIMULATOR SD1118 REMOVED FROM SERVICE AFTER FIRST TEST AND

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	05/25/2007	18:20:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	06/25/2007	11:58:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	07/26/2007	17:31:50	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:REMIXED SIMULATOR, . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	07/31/2007	17:03:05	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2006

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: USED 2 SEPARATE SIMULATORS FOR THE .05 .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	07/31/2007	20:25:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.000
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	0.000
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: . 05: Control Outside Tolerance. 08: Control Outside.

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source
FHP TROOP D - DELAND SKIPPER,ROGER,G	08/22/2007	17:08:25	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.049
0.20 g/210L Test #1 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.050
0.20 g/210L Test #2 (Retry)	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.049
0.20 g/210L Test #3 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.049
0.20 g/210L Test #4 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.049
0.20 g/210L Test #5 (Retry)	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.049
0.20 g/210L Test #6 (Retry)	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.081
			0.20 g/210L Test #7	0.049
			0.20 g/210L Test #7 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.050
			0.20 g/210L Test #8 (Retry)	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.050
			0.20 g/210L Test #9 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.049
			0.20 g/210L Test #10 (Retry)	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

20: Control Outside Tolerance. WRONG SIMULATOR USED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	08/29/2007	12:42:21	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.009
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.199
		0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:REMIXED NEW SIM. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	09/17/2007	17:55:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.000
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.116
0.20 g/210L Test #1 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.000
0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.000
0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.05 g/210L Test #3 (Retry)	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.000
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance, CHECKED SIM SEALS. 20: ControlINSPECTION COMPLETE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	10/16/2007	11:34:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	11/14/2007	17:16:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001119

Agency Operator	Date	Time	Software	Source	
FHP TROOP D - DELAND MORRISSEY,PATRICK,W	12/18/2007	10:12:52	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001120

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL VEIGA,SANDRA,F	08/28/2007	13:16:08	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

INSTRUMENT NOT IN SERVICE, TEST INSPECTION TO ENABLE INSPECTION.

Non-compliance:ENABLED PRIOR TO DEPT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001120

Agency Operator	Date	Time	Software	Source
F.H.P. VEIGA,SANDRA,F	12/13/2007	14:46:56	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001120

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001120

Agency Operator	Date	Time	Software	Source	
F.H.P. ADDISON,RAYMOND,E	12/18/2007	13:54:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632002I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	01/31/2007	03:19:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	02/23/2007	04:42:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	03/29/2007	20:27:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside HOSE DISCONNECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	04/30/2007	18:07:48	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	05/23/2007	18:12:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	06/23/2007	16:42:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	07/24/2007	01:03:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source
FHP TROOP E MIAMI VEIGA,SANDRA,F	07/31/2007	12:39:24	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	08/31/2007	03:51:12	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	09/29/2007	17:03:51	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	10/29/2007	23:35:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	11/23/2007	09:58:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001121

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	12/18/2007	13:54:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632002I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	01/31/2007	03:06:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	02/23/2007	04:30:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	03/29/2007	20:09:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	04/30/2007	17:51:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	05/23/2007	17:55:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	06/23/2007	17:00:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: HOSE NOT CONNECTED 1ST SAMPLE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	07/24/2007	01:17:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source
FHP TROOP E MIAMI VEIGA,SANDRA,F	07/31/2007	13:15:21	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

RAY ADDISON WAS PRESENT FOR INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001122

Agency Operator	Date	Time	Software	Source	
FHP TROOP E MIAMI ADDISON,RAYMOND,E	08/31/2007	04:11:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	11/06/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	01/29/2007	19:15:06	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededAmbient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	01/29/2007	19:19:14	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	01/29/2007	20:04:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	02/27/2007	10:00:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	03/29/2007	07:56:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	04/22/2007	20:22:33	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	05/29/2007	16:13:19	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	06/29/2007	14:53:30	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	07/30/2007	18:01:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source
FHP TROOP F FT MYERS VENTURI,GEORGE,L	08/29/2007	14:07:48	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	08/31/2007	10:48:28	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	08/31/2007	12:49:52	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

LOT 200905A FOR .05

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	09/21/2007	16:12:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	10/29/2007	21:15:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	11/26/2007	16:35:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001123

Agency Operator	Date	Time	Software	Source	
FHP TROOP F FT MYERS TAYLOR,DANIEL,W	12/30/2007	11:13:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	01/16/2007	07:44:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	612201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	02/18/2007	03:23:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	03/21/2007	13:19:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	04/15/2007	08:04:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	05/16/2007	11:13:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO TARTE,PHILLIP,A	06/19/2007	08:08:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	07/20/2007	12:48:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	08/08/2007	09:52:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	08/30/2007	11:24:37	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	09/21/2007	09:29:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source	
POLK COUNTY SO WIEGERT,KIMBRA,L	10/23/2007	09:47:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source
POLK COUNTY SO SUERETH,DONALD,P	12/20/2007	14:27:33	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001127

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0019
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Ambient FaiL. UNKNOWN SOURCE AREA FANNED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	01/15/2007	12:45:42	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	01/28/2007	22:32:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:ACETONE BAD.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	02/14/2007	15:24:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	03/08/2007	18:13:02	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	04/10/2007	18:12:55	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

THERE WAS AIR LEAK IN SIM. 08:SIR LEAK IN SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	05/22/2007	10:24:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

TOO MUCH MOUTH ALCOHOL USED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source
PUTNAM COUNTY SO MALHIOT,MATTHEW,E	06/05/2007	10:47:25	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.076
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JON,D	06/21/2007	10:28:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JON,D	06/21/2007	14:04:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JON,D	07/24/2007	20:17:45	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	08/01/2007	15:20:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JONATHAN,D	09/06/2007	09:45:08	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JON,D	10/29/2007	14:18:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:TOO MUCH ALCOHOL USED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JON,D	11/27/2007	11:35:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001128

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY SO KINNEY,JON,D	12/29/2007	10:24:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001129

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	01/28/2007	20:22:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001129

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	02/22/2007	19:13:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001129

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	03/09/2007	00:24:04	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001129

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	03/09/2007	01:01:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001129

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	04/12/2007	09:51:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001129

Agency Operator	Date	Time	Software	Source	
PUTNAM COUNTY S.O. KINNEY,JON,D	05/30/2007	09:32:51	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

SENT TO CMI FOR PRINTER PORT REPAIR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	01/24/2007	16:43:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	02/14/2007	13:44:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	03/29/2007	03:29:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	04/28/2007	19:36:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	05/15/2007	15:46:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	06/30/2007	02:05:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2007	**Expiration**	05/04/2007	**Expiration**	05/10/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	07/28/2007	18:36:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	08/30/2007	17:10:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.042
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.044
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.077
0.08 g/210L Test #3 (Retry)	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	09/29/2007	21:42:27	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source
VOLUSIA S.O. SKIPPER,ROGER,G	10/03/2007	21:08:33	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.08 g/210L Test #1 (Retry)	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.08 g/210L Test #3 (Retry)	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.08 g/210L Test #4 (Retry)	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.08 g/210L Test #5 (Retry)	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.08 g/210L Test #6 (Retry)	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.080
			0.08 g/210L Test #7 (Retry)	0.080
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.077
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8 (Retry)	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9 (Retry)	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10 (Retry)	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

A F / M A: IMPROPER SAMPLE/REPEATED. 08: RFI Detect RADIO KEYED BY INSTRUMENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	10/24/2007	19:29:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	11/29/2007	16:38:08	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08:FORGOT TO HOOK UP GAS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001130

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	12/19/2007	16:37:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.026
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.060
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	01/24/2007	16:38:25	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	618801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	02/14/2007	22:54:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	03/28/2007	18:59:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	04/28/2007	19:31:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2007	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Interferent Detect. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	05/15/2007	21:29:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	06/29/2007	20:42:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.194
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

LOW CONTROL TESTS DUE TO COLD ROOM

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	07/31/2007	01:33:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COLD ROOM, TEMP TOO LOW.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	08/30/2007	17:16:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	09/28/2007	00:16:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source
VOLUSIA S.O. SKIPPER,ROGER,G	10/03/2007	20:42:05	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	10/25/2007	18:07:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	11/29/2007	16:33:18	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001131

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	12/20/2007	02:40:56	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	01/24/2007	23:53:47	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	02/14/2007	13:49:16	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518801I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	03/28/2007	18:53:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	04/29/2007	03:36:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	05/15/2007	15:51:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	06/29/2007	20:05:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.063
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.070
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.063
0.08 g/210L Test: OK	No	0.08 g/210L Test #2 (Retry)	0.073
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.063
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

5. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.08 SAMPLE TOO LOW.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	06/30/2007	01:02:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2007	**Expiration**	05/10/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SECOND INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	07/28/2007	18:35:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.076
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.075
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.075
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

WRONG SIMULATOR ATTACHED'. COLD ROOM, LOW TEMP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	08/30/2007	22:06:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	09/28/2007	00:20:55	8100.27	Agency	Inspection

0.05 Lot#	200807A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705301I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/27/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
VOLUSIA S.O. SKIPPER,ROGER,G	10/03/2007	20:27:45	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	10/24/2007	19:26:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705301I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/27/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail UNKNOWN PROBLEM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	11/29/2007	23:15:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705301I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/27/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect. 00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001132

Agency Operator	Date	Time	Software	Source	
VOLUSIA S.O. IHNKEN,MATTHEW,D	12/19/2007	16:34:10	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705301I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/27/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.033
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.028
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.029
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	01/31/2007	14:30:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	02/27/2007	15:30:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/31/2007	07:00:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.067
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.074
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMULATOR LID NOT TIGHTENED DOWN ALL THE WAY, RESEALED. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	04/30/2007	18:31:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	05/31/2007	20:23:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMTOHY,B	06/30/2007	12:40:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	07/30/2007	07:10:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	08/29/2007	15:14:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	09/29/2007	18:36:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	10/28/2007	18:07:34	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	10/28/2007	18:11:30	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	10/28/2007	18:16:51	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	10/28/2007	19:02:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	10/30/2007	11:02:34	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	11/29/2007	08:43:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001133

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	12/27/2007	14:31:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMTOHY,B	01/31/2007	14:41:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	02/27/2007	15:22:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	03/31/2007	07:13:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	04/30/2007	18:26:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	05/31/2007	20:27:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	06/30/2007	12:37:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROBERTSON,TIMOTHY,B	07/30/2007	07:05:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	10/08/2007	12:39:35	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.046
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.046

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

A F / M A: TOO MUCH SOLUTION USED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	10/28/2007	18:18:22	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	10/28/2007	19:10:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	11/29/2007	08:11:55	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:FAILED TO USE MA SOLUTION. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	11/29/2007	08:59:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001134

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	12/27/2007	14:44:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	01/29/2007	20:24:57	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	615802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	03/07/2007	17:29:02	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	03/08/2007	23:49:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NON-COMPLIANCE ON 3-7-07/WEAK SOLUTION, SKIPPER NOTIFIED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	04/24/2007	03:44:45	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	05/23/2007	00:28:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	06/19/2007	20:56:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	07/13/2007	03:00:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source
DAYTONA BEACH P.D. SKIPPER,ROGER,G	08/28/2007	13:28:23	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	08/28/2007	17:14:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	09/30/2007	20:35:57	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	10/26/2007	20:29:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	08/01/2009	**Expiration**	12/14/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.061
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.070
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.066
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.071
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.071
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	10/26/2007	21:41:07	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	INT
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.042
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.069
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.186
0.20 g/210L Test #1 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

00: Interferent Detect. 05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	11/15/2007	01:49:23	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.181
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001135

Agency Operator	Date	Time	Software	Source	
DAYTONA BEACH P.D. WARD,JAMES,A	12/31/2007	20:24:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	01/30/2007	20:02:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	02/28/2007	22:17:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. HARRELSON,BOBBY,C	03/25/2007	22:00:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

8000 CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	04/27/2007	11:16:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.000
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Ambient Fail. Non-compliance:000 ON .20 TEST AND AMB FAIL FDLE NOTIFIED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
DELAND P.D. SKIPPER,ROGER,G	04/30/2007	11:26:04	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	04/30/2007	13:03:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	05/31/2007	05:29:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	06/30/2007	11:33:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	07/30/2007	16:06:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	08/30/2007	07:38:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. HARRELSON,BOBBY,C	09/28/2007	05:07:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	618801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

POWER TURNED OFF DURING FIRST TEST/SECOND TEST NO PROBLEM

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	10/31/2007	04:37:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724701G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: OUTSIDE TOL DUE TO SIM SWITCH OFF & COOLED . INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	11/30/2007	19:23:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	724701G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001136

Agency Operator	Date	Time	Software	Source	
DELAND P.D. PETRELLA,TRINITY,J	12/31/2007	16:09:39	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	724701G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSTRUMENT IS CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	01/30/2007	06:22:41	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	01/30/2007	06:56:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	02/27/2007	12:55:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	03/31/2007	14:47:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	04/30/2007	16:34:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail UNKNOWN SCOURSE, THEN PURGE FAIL. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	05/31/2007	19:12:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	06/30/2007	18:18:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	07/31/2007	14:38:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	09/04/2007	19:36:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	609301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source
FLAGLER COUNTY SO MALHIOT,MATTHEW,E	10/29/2007	14:36:05	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0011
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	10/30/2007	15:10:07	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	609301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	11/30/2007	13:48:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

CHANGE OF GAS CYLINDER THIS INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001137

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	12/31/2007	15:08:32	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	01/30/2007	07:35:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	02/27/2007	12:04:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	03/31/2007	15:21:49	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	04/30/2007	15:51:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	05/31/2007	19:58:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.OPEN MOUTH ALCOHOL COTAINER NEXT TO INTAKE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	06/30/2007	17:38:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	07/31/2007	15:11:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY S.O. BRAY,JOHN,A	09/04/2007	18:52:45	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	609301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source
FLAGLER COUNTY SO MALHIOT,MATTHEW,E	10/29/2007	14:47:37	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	04/05/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	10/30/2007	15:47:52	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	609301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	11/30/2007	12:57:49	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

GAS CYLINDER CHANGE THIS INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001138

Agency Operator	Date	Time	Software	Source	
FLAGLER COUNTY SO BRAY,JOHN,A	12/31/2007	15:43:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	611501G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/27/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	01/12/2007	12:43:50	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	01/26/2007	16:38:15	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

F.DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	02/23/2007	11:08:42	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	03/08/2007	10:54:30	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

LOOSE CONNECTION. P.NANZ WAS PRESENT AND OBSERVED INSPECTIO

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	03/30/2007	11:34:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	04/20/2007	16:48:38	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	04/27/2007	17:18:53	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	05/11/2007	10:09:11	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	05/25/2007	15:25:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	06/08/2007	10:58:17	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	06/22/2007	09:47:19	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	07/12/2007	15:09:49	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632102I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	08/03/2007	10:04:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	711401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	08/24/2007	15:50:11	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	711401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
BROWARD COUNTY SO VEIGA,SANDRA,F	08/29/2007	15:41:42	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.05 g/210L Test #4 (Retry)	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.05 g/210L Test #5 (Retry)	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.05 g/210L Test #6 (Retry)	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7 (Retry)	0.049
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.05 g/210L Test #8 (Retry)	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.05 g/210L Test #9 (Retry)	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.000
			0.05 g/210L Test #10 (Retry)	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0152
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

05: Control Outside Tolerance.SIMULATOR NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	09/07/2007	15:28:56	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	711401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	09/28/2007	11:05:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	10/12/2007	15:46:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	11/09/2007	12:52:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO NANZ,PATRICIA,M	11/30/2007	15:28:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

F DICIENZO WAS PRESENT AND OBSERVED THE ENTIRE INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001139

Agency Operator	Date	Time	Software	Source	
BROWARD COUNTY SO DICIENZO,FRANCA,R	12/07/2007	15:58:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/29/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PAT NANZ WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	01/23/2007	23:49:31	8100.27	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	02/05/2007	22:40:42	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.20 g/210L Test #3 (Retry)	0.199
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside ToleranCe.ONE SIMULATOR FOR BOTH,OUTSIDE OF TOLERANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	03/02/2007	00:37:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	04/08/2007	21:04:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	05/06/2007	22:21:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	06/26/2007	00:45:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: COT SIMULATOR NOT WARMED UP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. BELLESFIELD,SCOTT,J	08/02/2007	21:36:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
LAKE HELEN P.D. SKIPPER,ROGER,G	08/22/2007	11:22:35	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0013
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	09/19/2007	21:43:35	8100.27	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	09/19/2007	21:50:42	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:MOUTH ALCOHOL INTRODUCED FIRSTSequence Aborted. Non-compliance:INSPECTOR ERROR, MOUTH ALC INTRO. FIRST .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	09/19/2007	22:45:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.000
0.20 g/210L Test #1 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.000
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.190
0.20 g/210L Test #3 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: CONTROL OUTSIDE TOLERANCE. AIR LEAK CORRECTED e.AIR LEAK CORRECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	10/17/2007	16:55:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside/AIR LEAK CORRECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	11/14/2007	16:16:40	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	08/01/2009	**Expiration**	12/14/2008		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.08 OUT OF RANGE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	11/16/2007	11:44:47	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001149

Agency Operator	Date	Time	Software	Source	
LAKE HELEN P.D. MULLINS,ROBERT,B	12/18/2007	12:12:49	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.035
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.033
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. AIR LEAK CORRECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	01/23/2007	08:58:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	02/14/2007	08:23:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	03/06/2007	08:32:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	04/17/2007	08:37:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	05/08/2007	08:42:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. SMOKE IN OUTSIDE AIR BROUGHT IN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	06/05/2007	08:57:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
MANATEE COUNTY SO VENTURI,GEORGE,L	07/16/2007	13:01:55	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	08/01/2007	08:36:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	09/12/2007	09:12:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	10/02/2007	16:14:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	11/06/2007	09:34:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001150

Agency Operator	Date	Time	Software	Source	
MANATEE COUNTY SO ANDREWS,ROBERT,P	12/13/2007	14:23:53	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	01/19/2007	10:24:31	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	01/19/2007	10:28:49	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	01/19/2007	11:04:33	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502101G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	01/24/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	02/01/2007	11:47:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	03/09/2007	14:35:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det SIMULATOR NOT PROPERLY HEATED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	04/17/2007	14:44:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	05/30/2007	15:51:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	06/29/2007	15:31:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
HAINES CITY PD SUERETH,DONALD,P	07/10/2007	12:01:55	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.080
			0.20 g/210L Test #7	0.200
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.081
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08:OUT OF TOL/ASSISTING WITH 80-1007 AI. GAS NOT CONECTED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	07/10/2007	12:42:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	08/30/2007	14:16:41	8100.27	Agency	Inspection

0.05 Lot#	200606B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	09/28/2007	14:14:59	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	09/28/2007	15:00:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: DIDN'T HAVE ENOUGH MOUTH ALCOHOL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	10/29/2007	11:20:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect OFC WALKED IN WITH RADIO ON . OFC WALKED IN WITH RADIO ON

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD STEWART,ANGELA,R	11/30/2007	16:12:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001151

Agency Operator	Date	Time	Software	Source	
HAINES CITY PD RODGERS,KAREN,D	12/26/2007	16:53:10	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	01/22/2007	12:58:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	02/28/2007	10:25:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	03/28/2007	10:52:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	04/26/2007	11:57:35	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded. MOUTH ALCOHOL EXCEEDED RANGE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	04/26/2007	12:54:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	05/30/2007	18:25:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	06/30/2007	15:17:32	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	06/30/2007	15:24:34	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	06/30/2007	15:53:41	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.048
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.048
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededRange Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	06/30/2007	16:31:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	07/31/2007	21:39:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD SKIPPER,ROGER,G	08/28/2007	16:25:13	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/14/2009

 # Sims 5 **Avg Std Dev** 0.0009 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.070
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.064
0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test #2	0.205
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.058
0.08 g/210L Test #3 (Retry)	0.080
0.20 g/210L Test #3	0.205
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.060
0.08 g/210L Test #4 (Retry)	0.080
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.069
0.08 g/210L Test #5 (Retry)	0.080
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.082
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.075
0.08 g/210L Test #6 (Retry)	0.081
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.076
			0.08 g/210L Test #7 (Retry)	0.081
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.082
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.08 g/210L Test #8	0.078
			0.08 g/210L Test #8 (Retry)	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.08 g/210L Test #9 (Retry)	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.08 g/210L Test #10 (Retry)	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0076
			0.08 g/210L Test (Std Dev)	0.0010
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

08: Control Outside Tolerance SIM HEATER FAIL. SIM CHANGED.. 0.08 SIMULATOR REPLACED FOR 2ND 0.08 TESTS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	08/29/2007	01:42:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	09/20/2007	21:37:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	10/25/2007	17:30:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	11/13/2007	11:19:41	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001152

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	12/30/2007	20:32:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	01/23/2007	08:34:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	02/14/2007	09:49:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	03/21/2007	10:08:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	04/30/2007	17:15:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	05/01/2007	07:55:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	06/15/2007	08:15:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	07/31/2007	08:06:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	08/01/2007	07:44:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	09/24/2007	07:47:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	10/29/2007	09:24:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	11/02/2007	08:18:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source	
LEON COUNTY SO GIANGROSSO,JOE,P	12/17/2007	08:12:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source
LEON COUNTY SO GEDDINGS,MARGARET,M	12/17/2007	12:30:24	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001153

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	01/24/2007	23:49:05	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.086
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.085
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.085
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.084
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.085
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	02/14/2007	22:46:04	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.210
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.085
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	03/29/2007	03:23:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.217
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.205
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.053
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.083
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.215
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.212
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	04/29/2007	03:23:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.210
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	05/15/2007	21:22:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. IHNKEN,MATTHEW,D	06/14/2007	00:52:03	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	07/28/2007	04:33:58	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.036
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.038
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.047
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.039
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.048
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, SIMULATOR WAS NOT 34 C Non-compliance:SIMULATOR NOT 34 C.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	07/28/2007	05:26:27	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance , POSSIBLY A BAD BATCH OF .05Non-compliance:BOTH BATCHES OF .05 SOLUTION WERE AT 34C. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	07/28/2007	19:37:07	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.132
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.163
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.175
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance AIR LEAK ON SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source
OAK HILL P.D. SKIPPER,ROGER,G	08/21/2007	16:44:27	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.206
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.206
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.207
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.083
0.20 g/210L Test #5	0.207
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.083
0.20 g/210L Test #6	0.207
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.084
0.20 g/210L Test #7	0.207
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.084
			0.20 g/210L Test #8	0.207
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.084
			0.20 g/210L Test #9	0.208
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.084
			0.20 g/210L Test #10	0.207
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0016
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

OFFICER S. CHANDLER, A.I., PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	08/30/2007	21:58:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.053
0.08 g/210L Test #1	0.083
0.20 g/210L Test #1	0.209
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.054
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.209
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.054
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.210
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	09/27/2007	21:19:05	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.036
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.036
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.038
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance . AIR LEAK ON SIMULATOR. FDLE/ SKIPPER NOTIFIED. REINSPECT. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	09/27/2007	22:15:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.048
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2 (Retry)	0.078
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.049
0.08 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: RFI Detect, Improper Sample. Non-compliance:MULTIPLE SIM. LEAKS. SIMS TO BE CALIBRATED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	09/27/2007	23:18:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.199
		0.20 g/210L Test #3 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance BENT HOSE ON SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	10/14/2007	19:10:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	11/29/2007	20:16:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001154

Agency Operator	Date	Time	Software	Source	
OAK HILL P.D. CHANDLER,SHANE,L	12/25/2007	19:22:22	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612202G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect CELL PHONE ON IN STATION/ NOW OFF.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	01/31/2007	13:47:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	02/28/2007	13:58:32	8100.27	Agency	Inspection

0.05 Lot#	200605F	**0.08 Lot#**	200605C	**0.20 Lot#**	200612C	**0.08 Gas Lot#**	627201I
Expiration	05/10/2008	**Expiration**	05/04/2008	**Expiration**	12/14/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	03/30/2007	16:11:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	62701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	04/30/2007	17:11:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	05/31/2007	16:11:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	06/29/2007	16:09:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.189
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.189
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	06/29/2007	16:59:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	07/31/2007	16:33:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
ORMOND BEACH P.D. SKIPPER,ROGER,G	08/28/2007	10:41:01	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	08/31/2007	13:53:22	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	09/30/2007	18:50:47	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.188
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	10/30/2007	15:40:59	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	11/30/2007	18:10:48	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001156

Agency Operator	Date	Time	Software	Source	
ORMOND BEACH P.D. HALLER,ROBERT,M	12/31/2007	17:05:27	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	01/31/2007	14:29:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	02/28/2007	15:14:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
LAKE WALES PD SUERETH,DONALD,P	03/06/2007	13:07:22	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0017
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	04/30/2007	10:04:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	06/07/2007	14:52:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD QUINN,PATRICK,D	07/06/2007	10:49:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	08/31/2007	15:15:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	09/30/2007	16:58:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	11/29/2007	04:50:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001157

Agency Operator	Date	Time	Software	Source	
LAKE WALES PD KEEL,RONALD,L	12/28/2007	02:21:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. DOMINY,RAYMOND,F	01/02/2007	00:52:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	02/05/2007	20:54:00	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	\626903I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	03/15/2007	02:55:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.075
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

08: AIR LEAK SIMULATOR HOSE, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	03/15/2007	03:48:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. DOMINY,RAYMOND,F	04/05/2007	20:24:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	05/06/2007	20:08:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.047
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.077
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #3	0.077
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

PUSHED WRONG BUTTON SKIPPED .20.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	05/06/2007	21:06:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	06/02/2007	03:52:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	07/14/2007	20:32:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	08/02/2007	00:59:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
PONCE INLET P.D. SKIPPER,ROGER,G	08/21/2007	10:51:46	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. DOMINY,RAYMOND,F	09/06/2007	01:46:33	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.099
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.031
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.100
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.039
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.099
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

.05 OUT OF RANGE. .O5 OUT OF RANGE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. DOMINY,RAYMOND,F	09/12/2007	01:16:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. SCOTT,ALLEN,J	10/10/2007	22:05:50	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. DOMINY,RAYMOND,F	11/01/2007	23:20:10	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/13/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001158

Agency Operator	Date	Time	Software	Source	
PONCE INLET P.D. DOMINY,RAYMOND,F	12/14/2007	00:56:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/13/2008	**Expiration**	09/28/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	01/10/2007	13:38:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

AREA WAS NEAT AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	02/20/2007	11:08:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AREA WAS CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	03/15/2007	10:06:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A:WOULD NOT SHOW MOUTH ALCOHOL. AREA CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	04/17/2007	13:13:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A:DID NOT READ SLOPE NOT MET. AREA NEAT AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	05/20/2007	19:49:16	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:NOT ENOUGH TIME TO MIX, . Non-compliance:NOT READING INTERFERENT DETECT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	05/20/2007	20:56:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 3 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.186
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	INT
0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.192
0.20 g/210L Test #2 (Retry)	0.193
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.194
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:NOT MIXED ENOUGH. 20: Control Outside Tolerance. 08: Interferent Detect, Interferent Detect.
Non-compliance:AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	05/20/2007	21:56:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.008
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.187
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	INT
0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.193
0.20 g/210L Test #2 (Retry)	0.194
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.193
0.20 g/210L Test #3 (Retry)	0.194
Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance. 20: Control Outside Tolerance.INTERFERENT DETECT ON DRY GASNon-compliance: INTERFERENT DETECT ON DRY GAS .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	05/20/2007	22:56:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:NOT AIR TIGHT SEAL. 08: Interferent Detect, Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	05/21/2007	11:05:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

SKIPPER NOTIFIED FAILED X4 DUE TO FLOOR CLEANING PRODUCTS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	06/26/2007	20:35:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.072
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.020
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.073
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	No	0.20 g/210L Test #2	0.045
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2 (Retry)	0.206
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.073
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.076
		0.20 g/210L Test #3 (Retry)	0.216
		Interferent Detect Test #3	INT

Remarks:

SIM TEMPS TOO HIGHNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	06/27/2007	12:41:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AREA NEAT AND CLEAN. SIM TEMPS WERE TO HIGH ON PREV. INSP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	07/25/2007	15:58:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. AREA NEAT AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
ORANGE CITY P.D. SKIPPER,ROGER,G	08/22/2007	14:06:39	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	08/26/2007	19:31:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AREA WAS CLEAN AND DRY.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	09/25/2007	10:47:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AREA WAS NEAT AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	10/20/2007	17:02:54	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AREA WAS NEAT AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	11/19/2007	07:54:58	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AREA WAS NEAT AND CLEAN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001159

Agency Operator	Date	Time	Software	Source	
ORANGE CITY P.D. PITZER,CHRISTOPHER,J	12/25/2007	08:01:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

AREA WAS NEAT AND CLEAN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	01/23/2007	16:37:02	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	02/21/2007	09:10:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	03/28/2007	14:48:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	04/04/2007	18:08:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	05/14/2007	10:12:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	06/19/2007	16:18:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	07/08/2007	12:58:27	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

PURGE FAILNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	07/08/2007	14:04:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	08/20/2007	11:22:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	09/17/2007	19:30:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2006	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	10/09/2007	14:15:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	20605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
BREVARD S.O. SKIPPER,ROGER,G	10/09/2007	17:05:36	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	11/13/2007	12:31:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001166

Agency Operator	Date	Time	Software	Source	
BREVARD S.O. WAGNER,ROBERT,F	12/17/2007	11:06:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	01/18/2007	13:59:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	02/15/2007	14:08:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 4 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.073
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.071
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.071
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.071
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.071
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.046
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.072
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.071
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	02/19/2007	13:43:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	03/13/2007	12:15:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	04/17/2007	13:56:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.073
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.074
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.073
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.073
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.074
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	04/18/2007	14:06:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	05/18/2007	14:28:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	06/15/2007	12:36:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	07/23/2007	14:20:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	08/15/2007	13:52:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	09/17/2007	14:10:49	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
CLERMONT P.D. SKIPPER,ROGER,G	10/15/2007	11:06:17	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	10/18/2007	14:44:54	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	11/21/2007	13:56:09	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001174

Agency Operator	Date	Time	Software	Source	
CLERMONT P.D. LEE,JOHN,P	12/26/2007	14:21:53	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	03/05/2007	03:28:30	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	05/17/2007	20:21:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	06/11/2007	04:05:21	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:OPERTOR ERROR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,SGT. WILLIAM,E	07/12/2007	04:56:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717102I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,WILLIAM,E	08/09/2007	05:34:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source
LADY LAKE P.D. SKIPPER,ROGER,G	09/05/2007	10:54:49	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.083
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.053
0.08 g/210L Test #5	0.083
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.053
0.08 g/210L Test #6	0.084
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.054
0.08 g/210L Test #7	0.084
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.053

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.084
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.054
			0.08 g/210L Test #9	0.084
			0.20 g/210L Test #9	0.204
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.054
			0.08 g/210L Test #10	0.084
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,SGT. WILLIAM,E	09/06/2007	04:14:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,SGT. WILLIAM,E	10/04/2007	08:32:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/15/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,SGT WILLIAM,E	11/14/2007	19:33:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	717102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001177

Agency Operator	Date	Time	Software	Source	
LADY LAKE P.D. OLSEN,SGT. WILLIAM,E	12/09/2007	12:35:07	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	01/30/2007	19:50:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	02/13/2007	21:42:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	03/21/2007	16:43:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	04/10/2007	17:23:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	05/10/2007	23:25:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	06/22/2007	19:10:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	07/18/2007	18:10:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	08/22/2007	14:48:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source
MIAMI P.D. VEIGA,SANDRA,F	09/12/2007	14:58:15	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626102I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

# Sims	5	**Avg Std Dev**	0.0005	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

OFFICER PEDRO BELTRAN WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	09/12/2007	18:14:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	10/30/2007	14:53:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	11/14/2007	21:52:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	626102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001178

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	12/11/2007	20:59:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	01/30/2007	19:45:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	02/13/2007	21:03:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	03/21/2007	15:25:09	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	03/21/2007	16:38:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	04/10/2007	17:18:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	05/10/2007	23:29:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	06/22/2007	18:49:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	07/18/2007	15:50:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	08/22/2007	15:00:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
MIAMI P.D. VEIGA,SANDRA,F	09/12/2007	14:49:07	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626102I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

OFFICER PEDRO BELTRAN WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	09/12/2007	18:04:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TIME FAST BY THREE MINUTE. MOUTH ALCOHOL SOLUTION WEAK

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. QWW,QW,Q	12/03/2007	16:11:41	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. Q,Q,Q	12/03/2007	16:12:07	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. TECH,SOURCE,	12/03/2007	16:13:30	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
MIAMI PD VEIGA,SANDRA,F	12/13/2007	14:39:20	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	626102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.05 g/210L Test #1 (Retry)	0.049
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.05 g/210L Test #3 (Retry)	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.05 g/210L Test #4 (Retry)	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.05 g/210L Test #5 (Retry)	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.05 g/210L Test #6 (Retry)	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source
			0.05 g/210L Test #7 (Retry)	0.049
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.081
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.048
			0.05 g/210L Test #8 (Retry)	0.049
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.05 g/210L Test #9 (Retry)	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	
			0.05 g/210L Test #10 (Retry)	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0009
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

05: RFI Detect CELL PHONE RANG OUTSIDE THE DOOR CAUSED RFI.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001179

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	12/25/2007	15:44:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
BOCA RATON PD HACKNEY,DWITE,N	01/25/2007	14:41:15	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	01/31/2007	14:56:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

HOSE NOT CONN TO PORT 1ST SAMPLE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	02/12/2007	09:52:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: 1ST .20 SAMPLE OUT OF TOL, TEST RE-RUN .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	03/21/2007	09:59:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.188
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

20: 1ST TWO SAMPLE OUT OF TOL.TEST RE-RUN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	04/06/2007	10:03:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.190
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.190
		0.20 g/210L Test #3 (Retry)	0.191
		Interferent Detect Test #3	INT

Remarks:

20: 2ND .20 OUT OF TOL, TEST RE-RUN, Control Outside Tolera. Non-compliance: .02 OUT OF TOL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	04/06/2007	11:26:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	05/15/2007	10:27:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	06/14/2007	14:13:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	07/16/2007	11:07:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	08/20/2007	10:54:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	09/20/2007	15:09:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	10/23/2007	12:15:33	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	11/26/2007	13:27:12	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001183

Agency Operator	Date	Time	Software	Source	
BOCA RATON PD BROCK,JOHN,R	12/19/2007	10:03:42	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	733202I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	11/30/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	01/31/2007	12:10:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	02/26/2007	13:43:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	03/28/2007	21:37:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	04/24/2007	13:36:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	05/31/2007	14:16:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	06/29/2007	13:46:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	07/29/2007	22:47:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	714101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	08/30/2007	15:10:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/30/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HACKNEY,DWITE,N	09/11/2007	12:32:11	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	10/03/2008	**Expiration**	10/03/2008

# Sims	5	**Avg Std Dev**	0.0005	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.082
			0.20 g/210L Test #7	0.201
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.081
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.081
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0391
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: Control Outside Tolerance DRY GAS TUBE NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	09/28/2007	15:14:01	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	10/29/2007	20:25:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	11/29/2007	10:38:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001189

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	12/27/2007	21:19:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
BOYNTON BEACH PD HACKNEY,DWITE,N	01/22/2007	11:46:20	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.205
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.204
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.204
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.204
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	01/24/2007	12:40:08	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	510902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR JANUARY 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	02/06/2007	12:23:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	510902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/21/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR FEBRUARY 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	03/09/2007	11:33:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR MARCH 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	04/06/2007	11:50:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR APRIL 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	05/11/2007	10:20:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.035
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.034
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.035
		0.08 g/210L Test #3 (Retry)	0.075
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	05/11/2007	11:51:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR MAY2007.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	06/18/2007	11:06:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.042
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	07/20/2007	13:26:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION JULY 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	08/10/2007	13:17:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION AUGUST 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	09/14/2007	11:02:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR SEPTEMBER 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	10/19/2007	10:35:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR OCTOBER 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	11/20/2007	11:09:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR NOVEMBER 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001190

Agency Operator	Date	Time	Software	Source	
BOYNTON BEACH PD LANE,MICHAEL,A	12/21/2007	11:29:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION FOR DECEMBER 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	01/12/2007	07:46:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	02/26/2007	07:40:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

AF/MA:INADEQUATE SAMPLE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	03/23/2007	06:42:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	04/13/2007	06:50:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	05/24/2007	13:26:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	06/28/2007	09:12:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	07/27/2007	09:38:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	08/20/2007	08:49:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
PANAMA CITY BEACH PD GEDDINGS,MARGARET,M	09/10/2007	12:27:33	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	09/24/2007	09:40:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	10/11/2007	13:43:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:CHANGED SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	11/26/2007	07:32:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001192

Agency Operator	Date	Time	Software	Source	
PANAMA CITY BEACH PD GAILFOIL,MICHAEL,L	12/28/2007	06:47:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	01/01/2007	09:05:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	504601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC MOVED FROM B TO P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	02/01/2007	18:10:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC AS P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	03/14/2007	09:02:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	04/02/2007	16:52:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	05/01/2007	08:53:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

IJC P OBSERVED BY CPL PEABODY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	06/04/2007	17:07:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

IJC B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	07/02/2007	17:06:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC OBSERVED BY CPL. NEIDERHISER 1287.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	08/01/2007	09:18:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC OBSERVED BY CPL PEABODY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	09/05/2007	09:10:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. FORGOT TO REATTACH GAS LINE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source
COLLIER COUNTY SO VENTURI,GEORGE,L	09/13/2007	12:59:04	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	10/01/2007	11:49:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

TESTED AT SPEC OPS OBSERVED BY CPL. NEIDERHISER

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001201

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	11/05/2007	09:40:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.116
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.123
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance ADJUSTED TUBING..TESTED SOG

COBRA ©1997-2006 CMI, Inc.