Florida Department of Law Enforcement

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	01/30/2007	14:22:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	02/27/2007	11:36:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	03/30/2007	15:56:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	04/28/2007	22:12:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200606F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	05/26/2007	00:07:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	06/26/2007	10:17:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605F	**0.20 Lot#**	200612D	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	05/10/2008	**Expiration**	12/14/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	07/30/2007	09:35:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	08/27/2007	12:11:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	09/29/2007	10:57:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source
NASSAU COUNTY SO MALHIOT,MATTHEW,E	10/22/2007	12:31:25	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0010
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	10/31/2007	18:58:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	630702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	11/29/2007	07:16:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001281

Agency Operator	Date	Time	Software	Source	
NASSAU COUNTY SO NYE,TERRY,L	12/17/2007	09:39:59	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	630702I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/06/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	01/23/2007	22:08:29	8100.27	Agency	Inspection

0.05 Lot#	200606B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.000
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	01/23/2007	23:12:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.000
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.076
		Interferent Detect Test #3	INT

Remarks:

LEAKING HOSE, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	01/24/2007	00:06:21	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.076
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1 (Retry)	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.076
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.05 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test: OK	No	Interferent Detect Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3 (Retry)	INT

Remarks:

SIM LEAK. DIDNOT SWITCH SIMS, Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	01/24/2007	00:35:23	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	INT
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

WRONG SIMMULATOR, Interferent Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	01/26/2007	21:31:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	02/20/2007	08:22:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	03/25/2007	00:25:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	04/20/2007	16:41:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	05/10/2007	17:53:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source
U OF FLORIDA PD MALHIOT,MATTHEW,E	06/13/2007	11:24:26	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD MALLO,NICOLE,L	06/23/2007	17:07:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD MALLO,NICOLE,L	07/26/2007	11:31:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD MALLO,NICOLE,L	08/18/2007	21:07:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/15/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	09/26/2007	16:42:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD SASSER,WILLIAM,C	10/29/2007	17:07:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD MALLO,NICOLE,L	11/30/2007	01:56:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001282

Agency Operator	Date	Time	Software	Source	
U OF FLORIDA PD MALLO,NICOLE,L	12/23/2007	09:21:42	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD JONES,STEVEN,E	01/06/2007	21:29:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	02/28/2007	11:46:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	505901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	03/30/2007	10:32:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	04/12/2007	11:27:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	05/14/2007	15:46:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	06/13/2007	11:52:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	07/19/2007	11:36:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	08/28/2007	12:04:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.CELL PHONE TOO CLOSE TO INSTRUMENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	09/25/2007	11:54:38	8100.27	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	09/25/2007	14:10:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	10/22/2007	15:12:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source
HIGH SPRINGS PD MALHIOT,MATTHEW,E	10/31/2007	10:56:11	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	11/30/2007	09:47:37	8100.27	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	INT
Interferent Detect Test: Interferent Detect	Yes		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

.05 SOL OUT OF DATE. BEGINNING AGAIN WITH CORRECT SOL..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	11/30/2007	12:06:59	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: OUTSIDE TOLLERENCE, DID NOT GIVE ENOUGH TIME TO WARM UP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001283

Agency Operator	Date	Time	Software	Source	
HIGH SPRINGS PD FULWOOD,GORDON,H	12/28/2007	18:42:49	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	01/17/2007	09:33:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	02/23/2007	08:57:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	03/05/2007	12:49:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	04/22/2007	12:39:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.187
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.187
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.187
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	04/22/2007	13:15:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	05/20/2007	12:30:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	61501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	06/15/2007	09:10:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test #1 (Retry)	0.075
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.045
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.045
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
		0.20 g/210L Test #2	0.191
		0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.05 g/210L Test #3 (Retry)	0.045
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.075
		0.20 g/210L Test #3	0.190
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	07/13/2007	07:22:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	08/15/2007	10:29:22	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	08/15/2007	11:18:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	09/12/2007	09:26:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	10/24/2007	08:20:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source
GILCHRIST COUNTY SO MALHIOT,MATTHEW,E	10/24/2007	11:15:43	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	11/16/2007	07:44:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001284

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	12/10/2007	10:32:12	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	01/17/2007	10:47:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	02/23/2007	09:36:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	03/05/2007	13:55:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	04/22/2007	13:59:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	05/20/2007	13:43:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.048
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.048
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.048
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	06/15/2007	09:57:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	07/13/2007	08:02:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	08/15/2007	11:25:28	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	08/15/2007	12:01:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	09/12/2007	10:26:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	10/24/2007	09:42:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source
GILCHRIST COUNTY SO MALHIOT,MATTHEW,E	10/24/2007	11:28:25	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	11/16/2007	09:20:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001285

Agency Operator	Date	Time	Software	Source	
GILCHRIST COUNTY SO ACKER,LAWRENCE,J	12/10/2007	11:14:39	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	01/31/2007	22:52:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.RESIDUE IN SIM. CLEANED AND SECOND PASSED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	02/28/2007	22:13:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO MARKER,JAMES,E	03/31/2007	09:02:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	04/30/2007	21:10:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO MARKER,JAMES,E	05/31/2007	02:42:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	07/06/2007	10:05:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	09/20/2007	15:28:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source
BAKER COUNTY SO MALHIOT,MATTHEW,E	10/23/2007	13:41:51	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.045
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.046
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.046
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.046
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.046

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.045
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	10/31/2007	14:49:07	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	11/27/2007	09:29:46	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001286

Agency Operator	Date	Time	Software	Source	
BAKER COUNTY SO FAIRCLOTH,ADAM,W	12/31/2007	13:54:21	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.067
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.071
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.048
0.08 g/210L Test #3	0.073
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001287

Agency Operator	Date	Time	Software	Source
BAKER COUNTY SO MALHIOT,MATTHEW,E	10/23/2007	14:01:29	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001287

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	01/31/2007	14:51:36	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	03/01/2007	00:13:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	03/30/2007	16:23:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	06/04/2007	10:45:24	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.05 OUTSIDE TOLERANCE.RETEST W/ALT SIMS .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source
UNION COUNTY SO MALHIOT,MATTHEW,E	10/03/2007	13:14:08	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	INT
0.08 g/210L Dry Gas Std Test #1 (Retry)	INT
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049
0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.197
Interferent Detect Test #8	INT
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.196
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007

Remarks:

08: Interferent Detect PRINTER TOO CLOSE, Interferent Detec. UNKNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source
UNION COUNTY SO MALHIOT,MATTHEW,E	12/04/2007	12:51:58	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	12/18/2007	10:32:56	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	08/01/2009	**Expiration**	12/14/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:OUTSIDE TOL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001288

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	12/18/2007	11:07:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	01/31/2007	14:46:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	512501I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source
UNION COUNTY SO MALHIOT,MATTHEW,E	06/14/2007	15:11:45	8100.27	Department Inspection

0.05 Lot#	2O0605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO MALHIOT,MATTHEW,E	06/14/2007	15:44:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: TOO MUCH M A SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	07/23/2007	14:18:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	08/30/2007	22:09:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO FAIRCLOTH,ADAM,W	09/28/2007	12:04:17	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source	
UNION COUNTY SO MALHIOT,MATTHEW,E	10/03/2007	12:28:50	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source
UNION COUNTY SO MALHIOT,MATTHEW,E	12/04/2007	12:35:15	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001289

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	01/29/2007	10:08:52	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.210
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.216
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.219
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	01/29/2007	16:52:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	02/25/2007	15:45:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	03/28/2007	09:27:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	04/27/2007	17:01:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2006	**Expiration**	05/10/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	05/30/2007	10:48:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	06/27/2007	09:44:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	07/27/2007	10:44:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.187
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.189
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.189
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	07/31/2007	12:52:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	08/29/2007	11:12:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	710701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.072
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	09/22/2007	16:42:04	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source
STARKE PD MALHIOT,MATTHEW,E	10/23/2007	10:32:20	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	10/31/2007	11:36:46	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	11/28/2007	10:40:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

GOT CALLED TO TELEPHONE CALL AND MISSED RESULTS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001290

Agency Operator	Date	Time	Software	Source	
STARKE PD BROWN,WILLIAM,H	12/31/2007	10:06:19	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	01/18/2007	09:33:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	02/15/2007	09:51:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	03/24/2007	09:31:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	04/17/2007	09:51:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512501I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	05/23/2007	08:57:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	06/25/2007	21:54:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	07/24/2007	10:21:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	08/24/2007	09:59:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	09/21/2007	20:55:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	10/16/2007	11:22:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source
BRADFORD COUNTY SO MALHIOT,MATTHEW,E	10/23/2007	10:44:29	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	11/13/2007	10:18:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001291

Agency Operator	Date	Time	Software	Source	
BRADFORD COUNTY SO BENNETT JR,LOUIS,J	12/06/2007	09:59:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	03/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001292

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	01/31/2007	15:31:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001292

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	02/27/2007	12:10:48	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001292

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	03/28/2007	10:33:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08:DRY GAS NOT ATTACHED TO CAL INLET

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001292

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	04/30/2007	15:43:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD STOKER,BRYON,M	01/26/2007	23:06:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/05/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	02/16/2007	00:59:44	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	02/16/2007	02:01:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: . 08: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	03/18/2007	16:49:41	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	03/18/2007	17:13:58	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	03/26/2007	09:48:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	04/22/2007	09:58:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	05/14/2007	17:02:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	06/26/2007	08:21:58	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:MOUTH ALCOHOL READ/ NEED SLOPE NOT MET. Non-compliance:RESIDULE MOUTH ALCOHOL FAILURE, DEPT INSP NO.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	06/26/2007	10:29:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	07/16/2007	03:58:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail. 05: Ambient Fail, ROOM TEMP,MOVED VENTS AWAY.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	08/21/2007	02:48:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	09/19/2007	04:11:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	10/20/2007	09:23:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Ambient Fail,CHANGED AIR VENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source
WILLISTON PD MALHIOT,MATTHEW,E	11/20/2007	10:50:51	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	11/22/2007	09:08:22	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001293

Agency Operator	Date	Time	Software	Source	
WILLISTON PD MORRIS,FRED,A	12/16/2007	10:12:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	609301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	01/06/2007	01:24:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	02/03/2007	02:22:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	03/10/2007	02:29:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	04/07/2007	02:19:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKEHOFF,CRAIG,O	05/05/2007	03:10:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	06/02/2007	01:49:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	07/07/2007	00:21:35	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	08/11/2007	03:45:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	09/15/2007	02:10:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	606002I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source
SANTA ROSA COUNTY SO GEDDINGS,MARGARET,M	10/17/2007	13:52:23	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	10/22/2007	14:22:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	606002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	11/03/2007	01:49:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	606002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001302

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	12/08/2007	14:30:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	606002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source
SANTA ROSA SO GEDDINGS,MARGARET,M	08/20/2007	15:08:52	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source	
SANTA ROSA SO BRINKERHOFF,CRAIG,O	09/15/2007	00:33:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.210
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.213
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.085
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.210
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.213
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.085
		0.20 g/210L Test #3	0.215
		0.20 g/210L Test #3 (Retry)	0.213
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance FUSE ON WET BATH SIMULATOR NOT TIGHT. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source	
SANTA ROSA SO BRINKERHOFF,CRAIG,O	09/15/2007	01:07:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.209
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.210
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.085
		0.20 g/210L Test #3	0.207
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source	
SANTA ROSA SO BRINKERHOFF,CRAIG,O	10/22/2007	12:00:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.207
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source	
SANTA ROSA SO BRINKERHOFF,CRAIG,O	11/02/2007	23:37:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.051
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.084
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.207
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.057
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.084
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.057
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001303

Agency Operator	Date	Time	Software	Source	
SANTA ROSA COUNTY SO BRINKERHOFF,CRAIG,O	12/08/2007	12:12:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	01/25/2007	13:22:23	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	02/27/2007	02:21:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	03/23/2007	03:36:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	04/28/2007	07:32:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	05/26/2007	06:46:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range ExceededSequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	06/24/2007	18:30:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.000
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	06/27/2007	19:02:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	07/23/2007	00:38:38	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	08/25/2007	08:05:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	09/29/2007	06:54:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.083
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	10/19/2007	00:46:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source
MARIANNA PD GEDDINGS,MARGARET,M	10/24/2007	11:20:11	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0010
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	11/23/2007	09:07:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001305

Agency Operator	Date	Time	Software	Source	
MARIANNA PD MEARS,MICHAEL,C	12/22/2007	11:15:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.209
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001306

Agency Operator	Date	Time	Software	Source	
SNEADS PD GEDDINGS,MARGARET,M	04/30/2007	09:32:44	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001306

Agency Operator	Date	Time	Software	Source
OKALOOSA SO GEDDINGS,MARGARET,M	06/04/2007	11:08:40	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	525202I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.038
0.08 g/210L Test #1 (Retry)	0.078
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	0.000
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.039
0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	0.000
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.039
0.08 g/210L Test #3 (Retry)	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.040
0.08 g/210L Test #4 (Retry)	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.039
0.08 g/210L Test #5 (Retry)	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.040
0.08 g/210L Test #6 (Retry)	0.079

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001306

Agency Operator	Date	Time	Software	Source
			0.20 g/210L Test #6	0.198
			0.08 g/210L Dry Gas Std Test #6	0.077
			Interferent Detect Test #6	INT
			Interferent Detect Test #6 (Retry)	INT
			Alcohol Free Test #7	0.000
			0.05 g/210L Test #7	0.049
			0.08 g/210L Test #7	0.039
			0.08 g/210L Test #7 (Retry)	0.078
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.077
			Interferent Detect Test #7	INT
			Interferent Detect Test #7 (Retry)	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.040
			0.08 g/210L Test #8 (Retry)	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Interferent Detect Test #8 (Retry)	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.018
			0.08 g/210L Test #9 (Retry)	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.000
			0.08 g/210L Test #10 (Retry)	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0134
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Int Det:TIGHTENED LID ON SIMULATOR. 08: Control Outside Tolerance UNKNOWN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001306

Agency Operator	Date	Time	Software	Source	
OKALOOSA SO HULL,R,S	11/07/2007	06:32:54	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

'

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001306

Agency Operator	Date	Time	Software	Source	
OKALOOSA SO HULL,R,S	12/19/2007	18:35:09	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611801G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	05/01/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

FORGOT TO RECONNECT RETURN HOSE TO SIM AFTER ADDING ACETONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	01/11/2007	10:51:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.087
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.082
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.204
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.052
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.082
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.08 g/210L Test #3 (Retry)	0.082
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.BLEW OFF HEAD PRESSURE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	02/19/2007	13:38:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515902I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	03/06/2007	12:36:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.085
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.085
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,N	04/05/2007	10:51:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	05/09/2007	10:53:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect. 00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	06/12/2007	11:55:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	07/12/2007	15:07:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.RECIRCULATED ROOM AIR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	08/15/2007	14:55:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	09/13/2007	10:31:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	10/04/2007	11:12:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	11/13/2007	11:28:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

NONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source
MADISON COUNTY SO GEDDINGS,MARGARET,M	12/05/2007	12:44:07	8100.27	Department Inspection

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RFI CELL PHONE ONNon-compliance:RFI CELL PHONE ON.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source
MADISON COUNTY SO GEDDINGS,MARGARET,M	12/05/2007	14:16:24	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001307

Agency Operator	Date	Time	Software	Source	
MADISON COUNTY SO STRICKLAND,JAMES,M	12/10/2007	12:11:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.RECIRCULATED ROOM AIR..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	01/26/2007	15:56:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	02/25/2007	19:22:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	03/28/2007	19:29:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	04/25/2007	19:19:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2006	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	05/27/2007	17:48:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	06/28/2007	18:19:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	07/28/2007	20:16:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	08/25/2007	13:18:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source
BAY COUNTY SO GEDDINGS,MARGARET,M	09/10/2007	11:00:48	8100.27	Department Inspection

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededTOO MUCH ALCOHOL IN MOUTH RINSED . Non-compliance:HAD TOO MUCH MOUTH ALCOHOL IN MOUTH.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source
BAY COUNTY SO GEDDINGS,MARGARET,M	09/10/2007	13:42:31	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	09/23/2007	12:02:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	10/23/2007	11:17:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	11/27/2007	22:49:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001308

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	12/30/2007	14:26:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	01/26/2007	16:10:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	02/25/2007	20:20:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2006	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.075
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	03/28/2007	18:23:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	04/25/2007	20:17:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	05/27/2007	18:45:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	06/28/2007	20:14:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

PLACED IN STORAGE UNTIL JULY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	07/28/2007	18:29:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	08/25/2007	12:08:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source
BAY COUNTY SO GEDDINGS,MARGARET,M	09/10/2007	13:52:58	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.083
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.083
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.083
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.083
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.083
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.083
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	09/23/2007	12:17:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	10/23/2007	11:58:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	11/27/2007	23:34:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.084
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001309

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	12/30/2007	14:27:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.190
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	01/08/2007	09:24:00	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	05/06/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	02/05/2007	09:13:00	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512402I
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	05/06/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	03/05/2007	13:55:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512402I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	05/06/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	04/09/2007	09:28:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/06/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	05/07/2007	19:10:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:NOT DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	06/03/2007	06:45:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	06/28/2007	09:48:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	06/28/2007	12:07:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	07/27/2007	04:15:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	09/01/2007	07:07:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	620901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source
NICEVILLE PD GEDDINGS,MARGARET,M	11/08/2007	10:31:50	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	11/09/2007	11:25:43	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	620901I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001310

Agency Operator	Date	Time	Software	Source	
NICEVILLE PD CRUTHIRDS,JEREMY,D	12/07/2007	06:00:45	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	620901I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	08/02/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	01/10/2007	10:15:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	026101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	02/08/2007	11:56:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	03/08/2007	10:50:07	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:WEAK MOUTH ALCOHOL SOL.Range Exceeded. Non-compliance:WEAK MOUTH ALCOHOL SOL..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	03/08/2007	11:24:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2006	**Expiration**	05/04/2006	**Expiration**	05/10/2006	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	04/07/2007	09:58:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626610I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	05/11/2007	11:51:35	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

NOT READING ACETONE SUSPECT ACETONE HAS GON BAD . Non-compliance:SUSPECTED BAD BOTTLE OF ACETONE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	05/11/2007	13:14:55	8100.27	Agency	Inspection

0.05 Lot#	200621C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.LOOSE CONNECTION ON SIMULATOR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	06/15/2007	09:05:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source
WALTON COUNTY SO GEDDINGS,MARGARET,M	07/25/2007	17:38:37	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.206
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.208
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.209
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.209
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.207
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.208
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.208
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.207
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.208
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.208
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	07/27/2007	09:12:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	08/26/2007	10:26:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	09/18/2007	13:28:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed. A F / M A:WEAK SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	10/09/2007	13:31:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	11/14/2007	11:42:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001311

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	12/20/2007	10:06:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	01/10/2007	16:32:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	02/08/2007	15:11:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	02/09/2007	17:00:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.178
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

LID ON SIM JAR LOOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	03/08/2007	13:18:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2006	**Expiration**	05/04/2006	**Expiration**	05/10/2006	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	04/07/2007	14:08:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	05/03/2007	13:01:07	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	06/15/2007	08:54:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source
WALTON COUNTY SO GEDDINGS,MARGARET,M	07/25/2007	14:12:38	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.199
			0.08 g/210L Dry Gas Std Test #7	0.079
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.050
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0381
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: Control Outside ToleranceCONNECTED HOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	07/27/2007	09:01:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO HOWELL,TREVOR,J	08/26/2007	10:21:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	09/18/2007	12:46:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	10/09/2007	12:52:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	11/14/2007	11:02:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001312

Agency Operator	Date	Time	Software	Source	
WALTON COUNTY SO PRESTON,JOSEPH,D	12/20/2007	09:23:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	01/31/2007	22:56:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007		

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.077
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.196
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	03/11/2007	18:39:37	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:NOT DETECTING THE ACETONE AFTER FIRST DETECTION . Non-compliance: ACETONE NOT BEING DETECTED DURING TEST .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	03/18/2007	18:43:45	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:INTERMITANT DETECTION OF INTERFERENT . Non-compliance:INTERMITANT DETECTION OF INTEREFERENT .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD TECH,SOURCE,M	04/17/2007	13:04:24	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD TECH,SOURCE,D	04/23/2007	13:45:47	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD TECH,SOURCE,D	04/23/2007	13:47:31	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD TECH,SOURCE,J	04/23/2007	14:01:58	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD TECH,SOURCE,S	04/23/2007	16:10:27	8100.27	Agency	Inspection

0.05 Lot#	200605C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/03/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD TECH,SOURCE,S	04/23/2007	16:43:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/03/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source
SPRINGFIELD PD GEDDINGS,MARGARET,M	05/30/2007	09:28:32	8100.27	Department Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	525202I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.000
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.000
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.000
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.000
0.08 g/210L Test #4 (Retry)	0.078
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.000
0.08 g/210L Test #5 (Retry)	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.000
0.08 g/210L Test #6 (Retry)	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.000
			0.08 g/210L Test #7 (Retry)	0.079
			0.20 g/210L Test #7	0.196
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.000
			0.08 g/210L Test #8 (Retry)	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.074
			0.08 g/210L Test #9 (Retry)	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.078
			0.08 g/210L Test #10 (Retry)	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Test (Std Dev)	0.0320
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Time-Date changed. 08: Control Outside Tolerance TIGHTENED LID ON SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	06/07/2007	21:17:53	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: WEAK ACETONE CHANGED STOCK SOLUTION, . Non-compliance: WEAK ACETONE STOCK HAVE TO REPLACE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	06/08/2007	00:18:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.180
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.174
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.175
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	08/09/2007	08:55:29	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.050
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.050
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:BAD SEAL,USED WRONG SIMULATOR. Non-compliance:USED WRONG SIMULATOR WITH AN ALCOHOL SOLUTIO.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	08/09/2007	12:32:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008		

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.180
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.174
0.20 g/210L Test: OK	No	0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.174
		0.20 g/210L Test #3 (Retry)	0.191
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:BAD ACETONE SOLUTION. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.02 ALCOHOL TEST NOT IN COMPLIANCE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD GRIFFIN,DAVID,S	08/09/2007	16:14:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/15/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	INT
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Interferent Detect HAD RISIDUAL ACETONE ON SIMULATOR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD RIVERA,NICK,	09/20/2007	14:21:56	8100.27	Agency	Inspection

0.05 Lot# 200612C
Expiration 12/14/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		Interferent Detect Test #1	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1 (Retry)	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.043
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	Interferent Detect Test #3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test #3 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:INT NOT DET.WEAK SOLUTIONNon-compliance:WEAK SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source	
SPRINGFIELD PD RIVERA,NICK,	09/20/2007	16:48:01	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:INST MALF. Non-compliance: DOES NOT DETECT INT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source
SPRINGFIELD PD GEDDINGS,MARGARET,M	11/27/2007	10:18:47	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	719402I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001313

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

00: Ambient FailOPEN BOTTLES IN TRASH CAN, WILL REMOVE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERREL,PAUL,E	01/26/2007	14:52:14	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.046
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.046
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

WOULD NOT READ SOLUTION, Control Outside Tolerance. AIR LEAK IN RETURN VALVE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	01/26/2007	15:26:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	02/26/2007	20:38:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	03/28/2007	10:28:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	04/16/2007	10:43:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	05/22/2007	22:25:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AIR LEAK AROUND THE O RING.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	06/14/2007	01:04:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	07/22/2007	11:49:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	08/09/2007	16:48:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	09/24/2007	09:37:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source
OKEECHOBEE COUNTY SO HACKNEY,DWITE,N	09/24/2007	12:02:28	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.192
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.192
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.192
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.192
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.192
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.192
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.192
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source
OKEECHOBEE CO SO HACKNEY,DWITE,N	11/26/2007	10:31:27	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE CO SO FERRELL,PAUL,E	11/27/2007	17:43:55	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001320

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE CO SO FERRELL,PAUL,E	12/27/2007	07:00:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	01/26/2007	15:33:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

WOULD NOT READ SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	02/26/2007	20:46:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Improper Sample.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	03/19/2007	03:28:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	04/16/2007	10:54:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515902I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	05/22/2007	22:33:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	515902I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	06/14/2007	01:14:42	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	07/22/2007	11:58:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	08/09/2007	17:01:10	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2007	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

STILL HAD SOLUTION IN MOUTH WHEN CONDUCTING THE TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	09/24/2007	09:47:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source
OKEECHOBEE COUNTY SO HACKNEY,DWITE,N	09/24/2007	12:15:32	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.191
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.190
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.192
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0016
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	10/28/2007	21:07:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	11/27/2007	18:30:49	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001321

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE COUNTY SO FERRELL,PAUL,E	12/27/2007	07:04:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS WERE CLEAN AND DRY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source
OKEECHOBEE PD HACKNEY,DWITE,N	02/14/2007	12:51:16	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.084
0.20 g/210L Test #1	0.205
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.084
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.083
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0016
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source
OKEECHOBEE PD HACKNEY,DWITE,N	06/05/2007	12:06:53	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.207
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.207
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.206
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.205
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.206
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.206
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.205
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.205
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.206
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.206
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,L	06/10/2007	07:40:02	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,L	06/10/2007	08:16:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,L	07/31/2007	02:25:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,L	08/30/2007	15:18:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,L	09/13/2007	12:13:54	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,L	11/07/2007	14:12:33	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001322

Agency Operator	Date	Time	Software	Source	
OKEECHOBEE PD WILLIAMS,VICTORIA,	12/28/2007	16:14:15	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source
PORT ST LUCIE PD HACKNEY,DWITE,N	01/02/2007	15:01:41	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,IAN,	01/03/2007	11:44:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	02/28/2007	18:03:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	03/26/2007	13:13:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	04/11/2007	11:38:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	05/31/2007	12:36:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	06/27/2007	22:17:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	07/29/2007	16:29:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	08/28/2007	16:29:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	09/30/2007	15:23:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

INSP. ERROR, FORGOT TO REHOOK UP SIM AT START OF TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	10/31/2007	04:48:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	11/23/2007	22:47:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001323

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	12/31/2007	16:50:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

05: RFI CELL PHONE RANG.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	01/11/2007	03:23:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	02/15/2007	10:16:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	03/09/2007	15:56:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	04/04/2007	11:22:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	05/15/2007	10:36:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	06/17/2007	15:28:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.178
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.185
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	07/13/2007	10:37:48	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.186
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.188
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.075
0.20 g/210L Test: OK	No	0.20 g/210L Test #2	0.187
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2 (Retry)	0.190
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.074
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.187
		0.20 g/210L Test #3 (Retry)	0.190
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	07/13/2007	11:51:45	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.042
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.046
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.047
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	07/13/2007	12:52:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source
INDIAN RIVER CO SO HACKNEY,DWITE,N	07/18/2007	11:30:48	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	08/20/2007	10:17:17	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	09/07/2007	15:21:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.072
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.078
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.071
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.071
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 08: CELL PHONE INTERFERENCE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	10/08/2007	10:16:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	11/05/2007	11:37:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714902I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	05/31/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001328

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER CO SO DIMINO,LEANETTE,M	12/06/2007	11:30:54	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	1714902I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	05/31/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	01/11/2007	03:19:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	02/15/2007	10:25:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	03/09/2007	16:01:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	04/04/2007	11:28:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	05/15/2007	10:33:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	06/17/2007	15:33:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	07/13/2007	10:50:31	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source
INDIAN RIVER SO HACKNEY,DWITE,N	07/18/2007	11:34:56	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	615301I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/05/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	08/20/2007	10:12:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	09/07/2007	15:23:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.186
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.187
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	10/08/2007	10:14:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	11/05/2007	11:43:44	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	615301I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001329

Agency Operator	Date	Time	Software	Source	
INDIAN RIVER SO DIMINO,LEANETTE,M	12/06/2007	11:36:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	615301I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	01/30/2007	09:59:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	02/23/2007	11:22:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	03/30/2007	15:21:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	04/26/2007	10:50:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	05/30/2007	09:58:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	06/19/2007	11:05:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	07/31/2007	10:02:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	08/28/2007	11:00:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source
DESOTO COUNTY SO VENTURI,GEORGE,L	08/28/2007	14:33:51	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	09/28/2007	11:23:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	10/29/2007	10:09:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	11/27/2007	11:56:45	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001340

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	12/28/2007	10:03:50	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	01/30/2007	18:38:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	02/23/2007	11:31:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	03/30/2007	15:25:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	04/26/2007	10:44:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Ambient Fail. 08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	05/30/2007	10:00:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	06/19/2007	11:01:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	7005001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	07/31/2007	10:11:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	08/28/2007	11:03:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source
DESOTO COUNTY SO VENTURI,GEORGE,L	08/28/2007	14:21:03	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source
DESOTO COUNTY SO VENTURI,GEORGE,L	11/14/2007	16:15:40	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	11/27/2007	12:32:00	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001341

Agency Operator	Date	Time	Software	Source	
DESOTO COUNTY SO HARRIS,BRIAN,R	12/28/2007	10:20:10	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	01/15/2007	18:53:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	02/19/2007	18:20:01	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	03/19/2007	11:19:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	04/16/2007	09:18:02	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	05/10/2007	11:50:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	06/13/2007	13:13:18	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	07/03/2007	09:32:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

INSPECTION CONDUCTED AT NORTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	08/13/2007	13:33:57	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT N JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	09/17/2007	10:45:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT N COUNTY JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source
SARASOTA COUNTY SO VENTURI,GEORGE,L	09/24/2007	14:15:09	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	09/25/2007	09:24:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect/PORT RADIO//SENT OUT FOR PREV MAINT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source
SARASOTA COUNTY SO VENTURI,GEORGE,L	11/14/2007	15:52:11	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.076
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.076
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.076
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001344

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	12/03/2007	11:55:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT S JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	01/27/2007	10:13:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	02/25/2007	11:50:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	03/31/2007	16:03:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	04/26/2007	06:18:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.045
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	05/21/2007	01:43:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Purge Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	06/13/2007	04:47:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	7141011
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	07/15/2007	11:58:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	7141011
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	08/12/2007	11:14:00	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail, Ambient Fail. Non-compliance:AMBIENT FAIL TWICE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	08/12/2007	12:08:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

2ND ATTEMPT AFTER CONTACTING R.SKIPPER. 1ST HAD AMB FAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source
EUSTIS P.D. SKIPPER,ROGER,G	09/17/2007	13:45:57	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	09/24/2007	04:08:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	10/26/2007	04:39:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	11/28/2007	05:15:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001345

Agency Operator	Date	Time	Software	Source	
EUSTIS P.D. DONALDSON,ADAM,J	12/30/2007	11:16:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source
SARASOTA COUNTY SO VENTURI,GEORGE,L	01/08/2007	16:22:15	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	05/04/2007	10:18:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	06/11/2007	12:27:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	07/02/2007	12:33:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.207
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INSPECTION CONDUCTED IN RV

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	08/13/2007	11:09:28	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.085
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSPECTION CONDUCTED IN BAT1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	09/14/2007	15:42:45	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE IN BAT1.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	10/15/2007	14:15:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A:USED DIFF BOTTLE ON 2ND TRY.INSP DONE IN BAT1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	11/13/2007	09:12:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE IN BAT1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001346

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	12/03/2007	15:07:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE IN BAT1

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	01/12/2007	21:43:08	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	02/16/2007	19:03:25	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	03/16/2007	20:09:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	04/16/2007	11:16:29	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	05/04/2007	11:21:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	06/11/2007	11:44:45	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	07/02/2007	12:41:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

INSPECTION CONDUCTED IN RV

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	08/13/2007	10:16:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

INSPECTION CONDUCTED IN RV

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	09/14/2007	11:45:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT SOUTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source
SARASOTA COUNTY SO VENTURI,GEORGE,L	09/24/2007	14:25:28	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	10/16/2007	09:40:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT N JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	11/12/2007	11:48:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT S JAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001347

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	12/03/2007	10:53:21	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A:DIFFICULT TO ACTIVATE TONE AND MAINTAIN TONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	01/15/2007	17:20:04	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	02/19/2007	21:16:20	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	04/21/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	03/19/2007	09:22:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / MA:DIDN'T DETECT MOUTH ETOH,USED DIFF STOCK SOL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	04/16/2007	13:00:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	05/10/2007	09:24:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	06/11/2007	09:58:05	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	07/02/2007	10:31:11	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSPECTION CONDUCTED AT SOUTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	08/13/2007	08:36:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

INSPECTION COMPLETED AT SOUTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	09/14/2007	10:42:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT SOUTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source
SARASOTA COUNTY SO VENTURI,GEORGE,L	09/24/2007	14:50:15	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.080
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.080
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.080
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.195
			0.08 g/210L Dry Gas Std Test #7	0.079
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.080
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.080
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.080
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0252
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: Control Outside Tolerance: DGS HOSE NOT CONNECTED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	10/15/2007	08:46:38	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT SOUTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	11/13/2007	11:17:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT N JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001348

Agency Operator	Date	Time	Software	Source	
SARASOTA COUNTY SO BUTLER,CHRIS,T	12/04/2007	12:15:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT FDLE FT MYERS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	01/22/2007	11:10:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	02/20/2007	09:55:34	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	02/20/2007	10:58:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	03/14/2007	10:01:45	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	INT
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect, Interferent Detect. INSPECTIO RESTANon-compliance:INTERFERENCE DETECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	03/14/2007	11:06:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

FIRST INSPECTION ALCOHL FREE TEST 00 DETECTED INTERFERENT,

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	03/14/2007	11:17:28	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	04/16/2007	12:04:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	05/15/2007	15:52:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	06/06/2007	15:30:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

INT*.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	07/13/2007	13:56:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	08/20/2007	14:41:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source
FRUITLAND PARK P.D. SKIPPER,ROGER,G	09/05/2007	13:41:27	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	09/10/2007	11:15:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	10/02/2007	12:42:52	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	11/27/2007	09:56:18	8100.27	Agency	Inspection

0.05 Lot#	N	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001350

Agency Operator	Date	Time	Software	Source	
FRUITLAND PARK P.D. CUTTER,DENNIS,L	12/03/2007	15:48:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	01/18/2007	14:41:37	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	02/26/2007	19:05:31	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	04/18/2007	13:14:01	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516001I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	05/22/2007	12:58:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	516001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	06/27/2007	15:04:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	516001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	07/17/2007	16:25:00	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

MAYBE ACETONE IS WEAK. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES NAT PARK HARRIS,LAKEISHA,S	07/17/2007	18:45:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES N. P. VEIGA,SANDRA,F	10/04/2007	14:46:13	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

PUT INSTRUMENT IN ENABLE FOR DEPT INSPECTIONNon-compliance:INSTRUMENT BACK FROM CMI AI EXPIRED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source
EVERGLADES N. P. VEIGA,SANDRA,F	10/04/2007	16:42:36	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.081
0.08 g/210L Test #1 (Retry)	0.081
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	
0.08 g/210L Test #2 (Retry)	0.082
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3 (Retry)	0.082
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4 (Retry)	0.081
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5 (Retry)	0.082
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6 (Retry)	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7 (Retry)	0.082
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8 (Retry)	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9 (Retry)	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10 (Retry)	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

RFI DETECT DUE TO CELL PHONE. L. HARRISS PRESENT FOR INSPEC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES N. P. HARRIS,LAKEISHA,S	10/04/2007	17:27:57	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES N. P. HARRIS,LAKEISHA,S	11/30/2007	13:24:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001360

Agency Operator	Date	Time	Software	Source	
EVERGLADES N. P. HARRIS,LAKEISHA,S	12/27/2007	12:06:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	01/07/2007	10:48:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:NO M A DETECT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	02/04/2007	15:46:40	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:UNKNOWN. 00: Ambient FailUNKNOWN, Ambient Fail. Non-compliance:AMBIENT FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	02/05/2007	11:28:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	03/09/2007	12:43:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	04/05/2007	13:18:29	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail UNKNOWN, Ambient Fail. Non-compliance:AMBIENT FAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	04/05/2007	14:56:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	05/14/2007	13:15:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	06/04/2007	13:43:21	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:UNKNOWN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	07/11/2007	14:53:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	08/06/2007	15:56:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source
FL HIGHWAY PATROL HACKNEY,DWITE,N	09/06/2007	15:09:28	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Interferent Detect Test #1 (Retry)	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Interferent Detect Test #2 (Retry)	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Interferent Detect Test #3 (Retry)	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.046
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Interferent Detect Test #4 (Retry)	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Interferent Detect Test #5 (Retry)	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Interferent Detect Test #6 (Retry)	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.079
			Interferent Detect Test #7	INT
			Interferent Detect Test #7 (Retry)	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.046
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Interferent Detect Test #8 (Retry)	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Interferent Detect Test #9 (Retry)	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	0.000
			Interferent Detect Test #10 (Retry)	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Int Det:AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	09/06/2007	16:12:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	10/09/2007	12:33:59	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	11/02/2007	13:56:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001361

Agency Operator	Date	Time	Software	Source	
FL HIGHWAY PATROL DONKER,MICHAEL,R	12/02/2007	11:21:30	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	710701I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	01/09/2007	20:59:54	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SAMPLE GIVEN AT WRONG TIME.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	02/10/2007	11:47:42	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

INSUFFICENT ACETONE AMOUNT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	02/11/2007	19:47:24	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

RETEST OF FEBRUARY AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	03/09/2007	08:01:19	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	516002I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MARCH 2007 AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	04/21/2007	20:33:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	611501G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	05/05/2007	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: MOUTH ALCOHOL BOTTLE OPEN IN ROOM- APRIL INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	05/23/2007	12:39:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MAY 2007 AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	06/17/2007	10:14:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

JUNE AGENCY INSP

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	07/28/2007	15:25:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

JULY AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	08/11/2007	14:18:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AUGUST AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source
CHARLOTTE COUNTY SO VENTURI,GEORGE,L	08/30/2007	13:56:18	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	09/08/2007	13:05:31	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RETEST FOR RANGE NOT MET. RANGE NOT MEET.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	09/08/2007	13:23:58	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: No Sample Provided. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	09/08/2007	14:18:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION FOR SEPT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	10/07/2007	21:05:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION FOR OCTOBER

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	11/16/2007	21:49:18	8100.27	Agency	Inspection

0.05 Lot#	611501G	**0.08 Lot#**	611501G	**0.20 Lot#**	
Expiration	04/27/2008	**Expiration**	04/27/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.048
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.079
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #3	0.079
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

USED WRON LOT AND EXP INFOUSED WRONG LOT AND EXP INFO.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	11/16/2007	22:39:23	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611501G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION FOR NOV 2007

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001363

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	12/20/2007	17:35:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	611501G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION MONTH OF DECEMBER

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	01/26/2007	07:06:04	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	627201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.178
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.187
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	02/26/2007	02:00:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.083
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	03/26/2007	04:03:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	\
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

PRINTER SPOOLED AT SAME TIME

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	03/26/2007	05:56:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

00: Interferent Detect.ATTACHED WRONG SIMULATOR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	04/20/2007	06:43:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.TURNED RADIO ON . Non-compliance:2ND FAILURE ON 0.20. KICKED OUT OF AI PROC. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	05/24/2007	04:56:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source
GAINESVILLE PD MALHIOT,MATTHEW,E	06/04/2007	09:47:30	8100.27	Department Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK	Yes	0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2 (Retry)	0.049
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #2	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #3 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #3 (Retry)	0.049
0.05 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Test: OK	No	Alcohol Free Test #4 (Retry)	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #4 (Retry)	0.049
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #4	INT
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5 (Retry)	0.049
		Interferent Detect Test #5	INT
		Alcohol Free Test #6 (Retry)	0.000
		Interferent Detect Test #6	INT
		Alcohol Free Test #7 (Retry)	0.000
		Interferent Detect Test #7	INT
		Alcohol Free Test #8 (Retry)	0.000
		Interferent Detect Test #8	INT
		Alcohol Free Test #9 (Retry)	0.000
		Interferent Detect Test #9	INT
		Alcohol Free Test #10 (Retry)	0.000
		Interferent Detect Test #10	INT

Remarks:

00: Ambient Fail . 05: RFI Detect CELL PHONE IN AREA, RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source
GAINESVILLE PD MALHIOT,MATTHEW,E	06/04/2007	11:13:44	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	06/27/2007	23:23:51	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	06/28/2007	00:12:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	07/18/2007	05:14:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	08/22/2007	05:56:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	09/25/2007	08:07:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627201I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.049
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.049
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.049
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance/ATTACHED WRONG SIM..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	10/30/2007	07:22:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	724801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	11/28/2007	22:06:24	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.121
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.039
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.108
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2 (Retry)	0.043
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test #3	0.104
0.08 g/210L Test: OK	No	0.05 g/210L Test #3 (Retry)	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance/SIMULATOR HEATER PROBLEM. . Non-compliance:SIMULATOR G-6396 PROBLEM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	11/28/2007	23:05:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	724801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/06/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001364

Agency Operator	Date	Time	Software	Source	
GAINESVILLE PD RAULERSON,JOSEPH,H	12/27/2007	15:34:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	724801G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/06/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.184
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance/RE-SEALED JAR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	01/31/2007	15:39:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	02/27/2007	12:01:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	03/28/2007	10:22:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	04/30/2007	15:31:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.178
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.187
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.20 g/210L Test: OK	No	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.194
		Interferent Detect Test #3	INT

Remarks:

CHANGED HOSE ON SIMULATOR,SIM OK WOULD NOT GO INTO TOLERANCWOULD NOT PROVIDE 3 .20 SAMPLES.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	04/30/2007	16:39:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	516002I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	05/03/2007	15:44:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	515901I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/13/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

DEMOSTRATION FOR SAO

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	06/12/2007	10:44:26	8100.27	Agency	Inspection

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. NO ROOM CIRCULATION AND SIMULATORS PRESENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	06/12/2007	11:43:56	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	07/25/2007	15:56:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	08/31/2007	11:39:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	09/30/2007	12:54:29	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	10/31/2007	15:55:18	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source
LEVY COUNTY SO MALHIOT,MATTHEW,E	11/20/2007	13:26:22	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	11/27/2007	20:59:07	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001365

Agency Operator	Date	Time	Software	Source	
LEVY COUNTY SO POLIMENI,NEIL,A	12/31/2007	08:07:46	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	01/26/2007	20:47:55	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	611801G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/01/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	02/22/2007	23:30:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	03/11/2007	20:43:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	04/15/2007	20:42:31	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	04/15/2007	21:39:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	04/16/2007	12:15:48	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	05/13/2007	21:35:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	06/13/2007	20:44:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: SIM ACCIDENTALLY TURNED OFF/TURNED BACK ON/INSP RESTART. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/19/2007	21:29:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	08/10/2007	02:04:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	09/16/2007	23:09:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	10/21/2007	02:21:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	11/18/2007	00:41:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001367

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	12/16/2007	03:02:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	01/30/2007	09:18:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	02/28/2007	09:44:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	03/27/2007	09:45:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	04/26/2007	09:29:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	05/31/2007	09:24:08	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,R,D	06/26/2007	09:53:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	07/31/2007	09:25:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	08/21/2007	09:24:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	09/25/2007	10:35:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.075
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.190
0.20 g/210L Test #2 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.191
0.20 g/210L Test #3 (Retry)	0.192
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail. CHANGED HOSE CONN.. CHANGE INSTR.CONN. O-RING.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	10/30/2007	09:23:11	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source
MAYPORT NAVAL BASE MALHIOT,MATTHEW,E	11/28/2007	11:59:47	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	11/29/2007	07:37:18	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

REPLACE O-RING ON SIM.PORT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001368

Agency Operator	Date	Time	Software	Source	
MAYPORT NAVAL BASE THOMASON,ROBERT,D	12/27/2007	09:39:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	706101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	03/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	01/26/2007	20:01:44	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	02/22/2007	20:29:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	03/11/2007	23:04:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	04/15/2007	23:48:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source
PINELLAS COUNTY SO SUERETH,DONALD,P	04/16/2007	14:07:59	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	618801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/10/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	05/13/2007	22:13:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	06/13/2007	22:02:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	07/19/2007	22:30:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO PEACOCK,CHERYL,A	08/09/2007	22:50:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	09/16/2007	23:49:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	10/21/2007	23:18:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001415

Agency Operator	Date	Time	Software	Source	
PINELLAS COUNTY SO DEANE,M,	11/18/2007	20:25:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside ToleranceUNKNOWN CAUSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	01/17/2007	13:01:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	02/14/2007	12:38:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/21/2007	13:35:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	04/17/2007	11:01:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY SO SKIPPER,ROGER,G	04/18/2007	15:08:09	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	05/16/2007	13:00:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	06/13/2007	13:18:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	07/18/2007	12:58:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	08/08/2007	12:32:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	09/12/2007	13:09:11	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	10/03/2007	13:07:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	11/14/2007	12:32:10	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001416

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	12/12/2007	12:34:41	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	01/17/2007	12:16:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	02/14/2007	12:25:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:RECHECKED CONNECTIONS AFTER 1ST SAMPLE/REPEATED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/21/2007	13:11:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY S.O. SKIPPER,ROGER,G	04/05/2007	12:16:40	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	04/11/2007	12:48:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.189
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance./RECHECKED TUBE CONECTIONS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	05/16/2007	12:51:19	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	06/13/2007	13:09:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518702I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.193
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	07/18/2007	12:26:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	08/08/2007	12:20:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	09/12/2007	13:07:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	10/03/2007	13:08:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	11/14/2007	12:32:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001418

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	12/12/2007	12:35:16	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	01/17/2007	13:06:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	02/14/2007	12:22:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518801I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY SO MELVILLE,KELLY,M	03/21/2007	13:33:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	512402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	05/06/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source
ORANGE COUNTY S.O. SKIPPER,ROGER,G	04/05/2007	11:42:44	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0009 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0010
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	04/11/2007	12:50:42	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance/RECHECKED TUBE CONNECTIONS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	05/16/2007	12:37:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	06/13/2007	13:11:22	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	07/18/2007	12:39:49	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	08/08/2007	12:32:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	09/12/2007	13:21:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	10/03/2007	13:06:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	626903I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/28/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	11/14/2007	12:32:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001420

Agency Operator	Date	Time	Software	Source	
ORANGE COUNTY S.O. MELVILLE,KELLY,M	12/12/2007	12:34:53	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	01/31/2007	12:04:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/05/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	02/26/2007	16:35:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	03/28/2007	21:30:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	04/24/2007	16:13:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	05/31/2007	14:05:48	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	06/29/2007	16:03:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	07/29/2007	22:41:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	08/30/2007	18:15:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source
MARTIN COUNTY SO HACKNEY,DWITE,N	09/11/2007	11:58:10	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	09/28/2007	17:29:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	10/29/2007	20:18:34	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	11/29/2007	13:22:07	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001439

Agency Operator	Date	Time	Software	Source	
MARTIN COUNTY SO DECKARD,JOE,B	12/27/2007	21:13:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	619101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	01/30/2007	02:43:48	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.077
0.20 g/210L Test #1	0.189
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.076
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.076
0.08 g/210L Test #3 (Retry)	0.078
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	02/27/2007	02:58:30	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.044
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.072
0.08 g/210L Test #1 (Retry)	0.075
0.20 g/210L Test #1	0.185
0.20 g/210L Test #1 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.045
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.188
0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.045
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.074
0.08 g/210L Test #3 (Retry)	0.076
0.20 g/210L Test #3	0.190
0.20 g/210L Test #3 (Retry)	0.191
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.YES

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	03/07/2007	02:02:50	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #2	0.019
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Sequence Aborted. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	03/07/2007	02:54:34	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	52001I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.186
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.191
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	
Alcohol Free Test: OK	Yes	Interferent Detect Test #1 (Retry)	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.076
		0.20 g/210L Test #2	0.189
		0.20 g/210L Test #2 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.191
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	04/25/2007	02:04:39	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1 (Retry)	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	05/02/2007	02:36:35	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.045
0.05 g/210L Test #1 (Retry)	0.046
0.08 g/210L Test #1	0.072
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.182
0.20 g/210L Test #1 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.076
0.20 g/210L Test #2	0.186
0.20 g/210L Test #2 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.046
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.075
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.186
0.20 g/210L Test #3 (Retry)	0.190
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT

Remarks:

05: RFI Detect. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source
KENNEDY SPACE CENTER SKIPPER,ROGER,G	05/15/2007	10:55:17	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 4 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.052

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Ambient Fail FIRST AIRBLANK FOR .00 TEST RETEST..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	07/06/2007	16:01:19	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #3	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #3 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: Purge Fail, . SKIPPER NOTIFIEDNon-compliance: FAILED INT TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL W,W	07/06/2007	16:22:52	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2 (Retry)	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #3	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test #3 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Ambient Fail. Int Det: Purge Fail, SKIPPER NOTIFIED. Non-compliance: FAILED INT TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source
KENNEDY SPACE CENTER SKIPPER,ROGER,G	09/04/2007	12:52:34	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 05 March 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

INSPECTION AFTER REPAIR AND ANNUAL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	09/27/2007	01:44:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	20012D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test #1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect. 20: RFI Detect HAND HELD RADIO BEING USED IN HALLWAY. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	10/01/2007	02:32:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	11/13/2007	02:43:50	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001440

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER MC CARTHY,RUSSELL,W	12/21/2007	01:55:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. 08: RFI Detect RADIO USED IN HALL OUTSIDE INTOX ROOM.

COBRA ©1997-2006 CMI, Inc.