Florida Department of Law Enforcement

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	01/01/2007	23:48:52	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	02/03/2007	00:08:17	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	03/05/2007	23:40:27	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	04/03/2007	23:22:26	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	05/01/2007	00:51:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	07/26/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	06/01/2007	23:44:50	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	07/26/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	07/02/2007	23:39:45	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	07/26/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed. 05: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	08/07/2007	23:16:49	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	09/01/2007	03:02:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source
KENNEDY SPACE CENTER SKIPPER,ROGER,G	09/27/2007	10:22:15	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 4 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.076
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

00: Ambient Fail SAMPLE CHAMBER DID NOT PURGE M/A TEST. . 0.00 ALCOHOL FREE SIM TESTED REPEATED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	10/01/2007	23:38:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	71701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:LOOSE HOSE CONNECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	11/07/2007	00:30:10	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

TIME AND DATE CHANGE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001441

Agency Operator	Date	Time	Software	Source	
KENNEDY SPACE CENTER HIGHTOWER,AARON,R	12/01/2007	01:02:49	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	01/31/2007	11:27:53	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/26/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	02/02/2007	12:43:39	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	07/26/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	03/26/2007	12:42:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	04/11/2007	14:35:47	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	610001G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	05/16/2007	13:13:42	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	06/19/2007	11:28:37	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	07/18/2007	11:55:27	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK VELAZQUEZ,KIMBERLY,	08/22/2007	11:27:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK ESPINOZA,PABLO,	09/11/2007	13:39:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK VELAZQUEZ,KIMBERLY,	10/10/2007	10:23:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source
BISCAYNE NAT. PARK VEIGA,SANDRA,F	10/10/2007	17:20:50	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/12/2008

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.190
0.08 g/210L Dry Gas Std Test #2	0.083
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.191
0.08 g/210L Dry Gas Std Test #3	0.083
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.192
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.192
0.08 g/210L Dry Gas Std Test #5	0.083
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.192
0.08 g/210L Dry Gas Std Test #6	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.193
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.194
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

OFFICER K. VELAZQUEZ & PSA ESPINOZA WERE PRESENT FOR INSPEC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK VELAZQUEZ,KIMBERLY,	11/12/2007	13:04:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail/CLEARED AREA/PSA ESPINOZA PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001442

Agency Operator	Date	Time	Software	Source	
BISCAYNE NAT. PARK VELAZQUEZ,KIMBERLY,	12/11/2007	12:02:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	01/28/2007	12:40:40	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/26/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	02/20/2007	01:23:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	520201I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/26/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	03/25/2007	09:19:41	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source
CLAY COUNTY SO MALHIOT,MATTHEW,E	04/19/2007	10:21:56	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.046
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.194
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	04/22/2007	00:28:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	05/20/2007	05:17:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	06/15/2007	23:13:54	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	07/05/2007	21:12:50	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	08/06/2007	22:41:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	09/22/2007	13:22:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO MEADOR,RICKY,A	10/15/2007	21:56:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	11/07/2007	12:00:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

MONTHLY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001443

Agency Operator	Date	Time	Software	Source	
CLAY COUNTY SO HARTER,CARL,L	12/10/2007	13:40:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	710701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/18/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	01/27/2007	11:02:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	02/25/2007	12:10:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	03/25/2007	10:03:49	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Time-Date changed. Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	04/21/2007	10:09:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	05/23/2007	13:18:45	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	520101I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/22/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	06/30/2007	13:16:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance. RECHECKED CONNECTIONS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	07/20/2007	14:55:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.-RECHECKED CONNECTIONS/SEALS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	08/30/2007	14:17:02	8100.27	Agency	Inspection

 # Sims 3 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:TIGHTENED SIM JAR AFTER 1ST ATTEMPT (LOOSE), . Non-compliance:INT FAILURE_POSSIBLE POOR ACETONE SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	08/31/2007	13:14:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail-DID NOT DISCONNECT CONTAMINATED MOUTHPEICE. 20: Control Outside Tolerance:RECHECKED CONNECTIONS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source
HURLBURT FIELD PD GEDDINGS,MARGARET,M	09/06/2007	13:12:37	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	10/20/2007	10:50:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 3 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.074
0.08 g/210L Test #1 (Retry)	0.076
0.20 g/210L Test #1	0.187
0.20 g/210L Test #1 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test #2	0.195
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.08 g/210L Test #3 (Retry)	0.077
0.20 g/210L Test #3	0.196
0.20 g/210L Test #3 (Retry)	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance: RE-SEATED SIM ADAPTORS. 20: Control Outside Tolerance: RECHECKED SIM FOR AIR LEAK.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	11/18/2007	10:29:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705301I
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/27/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Ambient Fail. OPEN ARS NEAR BREATH TUBE

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001469

Agency Operator	Date	Time	Software	Source	
HURLBURT FIELD PD HEMPE,EDWARD,J	12/15/2007	08:52:01	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705301I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/27/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. GARRITY,THOMAS,A	01/31/2007	11:34:41	8100.27	Agency	Inspection

 # Sims 4 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. GARRITY,THOMAS,A	01/31/2007	12:13:39	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.047
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. GARRITY,THOMAS,A	02/02/2007	10:54:43	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. GARRITY,THOMAS,A	03/30/2007	12:08:05	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. GARRITY,THOMAS,A	04/17/2007	09:19:08	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	518802I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. LANGLEY,CHARLES,J	06/29/2007	02:50:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	518802I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.074
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.074
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. LANGLEY,CHARLES,J	06/29/2007	03:11:19	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. LANGLEY,CHARLES,J	06/29/2007	03:13:37	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. LANGLEY,CHARLES,J	06/29/2007	03:16:04	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. LANGLEY,CHARLES,J	06/29/2007	03:31:05	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK A.F.B. WAGNER,ROBERT,F	07/30/2007	18:45:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	518802I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/12/2007

 # Sims 5 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.022
0.05 g/210L Test #1 (Retry)	0.045
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.073
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.073
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.026
0.05 g/210L Test #2 (Retry)	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.073
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.073
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.029
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.073
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.072
Interferent Detect Test #3	INT

Remarks:

Time-Date changed. 05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. GAS CYLINDER EXPIRED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source
PATRICK AFB SFS SKIPPER,ROGER,G	10/02/2007	13:57:34	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.082
0.20 g/210L Test #1	0.201
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.202
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.082
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.202
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

INSPECTION AFTER REPAIR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK AFB SFS WAGNER,ROBERT,F	10/09/2007	15:19:34	8100.27	Agency	Inspection

0.05 Lot#	200605C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	706102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.082
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.083
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.084
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.084
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK AFB SFS WAGNER,ROBERT,F	11/26/2007	13:01:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	723402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/23/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK AFB SFS WAGNER,ROBERT,F	12/18/2007	08:04:00	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SIMULATOR OFF/COLD. INT DET COLD SIM.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK AFB SFS WAGNER,ROBERT,F	12/18/2007	08:31:53	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

ACETONE DEPLETED. ACETONE DEPLETED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001471

Agency Operator	Date	Time	Software	Source	
PATRICK AFB SFS WAGNER,ROBERT,F	12/18/2007	09:09:59	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	723402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/23/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: Purge Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	01/26/2007	16:33:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	02/25/2007	21:06:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	03/28/2007	19:13:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	04/25/2007	19:34:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	05/27/2007	18:02:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	06/28/2007	19:51:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	07/28/2007	19:17:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	08/25/2007	12:49:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source
BAY COUNTY SO GEDDINGS,MARGARET,M	09/10/2007	13:37:58	8100.27	Department Inspection

0.05 Lot#	200612D	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.078
			0.20 g/210L Test #7	0.196
			0.08 g/210L Dry Gas Std Test #7	0.080
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0253
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Control Outside Tolerance HOOKED UP HOSE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	09/23/2007	12:59:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	10/23/2007	12:56:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	11/28/2007	00:22:04	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001646

Agency Operator	Date	Time	Software	Source	
BAY COUNTY SO LANGFORD,WILLIAM,E	12/30/2007	15:18:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	01/24/2007	15:36:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	02/02/2007	17:42:16	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	03/02/2007	17:00:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	04/03/2007	16:54:25	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	05/03/2007	10:36:10	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	06/15/2007	16:36:03	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	07/13/2007	15:58:50	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	08/30/2007	09:07:26	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source
CAPE CORAL PD VENTURI,GEORGE,L	09/05/2007	12:53:47	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

# Sims	5	**Avg Std Dev**	0.0004	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	09/12/2007	17:27:51	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	10/05/2007	14:16:31	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	723501I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	08/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	11/20/2007	16:11:24	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	723501I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	08/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001647

Agency Operator	Date	Time	Software	Source	
CAPE CORAL PD CLAFLIN,DAVID,S	12/04/2007	17:38:54	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	723501I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	08/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	01/28/2007	21:39:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	02/28/2007	00:22:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	03/26/2007	22:06:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: MOIST. IN HOSE BETWEEN SIM. INSTRUMENT.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	04/28/2007	23:12:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	05/29/2007	09:42:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605C	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/04/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

RADIO INTERF..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD GOLDBERG,THOMAS,R	06/26/2007	10:16:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance- MOISTUREIN TUBE .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	07/27/2007	10:23:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008		

 # Sims 2 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.153
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.161
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.156
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect/CELL PHONE, Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	07/27/2007	14:22:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.203
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source
KENNETH CITY PD SUERETH,DONALD,P	08/12/2007	16:26:45	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

OFFICER VIENO PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	08/12/2007	17:18:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	09/19/2007	06:37:51	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD VIENO,MICHAEL,J	10/10/2007	06:19:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/15/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source
KENNETH CITY PD SUERETH,DONALD,P	12/18/2007	11:06:35	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0009 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.082
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.083
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.082
0.20 g/210L Test #6	0.203
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.053

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.203
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.054
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.053
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0018
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD SUERETH,DONALD,P	12/18/2007	11:08:45	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DIAGNOSTIC INTENTIONALLY FAILED TO CONDUCT SUBJ. TEST.Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001649

Agency Operator	Date	Time	Software	Source	
KENNETH CITY PD GOLDBERG,THOMAS,R	12/19/2007	06:25:09	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	701201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	01/15/2009

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	01/01/2007	08:38:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC MOVED FROM P TO B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	02/01/2007	18:14:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC AS B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	03/14/2007	09:30:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	04/02/2007	16:55:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC B

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	05/01/2007	08:57:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

IJC OBSERVED BY CPL. NEIDERHISER

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	06/04/2007	17:19:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

IJC P

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	07/02/2007	16:42:54	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC OBSERVED BY CPL PEABODY 2279

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	08/01/2007	08:50:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC CPL. NEIDERHISER OBSERVED TESTING

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	09/05/2007	10:02:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source
COLLIER COUNTY SO VENTURI,GEORGE,L	09/13/2007	13:15:28	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	10/01/2007	15:01:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC OBSERVED BY CPL PEABODY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO NEIDERHISER,JOHN,E	11/05/2007	15:01:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	622802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC OBSERVED BY CPL PEABODY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001651

Agency Operator	Date	Time	Software	Source	
COLLIER COUNTY SO PEABODY,MICHAEL,F	12/04/2007	09:07:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	622802I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

TESTED AT IJC.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	01/31/2007	16:02:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.054
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	02/28/2007	21:32:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	03/21/2007	17:00:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD CHANG,DAVID,	04/10/2007	16:11:46	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	05/09/2007	17:56:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	512601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/10/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	06/22/2007	19:17:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	07/18/2007	18:18:40	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	08/23/2007	19:13:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD PITA,JORGE,A	09/19/2007	16:48:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	10/09/2007	15:59:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source
WEST MIAMI PD VEIGA,SANDRA,F	10/15/2007	14:37:30	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0002 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.083
0.20 g/210L Test #1	0.207
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.083
0.20 g/210L Test #2	0.207
0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.208
0.08 g/210L Dry Gas Std Test #3	0.078
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.207
0.08 g/210L Dry Gas Std Test #4	0.078
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.083
0.20 g/210L Test #5	0.207
0.08 g/210L Dry Gas Std Test #5	0.078
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.083
0.20 g/210L Test #6	0.207
0.08 g/210L Dry Gas Std Test #6	0.078
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.083
			0.20 g/210L Test #7	0.208
			0.08 g/210L Dry Gas Std Test #7	0.078
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.051
			0.08 g/210L Test #8	0.083
			0.20 g/210L Test #8	0.207
			0.08 g/210L Dry Gas Std Test #8	0.077
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.207
			0.08 g/210L Dry Gas Std Test #9	0.077
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.207
			0.08 g/210L Dry Gas Std Test #10	0.078
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0000
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0245
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08:DRY/GAS HOSE WAS NOT CONNECTED

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	11/21/2007	22:31:27	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	626102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001652

Agency Operator	Date	Time	Software	Source	
WEST MIAMI PD BELTRAN,PEDRO,	12/11/2007	21:16:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	01/17/2007	22:09:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	02/15/2007	20:43:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	03/21/2007	23:17:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ALBERTSON,JASON,W	04/18/2007	23:16:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMS DRY AND CLEAN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD ARKOVICH,ROBBIE,S	05/03/2007	20:42:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	06/13/2007	21:44:48	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	07/11/2007	18:27:01	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	08/11/2007	01:13:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source
ST PETERSBURG PD SUERETH,DONALD,P	09/11/2007	15:46:11	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.203
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.052

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0014
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Ambient Fail. ROOM FANNED..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	09/26/2007	21:26:43	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	10/11/2007	12:18:03	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	11/19/2007	13:03:06	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001653

Agency Operator	Date	Time	Software	Source	
ST PETERSBURG PD WEISKOPF,MICHAEL,D	12/20/2007	21:20:20	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	635301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/21/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

IN COMPLIANCE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	01/31/2007	16:05:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	02/13/2007	21:47:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	03/21/2007	16:54:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD CHANG,DAVID,	04/10/2007	16:38:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	05/09/2007	18:14:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	06/22/2007	18:53:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	07/18/2007	18:03:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI PD BELTRAN,PEDRO,	08/23/2007	19:19:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source
MIAMI P.D. VEIGA,SANDRA,F	09/12/2007	15:15:59	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531401G
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	11/14/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.076
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

OFFICER PEDRO BELTRAN WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	09/12/2007	18:20:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	10/09/2007	15:13:16	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:SIMULATOR CONNECTOR TO INLET WAS LOOSE . Non-compliance:A.I. ERROR FOR NO PROPERLY CONNECTING SIM. .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	10/09/2007	16:02:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	11/14/2007	21:45:00	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	626102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001654

Agency Operator	Date	Time	Software	Source	
MIAMI P.D. BELTRAN,PEDRO,	12/11/2007	20:54:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:FORGOT TO ADD ACETONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMTOHY,B	01/31/2007	14:35:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/05/2007	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMOTHY,B	02/27/2007	15:24:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.183
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.185
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.184
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMTOHY,B	03/31/2007	07:04:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMOTHY,B	04/30/2007	18:22:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMOTHY,B	05/31/2007	20:33:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMOTHY,B	06/30/2007	13:44:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMOTHY,B	07/30/2007	07:21:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROBERTSON,TIMOTHY,B	08/29/2007	15:17:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROSADO,ANGELO,D	09/29/2007	18:20:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.047
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.078
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #3	0.078
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROSADO,ANGELO,D	09/29/2007	18:50:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROSADO,ANGELO,D	10/28/2007	18:09:03	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROSADO,ANGELO,D	10/28/2007	18:16:19	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST. JOHNS CO SO ROSADO,ANGELO,D	10/28/2007	19:02:24	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source
ST JOHN'S COUNTY SO MALHIOT,MATTHEW,E	10/30/2007	11:13:18	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.194
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.194
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.192
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	11/29/2007	08:43:33	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610201G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/17/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

FAILED TO USE MA SOLUTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001655

Agency Operator	Date	Time	Software	Source	
ST JOHN'S COUNTY SO ROSADO,ANGELO,D	12/27/2007	14:33:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source
NORTH MIAMI PD VENTURI,GEORGE,L	03/06/2007	14:59:45	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	525202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/14/2007

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.077
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD BELTRAN,PEDRO,	04/10/2007	17:13:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	525202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD BELTRAN,PEDRO,	05/09/2007	18:13:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	525202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, HOSE WAS NOT CONNECTED . FORGOT TO CONNECT DRY GAS HOSE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD BELTRAN,PEDRO,	06/27/2007	14:47:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	525202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD BELTRAN,PEDRO,	07/18/2007	18:12:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	525202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD BELTRAN,PEDRO,	08/23/2007	19:08:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	525202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD DEJESUS,RAMON,	09/12/2007	21:07:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	525202I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source
NORTH MIAMI PD VEIGA,SANDRA,F	10/25/2007	13:47:31	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714902I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/31/2009

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.193
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.193
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.193
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.194
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.193
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.193
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0010
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

OFFICER R. DEJESUS WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD DEJESUS,RAMON,	10/25/2007	14:29:46	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	7149021
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/31/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD DE JESUS,RAMON,	11/29/2007	20:30:03	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714902I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/31/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD DE JESUS,RAMON,	12/24/2007	13:25:06	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001656

Agency Operator	Date	Time	Software	Source	
NORTH MIAMI PD DE JESUS,RAMON,	12/24/2007	14:01:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714902I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/31/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	01/19/2007	19:35:05	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	02/02/2007	12:32:50	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200504D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531401G
Expiration	09/22/2007	**Expiration**	04/21/2007	**Expiration**	09/22/2007	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	03/08/2007	13:49:49	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	531401G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD VELAZQUEZ,KIMBERLY,	04/11/2007	14:05:47	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531401G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD VELAZQUEZ,KIMBERLY,	05/16/2007	13:15:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/14/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	06/19/2007	11:23:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD VELAZQUEZ,KIMBERLY,	07/18/2007	11:44:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	08/15/2007	11:48:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	09/03/2007	13:45:25	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD VELAZQUEZ,KIMBERLY,	10/10/2007	10:16:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

PSA ESPINOZA PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source
MICCOSUKEE PD VEIGA,SANDRA,F	10/10/2007	17:04:43	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	01/26/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.077
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.077
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.077
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.077
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.077
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.077
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

OFFICER K. VELAZQUEZ & PSA ESPINOZA WERE PRESENT FOR INSPEC

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	11/29/2007	22:41:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPIN0ZA,PABLO,	11/29/2007	22:45:01	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001657

Agency Operator	Date	Time	Software	Source	
MICCOSUKEE PD ESPINOZA,PABLO,	12/11/2007	11:29:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

OFC VELAZQUEZ PRESENT FOR AI

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD MALLO,NICOLE,L	04/27/2007	14:48:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD SASSER,WILLIAM,C	05/10/2007	16:52:42	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A]	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source
UNIV OF FLORIDA PD MALHIOT,MATTHEW,E	06/13/2007	11:44:07	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.196
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD MALLO,NICOLE,L	06/23/2007	17:56:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD MALLO,NICOLE,L	07/26/2007	12:21:41	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD MALLO,NICOLE,L	08/18/2007	22:01:21	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD SASSER,WILLIAM,C	09/26/2007	17:32:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	629302I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD SASSER,WILLIAM,C	10/29/2007	17:13:08	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD MALLO,NICOLE,L	11/30/2007	02:36:13	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001714

Agency Operator	Date	Time	Software	Source	
UNIV OF FLORIDA PD MALLO,NICOLE,L	12/23/2007	10:16:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	724801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	09/06/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A:TOO LITTLE MOUTH ALCOHOL/STILL HAD M/A FROM PRIOR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	01/15/2007	07:33:06	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	02/12/2007	06:59:54	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	03/25/2007	07:35:45	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	04/24/2007	17:38:12	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	05/19/2007	07:12:13	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515901I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	06/28/2007	07:31:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	07/25/2007	02:38:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Sequence Aborted.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	07/25/2007	03:32:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	08/27/2007	07:31:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	09/05/2007	09:01:23	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source
OSCEOLA CORRECTIONS HACKNEY,DWITE,N	09/05/2007	15:06:37	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.081
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.081
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.081
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	10/26/2007	07:31:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	705001I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	11/28/2007	07:01:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001715

Agency Operator	Date	Time	Software	Source	
OSCEOLA CORRECTIONS CAVINEE,KEVIN,A	12/17/2007	06:50:43	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	705001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	02/22/2009

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Improper Sample.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD FULWOOD,GORDON,H	04/10/2007	16:39:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518701I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD FULWOOD,GORDON,H	05/16/2007	15:23:30	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	518701I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/08/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD FULWOOD,GORDON,H	06/13/2007	15:03:31	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD FULWOOD,GORDON,H	07/20/2007	10:41:04	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD FULWOOD,GORDON,H	08/31/2007	17:33:10	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD PARKS,DONALD,M	09/27/2007	20:54:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD PARKS,DONALD,M	10/29/2007	20:45:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632002I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source
ALACHUA PD MALHIOT,MATTHEW,E	10/31/2007	11:08:50	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0003 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD PARKS,DONALD,M	11/28/2007	01:08:42	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632002I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001720

Agency Operator	Date	Time	Software	Source	
ALACHUA PD PARKS,DONALD,M	12/28/2007	20:40:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632002I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	01/25/2007	23:22:33	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	502701G
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	02/02/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	02/22/2007	23:42:43	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

TOO MUCH MOUTH ALCOHOL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	03/23/2007	12:16:31	8100.27	Agency	Inspection

0.05 Lot#	200504C	**0.08 Lot#**	200504D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	510902I
Expiration	04/21/2007	**Expiration**	04/21/2007	**Expiration**	05/05/2007	**Expiration**	04/21/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	04/26/2007	20:27:34	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/05/2007	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	05/30/2007	19:17:54	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MORGAN,DARIN,P	06/12/2007	16:55:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	07/26/2007	18:54:55	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed. A F / M A:BAD MOUTH PIECE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	08/30/2007	18:35:41	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. A F / M A:OPER ERROR. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	08/30/2007	19:24:40	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	535402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/22/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

OPER ERR.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
PALM BAY PD HACKNEY,DWITE,N	09/20/2007	10:21:48	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.081
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.195
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.196
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.196
			0.08 g/210L Dry Gas Std Test #9	0.081
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	09/20/2007	12:58:18	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	10/29/2007	13:22:49	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	11/22/2007	20:47:26	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance/LOOSE CONNECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001721

Agency Operator	Date	Time	Software	Source	
PALM BAY PD MARCINIK,DAVID,E	12/27/2007	19:44:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	714102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	01/26/2007	11:01:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	02/02/2007	12:12:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	504502I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	02/17/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	03/02/2007	13:01:40	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	04/03/2007	13:34:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	05/02/2007	16:26:51	8100.27	Agency	Inspection

0.05 Lot#	200505B	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	06/15/2007	10:34:18	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	07/13/2007	13:15:51	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	612201G	**0.08 Gas Lot#**	701201G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2008	**Expiration**	01/15/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	08/16/2007	12:27:15	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance GAS TUBE NOT ATTACHED . OPERATOR ERROR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
LEE COUNTY SO VENTURI,GEORGE,L	08/22/2007	15:52:50	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.199
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	09/12/2007	12:01:42	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612201G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	10/05/2007	11:10:52	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612202G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance TANK NOT ATTACHED. OP ERROR

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	11/20/2007	10:47:53	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001722

Agency Operator	Date	Time	Software	Source	
LEE COUNTY SO CLAFLIN,DAVID,S	12/11/2007	15:29:58	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	702401G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	01/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001723

Agency Operator	Date	Time	Software	Source
VENICE PD VENTURI,GEORGE,L	01/24/2007	17:47:54	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

# Sims	5	**Avg Std Dev**	0.0002	**Complies**	No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001723

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0000
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COMPLIES WITH 11D-8 F.A.C..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001723

Agency Operator	Date	Time	Software	Source	
VENICE PD BUTLER,CHRIS,T	10/15/2007	11:58:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.PUT INTO SERVICE AT S JAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001723

Agency Operator	Date	Time	Software	Source	
VENICE PD BUTLER,CHRIS,T	11/12/2007	12:34:15	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	608301G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/27/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT S JAIL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001723

Agency Operator	Date	Time	Software	Source	
VENICE PD BUTLER,CHRIS,T	12/03/2007	09:26:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	723501I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	08/26/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

INSPECTION DONE AT SOUTH JAIL

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	01/18/2007	10:36:58	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	02/21/2007	14:02:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	03/20/2007	12:18:38	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

SIMULATOR TURNED OFF TEMP TOO LOW.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	04/11/2007	11:57:52	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	05/25/2007	11:33:44	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.206
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

LEAK AT SIMULATOR PORT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	06/27/2007	11:16:47	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518602I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/07/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
HILLSBOROUGH SO SUERETH,DONALD,P	07/18/2007	14:02:09	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 4 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.199
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.201
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	07/19/2007	11:41:25	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	08/28/2007	11:25:25	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632401I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	09/21/2007	10:42:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: TUBING LOOSE AT SIMULATOR PORT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	10/15/2007	10:01:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	11/21/2007	09:30:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	362401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001731

Agency Operator	Date	Time	Software	Source	
HILLSBOROUGH SO GRAY,BEVERLY,A	12/20/2007	10:10:09	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632401I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source	
SOUTH MIAMI P.D. VEIGA,SANDRA,F	09/12/2007	12:52:06	8100.26	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

NOT IN SERVICENon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source
SOUTH MIAMI P.D. VEIGA,SANDRA,F	09/12/2007	16:13:12	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	04/08/2009

# Sims	5	**Avg Std Dev**	0.0006	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.194
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.047
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.047
0.08 g/210L Test #3	0.077
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.077
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0008
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

OFFICER PEDRO BELTRAN WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source	
SOUTH MIAMI P.D. BELTRAN,PEDRO,	09/12/2007	18:30:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source	
SOUTH MIAMI P.D. WEISSBERG,M,	10/09/2007	14:56:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source	
SOUTH MIAMI P.D. WEISSBERG,M,	11/14/2007	17:42:57	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	531505G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001737

Agency Operator	Date	Time	Software	Source	
SOUTH MIAMI P.D. WEISSBERG,MICHAEL,W	12/11/2007	21:58:27	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A:DID NOT USE MOUTH ALCOHOL SOLN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	01/09/2007	21:04:25	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

JANUARY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	02/10/2007	11:51:06	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	02/11/2007	18:48:20	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

RETEST OF FEBRUARY AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	03/09/2007	08:10:49	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/05/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MARCH 2007 AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO PUTMAN,BRIAN,A	04/21/2007	19:41:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200612D	**0.20 Lot#**	200505C	**0.08 Gas Lot#**	531505G
Expiration	09/22/2007	**Expiration**	12/14/2008	**Expiration**	05/05/2007	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

APRIL INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	05/23/2007	12:32:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MAY 2007 AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	06/17/2007	10:07:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200509B	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	5315O5G
Expiration	12/14/2008	**Expiration**	09/22/2007	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

JUNE AGENCY INSP

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	07/28/2007	13:37:26	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

FAIL PURGE TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	07/28/2007	15:22:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.043
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	08/11/2007	14:14:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 3 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AUGUST AGENCY INSPECTION

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
CHARLOTTE COUNTY SO VENTURI,GEORGE,L	08/30/2007	14:23:16	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.077
0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.195
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.195
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.195
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.078
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.047
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.046
			0.08 g/210L Test #10	0.078
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0003
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO SPEICHER,DALE,A	09/08/2007	14:36:14	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	531505G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	11/15/2007

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

OPERATOR ERROR, REMOVED FROM SERVICE FOR PM

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
CHARLOTTE COUNTY SO VENTURI,GEORGE,L	11/14/2007	16:04:10	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605C	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/20/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	11/16/2007	21:45:34	8100.27	Agency	Inspection

0.05 Lot# 610301G
Expiration 04/18/2008

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.048
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.049
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

USED WRONG LOT NUMBERS AND EXPIRATION DATES. REDO AG INSP USED WRONG LOT / EXP INFORMATION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	11/16/2007	22:26:39	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

RETESTED TEST PASS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001739

Agency Operator	Date	Time	Software	Source	
CHARLOTTE COUNTY SO WRIGHT,TAMMY,M	12/20/2007	17:31:39	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AGENCY INSPECTION MONTH OF DECEMBER

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	01/24/2007	16:59:37	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	01/26/2007	06:51:53	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	02/16/2007	12:52:36	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

A F / M A: No Sample Provided.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
CLEARWATER PD SUERETH,DONALD,P	02/20/2007	13:50:05	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	518602I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/07/2007

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.196
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.199
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0011
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	03/23/2007	04:50:46	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	615301I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	04/10/2007	18:59:10	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	05/25/2007	04:04:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.179
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.180
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.182
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	06/22/2007	09:59:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

MOUTH ALCOHOL LOT 2006-A ACETONE LOT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	07/27/2007	05:06:29	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615301I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/05/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILLER,THEODORE,J	08/16/2007	17:05:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

MOUTH ALCOHOL LOT 2006-A ACETONE LOT 2006-D

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source	
CLEARWATER PD MILNE,THOMAS,P	09/11/2007	15:32:44	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	615802I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

INST TAKEN O/S FOR CALIBRATION CHECKS

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
CLEARWATER PD SUERETH,DONALD,P	11/08/2007	14:08:12	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.047
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0015
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
CLEARWATER PD SUERETH,DONALD,P	11/28/2007	17:39:53	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605C	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0008 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.046
0.08 g/210L Test #1	0.075
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.046
0.08 g/210L Test #2	0.075
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.076
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.000
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.047
0.08 g/210L Test #4	0.076
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.000
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.047
0.08 g/210L Test #5	0.076
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.000
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.047
0.08 g/210L Test #6	0.077
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.000
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.047

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001740

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.077
			0.20 g/210L Test #7	0.197
			0.08 g/210L Dry Gas Std Test #7	0.000
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.047
			0.08 g/210L Test #8	0.076
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.000
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.000
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.000
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0007
			0.20 g/210L Test (Std Dev)	0.0016
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Control Outside Tolerance. GAS HOSE DISCONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	01/12/2007	12:18:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	610001G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	02/23/2007	08:49:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	03/16/2007	10:46:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	04/13/2007	11:44:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	632102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	05/04/2007	09:22:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	06/08/2007	12:11:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	706101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/05/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	07/13/2007	11:02:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
PALM BEACH CO SO HACKNEY,DWITE,N	07/26/2007	14:53:09	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	627201I
Expiration	12/14/2008	**Expiration**	12/14/2006	**Expiration**	12/05/2008	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.048
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	08/17/2007	10:59:39	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632102I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	09/14/2007	12:13:28	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	10/19/2007	10:19:12	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610001G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	11/21/2007	11:23:02	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714901I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/31/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001741

Agency Operator	Date	Time	Software	Source	
PALM BEACH CO SO CROUCHER,GREGORY,P	12/14/2007	12:29:09	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	610001G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	01/23/2007	09:17:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	
Expiration	05/04/2008	**Expiration**	05/04/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.046
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.075
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.046
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.076
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: Control Outside Tolerance. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
BRADENTON PD VENTURI,GEORGE,L	03/21/2007	14:09:27	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	626101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	09/20/2008

 # Sims 4 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.078
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.000
0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.077
0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.078
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.077
0.08 g/210L Dry Gas Std Test #4 (Retry)	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.077
0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.079
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #7	0.079
			0.20 g/210L Test #7	0.198
			0.08 g/210L Dry Gas Std Test #7	0.078
			0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
			Interferent Detect Test #7	INT
			Alcohol Free Test #8	0.000
			0.05 g/210L Test #8	0.049
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.079
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0245
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

08: Control Outside Tolerance: DGS HOSE NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	04/17/2007	11:51:11	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	05/08/2007	14:23:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	06/05/2007	11:42:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	07/03/2007	12:58:28	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	08/01/2007	12:36:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	09/12/2007	09:53:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	10/02/2007	16:06:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	606002I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	03/03/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	11/06/2007	16:24:38	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	12/13/2007	13:02:51	8100.27	Agency	Inspection

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	12/13/2007	13:33:28	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008		

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.077
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.197
0.05 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001742

Agency Operator	Date	Time	Software	Source	
BRADENTON PD ANDREWS,ROBERT,P	12/13/2007	14:09:05	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	01/19/2007	13:37:06	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	02/13/2007	12:41:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
LAKELAND PD SUERETH,DONALD,P	03/05/2007	15:40:42	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

# Sims	4	**Avg Std Dev**	0.0005	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.082
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.082
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.082
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.052
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.082
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.082
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.082
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.053
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.199
0.08 g/210L Dry Gas Std Test #6	0.082
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.053
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.199
0.08 g/210L Dry Gas Std Test #7	0.082
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.053

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.082
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.082
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.082
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	03/14/2007	12:10:18	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	04/27/2007	11:38:36	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/24/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	05/18/2007	10:44:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	06/26/2007	12:29:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	07/24/2007	12:05:34	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	08/27/2007	12:42:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	09/24/2007	12:23:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	10/23/2007	11:44:15	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	11/26/2007	12:13:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001865

Agency Operator	Date	Time	Software	Source	
LAKELAND PD KNOBLOCH,WILLIAM,T	12/26/2007	12:47:08	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	619101I
Expiration	08/01/2009	**Expiration**	05/10/2008	**Expiration**	05/10/2008	**Expiration**	07/12/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL SMITH,ELMER,H	01/31/2007	22:25:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

20: CONTROL OUTSIDE TOLERANCE LOOSE HOOK UP .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL SMITH,ELMER,H	02/28/2007	20:39:16	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

TANK NOT CONNECTED ON FIRST TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL SMITH,ELMER,H	03/31/2007	19:15:56	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	04/22/2007	05:48:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.188
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.190
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.190
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	06/03/2007	07:04:32	8100.27	Agency	Inspection

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	06/03/2007	09:08:24	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	07/31/2007	15:33:22	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
US AIR FORCE MACDILL SUERETH,DONALD,P	08/01/2007	09:58:47	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	709601I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.195
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.051
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.196
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.197
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.196
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.052
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.197
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0007
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	09/12/2007	15:33:42	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	10/19/2007	08:52:37	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001866

Agency Operator	Date	Time	Software	Source	
US AIR FORCE MACDILL RICE,NYLES,L	12/04/2007	00:13:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	535402I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	12/22/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.185
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.187
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.189
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source
PORT ST LUCIE PD HACKNEY,DWITE,N	01/02/2007	15:40:03	8100.27	Department Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	518702I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	07/08/2007

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.077
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.077
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.077
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.077
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.077
			0.20 g/210L Test #9	0.197
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.077
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,IAN,	01/03/2007	11:38:30	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	515902I
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	06/13/2007

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	02/28/2007	17:58:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	03/26/2007	13:16:39	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	04/11/2007	11:42:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	05/31/2007	12:32:52	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	06/27/2007	22:23:12	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	07/29/2007	16:40:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	08/28/2007	16:38:25	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.075
		Interferent Detect Test #3	INT

Remarks:

DID NOT HAVE SIMULATOR HOOKED UP FOR 1ST PART OF 1ST SAMPLE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	09/30/2007	15:28:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.071
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.071
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

INLET TUBE NOT SEATED SNUG.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	10/31/2007	04:54:32	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	11/23/2007	23:00:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-001962

Agency Operator	Date	Time	Software	Source	
PORT ST LUCIE PD HARRIS,OFC,I	12/31/2007	16:42:41	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	607401G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	03/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	01/19/2007	09:08:43	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612501G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	No	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. 08: Control Outside DUE TO AIR SEAL NOT MAINTAINED .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	02/16/2007	10:25:42	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612501G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	03/15/2007	11:18:18	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	612501G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	09/22/2007	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Improper Sample.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	04/25/2007	09:55:02	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	09/22/2007	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	05/25/2007	08:34:26	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	06/22/2007	09:38:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	07/27/2007	08:02:20	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	08/28/2007	08:54:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	09/27/2007	08:58:14	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	05/10/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range ExceededRange Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source
NAS JACKSONVILLE MALHIOT,MATTHEW,E	10/25/2007	09:55:51	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	627101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	10/03/2008

 # Sims 5 **Avg Std Dev** 0.0005 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.051
0.08 g/210L Test #1	0.083
0.20 g/210L Test #1	0.203
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.052
0.08 g/210L Test #2	0.084
0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.083
0.20 g/210L Test #3	0.201
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.053
0.08 g/210L Test #4	0.083
0.20 g/210L Test #4	0.202
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.084
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.084
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.083
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.084
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.083
			0.20 g/210L Test #9	0.202
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.053
			0.08 g/210L Test #10	0.083
			0.20 g/210L Test #10	0.202
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	10/29/2007	08:58:03	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612501G
Expiration	05/04/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	11/20/2007	09:46:57	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002097

Agency Operator	Date	Time	Software	Source	
NAS JACKSONVILLE HEMPHILL,CLOTE,	12/28/2007	08:39:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	612501G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	05/08/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	03/08/2007	13:59:12	8100.27	Agency	Inspection

0.05 Lot# 200605B
Expiration 05/04/2008

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.069
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.073
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.074
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	03/08/2007	14:01:25	8100.27	Agency	Inspection

 # Sims 0 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	03/08/2007	14:03:41	8100.27	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	03/08/2007	14:11:26	8100.27	Agency	Inspection

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	03/08/2007	14:33:47	8100.27	Agency	Inspection

 # Sims 1 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test #1	INT
Minimum Sample Volume Check: OK		Interferent Detect Test #1 (Retry)	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	No	Interferent Detect Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2 (Retry)	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test #3 (Retry)	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:Y, . Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	06/15/2007	11:12:23	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 2 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		Interferent Detect Test #1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.051
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.080
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test #2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #3	0.080
0.20 g/210L Test: OK	No	Interferent Detect Test #3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002170

Agency Operator	Date	Time	Software	Source	
NAS KEY WEST BLAIR,PETER,B	06/15/2007	12:25:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	617202I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/23/2008

 # Sims 4 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det: RFI Detect DUE TO CELL PHONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	01/24/2007	09:08:58	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	02/20/2007	08:04:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	03/14/2007	08:29:21	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	04/19/2007	07:53:00	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Subject Test Refused.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	05/23/2007	07:44:02	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	06/14/2007	14:17:43	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

00: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	07/12/2007	08:05:45	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	08/17/2007	10:52:13	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	09/05/2007	09:56:57	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source
FSU POLICE GEDDINGS,MARGARET,M	10/18/2007	16:41:39	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1 (Retry)	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.197
0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test #1	INT
Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.077
Interferent Detect Test #2	INT
Alcohol Free Test #3 (Retry)	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4 (Retry)	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.197
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5 (Retry)	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.079
0.20 g/210L Test #5	0.198
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6 (Retry)	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.198
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7 (Retry)	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8 (Retry)	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.198
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9 (Retry)	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10 (Retry)	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.198
			0.08 g/210L Dry Gas Std Test #10	0.078
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Ambient FailNO VENTILATION IN ROOM OPEN DOORS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	10/29/2007	10:16:51	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	11/02/2007	07:57:19	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Sequence Aborted. 00: Ambient Fail. 20: Ambient Fail.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002171

Agency Operator	Date	Time	Software	Source	
FSU POLICE GIANGROSSO,JOE,P	12/13/2007	08:12:55	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	617202I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	01/22/2007	13:07:33	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	02/28/2007	10:35:07	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

Time-Date changed.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	03/28/2007	10:49:59	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	04/26/2007	12:28:05	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	05/30/2007	18:22:01	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	06/30/2007	16:55:50	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	620901I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	08/02/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

20: RFI Detect/CELL PHONE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	07/31/2007	21:50:38	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	717102I
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.05 g/210L Test #3 (Retry)	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

05: RFI Detect/CELLULAR PHONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source
NEW SYMRNA BEACH PD SKIPPER,ROGER,G	08/28/2007	16:39:44	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	719402I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	07/15/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.052
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.204
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.052
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.204
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.204
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.052
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.052
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.204
0.08 g/210L Dry Gas Std Test #6	0.081
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.052
0.08 g/210L Test #7	0.082
0.20 g/210L Test #7	0.204
0.08 g/210L Dry Gas Std Test #7	0.081
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.052

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.204
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.052
			0.08 g/210L Test #9	0.082
			0.20 g/210L Test #9	0.204
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.053
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.204
			0.08 g/210L Dry Gas Std Test #10	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD WILLIAMS,DAVID,B	08/29/2007	01:41:06	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
		Interferent Detect Test #3	INT

Remarks:

08: DRY GAS NOT CONNECTED.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	09/20/2007	21:42:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	10/25/2007	17:26:03	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002231

Agency Operator	Date	Time	Software	Source	
NEW SYMRNA BEACH PD ROOS,CHRISTOPHER,R	11/13/2007	11:14:14	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	717102I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	06/21/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

05: SIMULATOR WAS NOT HOOKED UP DURING FIRST .05 TEST.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	01/19/2007	23:59:55	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	622802I
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	08/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	02/27/2007	07:38:21	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	03/27/2007	09:15:56	8100.27	Agency	Inspection

0.05 Lot# 200509A
Expiration 09/22/2007

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		Interferent Detect Test #1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.050
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.050
Alcohol Free Test: OK	Yes	Interferent Detect Test #3	INT
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	03/27/2007	09:56:44	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

PREVIOUS TEST MISTAKENLY TERMINATED DUE WRONG KEY STROKE

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	04/24/2007	09:29:51	8100.27	Agency	Inspection

0.05 Lot#	200509A	**0.08 Lot#**	200509B	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	634801G
Expiration	09/22/2007	**Expiration**	09/22/2007	**Expiration**	09/22/2007	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	05/30/2007	07:15:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	06/20/2007	07:42:35	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	07/31/2007	00:10:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	08/28/2007	07:28:05	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1 (Retry)	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	0.000
		Interferent Detect Test #2 (Retry)	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	0.000
		Interferent Detect Test #3 (Retry)	INT

Remarks:

Int Det:SIMULATOR LID TIGHTENED FOR BETTER SEAL.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	09/26/2007	07:53:00	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	634801G
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	12/18/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.034
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.038
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.045
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: SIMULATOR TURNED OFF BEFORE TEST,REHEATED AND TEST RUN.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
MIAMI BEACH PD VEIGA,SANDRA,F	10/25/2007	13:29:01	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.190
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.191
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.192
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.193
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.194
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.048
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.194
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.194
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.048

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.195
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.048
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.195
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.048
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.195
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0003
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0017
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

OFFICER KEVIN MILLAN WAS PRESENT FOR INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	10/26/2007	00:05:46	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	11/29/2007	08:00:36	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	714101I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	05/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

AI DIONNE PRESENT

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002232

Agency Operator	Date	Time	Software	Source	
MIAMI BEACH PD MILLAN,KEVIN,	12/20/2007	00:59:45	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	724701G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	09/06/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Interferent Detect Test #3	INT

Remarks:

AI M. DIONNE PRESENT FOR INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source
HIALEAH P.D. VEIGA,SANDRA,F	06/14/2007	13:54:15	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 5 **Avg Std Dev** **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.048
0.08 g/210L Test #1	0.075
0.08 g/210L Test #1 (Retry)	0.075
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.074
0.08 g/210L Test #2 (Retry)	0.074
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.080
0.08 g/210L Test #3 (Retry)	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.08 g/210L Test #4 (Retry)	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.048
0.08 g/210L Test #5	0.079
0.08 g/210L Test #5 (Retry)	0.075
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.073
0.08 g/210L Test #6 (Retry)	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.073
0.08 g/210L Test #7 (Retry)	0.075
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049
0.08 g/210L Test #8	0.073
0.08 g/210L Test #8 (Retry)	0.074
Interferent Detect Test #8	INT
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.080
0.08 g/210L Test #9 (Retry)	0.075

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.049
			0.08 g/210L Test #10	0.080
			0.08 g/210L Test #10 (Retry)	0.081
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0033
			0.08 g/210L Test (Std Dev)	0.0027

Remarks:

08: Control Outside Tolerance. LOOSE CONNECTION., Control O. RE RETURN TO FACTORYNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source	
HIALEAH P. D. VEIGA,SANDRA,F	10/08/2007	14:58:15	8100.27	Agency	Inspection

 # Sims 0 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RETURNED FROM CMI, NO AI INSPECTIONNon-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source
HIALEAH P. D. VEIGA,SANDRA,F	10/08/2007	17:23:47	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200605C	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	05/04/2008	**Expiration**	12/05/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0007 **Complies** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.192
0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test #2	0.000
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.080
0.20 g/210L Test #3	0.195
0.08 g/210L Dry Gas Std Test #3	0.000
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.196
0.08 g/210L Dry Gas Std Test #4	0.000
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.196
0.08 g/210L Dry Gas Std Test #5	0.000
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.197
0.08 g/210L Dry Gas Std Test #6	0.000
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.000
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.197
			0.08 g/210L Dry Gas Std Test #8	0.000
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.198
			0.08 g/210L Dry Gas Std Test #9	0.000
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.000
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0006
			0.20 g/210L Test (Std Dev)	0.0020
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

08: DRY GAS NOT PROPERLY CONNECTED Non-compliance:DRY GAS.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source
HIALEAH P. D. VEIGA,SANDRA,F	10/15/2007	14:35:30	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/08/2009

 # Sims 5 **Avg Std Dev** 0.0004 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.081
0.20 g/210L Test #1	0.202
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.051
0.08 g/210L Test #2	0.081
0.20 g/210L Test #2	0.203
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.203
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.051
0.08 g/210L Test #4	0.081
0.20 g/210L Test #4	0.203
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.051
0.08 g/210L Test #5	0.081
0.20 g/210L Test #5	0.203
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.051
0.08 g/210L Test #6	0.081
0.20 g/210L Test #6	0.202
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.202
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.082
			0.20 g/210L Test #8	0.202
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.203
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.082
			0.20 g/210L Test #10	0.203
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

M/A NEED MORE SOLUTION.OFF A. ALVAREZ WAS PRESENT DURING IN

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002462

Agency Operator	Date	Time	Software	Source	
HIALEAH P. D. MONTALVO,M,	12/11/2007	10:47:17	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	705001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	02/22/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source
HIALEAH P.D. VEIGA,SANDRA,F	06/14/2007	14:37:52	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	09/22/2007	**Expiration**	11/20/2008

# Sims	5	**Avg Std Dev**	0.0004	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.200
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.201
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.049
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.079
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.078
0.20 g/210L Test #5	0.201
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.078
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.079
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.078
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.079
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0004
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	07/25/2007	10:47:29	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632201I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.065
0.05 g/210L Test #1 (Retry)	0.055
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.173
0.20 g/210L Test #1 (Retry)	0.196
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.056
0.05 g/210L Test #2 (Retry)	0.054
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.193
0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.054
0.05 g/210L Test #3 (Retry)	0.053
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.197
0.20 g/210L Test #3 (Retry)	0.197
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT

Remarks:

A F / M A: . 05: Control Outside Tolerance. 20: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	08/30/2007	13:44:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

N/A

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	09/25/2007	08:23:58	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test #1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.066
Alcohol Free Test: OK	Yes	Interferent Detect Test #2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.20 g/210L Test: OK	No	0.08 g/210L Test #3	0.068
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance, RFI Detect. RFI FROM POLICE RADIO.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. MONTALVO,M,	09/25/2007	10:40:53	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	632001I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/22/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. MONTALVO,M,	10/12/2007	01:03:35	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	11/19/2007	08:52:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	632001I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/20/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	12/12/2007	08:33:33	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL. Non-compliance:TOO MUCH MOUTH ALCOHOL, OPERATOR ERROR..

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002463

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	12/12/2007	09:15:02	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	607401G
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	03/23/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source
HIALEAH P.D. VEIGA,SANDRA,F	06/14/2007	13:39:34	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200509C	**0.08 Gas Lot#**	632101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	09/22/2007	**Expiration**	11/20/2008

# Sims	5	**Avg Std Dev**	0.0003	**Complies**	Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.049
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.049
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.199
0.08 g/210L Dry Gas Std Test #3	0.079
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.049
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.199
0.08 g/210L Dry Gas Std Test #4	0.079
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.079
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.079
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.050
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.079
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.050

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.200
			0.08 g/210L Dry Gas Std Test #8	0.079
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.050
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.200
			0.08 g/210L Dry Gas Std Test #9	0.079
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.200
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0006
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	07/25/2007	09:39:47	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.053
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.055
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

A F / M A: Range Exceeded. 05: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. MONTALVO,MANUEL,	08/28/2007	13:45:32	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

08: Control Outside Tolerance.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source
HIALEAH P.D. VEIGA,SANDRA,F	09/04/2007	13:21:14	8100.27	Department Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	709601I
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	04/08/2009

 # Sims 4 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.047
0.08 g/210L Test #1	0.079
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.048
0.08 g/210L Test #2	0.079
0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test #2	0.078
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.048
0.08 g/210L Test #3	0.079
0.20 g/210L Test #3	0.200
0.08 g/210L Dry Gas Std Test #3	0.078
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.048
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.200
0.08 g/210L Dry Gas Std Test #4	0.078
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.049
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.200
0.08 g/210L Dry Gas Std Test #5	0.078
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.049
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.201
0.08 g/210L Dry Gas Std Test #6	0.078
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.080
0.20 g/210L Test #7	0.201
0.08 g/210L Dry Gas Std Test #7	0.078
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.080
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.078
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.049
			0.08 g/210L Test #9	0.080
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.078
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.050
			0.08 g/210L Test #10	0.080
			0.20 g/210L Test #10	0.201
			0.08 g/210L Dry Gas Std Test #10	0.079
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0008
			0.08 g/210L Test (Std Dev)	0.0004
			0.20 g/210L Test (Std Dev)	0.0009
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

OFFICER A. ALVAREZ WAS PRESENT DURING THE INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	09/24/2007	15:33:20	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008		

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		Interferent Detect Test #3	INT

Remarks:

20: Control Outside Tolerance, RFI Detect. Non-compliance: .

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. ALVAREZ,AMADO,L	09/25/2007	09:19:46	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200612D	**0.20 Lot#**	200612A	**0.08 Gas Lot#**	631101I
Expiration	12/14/2008	**Expiration**	12/14/2008	**Expiration**	12/05/2008	**Expiration**	11/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Interferent Detect Test #2	INT
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

08: RFI Detect.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. MONTALVO,M,	10/12/2007	02:40:25	8100.27	Agency	Inspection

 # Sims 5 **Complies** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance:OFFICER CAME IN WITH A CELL-PHONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. MONTALVO,M,	10/12/2007	03:12:17	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	631101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002464

Agency Operator	Date	Time	Software	Source	
HIALEAH P.D. RODRIGUEZ,MANUEL,	11/19/2007	16:14:23	8100.27	Agency	Inspection

0.05 Lot#	2007708A	**0.08 Lot#**	200708C	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	631101I
Expiration	08/01/2009	**Expiration**	08/01/2009	**Expiration**	08/01/2009	**Expiration**	11/09/2008

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source
SURFSIDE P.D. VEIGA,SANDRA,F	07/18/2007	16:07:56	8100.27	Department Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Avg Std Dev** 0.0006 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.050
0.08 g/210L Test #1	0.080
0.20 g/210L Test #1	0.198
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.050
0.08 g/210L Test #2	0.080
0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test #2	0.079
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.050
0.08 g/210L Test #3	0.081
0.20 g/210L Test #3	0.198
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT
Alcohol Free Test #4	0.000
0.05 g/210L Test #4	0.050
0.08 g/210L Test #4	0.080
0.20 g/210L Test #4	0.198
0.08 g/210L Dry Gas Std Test #4	0.080
Interferent Detect Test #4	INT
Alcohol Free Test #5	0.000
0.05 g/210L Test #5	0.050
0.08 g/210L Test #5	0.080
0.20 g/210L Test #5	0.199
0.08 g/210L Dry Gas Std Test #5	0.080
Interferent Detect Test #5	INT
Alcohol Free Test #6	0.000
0.05 g/210L Test #6	0.050
0.08 g/210L Test #6	0.080
0.20 g/210L Test #6	0.200
0.08 g/210L Dry Gas Std Test #6	0.080
Interferent Detect Test #6	INT
Alcohol Free Test #7	0.000
0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.081
0.20 g/210L Test #7	0.200
0.08 g/210L Dry Gas Std Test #7	0.080
Interferent Detect Test #7	INT
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051

Printed: 22 February 2008

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source
			0.08 g/210L Test #8	0.081
			0.20 g/210L Test #8	0.201
			0.08 g/210L Dry Gas Std Test #8	0.080
			Interferent Detect Test #8	INT
			Alcohol Free Test #9	0.000
			0.05 g/210L Test #9	0.051
			0.08 g/210L Test #9	0.081
			0.20 g/210L Test #9	0.201
			0.08 g/210L Dry Gas Std Test #9	0.080
			Interferent Detect Test #9	INT
			Alcohol Free Test #10	0.000
			0.05 g/210L Test #10	0.051
			0.08 g/210L Test #10	0.081
			0.20 g/210L Test #10	0.199
			0.08 g/210L Dry Gas Std Test #10	0.080
			Interferent Detect Test #10	INT
			0.05 g/210L Test (Std Dev)	0.0005
			0.08 g/210L Test (Std Dev)	0.0005
			0.20 g/210L Test (Std Dev)	0.0012
			0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

OFFICER E. PINEDA WAS PRESENT DURING INSPECTION.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source	
SURFSIDE P.D. PINEDA,EDWIN,	07/18/2007	17:46:17	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

OFC. K. VELAZQUEZ PRESENT FOR A.I.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source	
SURFSIDE P.D. PINEDA,EDWIN,	08/22/2007	11:35:16	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source	
SURFSIDE P.D. PINEDA,EDWIN,	09/26/2007	13:40:26	8100.27	Agency	Inspection

0.05 Lot#	200708A	**0.08 Lot#**	200612D	**0.20 Lot#**	200708D	**0.08 Gas Lot#**	711301G
Expiration	08/01/2009	**Expiration**	12/14/2008	**Expiration**	08/01/2009	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	INT
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test #1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test #2	INT
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

00: Interferent Detect HOOKUP WRONG SIMULATOR W/ACETONE.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source	
SURFSIDE P.D. PINEDA,EDWIN,	10/30/2007	13:40:06	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605C	**0.08 Gas Lot#**	711301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 4 **Complies** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test #1	0.000
0.05 g/210L Test #1	0.040
0.05 g/210L Test #1 (Retry)	0.047
0.08 g/210L Test #1	0.076
0.20 g/210L Test #1	0.188
0.20 g/210L Test #1 (Retry)	0.193
0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test #1	INT
Alcohol Free Test #2	0.000
0.05 g/210L Test #2	0.041
0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test #2	0.078
0.20 g/210L Test #2	0.191
0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test #2	0.080
Interferent Detect Test #2	INT
Alcohol Free Test #3	0.000
0.05 g/210L Test #3	0.033
0.05 g/210L Test #3 (Retry)	0.047
0.08 g/210L Test #3	0.078
0.20 g/210L Test #3	0.192
0.20 g/210L Test #3 (Retry)	0.194
0.08 g/210L Dry Gas Std Test #3	0.080
Interferent Detect Test #3	INT

Remarks:

A F / M A:BLEW ALCOHOL FREE. 05: Control Outside/NEW SIM. IMPROPER ATTACHMENT. 20: Control Outside/TIGHTENED SIM. TOP.

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source	
SURFSIDE P.D. PINEDA,EDWIN,	11/12/2007	12:46:36	8100.27	Agency	Inspection

0.05 Lot#	200612C	**0.08 Lot#**	200605C	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	12/14/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.

Inspection Tests

Instrument Serial #80-002566

Agency Operator	Date	Time	Software	Source	
SURFSIDE P.D. CHANG,DAVID,	12/25/2007	15:48:47	8100.27	Agency	Inspection

0.05 Lot#	200605B	**0.08 Lot#**	N	**0.20 Lot#**	200605F	**0.08 Gas Lot#**	711301G
Expiration	05/04/2008	**Expiration**	05/04/2008	**Expiration**	05/10/2008	**Expiration**	04/25/2009

 # Sims 5 **Complies** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test #1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test #2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080
		Interferent Detect Test #3	INT

Remarks:

COBRA ©1997-2006 CMI, Inc.